



PARAGON ANALYTICS

225 Commerce Drive ♦ Fort Collins, CO 80524 ♦ (800) 443-1511 ♦ (970) 490-1511 ♦ FAX (970) 490-1522

June 30, 2006

Mr. James Hix
Cordilleran Compliance Services, Inc.
5550 Marshall Street
Arvada CO 80002

Re: Paragon Workorder: 06-05-177
Client Project Name: Battlement Mesa
Client Project Number: E04243

Dear Mr. Hix:

Two water samples were received from Cordilleran Compliance Services, Inc. on May 24, 2006. The samples were scheduled for Tritium (pages 1-78) and Gamma Spectroscopy (pages 1-183) analysis.

The results for this analysis are contained in the enclosed reports.

Thank you for your confidence in Paragon Analytics. Should you have any questions, please call.

Sincerely,

Paragon Analytics
Lance Steere
Project Manager

LRS/mh
Enclosure: Report

Paragon Analytics

Sample Number(s) Cross-Reference Table

Paragon OrderNum: 0605177

Client Name: Cordilleran Compliance Services, Inc.

Client Project Name: Battlement Mesa

Client Project Number: E04243

Client PO Number:

| Client Sample | Lab Sample Number | COC Number | Matrix | Date Collected | Time Collected |
|---------------|-------------------|------------|--------|----------------|----------------|
| BM3613 PW | 0605177-1 | | WATER | 17-May-06 | 8:58 |
| BM2642 | 0605177-2 | | WATER | 20-May-06 | 12:15 |



Paragon Analyticals
A Division of DataChem Laboratories, Inc.

225 Commerce Drive Fort Collins, CO 80524
800-443-1511 or (970) 490-1511 (970) 490-1522 Fax

Accession Number (LAB ID) 0506177 Date 5/24/06
Chain-of-Custody Page 0605177 of 1 Originator: Retain pink copy

Project Name / No.: Presco/E04243 Sampler(s):
Report To: James Hix
Phone: 303.237.2072
Fax: 303.237.2459
E-mail: james.hix@cardcomp.com
Company: Cardilleran Compliance Services
Address: 5550 Marshall Street
Arvado, CO 80002

Circle method (right); provide additional information as needed (comments).

| Sample ID | Date | Time | Lab ID | Matrix | Preservative | (Indicate type) | No. of Containers |
|---------------|----------------|-------------|----------|--------|--------------|-----------------|-------------------|
| <u>BM2642</u> | <u>5/23/06</u> | <u>1215</u> | <u>2</u> | | | | |

| Method | Standard or Rush (Due) | Turnaround | Dispose: Date | or Return to Client |
|-----------------------------|-------------------------------------|------------|---------------|---------------------|
| VOCs | SW8260B E624 E5242 | | | |
| BTEX (only) | SW8021B | | | |
| SVOCs | SW8270C E625 | | | |
| OC Pesticides | SW8081A E608 | | | |
| PCBs | SW8082 E608 | | | |
| Herbicides | SW8151A | | | |
| Explosives | SW6330 | | | |
| TCLP Organics SW1311 | SW8260B R270C 8081A 8151A | | | |
| TCLP Metals SW1311 | SW6010B SW7471 | | | |
| Total Metals by ICP | SW6010B 7470 7471 E200 ILMO | | | |
| Dissolved Metals by ICP | SW6010B 7470 E200 ILMO | | | |
| Total Metals by ICP/MMS | SW6020 E200B ILMO | | | |
| Dissolved Metals by ICP/MMS | SW6020 E200B ILMO | | | |
| Hexavalent Chromium | SW7196A Alkaline Digest? Y / N | | | |
| Inorganic Anions | SW9056 E300.0 (specify in comments) | | | |
| Solids: | Total E160.3 TDS E160.1 TSS E160.2 | | | |
| pH | SW9040B SW9045C | | | |
| TPH | GRO DRO SW8015B (both) | | | |
| Gross Alpha / Beta | SW9310 E900.0 | | | |
| Actinides by Paragon SOP | Pu / U / Am / Th / Cm / | | | |
| Tritium | E906.0 | | | |
| Total Alpha-Emitting Radium | SW9315 E903.0 | | | |
| Radium 226 E903.1 | Radium 226 SW9320 E904.0 | | | |
| Strontium 90 D5811-00 | Strontium 90 D5811-00 | | | |
| Gamma Isotopes | E901.1 | | | |

[Handwritten Signature]
5/23/06

Relinquished By: *[Signature]*
Signature _____
Printed Name _____
Date _____ Time _____
Company _____

Received By: *[Signature]*
Signature _____
Printed Name L. Kozar
Date 5/24/06 Time 0925
Company Paragon

* Time Zone: EST CST MST PST Matrix Key: O = oil, S = sol, NS = non-sol solid, W = water, L = liquid, E = extract, F = filter
Comments:

CONDITION OF SAMPLE UPON RECEIPT FORM

Paragon Analytics

Client: Cardi Heran

Workorder No: 0603177

Project Manager: _____

Initials: _____

Date: _____

| | | | |
|---|---------------------------------------|--------------------------------------|-------------------------------------|
| 1. Does this project require any special handling in addition to standard Paragon procedures? | | YES | <input checked="" type="radio"/> NO |
| 2. Are custody seals on shipping containers intact? | NONE | <input checked="" type="radio"/> YES | NO |
| 3. Are Custody seals on sample containers intact? | <input checked="" type="radio"/> NONE | YES | NO |
| 4. Is there a COC (Chain-of-Custody) present or other representative documents? | | <input checked="" type="radio"/> YES | NO |
| 5. Are the COC and bottle labels complete and legible? | | <input checked="" type="radio"/> YES | NO |
| 6. Is the COC in agreement with samples received? (IDs, dates, times, no. of samples, no. of containers, matrix, requested analyses, etc.) | | <input checked="" type="radio"/> YES | NO |
| 7. Were airbills / shipping documents present and/or removable? | DROP OFF | <input checked="" type="radio"/> YES | NO |
| 8. Are all aqueous samples requiring preservation preserved correctly? (excluding volatiles) | N/A | YES | <input checked="" type="radio"/> NO |
| 9. Are all aqueous non-preserved samples pH 4-9? | N/A | <input checked="" type="radio"/> YES | NO |
| 10. Is there sufficient sample for the requested analyses? | | YES | <input checked="" type="radio"/> NO |
| 11. Were all samples placed in the proper containers for the requested analyses? | | <input checked="" type="radio"/> YES | NO |
| 12. Are all samples within holding times for the requested analyses? | | <input checked="" type="radio"/> YES | NO |
| 13. Were all sample containers received intact? (not broken or leaking, etc.) | | <input checked="" type="radio"/> YES | NO |
| 14. Are all samples requiring no headspace (VOC, GRO, Rx CN/S, radon), headspace free? Size of bubble: _____ < green pea _____ > green pea | <input checked="" type="radio"/> N/A | YES | NO |
| 15. Were samples checked for and free from the presence of residual chlorine? (Applicable when PM has indicated samples are from a chlorinated water source; note if field preservation with sodium thiosulfate was not observed.) | <input checked="" type="radio"/> N/A | YES | NO |
| 16. Were the samples shipped on ice? | | <input checked="" type="radio"/> YES | NO |
| 17. Were cooler temperatures measured at 0.1-6.0°C? IR gun used*: <input checked="" type="radio"/> #2 #4 RAD ONLY | | <input checked="" type="radio"/> YES | NO |
| Cooler #: <u>1</u> | | | |
| Temperature (°C): <u>4.4</u> | | | |
| No. of custody seals on cooler: <u>2</u> | | | |
| External µR/hr reading: <u>14</u> | | | |
| Background µR/hr reading: <u>12</u> | | | |
| Were external µR/hr readings ≤ two times background and within DOT acceptance criteria? <input checked="" type="radio"/> YES / NO / NA (If no, see Form 008.) | | | |

Additional Information: PROVIDE DETAILS BELOW FOR A NO RESPONSE TO ANY QUESTION ABOVE, EXCEPT #1 AND #16.

* possible limited volume for tritium only 25µl received
Volume OK

△ 1 liter Poly for RAD ANALYSIS received at PH 7
added 4ml HNO3 lot #B19075 @ 1000 5:18:06 final ph 1.6
SL

v. sl. sed in & --- include?

If applicable, was the client contacted? YES / NO / NA Contact: _____ Date/Time: _____

Project Manager Signature / Date: [Signature] 5/18/06

*IR Gun #2: Oakton, SN 29922500201-0066

*IR Gun #4: Oakton, SN 2372220101-0002

CONDITION OF SAMPLE UPON RECEIPT FORM

Paragon Analytics

Client: Cordilleran

Workorder No: 06056177-0605177

Project Manager: LS

Initials: Jur Date: 5/24/06

| | | | |
|---|--|--------------------------------------|---|
| 1. Does this project require any special handling in addition to standard Paragon procedures? | | YES | <input checked="" type="radio"/> NO |
| 2. Are custody seals on shipping containers intact? | <input checked="" type="radio"/> NONE | YES | NO |
| 3. Are Custody seals on sample containers intact? | <input checked="" type="radio"/> NONE | YES | NO |
| 4. Is there a COC (Chain-of-Custody) present or other representative documents? | | <input checked="" type="radio"/> YES | NO |
| 5. Are the COC and bottle labels complete and legible? | | <input checked="" type="radio"/> YES | NO |
| 6. Is the COC in agreement with samples received? (IDs, dates, times, no. of samples, no. of containers, matrix, requested analyses, etc.) | | YES | <input checked="" type="radio"/> NO |
| 7. Were airbills / shipping documents present and/or removable? | DROP OFF | <input checked="" type="radio"/> YES | NO |
| 8. Are all aqueous samples requiring preservation preserved correctly? (excluding volatiles) | N/A | YES | <input checked="" type="radio"/> NO |
| 9. Are all aqueous non-preserved samples pH 4-9? | N/A | <input checked="" type="radio"/> YES | NO |
| 10. Is there sufficient sample for the requested analyses? | | <input checked="" type="radio"/> YES | NO |
| 11. Were all samples placed in the proper containers for the requested analyses? | | <input checked="" type="radio"/> YES | NO |
| 12. Are all samples within holding times for the requested analyses? | | <input checked="" type="radio"/> YES | NO |
| 13. Were all sample containers received intact? (not broken or leaking, etc.) | | <input checked="" type="radio"/> YES | NO |
| 14. Are all samples requiring no headspace (VOC, GRO, Rx CN/S, radon), headspace free? Size of bubble: ___ < green pea ___ > green pea | <input checked="" type="radio"/> N/A | YES | NO |
| 15. Were samples checked for and free from the presence of residual chlorine? (Applicable when PM has indicated samples are from a chlorinated water source; note if field preservation with sodium thiosulfate was not observed.) | <input checked="" type="radio"/> N/A | YES | NO |
| 16. Were the samples shipped on ice? | | <input checked="" type="radio"/> YES | NO |
| 17. Were cooler temperatures measured at 0.1-6.0°C? | IR gun used*: #2 <input checked="" type="radio"/> #4 | RAD ONLY | <input checked="" type="radio"/> YES NO |
| Cooler #: <u>1</u> | | | |
| Temperature (°C): <u>1.8</u> | | | |
| No. of custody seals on cooler: <u>0</u> | | | |
| DOT Survey/Acceptance Information | External µR/hr reading: <u>12</u> | | |
| | Background µR/hr reading: <u>12</u> | | |
| Were external µR/hr readings ≤ two times background and within DOT acceptance criteria? <input checked="" type="radio"/> YES / NO / NA (If no, see Form 008.) | | | |

Additional Information: PROVIDE DETAILS BELOW FOR A NO RESPONSE TO ANY QUESTION ABOVE, EXCEPT #1 AND #16.

add on - sample # 2

1. See page 2
2. Bottle sent for tritium labelled as Gamma Preserved bottle sent for Gamma labelled as tritium.

If applicable, was the client contacted? YES / NO / NA Contact: [Signature] Date/Time: _____

Project Manager Signature / Date: [Signature] 5/25/06

*IR Gun #2: Oakton, SN 29922500201-0066

*IR Gun #4: Oakton, SN 2372220101-0002

CONDITION OF SAMPLE UPON RECEIPT FORM

Paragon Analytics

Client: Cardillera

Workorder No: 0605177
~~0504177~~ jmw 5/24/02

Project Manager: LS

Initials: jmw Date: 5/24/02

Additional Information:

N/A

Was the laboratory directed to proceed with the analysis of any samples yielding the presence of residual chlorine? YES / NO / NA

NOTE:

No pH adjustments shall be made without prior consent of Project Manager. After pH adjustments, hold metals and radchem samples ≥ 16 hrs. before analysis.

Was the pH of any sample adjusted by the laboratory? YES (See Table below) / NO

pH Excursion:

| Paragon Sample ID | Client Sample ID | Initial pH | Final pH | Reagent Used | Volume Added (mL) | Lot No. of Reagent | Requested Analysis | Initials / Date / Time |
|-------------------|------------------|------------|----------|------------------|-------------------|--------------------|--------------------|------------------------|
| 0605177-2-2 | BM2642 | 7 | 1.6 | HNO ₃ | 3.0 | B19075 | Gamma | jmw 5/24/02 1050 |
| | | | | | | | | |
| | | | | | | | | |
| | | | | | | | | |
| | | | | | | | | |
| | | | | | | | | |
| | | | | | | | | |
| | | | | | | | | |
| | | | | | | | | |
| | | | | | | | | |
| | | | | | | | | |

If applicable, was the client contacted? YES / NO / NA Contact: _____ Date/Time: _____

Project Manager Signature / Date: LS 5/25

FedEx US Airbill
Express

Tracking Number **8573 0923 5636**

0200

060527
Recipient's Copy

1 From
Date 05/17/06
Sender's Name James Hix Phone 303 557-2072
Company William Compliance Service
Address 500 Marshall Street
City Arvada State CO ZIP 80002

2 Your Internal Billing Reference FD4243

3 To
Recipient's Name James Hix Phone 300 443-1511
Company William Compliance Service
Recipient's Address 225 W. 1st Avenue Drive
City Front Range State CO ZIP 80002



4a Express Package Service
 FedEx Priority Overnight
Next business morning.* Friday shipments will be delivered on Monday unless SATURDAY Delivery is selected.
 FedEx Standard Overnight
Next business afternoon.* Saturday Delivery NOT available.
 FedEx First Overnight
Earliest next business morning.* Saturday Delivery NOT available.
 FedEx 2Day
Second business day.** Thursday shipments will be delivered on Monday unless SATURDAY Delivery is selected.
 FedEx Express Saver
Third business day.* Saturday Delivery NOT available.
*To most locations.

4.9

4b Express Freight Service
 FedEx 1Day Freight
Next business day.** Friday shipments will be delivered on Monday unless SATURDAY Delivery is selected.
 FedEx 2Day Freight
Second business day.** Thursday shipments will be delivered on Monday unless SATURDAY Delivery is selected.
*To most locations.

5 Packaging
 FedEx Envelope*
 FedEx Pak*
Includes FedEx Small Pak, FedEx Large Pak, and FedEx Sturdy Pak.
 FedEx Box
 FedEx Tube
 Other
*Declared value limit \$500.

6 Special Handling
 SATURDAY Delivery
Not available for FedEx Standard Overnight, FedEx First Overnight, FedEx Express Saver, or FedEx 3Day Freight.
 HOLD Weekday at FedEx Location
Not available for FedEx First Overnight.
 HOLD Saturday at FedEx Location
Available ONLY for FedEx Priority Overnight and FedEx 2Day to select locations.
Does this shipment contain dangerous goods?
 No **Yes**
One box must be checked. As per attached Shipper's Declaration. Shipper's Declaration not required.
 Dry Ice
Dry Ice, 5, UN 1845
 Cargo Aircraft Only

7 Payment Bill to:
 Sender (Acct. No. in Section 1 will be billed.)
 Recipient
 Third Party
 Credit Card
 Cash/Check
Obtain Recip. Acct. No.

Total Packages 1 Total Weight 1.92 lbs Total Declared Value 0.00 Total Charges 0.00
*Our liability is limited to \$100 unless you declare a higher value. See back for details.

8 NEW Residential Delivery Signature Options *If you require a signature, check Direct or Indirect.
 No Signature Required
Package may be left without obtaining a signature for delivery.
 Direct Signature
Anyone at recipient's address may sign for delivery. Fee applies.
 Indirect Signature
If no one is available at recipient's address, anyone at a neighboring address may sign for delivery. Fee applies.

520

0605177
~~0506177~~
JUN 5/24/07

ORIGIN ID: GJTA (9) 263-7800
CANDY SMITH
CORDILLERAN COMPLICE SVCS, INC.
826 21 ROAD

Ship Date: 23MAY06
Actual Wgt: 25.0 LB MAN
System#: 144334/CAFE2285
Account: S 235727234

GRAND JUNCTION, CO 1505
UNITED STATES US

TO

(800) 443-1511

PARAGON ANALYTICS
225 COMMERCE RIVE


FedEx
Express

FORT COLLINS, CO 80524



12
1.8

REF: E04243



Delivery Address
Zip Code

BILL SENDER

PRIORITY OVRNIGHT

WED

Deliver By:
24MAY06


TRK# 7290 1771 5077 Form 0201

DEN RA

80524 -C-US

72 FTCA

Part# 156146-434 NRIT 9-05





Paragon Analytics

Radiochemistry Case Narrative

Tritium

Cordilleran Compliance Services, Inc.

Battlement Mesa / E04243

Paragon WO 0605177

1. This report consists of the analytical results and supporting documentation for two water samples received by Paragon on 5/24/06.
2. These samples were prepared according to Paragon Analytics procedure SOP700R9.
3. The samples were analyzed for the presence of tritium according to Paragon Analytics procedure SOP704R7. The analyses were completed on 6/21/06
4. The analysis results for these samples are reported in units of pCi/L. The samples were not filtered prior to analysis.
5. The daily check for the ^{14}C window parameter failed low on 6/10/06, 6/11/06, and 6/12/06. The batch Q.C. is in control for this workorder and the data quality is believed to be unaffected. Please refer to QASS 301252 for further information regarding this incident.
6. There are cases where the magnitude of negative activity is greater than the 2σ TPU. Under typical conditions, where background data is normally distributed and analyzed by paired observations, this event is likely to occur at least 2.5% of the time. Review of the data does not indicate a problem with the instrument or reporting systems and results are reported without further qualification.
7. No anomalous situations were encountered during the preparation or analysis of these samples. All quality control criteria were met.

The data contained in the following report have been reviewed and approved by the personnel listed below. In addition, Paragon Analytics certifies that the analyses reported herein are true, complete and correct within the limits of the methods employed.



Mike Clemmer
Radiochemistry Instrument Technician

6/22/06
Date



Julie Edlyson
Radiochemistry Final Data Review

6/22/06
Date

Paragon Analytics

A Division of DataChem Laboratories, Inc.

000001

PARAGON ANALYTICS, INC.
Radiochemistry Data Package

Section 1

**SAMPLE RESULTS
SUMMARY**

Tritium Analysis By Liquid Scintillation Sample Results Summary

Client Name: Cordilleran Compliance Services, Inc. Laboratory Name: Paragon Analytics Page: 1 of 1
 Client Project Name: Battlement Mesa PAI Work Order: 0605177 Reported on: Thursday, June 22, 2006
 Client Project Number: E04243 10:10:34 AM

| Lab Sample ID | Client Sample ID | Sample Type | Nuclide | Result +/- 2 s TPU | MDC | Units | Matrix | Prep Batch | Date Analyzed | Flags |
|---------------|------------------|-------------|---------|--------------------|-----|-------|--------|------------|---------------|-------|
| 0605177-1 | BM3613 PW | Sample | H-3 | -200 +/- 190 | 330 | pCi/l | WATER | 3H060522-1 | 6/10/2006 | U |
| 0605177-2 | BM2642 | Sample | H-3 | -170 +/- 210 | 350 | pCi/l | WATER | 3H060606-2 | 6/21/2006 | U |

Comments:

Data Package ID: H30605177-1

Qualifiers/Flags:
 U - Result is less than the sample specific MDC
 LT - Result is less than Requested MDC, greater than sample specific MDC.
 Y1 - Chemical Yield is in control at 100-110%. Quantitative Yield is assumed.
 Y2 - Chemical Yield outside default limits
 M - The requested MDC was not met.
 M3 - The requested MDC was not met, but the reported activity is greater than the reported MDC

Abbreviations:
 TPU - Total Propagated Uncertainty (see PAI SOP 743)
 MDC - Minimum Detectable Concentration (see PAI SOP 709)
 BDL - Below Detection Limit

000003

2

PARAGON ANALYTICS, INC.
Radiochemistry Data Package

Section 2

**QC RESULTS
SUMMARY**

000004

Tritium Analysis By Liquid Scintillation

PAI 704 Rev 7

Method Blank Results

Lab Name: Paragon Analytics

Work Order Number: 0605177

Client Name: Cordilleran Compliance Services, Inc.

ClientProject ID: Battlement Mesa E04243

Lab ID: 3H060522-1MB

Sample Matrix: WATER

Prep Batch: 3H060522-1

Final Aliquot: 10.0 ml

Prep SOP: PAI 700 Rev 9

QCBatchID: 3H060522-1-2

Result Units: pCi/l

Date Collected: 22-May-06

Run ID: 3H060522-1A

File Name: Manual Entry

Date Prepared: 22-May-06

Count Time: 100 minutes

Date Analyzed: 10-Jun-06

| CASNO | Target Nuclide | Result +/- 2 s TPU | MDC | Lab Qualifier |
|------------|----------------|--------------------|-----|---------------|
| 10028-17-8 | H-3 | -90 +/- 190 | 330 | U |

Comments:

Qualifiers/Flags:

U - Result is less than the sample specific MDC.

Y1 - Chemical Yield is in control at 100-110%. Quantitative Yield is assumed.

Y2 - Chemical Yield outside default limits.

LT - Result is less than Requested MDC, greater than sample specific MDC.

Abbreviations:

TPU - Total Propagated Uncertainty (see PAI SOP 743)

MDC - Minimum Detectable Concentration (see PAI SOP 709)

BDL - Below Detection Limit

M - Requested MDC not met.

B - Analyte concentration greater than MDC.

B3 - Analyte concentration greater than MDC but less than Requested MDC.

Data Package ID: H30605177-1

Tritium Analysis By Liquid Scintillation

PAI 704 Rev 7

Method Blank Results

Lab Name: Paragon Analytics

Work Order Number: 0605177

Client Name: Cordilleran Compliance Services, Inc.

ClientProject ID: Battlement Mesa E04243

Lab ID: 3H060606-2MB

Sample Matrix: WATER
Prep SOP: PAI 700 Rev 9
Date Collected: 06-Jun-06
Date Prepared: 06-Jun-06
Date Analyzed: 21-Jun-06

Prep Batch: 3H060606-2
QCBatchID: 3H060606-2-1
Run ID: 3H060606-2A
Count Time: 90 minutes

Final Aliquot: 10.0 ml
Result Units: pCi/l
File Name: YU0562001

| CASNO | Target Nuclide | Result +/- 2 s TPU | MDC | Lab Qualifier |
|------------|----------------|--------------------|-----|---------------|
| 10028-17-8 | H-3 | -10 +/- 210 | 350 | U |

Comments:

Qualifiers/Flags:

U - Result is less than the sample specific MDC.
Y1 - Chemical Yield is in control at 100-110% Quantitative Yield is assumed.
Y2 - Chemical Yield outside default limits.
LT - Result is less than Requested MDC, greater than sample specific MDC.

Abbreviations:

TPU - Total Propagated Uncertainty (see PAI SOP 743)
MDC - Minimum Detectable Concentration (see PAI SOP 709)
BDL - Below Detection Limit

M - Requested MDC not met.
B - Analyte concentration greater than MDC.
B3 - Analyte concentration greater than MDC but less than Requested MDC.

Data Package ID: H30605177-1

Tritium Analysis By Liquid Scintillation

PAI 704 Rev 7

Laboratory Control Sample(s)

Lab Name: Paragon Analytics

Work Order Number: 0605177

Client Name: Cordilleran Compliance Services, Inc.

ClientProject ID: Battlement Mesa E04243

Lab ID: 3H060522-1LCS

Sample Matrix: WATER

Prep Batch: 3H060522-1

Final Aliquot: 9.90 ml

Prep SOP: PAI 700 Rev 9

QCBatchID: 3H060522-1-2

Result Units: pCi/l

Date Collected: 22-May-06

Run ID: 3H060522-1A

File Name: Manual Entry

Date Prepared: 22-May-06

Count Time: 93.35 minutes

Date Analyzed: 10-Jun-06

| CASNO | Target Nuclide | Results +/- 2s TPU | MDC | Spike Added | % Rec | Control Limits | Lab Qualifier |
|------------|----------------|--------------------|-----|-------------|-------|----------------|---------------|
| 10028-17-8 | H-3 | 10600 +/- 1600 | 300 | 11000 | 96.4 | 85 - 115 | P |

Comments:

Qualifiers/Flags:

- U - Result is less than the sample specific MDC.
- LT - Result is less than Requested MDC, greater than sample specific MDC.
- Y1 - Chemical Yield is in control at 100-110%. Quantitative Yield is assumed.
- Y2 - Chemical Yield outside default limits.
- L - LCS Recovery below lower control limit.
- H - LCS Recovery above upper control limit.
- P - LCS Recovery within control limits.
- M - The requested MDC was not met
- M3 - The requested MDC was not met, but thereported activity is greater than the reported MDC.

Abbreviations:

- TPU - Total Propagated Uncertainty (see PAI SOP 743)
- MDC - Minimum Detectable Concentration (see PAI SOP 709)

Data Package ID: *H30605177-1*

Tritium Analysis By Liquid Scintillation

PAI 704 Rev 7

Laboratory Control Sample(s)

Lab Name: Paragon Analytics

Work Order Number: 0605177

Client Name: Cordilleran Compliance Services, Inc.

ClientProject ID: Battlement Mesa E04243

Lab ID: 3H060606-2LCS

Sample Matrix: WATER

Prep Batch: 3H060606-2

Final Aliquot: 9.90 ml

Prep SOP: PAI 700 Rev 9

QCBatchID: 3H060606-2-1

Result Units: pCi/l

Date Collected: 06-Jun-06

Run ID: 3H060606-2A

File Name: YU0562001

Date Prepared: 06-Jun-06

Count Time: 90 minutes

Date Analyzed: 21-Jun-06

| CASNO | Target Nuclide | Results +/- 2s TPU | MDC | Spike Added | % Rec | Control Limits | Lab Qualifier |
|------------|----------------|--------------------|-----|-------------|-------|----------------|---------------|
| 10028-17-8 | H-3 | 10300 +/- 1600 | 400 | 11000 | 94.1 | 85 - 115 | P |

Comments:

Qualifiers/Flags:

U - Result is less than the sample specific MDC.

LT - Result is less than Requested MDC, greater than sample specific MDC.

Y1 - Chemical Yield is in control at 100-110%. Quantitative Yield is assumed.

Y2 - Chemical Yield outside default limits.

L - LCS Recovery below lower control limit.

H - LCS Recovery above upper control limit.

P - LCS Recovery within control limits.

M - The requested MDC was not met.

M3 - The requested MDC was not met, but thereported activity is greater than the reported MDC.

Abbreviations:

TPU - Total Propagated Uncertainty (see PAI SOP 743)

MDC - Minimum Detectable Concentration (see PAI SOP 709)

Data Package ID: H30605177-1

PARAGON ANALYTICS, INC.
Radiochemistry Data Package

3

Section 3

**INDIVIDUAL
SAMPLE RESULTS**

000009

Tritium Analysis By Liquid Scintillation

PAI 704 Rev 7

Sample Results

Lab Name: Paragon Analytics

Work Order Number: 0605177

Client Name: Cordilleran Compliance Services, Inc.

ClientProject ID: Battlement Mesa E04243

| | | | |
|---------------------|---------------------------|--------------------------|-------------------------|
| Field ID: BM3613 PW | Sample Matrix: WATER | Prep Batch: 3H060522-1 | Final Aliquot: 10.0 ml |
| Lab ID: 0605177-1 | Prep SOP: PAI 700 Rev 9 | QCBatchID: 3H060522-1-2 | Prep Basis: Unfiltered |
| | Date Collected: 17-May-06 | Run ID: 3H060522-1A | Moisture(%): 100.000 |
| | Date Prepared: 22-May-06 | Count Time: 100 minutes | Result Units: pCi/l |
| | Date Analyzed: 10-Jun-06 | Report Basis: Unfiltered | File Name: Manual Entry |

| CASNO | Target Nuclide | Result +/- 2 s TPU | MDC | Lab Qualifier |
|------------|----------------|--------------------|-----|---------------|
| 10028-17-8 | H-3 | -200 +/- 190 | 330 | U |

Comments:

Qualifiers/Flags:

U - Result is less than the sample specific MDC.

Y1 - Chemical Yield is in control at 100-110%. Quantitative Yield is assumed.

Y2 - Chemical Yield outside default limits.

LT - Result is less than Requested MDC, greater than sample specific MDC.

M3 - The requested MDC was not met, but the reported activity is greater than the reported MDC

M - The requested MDC was not met.

Abbreviations:

TPU - Total Propagated Uncertainty (see PAI SOP 743)

MDC - Minimum Detectable Concentration (see PAI SOP 709)

BDL - Below Detection Limit

Data Package ID: H30605177-1

Tritium Analysis By Liquid Scintillation

PAI 704 Rev 7
Sample Results

Lab Name: Paragon Analytics
Work Order Number: 0605177
Client Name: Cordilleran Compliance Services, Inc.
ClientProject ID: Battlement Mesa E04243

| | | | |
|-------------------|---------------------------|--------------------------|------------------------|
| Field ID: BM2642 | Sample Matrix: WATER | Prep Batch: 3H060606-2 | Final Aliquot: 10.0 ml |
| Lab ID: 0605177-2 | Prep SOP: PAI 700 Rev 9 | QCBatchID: 3H060606-2-1 | Prep Basis: Unfiltered |
| | Date Collected: 20-May-06 | Run ID: 3H060606-2A | Moisture(%): 100.000 |
| | Date Prepared: 06-Jun-06 | Count Time: 90 minutes | Result Units: pCi/l |
| | Date Analyzed: 21-Jun-06 | Report Basis: Unfiltered | File Name: YU0562001 |

| CASNO | Target Nuclide | Result +/- 2 s TPU | MDC | Lab Qualifier |
|------------|----------------|--------------------|-----|---------------|
| 10028-17-8 | H-3 | -170 +/- 210 | 350 | U |

Comments:

Qualifiers/Flags:

U - Result is less than the sample specific MDC.

Y1 - Chemical Yield is in control at 100-110%. Quantitative Yield is assumed.

Y2 - Chemical Yield outside default limits.

LT - Result is less than Requested MDC, greater than sample specific MDC.

M3 - The requested MDC was not met, but the reported activity is greater than the reported MDC.

M - The requested MDC was not met.

Abbreviations:

TPU - Total Propagated Uncertainty (see PAI SOP 743)

MDC - Minimum Detectable Concentration (see PAI SOP 709)

BDL - Below Detection Limit

Data Package ID: H30605177-1

PARAGON ANALYTICS, INC.
Radiochemistry Data Package

Section 4

4

RAW DATA

000012

Tritium Analysis By Liquid Scintillation Raw Data Report

Laboratory Name: Paragon Analytics
 PAI Work Order: 0605177

Prep SOP: PAI 700
 Analytical SOP: PAI 704

Reported on: Wednesday, June 14, 2006
 12:55:35 PM

| Sample ID QC Type | Nuclide Type | Sample Date/Time | Prep Batch QC BatchID | Ingrwth Date/Time | Quench Factor %Lum | Matrix %Moist | Samp Alq Analy Alq | Inst ID Det ID | AnRunID File Name | Count Date/Time | GrossCPM BkgCPM | BaseEff ProgEff | Yield | CntDur(min) | Activity +/- 2 s TPU | MDC DeclEv | ReportUnits ReportBasis | DER RPD | %Spk. Recov Flags |
|----------------------|---------------------|--------------------------|----------------------------|----------------------|-----------------------|------------------|-----------------------|-------------------|-----------------------------|-----------------------|--------------------|--------------------|-------------|-------------|-------------------------|---------------|----------------------------|------------|----------------------|
| 0605177-1 SMP | H-3 Trg. Analyte | 5/17/2006 8:58:00 AM | 3H060522-1 3H060522-1-2 | NA | 140.3 0.1 | WATER 100 | 50 ml 10 ml | L56500 44-06 | 3H060522-1A Manual Entry | 6/10/2006 3:23 AM | 9.850 10.830 | 21.58% NA | 100 NA | 100 | -200 190 | 330 NA | pCi/l Unfiltered | NA NA | NA U |
| 0605177-2 SMP | H-3 Trg. Analyte | 5/20/2006 12:15:00 PM | 3H060606-2 3H060606-2-1 | NA | 143.6 0.22 | WATER 100 | 50 ml 10 ml | L56500 42-02 | 3H060606-2A YU0562001 | 6/21/2006 12:11 AM | 10.380 11.200 | 21.58% NA | 90 NA | 90 | -170 210 | 350 NA | pCi/l Unfiltered | NA NA | NA U |
| 3H060522-1 MB | H-3 Trg. Analyte | 5/22/2006 2:30:19 PM | 3H060522-1 3H060522-1-2 | NA | 144.1 0.06 | WATER NA | 50 ml 10 ml | L56500 44-12 | 3H060522-1A Manual Entry | 6/10/2006 1:16 PM | 10.420 10.830 | 21.58% NA | 100 NA | 100 | -90 190 | 330 NA | pCi/l Unfiltered | NA NA | NA U |
| 3H060522-1 LCS | H-3 Trg. Analyte | 5/22/2006 2:30:19 PM | 3H060522-1 3H060522-1-2 | NA | 143.6 0.04 | WATER NA | 50 ml 9.9 ml | L56500 45-01 | 3H060522-1A Manual Entry | 6/10/2006 2:51 PM | 61.070 10.830 | 21.58% NA | 93.35 NA | 93.35 | 10600 1600 | 300 NA | pCi/l Unfiltered | NA NA | 96.4 P |
| 3H060606-2 MB | H-3 Trg. Analyte | 6/6/2006 8:33:12 AM | 3H060606-2 3H060606-2-1 | NA | 144.1 0.23 | WATER NA | 50 ml 10 ml | L56500 42-08 | 3H060606-2A YU0562001 | 6/21/2006 9:11 AM | 11.170 11.200 | 21.58% NA | 90 NA | 90 | -10 210 | 350 NA | pCi/l Unfiltered | NA NA | NA U |
| 3H060606-2 LCS | H-3 Trg. Analyte | 6/6/2006 8:33:12 AM | 3H060606-2 3H060606-2-1 | NA | 144 0.12 | WATER NA | 50 ml 9.9 ml | L56500 42-09 | 3H060606-2A YU0562001 | 6/21/2006 10:41 AM | 60.170 11.200 | 21.58% NA | 90 NA | 90 | 10300 1600 | 400 NA | pCi/l Unfiltered | NA NA | 94.1 P |

Comments:

Data Package ID: H30605177-1

Qualifiers/Flags:

- U - Result is less than the sample specific MDC.
 - Y1 - Chemical Yield is in control at 100-110%. Quantitative yield is assumed.
 - Y2 - Chemical Yield outside default limits.
 - W - DER is greater than Warning Limit of 1.42
 - D - DER is greater than Control Limit of 2.13
 - + - Duplicate RPD not within limits.
 - LT - Result is less than Request MDC, greater than sample specific MDC
 - * - Aliquot Basis is 'As Received' while the Report Basis is 'Dry Weight'
 - # - Aliquot Basis is 'Dry Weight' while the Report Basis is 'As Received'
- M - Requested MDC not met.
 M3 - The requested MDC was not met, but the reported activity is greater than the reported MDC.
 L - LCS Recovery below lower control limit
 H - LCS Recovery above upper control limit
 P - LCS Matrix Spike Recovery within control limits
 N - Matrix Spike Recovery outside control limits
 NC - Not Calculated for duplicate results less than 5 times MDC
 B - Analyte concentration greater than MDC
 B3 - Analyte concentration greater than MDC but less than Requested MDC
- Notes:**
- 1) The Tracer results are not yield corrected (i.e. activity measured not activity added).
 - 2) Where sample time is not available, 12:00 PM (Mountain) is used for decay correction.
- Abbreviations:**
- TR - Tracer
 - TA - Target Analyte
 - TPU - Total Propagated Uncertainty (see PAI SOP 743)
 - MDC - Minimum Detectable Concentration (see PAI SOP 709)
 - DER - Duplicate Error Ratio
 - BDL - Below Detection Limit

060013

ID: H3- 10 ML GED.

9 JUN 2006 17:23

USER: 8 COMMENT: LS&500

PRESET TIME : 100.00
 DATA CALC : CPM H# : YES SAMPLE REPEATS : 1 PRINTER : STD
 COUNT BLANK : NO IC# : NO REPLICATES : 1 RS232 : OFF
 TWO PHASE : NO ACC : NO CYCLE REPEATS : 1 DISK : EDIT
 SCINTILLATOR: LIQUID LUMEX: NO LOW SAMPLE REJ: 0
 LOW LEVEL : YES HALF LIFE CORRECTION DATE: none

CHAN: 100.0 - 320.0 %ERROR: 2.65 FACTOR: 1.000000 BKG. SUB: 0
 CHAN: 380.0 - 990.0 %ERROR: 20.00 FACTOR: 1.000000 BKG. SUB: 0

| SAM NO | POS | TIME MIN | H# | WIND1 | | WIND2 | | LUMEX % | ELAPSED TIME |
|-----------|-------|-------------|-------|-------|--------|-------|--------|------------|-----------------|
| | | | | CPM | %ERROR | CPM | %ERROR | | |
| 1 | 44-1 | 100.00 | 144.0 | 10.67 | 6.12 | 41.07 | 3.12 | 0.07 | 101.37 |
| 2 | 44-2 | 100.00 | 144.5 | 10.45 | 6.19 | 40.80 | 3.13 | 0.06 | 202.79 |
| 3 | 44-3 | 100.00 | 142.4 | 10.91 | 6.06 | 40.37 | 3.15 | 0.09 | 304.25 |
| 4 | 44-4 | 100.00 | 143.8 | 11.22 | 5.97 | 41.47 | 3.11 | 0.08 | 405.69 |
| 5 | 44-5 | 91.80 | 143.5 | 62.05 | 2.65 | 40.98 | 3.26 | 0.04 | 498.87 |
| 6 | 44-6 | 100.00 | 140.3 | 9.86 | 6.37 | 41.14 | 3.12 | 0.10 | 600.33 |
| 7 | 44-7 | 100.00 | 143.8 | 10.53 | 6.16 | 41.23 | 3.11 | 0.07 | 701.77 |
| 8 | 44-8 | 100.00 | 142.9 | 10.83 | 6.08 | 40.16 | 3.16 | 0.08 | 803.20 |
| 9 | 44-9 | 100.00 | 144.0 | 10.39 | 6.20 | 41.33 | 3.11 | 0.07 | 904.64 |
| 10 | 44-10 | 92.65 | 144.2 | 61.51 | 2.65 | 40.94 | 3.25 | 0.03 | 998.64 |
| 11 | 44-11 | 92.30 | 144.0 | 61.73 | 2.65 | 41.34 | 3.24 | 0.03 | 1092.31 |
| 12 | 44-12 | 100.00 | 144.1 | 10.42 | 6.20 | 42.70 | 3.06 | 0.06 | 1193.75 |
| 13 | 45-1 | 93.35 | 143.6 | 61.07 | 2.65 | 41.97 | 3.20 | 0.04 | 1288.59 |
| 14 | 45-2 | 100.00 | 142.8 | 10.98 | 6.04 | 41.17 | 3.12 | 0.06 | 1390.01 |

000014

ID: H3- 10 ML GEO. 20 JUN 2006 22:41
 USER: 5 COMMENT: LS6500
 PRESET TIME : 90.00
 DATA CALL : CPM HH : YES SAMPLE REPEATS: 1 PRINTER : STD
 COUNT BLANK : NO IC# : NO REPLICATES : 1 PS201 : OFF
 TWO PHASE : NO ACC : NO CYCLE REPEATS : 1 DISK : EDIT
 SCINTILLATOR: LIQUID LUMEX: NO LOW SAMPLE REJ: 0
 LOW LEVEL : YES HALF LIFE CORRECTION DATE: none

CHAN: 100.0 - 320.0 %ERROR: 2.50 FACTOR: 1.000000 BKG. SUB: 0
 CHAN: 380.0 - 990.0 %ERROR: 2.00 FACTOR: 1.000000 BKG. SUB: 0

| SAM NO | POS | TIME MIN | HH | MULTIPL | | MULTIPL2 | | LUMEX % | ELAPSED TIME |
|--------|--------|----------|-------|---------|--------|----------|--------|---------|--------------|
| | | | | CPM | %ERROR | CPM | %ERROR | | |
| • 1 | 42-1 | 90.00 | 175.0 | 11.46 | 6.23 | 40.96 | 3.29 | 0.24 | 91.39 |
| 0.22 | 161.81 | | | | | | | | |

INTERRUPTING TO PRINT INTERRUPT DATA 21 JUN 2006 16:26

FILE = 4U056Z001
 MC 6/22/06

000015

| SAM NO | POS | TIME MIN | HH | WETNED1 | | WETNED2 | | LUMEX % | ELAPSED TIME |
|-----------|-------|-------------|-------|---------|--------|---------|--------|------------|-----------------|
| | | | | CPM | %ERROR | CPM | %ERROR | | |
| 3 | 42-3 | 90.00 | 144.8 | 10.61 | 6.47 | 39.86 | 3.34 | 0.19 | 1105.33 |
| 4 | 42-4 | 90.00 | 145.4 | 10.96 | 6.37 | 41.31 | 3.28 | 0.17 | 1196.74 |
| 5 | 42-5 | 90.00 | 144.2 | 11.02 | 6.35 | 41.18 | 3.29 | 0.19 | 1288.17 |
| 6 | 42-6 | 90.00 | 145.1 | 61.10 | 2.70 | 41.23 | 3.28 | 0.10 | 1379.60 |
| 7 | 42-7 | 90.00 | 144.7 | 58.90 | 2.75 | 40.71 | 3.30 | 0.11 | 1471.02 |
| 8 | 42-8 | 90.00 | 144.1 | 11.17 | 6.31 | 41.09 | 3.29 | 0.23 | 1562.45 |
| 9 | 42-9 | 90.00 | 144.0 | 60.17 | 2.72 | 41.70 | 3.26 | 0.12 | 1653.89 |
| 10 | 42-10 | 90.00 | 144.2 | 11.12 | 6.32 | 41.57 | 3.27 | 0.22 | 1745.33 |

FILE = YU 0562001
MC
6/22/06

U0562001

BSF Version : 3
Instrument Type : LS 6000
Data Capture Date : 20 Jun 2006 22:41:48
User Filename : A:\USER05\U0562001.BSF
USER# : 05
ID : H3- 10 ML GEO.
Comments : LS6500
Preset Count Time : 90.00
Calculation Mode : CPM
Sample Repeats : 1
Printer Output Mode : STD
Blank count : NO
Replicates : 1
RS232 Output Mode : OFF
Two-Phase Selected : NO
Color Correction : NO
Cycle Repeats : 1
Data Buffer Output Mode : EDIT
Scintillator Choice : LIQUID
Lum-Ex Selected : NO
Low Sample Reject Count : 0
Low Level Option : YES
Half Life Corr Date : NONE
Limits Window 1 : 100.00
Preset %Error Iso1 : 2.50
Norm Multiplier Iso1 : 1.00
Background CPM 1 : 0.00
Limits Window 2 : 380.00
Preset %Error Iso2 : 20.00
Norm Multiplier Iso2 : 1.00
Background CPM 2 : 0.00

1,42-01,90,145,11.46,6.23,40.96,3.29,0.24,91.39
2,42-02,90,143.6,10.38,6.54,42.03,3.25,0.22,182.81
3,42-03,90,144.8,10.61,6.47,39.86,3.34,0.19,1105.3
4,42-04,90,145.4,10.96,6.37,41.31,3.28,0.17,1196.7
5,42-05,90,144.2,11.02,6.35,41.18,3.29,0.19,1288.2
6,42-06,90,145.1,61.1,2.7,41.23,3.28,0.1,1379.6
7,42-07,90,144.7,58.9,2.75,40.71,3.3,0.11,1471
8,42-08,90,144.1,11.17,6.31,41.09,3.29,0.23,1562.4
9,42-09,90,144,60.17,2.72,41.7,3.26,0.12,1653.9
10,42-10,90,144.2,11.12,6.32,41.57,3.27,0.22,1745.3

FILE => U0562001
MC 6/22/06

Instrument ID: LS 6500

LSC Run Log

| Date | Sample ID | Count/Time (min.) | Rack & Position | Test | User # | Batch ID | Position Check | Initials | Comments |
|--------|---------------|-------------------|-----------------|---------|--------|------------|----------------|----------|----------|
| 6/9/06 | 31060522-1CB1 | 100 | 44 - 1 | 13-10ML | 5 | 31060522-1 | MC | MC | NA |
| | 0605170-9 | | | | | | | | |
| | 197-1 | | | | | | | | |
| | 198-2 | | | | | | | | |
| | 31060522-1CBZ | | | | | | | | |
| | 0605198-2B | | | | | | | | |
| | 2MS | 92.65 | | | | | | | |
| | 2MSB | 92.3 | | | | | | | |
| | 31060522-1MB | 100 | | | | | | | |
| | 11CS | 93.35 | | | | | | | |
| | 31060522-1CB3 | 100 | | | | | | | |
| 6/10 | DAILY Q.C. | 10 | | | | | | | |
| | 31060601-1CB1 | 10 | | | | | | | |
| | 0604900-1 | | | | | | | | |
| | 2M | | | | | | | | |
| | 4B | | | | | | | | |
| | 5B | | | | | | | | |
| | 6 | | | | | | | | |
| | 7D | | | | | | | | |
| | 8 | | | | | | | | |
| | 9 | | | | | | | | |
| | 10 | | | | | | | | |
| | 11 | | | | | | | | |
| | 10 | | | | | | | | |
| | 11 | | | | | | | | |
| | 31060601-1CBZ | | | | | | | | |
| | 0604900-1Z | | | | | | | | |

306255

Reviewed by / Date MC 6/22/06

000018

Instrument ID: LS 6500

LSC Run Log

| Date | Sample ID | CountTime (min.) | Rack & Position | Test | User # | Batch ID | Position Check | Initials | Comments |
|---------|---------------|------------------|-----------------|---------|--------|------------|----------------|----------|----------|
| 6/19/06 | 34060601-3CB1 | 230 | 51-1 | H3-5mL | 3 | 34060601-3 | MC | MC | N/A |
| | 0605175-1 | 127.1 | -2 | | | | | | |
| | 2 | 8.35 | -3 | | | | | | |
| | 3 | 8.1 | -4 | | | | | | |
| | 176-1 | 230 | -5 | | | | | | |
| | 1-2 | | -6 | | | | | | |
| | 34060601-3CBZ | | -7 | | | | | | |
| | 0605175-2D | | -8 | | | | | | |
| | 0605194-1 | 190.15 | -9 | | | | | | |
| | 2 | 194.85 | -10 | | | | | | |
| | 3 | 161.45 | -11 | | | | | | |
| | 34060601-3MB | 230 | -12 | | | | | | |
| | 0605175-1 | 19.05 | 5763-1 | | | | | | |
| | 1-2 | 230 | 1-2 | | | | | | |
| 6/20/06 | DAILY Q.C. | 10 | 1-1-4 | | 1 | | MC | MC | |
| 6/21/06 | 34060606-ZCB1 | 90 | 42-1 | H3-10mL | 5 | 34060606-2 | | | |
| | 0605177-2 | | -2 | | | | | | |
| | 0605243-1 | | -3 | | | | | | |
| | -1D | | -4 | | | | | | |
| | 34060606-ZCBZ | | -5 | | | | | | |
| | 0605243-1MS | | -6 | | | | | | |
| | 1MSD | | -7 | | | | | | |
| | 34060606-ZMB | | -8 | | | | | | |
| | -2LCS | | -9 | | | | | | |
| | -2CBZ | | -10 | | | | | | |
| 6/22/06 | DAILY Q.C. | 10 | 1-1-4 | | 1 | | MC | MC | |
| | 34060630-1CB1 | | 56-1 | H3-5mL | 9 | 34060630-1 | | | |
| | 0605209-1D | | -2 | | | | | | |
| | 2 | | -3 | | | | | | |
| | | | -4 | | | | | | |

306259

MC 6/22/06

Reviewed by / Date

000019

PARAGON ANALYTICS, INC.
Radiochemistry Data Package

Section 5

**QUALITY ASSURANCE
SUMMARY REPORTS**

5

000020



QUALITY ASSURANCE SUMMARY SHEET

PAI W.O. # / BATCH
TEST ALL LS6500 ANALYSES
METHOD LIO. SCINT.
SOP/REV (PREP)
SOP/REV (ANAL) 7CHR7

Briefly document any QA or other problems or deviations associated with the analysis of samples. Problems could result from: log-in, color, odor, dilution, consistency, scheduling, equipment, or instrumentation, or may include documentation of minor deviations necessary due to unique DQO's or sample characteristics.

BEGINNING ON 6/9/06, THE ¹⁴C COUNT RATE
FAILED DAILY CHECKS. THIS FAILURE
CONTINUED ON 6/10/06, 6/11/06, AND 6/12/06.
THE ³H AND BLANK DAILY CHECK
PARAMETERS WERE IN CONTROL FOR
6/9/06, 6/10/06, 6/11/06, AND 6/12/06. THE
DATA QUALITY FOR ALL ³H DATA
ACQUIRED DURING THIS TIME PERIOD
IS BELIEVED TO BE UNAFFECTED.
THE BECKMAN LS6500 INSTRUMENT
WAS GAIN RE-CALIBRATED ON 6/12/06
AND THE SUBSEQUENT DAILY CHECK
RUN PASSED FOR ALL REQUIRED MC 6/13/06
PARAMETERS EXCEPT
FOR THE ¹⁴C COUNT RATE. RESULTS ARE
SUBMITTED WITHOUT FURTHER
QUALIFICATION. MC 6/12/06

MC
6/13/06

TECHNICIAN/ANALYST  DATE 6/12/06
DEPARTMENT MANAGER  DATE 6/13/06

301252

PARAGON ANALYTICS, INC.
Radiochemistry Data Package

Section 6

**LABORATORY
BENCH SHEETS**

6

000022

Radiochemistry Instrument Worksheet

Paragon Analytics

Prep Batch: 3H060522-1

Prep Procedure: H3

10-ML

RG 6/28/06
NA 301252

| Prep Num | LabID | QC Type | Init Alq | Fin Alq | Units | Cnt 1 File/Inst | Cnt 1 Rack-Pos | Cnt 1 Pos Chk By | Cnt 2 File/Inst | Cnt 2 Rack-Pos | Cnt 2 Pos Chk By | Cnt 3 File/Inst | Cnt 3 Rack-Pos | Cnt 3 Pos Chk By | Notes |
|----------|---------------|---------|----------|----------|-------|-----------------|----------------|------------------|-----------------|----------------|------------------|-----------------|----------------|------------------|---------------------|
| 1 | 0605170-9 | SMP | 50 | 10 | ml | L66500 | 44-2 | MC | L66500 | 44-2 | MC | X | | | C RECOUNT X RECOUNT |
| 1 | 0605170-9 | DUP | 50 | 10 | ml | | -3 | | | | | | | | |
| 1 | 0605170-19 | SMP | 50 | 10 | ml | | -4 | | | | | | | | |
| 1 | 0605170-19 | MS | 50 | 9.900990 | ml | | -5 | | | | | | | | |
| 1 | 0605177-1 | SMP | 50 | 10 | ml | | -6 | | | | | | | | |
| 1 | 0605198-2 | SMP | 50 | 10 | ml | | -7 | | | | | | | | |
| 1 | 0605198-2 | DUP | 50 | 10 | ml | | -9 | | | | | | | | |
| 1 | 0605198-2 | MS | 50 | 9.900990 | ml | | -10 | | | | | | | | |
| 1 | 0605198-2 | MSD | 50 | 9.900990 | ml | | -11 | | | | | | | | |
| 1 | 3H060522-1CB1 | MB | 50 | 10 | ml | | 44-1 | | | | | | | | X RECOUNT |
| 1 | 3H060522-1CB2 | MB | 50 | 10 | ml | | 45-2 | | | | | | | | |
| 1 | 3H060522-1CB3 | MB | 50 | 10 | ml | | 44-12 | | | | | | | | |
| 1 | 3H060522-1 | MB | 50 | 10 | ml | | 45-1 | | | | | | | | |
| 1 | 3H060522-1 | LCS | 50 | 9.900990 | ml | | | | | | | | | | |

| Spike Solution Information | | | | | | |
|----------------------------|---------|-------------|-----------|--------|-----------|---------------|
| Soln # | Nuclide | SolnID | Prep Conc | Units | Prep Date | Aliquot Units |
| S1 | H-3 | 486.3020.23 | 2,446.059 | DPM/ml | 05/22/06 | 0.5 ml |
| | | | | | | ST-002 |

WRONG USER (5-MLU) USED
ML 6/13/06

X USE # OUTSIDE CALIBRATION RANGE FOR SAMPLES. NO SINGLE-POINT CALIBRATION VALID FOR USE. SAMPLES AND Q.C. WILL BE QUENCHED WITH NITROMETHANE AND RE-COUNTED. MC 6/13/06

000023

Radiochemistry Prep Worksheet

Prep Batch: 3H060522-1

Paragon Analytics

Prep Procedure: H3

Reviewed By: KAA *KA* Review Date: 5/23/2006

Non-Routine Pre-Treatment? Y / N Batch: *NA* Re-Prep? Y / N Prep QASS / NCR? Y / N
 Prep Analyst: Kimberly Arnett *KA* Balance: Cocktail: UG LLT
 Prep Date: 5/22/2006 Balance: Cocktail Pipet: T-002
 Matrix Class: liquid Prep Dept: RS Aliquot Pipet: RS-007

| Samp Num | Prep Num | LabID | QC Type | Dish No. | Init Alq ml | Fin Alq ml | Prep Basis | Water Added(ml) | Moisture(%) | Analysis Vol.(ml) | Standards | Prep Notes |
|----------|----------|---------------|---------|----------|-------------|------------|------------|-----------------|-------------|-------------------|-----------|------------|
| 1 | 1 | 0605170-9 | SMP | 50 | 10 | 10 | Unfiltered | 0 | 100 | 10 | | |
| 2 | 1 | 0605170-9 | DUP | 50 | 10 | 10 | Unfiltered | 0 | 100 | 10 | | |
| 3 | 1 | 0605170-19 | SMP | 50 | 10 | 10 | Unfiltered | 0 | 100 | 10 | | |
| 4 | 1 | 0605170-19 | MS | 50 | 9.900990 | 10 | Unfiltered | 0 | 100 | 10 | | |
| 5 | 1 | 0605177-1 | SMP | 50 | 10 | 10 | Unfiltered | 0 | 100 | 10 | | |
| 6 | 1 | 0605198-2 | SMP | 50 | 10 | 10 | Unfiltered | 0 | 100 | 10 | | |
| 7 | 1 | 0605198-2 | DUP | 50 | 10 | 10 | Unfiltered | 0 | 100 | 10 | | |
| 8 | 1 | 0605198-2 | MS | 50 | 9.900990 | 10 | Unfiltered | 0 | 100 | 10 | | |
| 9 | 1 | 0605198-2 | MSD | 50 | 9.900990 | 10 | Unfiltered | 0 | 100 | 10 | | |
| 10 | 1 | 3H060522-1CB1 | MB | 50 | 10 | 10 | Unfiltered | 0 | 100 | 10 | | |
| 11 | 1 | 3H060522-1CB2 | MB | 50 | 10 | 10 | Unfiltered | 0 | 100 | 10 | | |
| 12 | 1 | 3H060522-1CB3 | MB | 50 | 10 | 10 | Unfiltered | 0 | 100 | 10 | | |
| 13 | 1 | 3H060522-1 | MB | 50 | 10 | 10 | Unfiltered | 0 | 100 | 10 | | |
| 14 | 1 | 3H060522-1 | LCS | 50 | 9.900990 | 10 | Unfiltered | 0 | 100 | 10 | | |

Spiked By: Kimberly Arnett Date: 5/23/2006
 Witnessed By: Jeff Kujawa Date: 5/23/2006
 Relinquished By: *KA* Date: 5/23/06
 Received By: *KA* Date: 5/23/06

| Spill Solution Information | | | | | | |
|----------------------------|---------|-------------|-----------|--------|-----------|----------|
| Soln # | Nuclide | SolnID | Prep Conc | Units | Prep Date | Pipet ID |
| S1 | H-3 | 486.3020.23 | 2,446.059 | DPM/ml | 05/22/06 | ST-002 |

000024

Comments

Radiochemistry Prep Worksheet

Prep Batch: 3H060522-1

Paragon Analytics

Prep Procedure: H3

Prep Batch Not Validated!!!

Reviewed By:

Review Date:

Non-Routine Pre-Treatment? Y / (N) NA Batch: NA Re-Prep? Y / (N) NA Prep QASS / NCR? Y / (N) NA

Prep SOP: PAI 700 Rev: 9

Prep Analyst: Kimberly Arnett NA

Balance:

Prep Date: 5/22/06

Cocktail: UGLLT

Matrix Class: liquid

Cocktail Pipet: T-002

Aliquot Pipet: RS-007

| Sampl Num | Prep Num | LabID | QC Type | Dish No. | Init Alq ml | Fin Alq ml | Prep Basis | Water Added(ml) | Moisture(%) | Analysis Vol.(ml) | Standards | Prep Notes |
|--------------|-------------|---------------|------------|-------------|----------------|---------------|------------|-----------------|-------------|-------------------|-----------|------------|
| 1 | 1 | 0605170-9 | SMP | 50 | 10 | 10 | Unfiltered | 0 | 100 | 10 | | |
| 2 | 1 | 0605170-9 | DUP | 50 | 10 | 10 | Unfiltered | 0 | 100 | 10 | | |
| 3 | 1 | 0605170-19 | SMP | 50 | 10 | 10 | Unfiltered | 0 | 100 | 10 | | |
| 4 | 1 | 0605170-19 | MS | 50 | 9.900990 | 9.900990 | Unfiltered | 0 | 100 | 10 | S1 | |
| 5 | 1 | 0605177-1 | SMP | 50 | 10 | 10 | Unfiltered | 0 | 100 | 10 | | |
| 6 | 1 | 0605198-2 | SMP | 50 | 10 | 10 | Unfiltered | 0 | 100 | 10 | | |
| 7 | 1 | 0605198-2 | DUP | 50 | 10 | 10 | Unfiltered | 0 | 100 | 10 | | |
| 8 | 1 | 0605198-2 | MS | 50 | 9.900990 | 9.900990 | Unfiltered | 0 | 100 | 10 | S1 | |
| 9 | 1 | 0605198-2 | MSD | 50 | 9.900990 | 9.900990 | Unfiltered | 0 | 100 | 10 | S1 | |
| 10 | 1 | 3H060522-1CB1 | MB | 50 | 10 | 10 | Unfiltered | 0 | 100 | 10 | | |
| 11 | 1 | 3H060522-1CB2 | MB | 50 | 10 | 10 | Unfiltered | 0 | 100 | 10 | | |
| 12 | 1 | 3H060522-1CB3 | MB | 50 | 10 | 10 | Unfiltered | 0 | 100 | 10 | | |
| 13 | 1 | 3H060522-1 | MB | 50 | 10 | 10 | Unfiltered | 0 | 100 | 10 | | |
| 14 | 1 | 3H060522-1 | MB | 50 | 10 | 10 | Unfiltered | 0 | 100 | 10 | | |
| 15 | 1 | 3H060522-1 | LCS | 50 | 9.900990 | 9.900990 | Unfiltered | 0 | 100 | 10 | S1 | |
| 16 | 1 | 3H060522-1 | LCS | 50 | 9.900990 | 9.900990 | Unfiltered | 0 | 100 | 10 | S1 | |

Spiked By: Kimberly Arnett NA Date: 5/23/06

Witnessed By: John Date: 5/23/06

Relinquished By:

Date:

Received By:

Date:

| Spike Solution Information | | | | | | |
|----------------------------|---------|-------------|-----------|--------|-----------|----------|
| Soln # | Nuclide | SolnID | Prep Conc | Units | Prep Date | Pipet ID |
| S1 | H-3 | 486.3020.23 | 2,446.059 | DPM/ml | 05/22/06 | ST-002 |

5/21
4/15/07

000025

Comments
UG LLT LOT #97-060101

TRITIUM RUN LOG

297942

Logbook/Page No.

SOP 700 Rev 9

Form 1306r1.doc (3/30/2003)

| Date | Workorder/ Sample Number | Column ID | Flask ID | Run No. | Comments | Technician's Initials |
|---------|-----------------------------|--------------|-------------|------------|------------------|--------------------------|
| 5/8/06 | 060425B -25 | A | T11 | 3 | BATCH 3H060505-2 | KAA |
| ↓ | 060502B -11 | B | 605 | ↓ | | |
| 5/9/06 | 0604225 - 3MS | A | T9 | 1 | | |
| ↓ | -3MSD | B | 609 | ↓ | | |
| ↓ | 0604232 -16MS | C | T2 | ↓ | | |
| ↓ | -16MSD | D | 602 | ↓ | | |
| ↓ | 0604256 -8MS | E | 103 | ↓ | | |
| ↓ | -8MSD | F | 600 | ↓ | | |
| 5/8/06 | 3H060505-2 -MB | C | T3 | 3 | | |
| 5/9/06 | ↓ -LCS | A | T1 | 2 | | |
| 5/15/06 | 0604225 -1 | A | 103 | 1 | BATCH 3H060515-1 | KAA |
| ↓ | -1Dup | B | T8 | ↓ | | |
| ↓ | 0605107 -1 | C | 102 | ↓ | | |
| ↓ | -2 | D | T4 | ↓ | | |
| ↓ | -3 | E | 604 | ↓ | | |
| 5/16/06 | ↓ -4 | A | T7 | 2 | | |
| ↓ | -5 | B | 101 | ↓ | | |
| ↓ | -6 | C | 605 | ↓ | | |
| ↓ | -7 | D | T3 | ↓ | | |
| ↓ | 3H060515-1 -MB | E | T11 | ↓ | | |
| ↓ | -LCS | A | 608 | 3 | | |
| ↓ | See 5/16/06 0605097 -1MS | B | T10 | ↓ | | |
| ↓ | -1MSD | C | 606 | ↓ | | |
| 5/22/06 | 0605170 -9 | A | T7 | 1 | BATCH 3H060522-1 | KAA |

Reviewed by / Date KAA 5/22/06

000026

TRITIUM RUN LOG


297943 Logbook/Page No.

SOP 700 Rev 9

Form 1306r1.doc (3/30/2003)

| Date | Workorder/ Sample Number | Column ID | Flask ID | Run No. | Comments | Technician Initials |
|---------|-----------------------------|--------------|-------------|------------|------------------|------------------------|
| 5/22/06 | 0605190 - 9dup | B T4 | T4 | 1 | Batch 3H060522-1 | YAA |
| | ↓ -A | C | 603 | ↓ | | |
| | 0605197 -1 | D | 111 | ↓ | | |
| | 0605198 -2 | E | T8 | ↓ | | |
| | ↓ -2dup | A | 604 | 2 | | |
| | 3H060522-1-M8 | B | 102 | ↓ | | |
| 5/23/06 | 0605190 -1ms | A | 666 | 1 | | |
| | 0605198 -2ms | B | T11 | ↓ | | |
| | ↓ -2msD | C | T6 | ↓ | | |
| | 3H060522-1 LCS | D | 605 | ↓ | | |

YAA
 5/22/06
 YAA
 5/23/06



Reviewed by / Date YAA 5/23/06

000027

SAMPLE CONDITION FORM (LIQUID)

ANALYST: KAA

ANALYSIS DATE: 5/22/06

METHOD: 3H

| WORK ORDER | SAMPLE ID | SAMPLE CONDITION | | |
|------------|-----------|------------------|--------|------------------|
| | | pH | Color | Remarks |
| 0605190 | 9 | 7.0 | clear | MA |
| ↓ | 19 | ↓ | ↓ | ↓ |
| 0605197 | 1 | ↓ | yellow | suspended matter |
| 0605198 | 2 | ↓ | clear | MA |
| | | | | |

Radiochemistry Instrument Worksheet

Paragon Analytics

Prep Batch: 3H060606-2

Prep Procedure: H3

10-ML

Analytical QASS / NCR? Y / (N) N/A

| Prep Num | LabID | QC Type | Init Aliq | Fin Aliq | Units | Cnt 1 File/Inst | Cnt 1 Rack-Pos | Cnt 1 Pos Chk By | Cnt 2 File/Inst | Cnt 2 Rack-Pos | Cnt 2 Pos Chk By | Cnt 3 File/Inst | Cnt 3 Rack-Pos | Cnt 3 Pos Chk By | Notes |
|----------|---------------|---------|-----------|-----------|-------|-----------------|----------------|------------------|-----------------|----------------|------------------|-----------------|----------------|------------------|------------|
| 1 | 0605177-2 | SMP | 50 | 10 | ml | RV062001 | 42-2 | MC | | | | | | | |
| 1 | 0605243-1 | SMP | 50 | 10 | ml | | 3 | | | | | | | | |
| 1 | 0605243-1 | DUP | 50 | 10 | ml | | 4 | | | | | | | | |
| 1 | 0605243-1 | MS | 50 | 9,900,990 | ml | | 6 | | | | | | | | |
| 1 | 0605243-1 | MSD | 50 | 9,900,990 | ml | | 7 | | | | | | | | |
| 1 | 3H060606-2CB1 | MB | 50 | 10 | ml | | 42-1 | | | | | | | | RG 4/28/06 |
| 1 | 3H060606-2CB2 | MB | 50 | 10 | ml | | 5 | | | | | | | | |
| 1 | 3H060606-2CB3 | MB | 50 | 10 | ml | | 10 | | | | | | | | |
| 1 | 3H060606-2 | MB | 50 | 10 | ml | | 8 | | | | | | | | |
| 1 | 3H060606-2 | LCS | 50 | 9,900,990 | ml | | 9 | | | | | | | | |

| Spike Solution Information | | | | | | |
|----------------------------|---------|-------------|-----------|--------|-----------|---------------|
| Soln # | Nuclide | SolnID | Prep Conc | Units | Prep Date | Aliquot Units |
| S1 | H-3 | 486-3020-23 | 2,440.512 | DPM/ml | 06/06/06 | 0.5 ml |

Fraction Barcodes

| | | | |
|-----------------|------------------|-----------------|------------------|
| 0605243-1 | *3H060606-2PS2* | 0605177-2 | *3H060606-2PS1* |
| 0605243-1DUP | *3H060606-2PS3* | 3H060606-2LCS | *3H060606-2PS10* |
| 3H060606-2LCS | *3H060606-2PS10* | 3H060606-2MB | *3H060606-2PS9* |
| 3H060606-2CB1MB | *3H060606-2PS6* | 3H060606-2MB | *3H060606-2PS9* |
| 3H060606-2CB3MB | *3H060606-2PS8* | 3H060606-2CB2MB | *3H060606-2PS7* |
| 0605243-1MS | *3H060606-2PS4* | 0605243-1MSD | *3H060606-2PS5* |

000029

Radiochemistry Prep Worksheet

Paragon Analytics

Prep Batch: 3H060606-2

Prep Procedure: H3

Reviewed By: KAAVAK Review Date: 6/6/06

Non-Routine Pre-Treatment? Y / (N) Batch: MA

Re-Prep? Y / (N) Batch: MA

Prep QASS / NCR? Y / (N) MA

Prep SOP: PAI 700 Rev: 9

Prep Analyst: Kimberly Arnett

Cocktail: UG LLT

Prep Date: 6/6/06

Balance:

Cocktail Pipet: T-002

Prep Dept: RS

Aliquot Pipet: RS-007

Matrix Class: liquid

| Sampl Num | Prep Num | LabID | QC Type | Dish No. | Init Aliq ml | Fin Aliq ml | Prep Basis | Water Added(ml) | Moisture(%) | Analysis Vol.(ml) | Standards | Prep Notes |
|--------------|-------------|---------------|------------|-------------|-----------------|----------------|------------|-----------------|-------------|-------------------|-----------|------------|
| 1 | 1 | 0605177-2 | SMP | | 50 | 10 | Unfiltered | 0 | 100 | 10 | | |
| 2 | 1 | 0605243-1 | SMP | | 50 | 10 | Unfiltered | 0 | 100 | 10 | | |
| 3 | 1 | 0605243-1 | DUP | | 50 | 10 | Unfiltered | 0 | 100 | 10 | | |
| 4 | 1 | 0605243-1 | MS | | 50 | 9,900990 | Unfiltered | 0 | 100 | 10 | S1 | |
| 5 | 1 | 0605243-1 | MSD | | 50 | 9,900990 | Unfiltered | 0 | 100 | 10 | S1 | |
| 6 | 1 | 3H060606-2CB1 | MB | | 50 | 10 | Unfiltered | 0 | 100 | 10 | | |
| 7 | 1 | 3H060606-2CB2 | MB | | 50 | 10 | Unfiltered | 0 | 100 | 10 | | |
| 8 | 1 | 3H060606-2CB3 | MB | | 50 | 10 | Unfiltered | 0 | 100 | 10 | | |
| 9 | 1 | 3H060606-2 | MB | | 50 | 10 | Unfiltered | 0 | 100 | 10 | | |
| 10 | 1 | 3H060606-2 | LCS | | 50 | 9,900990 | Unfiltered | 0 | 100 | 10 | S1 | |

Spiked By: Kimberly Arnett Date: 6/6/06

Witnessed By: Jeff Kujawa Date: 6/6/06

| Spike Solution Information | | | | | |
|----------------------------|---------|-------------|-----------|--------|----------|
| Soln # | Nuclide | SolnID | Prep Conc | Units | Pipet ID |
| S1 | H-3 | 486.3020.23 | 2,440.512 | DPM/ml | 06/06/06 |
| | | | | 0.5 | ml |
| | | | | | ST-002 |

000030

Comments

UG LLT LOT #97-060101

Radiochemistry Prep Worksheet

Prep Batch: 3H060606-2

Paragon Analytics

Prep Batch Not Validated!!!

Prep Procedure: H3

Review Date:

Reviewed By:

Non-Routine Pre-Treatment? Y / (N) Batch: MA Re-Prep? Y / (N) Batch: MA Prep QASS / NCR? Y / (N) MA

Prep SOP: PAI 700 Rev: 9
 Prep SOP: NONE
 Matrix Class: liquid

Prep Analyst: Kimberly Arnett MA
 Prep Date: 6/6/06
 Prep Dept: RS

Balance: MA
 Balance: MA
 Cocktail: UG LLT
 Cocktail Pipet: T-002
 Aliquot Pipet: RS-007

| Samp Num | Prep Num | LabID | QC Type | Dish No. | Init Alq ml | Fin Alq ml | Prep Basis | Water Added(ml) | Moisture(%) | Analysis Vol.(ml) | Standards | Prep Notes |
|----------|----------|---------------|---------|----------|-------------|------------|------------|-----------------|-------------|-------------------|-----------|------------|
| 1 | 1 | 0605177-2 | SMP | 50 | 10 | 10 | Unfiltered | 0 | 100 | 10 | | |
| 2 | 1 | 0605243-1 | SMP | 50 | 10 | 10 | Unfiltered | 0 | 100 | 10 | | |
| 3 | 1 | 0605243-1 | DUP | 50 | 10 | 10 | Unfiltered | 0 | 100 | 10 | | |
| 4 | 1 | 0605243-1 | MS | 50 | 9.900990 | 9.900990 | Unfiltered | 0 | 100 | 10 | S1 | |
| 5 | 1 | 0605243-1 | MSD | 50 | 9.900990 | 9.900990 | Unfiltered | 0 | 100 | 10 | S1 | |
| 6 | 1 | 3H060606-2CB1 | MB | 50 | 10 | 10 | Unfiltered | 0 | 100 | 10 | | |
| 7 | 1 | 3H060606-2CB2 | MB | 50 | 10 | 10 | Unfiltered | 0 | 100 | 10 | | |
| 8 | 1 | 3H060606-2CB3 | MB | 50 | 10 | 10 | Unfiltered | 0 | 100 | 10 | | |
| 9 | 1 | 3H060606-2 | MB | 50 | 10 | 10 | Unfiltered | 0 | 100 | 10 | | |
| 10 | 1 | 3H060606-2 | LCS | 50 | 9.900990 | 9.900990 | Unfiltered | 0 | 100 | 10 | S1 | |

Spiked By: Kimberly Arnett MA Date: 6/6/06
 Witnessed By: [Signature] Date: 6/6/06

| Soln # | Nuclide | SolnID | Prep Conc | Units | Prep Date | Aliquot | Units | Pipet ID |
|--------|---------|-------------|-----------|--------|-----------|---------|-------|----------|
| S1 | H-3 | 486.3020.23 | 2,440.512 | DPW/ml | 06/06/06 | 0.5 | ml | ST-002 |

Est. 4/15/07

000031

Comments: UG LLT LOT #97-060101

297945

TRITIUM RUN LOG

Logbook/Page No.

SOP 700 Rev 9

Form 1306r1.doc (3/30/2003)

| Date | Workorder/ Sample Number | Column ID | Flask ID | Run No. | Comments | Technician's Initials |
|------------|-----------------------------|--------------|-------------|------------|------------------|--------------------------|
| 6/1/06 | 0604900 -24 | C | 101 | 5 | Batch 3H060601-1 | KAA |
| ↓ | 3H060601-1 -MB | D | 111 | ↓ | 3H060601-1 | ↓ |
| ↓ | 3H060601-2 -MB | E | 608 | ↓ | 3H060601-2 | ↓ |
| 6/2/06 | 0604900 -20MS | A | T5 | 1 | 3H060601-1 | ↓ |
| ↓ | ↓ -24MS | B | T1 | ↓ | 3H060601-2 | ↓ |
| ↓ | 3H060601-1 -LCS | C | T9 | ↓ | 3H060601-1 | ↓ |
| ↓ | 3H060601-2 -LCS | D | 103 | ↓ | 3H060601-2 | ↓ |
| 6/6/06 | 0605177 -2 | A | 605 | 1 | Batch 3H060606-2 | KAA |
| ↓ | 0605243 -1 | B | 609 | ↓ | ↓ | ↓ |
| ↓ | ↓ -1Duo | C | T6 | ↓ | ↓ | ↓ |
| ↓ | ↓ -1MS | D | 666 | ↓ | ↓ | ↓ |
| ↓ | ↓ -1MSD | E | T3 | ↓ | ↓ | ↓ |
| ↓ | 3H060606-2 -MB | F | 604 | ↓ | ↓ | ↓ |
| ↓ | ↓ -LCS | A | T4 | 2 | ↓ | ↓ |
| KAA 6/6/06 | | | | | | |

Reviewed by / Date KAA 6/6/06

000032

SAMPLE CONDITION FORM (LIQUID)

ANALYST: *USA*

ANALYSIS DATE: *6/6/06*

METHOD: *3H*

| WORK ORDER | SAMPLE ID | SAMPLE CONDITION | | |
|----------------|--------------|------------------|---------------|-----------|
| | | pH | Color | Remarks |
| <i>0605177</i> | <i>2</i> | <i>7.0</i> | <i>yellow</i> | <i>MA</i> |
| <i>0605243</i> | <i>1</i> | <i>7.0</i> | <i>↓</i> | <i>↓</i> |
| | | | | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |

USA 6/6/06

PARAGON ANALYTICS, INC.
Radiochemistry Data Package

Section 7

**STANDARDS
TRACEABILITY
DOCUMENTS**



000034

PREPARE A WORKING DILUTION OF 486.1284.85 AT ≈ 2500 dpm/ml

1. ^{0.31.02} TRANSFER 486.1284.85 TO AN AMBER GLASS BOTTLE
 MASS OF BOTTLE w/ LID: 259.20g BAL26
 MASS OF BOTTLE + STD: 261.25g
 NET MASS OF STD: 2.05g

2. ^{0.31.02} DILUTE TO 500 ml WITH DI WATER (SEE BELOW FOR ρ DETERMINATION)
 MASS OF BOTTLE + STD + H₂O: 760.0g BAL26
 MASS OF BOTTLE (from above): 259.20g
 NET MASS OF DILUTED STD: 500.8g

3. DETERMINE DENSITY OF DI WATER
 MASS OF 100 ml VOLUMETRIC FLASK: 56.4421g BAL12
 MASS OF FLASK + H₂O: 156.1980g
 NET MASS OF H₂O: 99.7559g
 $\rho = 0.997559$ g/ml

4. FINAL ACTIVITY CALCULATION

$$\left(\frac{127029.4 \text{ pCi/g}}{486.1284.85} \right) \left(\frac{0.9976 \text{ g/ml}}{500.8 \text{ g}} \right) (2.05 \text{ g}) (2.22 \text{ dpm/g}) = 3871.3 \text{ dpm/ml}$$

Std ID: 486.3020.23

Description: H-3

Expiration: 4/15/07
 Activity: 3871.30 dpm/mL

2s Uncertainty: 193.57 dpm/mL

Ref. Date: 3/23/98

Ref Time: N/A

Prep Date: 3/24/06 Prep by: AF


Matrix/Comp: DI H₂O

Half Life (y): 1.24E+01


| Reverification Log | | |
|--------------------|----------|-----------------|
| Analysis Date | Initials | Expiration Date |
| | | |
| | | |
| | | |
| | | |
| | | |
| | | |

Continued on Page

Read and Understood By



3/24/06



4/17/06

Signature

Date

Signature

Date

Tritium Spike Solution 486 1284.85

soln # 486 1284.84 was diluted to form a soln approx 100,000 dpm/mL as follows:

VOA tare 27.9711g

VOA + HS (486 1284.84) 42.6011g

Final (dil w/ DI H₂O) 62.7170g

dil factor

$$\begin{array}{r}
 42.6011 \text{ g} \\
 - 27.9711 \text{ g} \\
 \hline
 14.63 \text{ g STD} \\
 \times 10.14184 \text{ pCi/g} \\
 \hline
 148.37521 \text{ pCi} \\
 \div (62.7170 \text{ g} - 27.9711 \text{ g}) \\
 \hline
 427.029.4 \text{ pCi/g}
 \end{array}$$

RG 8/17/04
Std ID: 486.1284.85

Description: 3H
 Expiration: 7/30/05
 Activity: 948004.38 dpm/mL
 2s Uncertainty: 4740.02 dpm/mL
 Ref. Date: 3/23/98
 Ref Time: N/A
 Prep Date: 7/29/04
 Matrix/Comp. DI H₂O
 Half Life (y): 1.23E+01
 Prep by: JLK

Verified 5/27/98 JUS

Standard re-verified 3/25/03. RG 3/28/03.
Expires 3/25/04.

Standard re-verified. New exp date: 7/30/05 RG 8/17/04

see copy

5/3/00
5/4/00

DATE

Read and Understood By

JUS 5/22/98

DATE

RG 9/16/05
 Std ID: 486.1284.85
 Description: 3H
 Expiration: 8/11/06
 Activity: 948004.38 dpm/mL
 2s Uncertainty: 47400.22 dpm/mL
 Ref. Date: 3/23/98
 Ref Time: N/A
 Prep Date: 5/3/00
 Matrix/Comp. DI Water
 Half Life (y): 1.23E+01
 rep by: RTS

Std ID: 486.1284.85

RG 3/28/03
 Description: H-3
 Activity: 948004.38 dpm/ml
 2s Uncertainty: 4740.02 dpm/ml
 Ref. Date: 3/23/98
 Ref Time: na
 Prep Date: 5/3/00
 Expiration: 3/23/03 3/25/04
 Matrix/Comp. DI H₂O
 Half Life (y): 1.23E+01
 Prep by: RTS
 RG 3/28/03

MAY REQUIRE NCR FOR ICPT WORK.

Std ID: 486.1284.85

RG 8/17/04
 Description: Tritium in Water Stock Solution
 Activity: 427029 pCi/g or mL
 Uncertainty: pCi/g or mL
 Ref. Date: 3/30/98 3/23/98 1-7-02
 Ref Time: 10:00 AM 3/30/98 CEV 10/4/02
 Prep Date: 3/30/98 Prep by: RTS
 Expiration: 3/29/03 3/23/98 CEV 10/4/02
 Matrix/Comp. DI Water 3/29/03 CEV 10/4/02
 Short Lived - Decay Correct with each use

PAI solution number 486.1284.84 (Analytical SIR: 55543-827)

was transferred to a 40 mL TGA vial.

Certificate of Purity not available on 3/30/98 - will obtain tomorrow.

Activity 51.054 uCi
Volume 50.34 grams
Matrix H₂O

RG 8/13/04
Std ID: 486.1284.84

Std ID: 486.1284.84

Description: 3H
Expiration: 7/30/05
Activity: 1014184.00 pCi/g or ml
2s Uncertainty: 50709.00 pCi/g or ml
Ref. Date: 3/23/98
Ref Time: N/A
Prep Date: 7/29/04 Prep by: JLK
Matrix/Comp. DI Water
Half Life (y): 1.23E+01

Description: H-3
Activity: 1014184.00 pCi/g or ml
2s Uncertainty: 50709.00 pCi/g or ml
Ref. Date: 3/23/98
Ref Time: 0.417
Prep Date: 3/30/98 Prep by: RTS
Expiration: 7/3/04
Matrix/Comp. DI Water
Half Life (y): 1.24E+01

RG 8/13/04

RG 8/13/04

RG 8/13/04

Standard re-verified on 7/30/04. Expires 7/30/05.
RG 8/13/04

Std ID: 486.1284.84

Description: Tritium in Water Stock Solution
Activity: 1014184 pCi/g or mL
Uncertainty: pCi/g or mL
Ref. Date: ~~3/23/98~~ 3/23/98 7-7-02
Ref Time: 10:00
Prep Date 3/30/98 Prep by: RTS
Expiration 3/29/03
Matrix/Comp. DI Water
Short Lived - Decay Correct with each use

Verified 5/21/98
JCS

Standard re-verified on 7/3/04
Expires 7/3/04 7/7/03

Continued on Page

[Signature]
Signed Date 5/5/00

Read and Understood By
[Signature]
Signed Date 5/23/98

[Signature]
Signed Date 5/23/98



ANALYTICS

1380 Seaboard Industrial Blvd.
Atlanta, Georgia 30318 · U.S.A.

Phone (404) 352-8677
Fax (404) 352-2837

*PAID 486.
rec'd 3-24-98*

CERTIFICATE OF CALIBRATION
Standard Radionuclide Source

55543-307

H-3 in Water 50 mL in Flame Sealed Vial

This standard radionuclide source was prepared using an aliquot measured gravimetrically from a master radionuclide solution standard. The master radionuclide solution standard was calibrated by the Department Des Applications Et De La Metrologie Des Rayonnements Ionisants (DAMRI), Paris, France, as Number 24057. The calibration was checked by liquid scintillation counting after source preparation.

ANALYTICS maintains traceability to the National Institute of Standards and Technology through Measurements Assurance Programs as described in USNRC Reg. Guide 4.15, Revision 1.

| | |
|--------------------|---------------------------|
| ISOTOPE: | H-3 |
| ACTIVITY (dps): | 1.889 E+06 |
| CALIBRATION DATE: | March 23, 1998 12:00 EST. |
| HALF-LIFE: | 12.43 years |
| TOTAL UNCERTAINTY: | 5.0% |

50.34 grams of water.

P O NUMBER 21143, Item 1

PREPARED BY: *M. D. Currie*
M. D. Currie, Radiochemist

Q A APPROVED: *LC [signature] 3/23/98*

000038

PARAGON ANALYTICS, INC.
Radiochemistry Data Package

Section 8

CHAIN OF CUSTODY

8

000039

Paragon Analytics

Sample Number(s) Cross-Reference Table

Paragon OrderNum: 0605177

Client Name: Cordilleran Compliance Services, Inc.

Client Project Name: Battlement Mesa

Client Project Number: E04243

Client PO Number:

| Client Sample | Lab Sample Number | COC Number | Matrix | Date Collected | Time Collected |
|---------------|-------------------|------------|--------|----------------|----------------|
| BM3613 PW | 0605177-1 | | WATER | 17-May-06 | 8:58 |
| BM2642 | 0605177-2 | | WATER | 20-May-06 | 12:15 |



Paragon Analytics

A Division of DataChem Laboratories, Inc.

225 Commerce Drive Fort Collins, CO 80524
800-443-1511 or (970) 490-1511 (970) 490-1522 Fax

Project Name / No.: PR25CC/E042473 Sampler(s): _____

Report To: James Hix
Phone: 303.237.2072
Fax: 303.237.2459
E-mail: James.hix@datacomp.com
Company: Cardilleran Compliance Services
Address: 5550 Marshall Street
Arvada, CO 80002

Circle method (right); provide additional information as needed (comments).

| Sample ID | Date | Time | Lab ID | Matrix | Preservative | (Indicate type) | No. of Containers |
|-----------|---------|-------|--------|--------|--------------|-----------------|-------------------|
| BM2642 | 5/23/06 | 12:15 | 2 | | | | 2 |

Chain-of-Custody Date _____ Page _____ of _____

Standard or Rush (Due) _____ or Return to Client

| Method | Sample ID | Result |
|-----------------------------|-------------------------------------|--------|
| VOCs | SW8260B E624 E524 2 | |
| BTEX (only) | SW8021B | |
| SVOCs | SW8270C E625 | |
| OC Pesticides | SW8081A E608 | |
| PCBs | SW80802 E608 | |
| Herbicides | SW8151A | |
| Explosives | SW8030 | |
| TCLP Organics SW1311 | SW8260B E270C B081A B151A | |
| TCLP Metals SW1311 | SW6010B SW271 | |
| Total Metals by ICP | SW6010B 7470 7471 E200 ILMO | |
| Dissolved Metals by ICP | SW6010B 7470 E200 ILMO | |
| Total Metals by ICP/MMS | SW6020 E200.8 ILMO | |
| Dissolved Metals by ICP/MMS | SW6020 E200.8 ILMO | |
| Hexavalent Chromium | SW7198A Alkaline Digest? Y / N | |
| Inorganic Anions | SW9056 E300.0 (Specify in comments) | |
| Solids: | | |
| TPH | GRO BRO SW8015B (both) | |
| pH | SW9040B SW9045C | |
| Gross Alpha / Beta | SW9310 E900.0 | |
| Actinides by Peragon SOP | Pu / U / Am / Th / Cm / | |
| Tritium | E906.0 | |
| Total Alpha-Emitting Radium | SW9315 E903.0 | |
| Radium 226 E903.1 | Radium 226 SW9320 E904.0 | |
| Strontium 90 | SW811-00 | |
| Gamma Isotopes | E901.1 | |

Time Zone: EST CST MST PST Matrix Key: O = oil, S = soil NS = non-soil solid, W = water, L = liquid, E = extract, F = filter
Comments:

Relinquished By: _____ (1)
Signature _____
Printed Name _____
Date 5/23/06 Time 16:50
Company Cardilleran

Relinquished By: _____ (2)
Signature _____
Printed Name _____
Date _____ Time _____
Company _____

Received By: _____ (1)
Signature _____
Printed Name J. James
Date 5/24/06 Time 09:25
Company Paragon

Received By: _____ (2)
Signature _____
Printed Name _____
Date _____ Time _____
Company _____

0605177
0506277 pm 5/24/06



| Project Name / No.: | Sampler(s): | Turnaround (circle one): | Standard or Rush (Due _____): | Dispose: Date _____ | or Return to Client _____ | | | | | |
|--|-------------|--------------------------|-------------------------------|---------------------|---------------------------------|-------------------|---|--|--|--|
| Report To: James Hix Phone: 303.237.2072 Fax: 303.237.2659 E-mail: jameshix@cordcomp.com Company: Cordilleran Compliance Services Address: 5550 Marshall Street Arvada, CO 80002 | | | | | | | | | | |
| Sample ID | Date | Time * | Lab ID | Matrix | Preservative (indicate type) | No. of Containers | Circle method (right); provide additional information as needed (comments). | | | |
| BM3613 PW | 5/17/06 | 858 | 1 | W | None | 2 | | | | |
| * Time Zone: EST CST MST PST Matrix Key: O = oil, S = soil, NS = non-soil solid, W = water, L = liquid, E = extract F = filter Comments: | | | | | | | | | | |
| Relinquished By: Signature <u>James Hix</u> (1) Relinquished By: Signature _____ (2) Printed Name <u>James Hix</u> Printed Name _____ Date <u>05/12/06</u> Time _____ Date _____ Time _____ Company <u>Cordilleran Compliance</u> Company _____ | | | | | | | | | | |
| Received By: Signature <u>Sheri LaHart</u> (1) Received By: Signature _____ (2) Printed Name <u>Sheri LaHart</u> Printed Name _____ Date <u>5-18-06</u> Time <u>0925</u> Date _____ Time _____ Company <u>paragon</u> Company _____ | | | | | | | | | | |

CONDITION OF SAMPLE UPON RECEIPT FORM

Paragon Analytics

Client: Cardi Heran

Workorder No: 0605177

Project Manager: _____

Initials: _____ Date: _____

| | | | |
|---|--|--------------------------------------|---|
| 1. Does this project require any special handling in addition to standard Paragon procedures? | | YES | <input checked="" type="radio"/> NO |
| 2. Are custody seals on shipping containers intact? | NONE | <input checked="" type="radio"/> YES | NO |
| 3. Are Custody seals on sample containers intact? | <input checked="" type="radio"/> NONE | YES | NO |
| 4. Is there a COC (Chain-of-Custody) present or other representative documents? | | <input checked="" type="radio"/> YES | NO |
| 5. Are the COC and bottle labels complete and legible? | | <input checked="" type="radio"/> YES | NO |
| 6. Is the COC in agreement with samples received? (IDs, dates, times, no. of samples, no. of containers, matrix, requested analyses, etc.) | | <input checked="" type="radio"/> YES | NO |
| 7. Were airbills / shipping documents present and/or removable? | DROP OFF | <input checked="" type="radio"/> YES | NO |
| 8. Are all aqueous samples requiring preservation preserved correctly? (excluding volatiles) | N/A | YES | <input checked="" type="radio"/> NO |
| 9. Are all aqueous non-preserved samples pH 4-9? | N/A | <input checked="" type="radio"/> YES | NO |
| 10. Is there sufficient sample for the requested analyses? | | YES | <input checked="" type="radio"/> NO |
| 11. Were all samples placed in the proper containers for the requested analyses? | | <input checked="" type="radio"/> YES | NO |
| 12. Are all samples within holding times for the requested analyses? | | <input checked="" type="radio"/> YES | NO |
| 13. Were all sample containers received intact? (not broken or leaking, etc.) | | <input checked="" type="radio"/> YES | NO |
| 14. Are all samples requiring no headspace (VOC, GRO, Rx CN/S, radon), headspace free? Size of bubble: _____ < green pea _____ > green pea | <input checked="" type="radio"/> N/A | YES | NO |
| 15. Were samples checked for and free from the presence of residual chlorine? (Applicable when PM has indicated samples are from a chlorinated water source; note if field preservation with sodium thiosulfate was not observed.) | <input checked="" type="radio"/> N/A | YES | NO |
| 16. Were the samples shipped on ice? | | <input checked="" type="radio"/> YES | NO |
| 17. Were cooler temperatures measured at 0.1-6.0°C? | IR gun used*: <input checked="" type="radio"/> #2 #4 | RAD ONLY | <input checked="" type="radio"/> YES NO |
| Cooler #: <u>1</u> | | | |
| Temperature (°C): <u>4.4</u> | | | |
| No. of custody seals on cooler: <u>2</u> | | | |
| External µR/hr reading: <u>14</u> | | | |
| Background µR/hr reading: <u>12</u> | | | |
| Were external µR/hr readings ≤ two times background and within DOT acceptance criteria? <input checked="" type="radio"/> YES / NO / NA (If no, see Form 008.) | | | |

Additional Information: PROVIDE DETAILS BELOW FOR A NO RESPONSE TO ANY QUESTION ABOVE, EXCEPT #1 AND #16.

* possible limited volume for tritium only 125µl received
Volume OK ✓

△ 1 liter Poly for RAD ANALYSIS received at PH 7
added 4ml HNO3 lot #B19075 @ 1000 5.18.06 final ph 1.6
SL

v. sl. Sed. m & --- include?

If applicable, was the client contacted? YES / NO / NA Contact: _____ Date/Time: _____

Project Manager Signature / Date: [Signature] 5/18/06

*IR Gun #2: Oakton, SN 29922500201-0066

*IR Gun #4: Oakton, SN 2372220101-0002

FedEx US Airbill
Express

FedEx Tracking Number: 8573 0923 5636

1 From
Date: 2/17/06
Sender's Name: James Hix Phone: 703-97-8072
Company: J. H. Hix Properties
Address: 3800 Marshall Street
City: Alameda State: CA ZIP: 94002

2. Your Internal Billing Reference E04293

3 To
Recipient's Name: Sample Recipient Phone: 800 443-1511 / 703-490-1511
Company: Precision Analytics
Recipient's Address: 225 Commerce Drive
City: Fort Collins State: CO ZIP: 98502



8573 0923 5636

Recipient's Copy

4a Express Package Service
 FedEx Priority Overnight
 FedEx Standard Overnight
 FedEx First Overnight
 FedEx 2Day
 FedEx Express Saver
 FedEx 1Day Freight
 FedEx 2Day Freight
 FedEx 3Day Freight

4b Express Freight Service
 FedEx 1Day Freight
 FedEx 2Day Freight
 FedEx 3Day Freight

5 Packaging
 FedEx Envelope
 FedEx Pak
 FedEx Box
 FedEx Tube
 Other

6 Special Handling
 SATURDAY Delivery
 HOLD Weekday at FedEx Location
 HOLD Saturday at FedEx Location
Do this shipment contain dangerous goods?
 No
 Yes
 Dry Ice
 Cargo Aircraft Only

7 Payment
Sender's Acct. No. Recipient Third Party Credit Card Cash/Check
Total Packages: 1 Total Weight: 1.02 lbs Declared Value: \$0.00
NEW Resident Delivery Signature Options
 No Signature Required
 Direct Signature
 Indirect Signature

000044

CONDITION OF SAMPLE UPON RECEIPT FORM

Paragon Analytics

Client: Cordilleran

Workorder No: 06056177-0605177

Project Manager: LS

Initials: JWZ Date: 5/24/06

| | | | |
|---|--|--------------------------------------|--------------------------------------|
| 1. Does this project require any special handling in addition to standard Paragon procedures? | | YES | <input checked="" type="radio"/> NO |
| 2. Are custody seals on shipping containers intact? | <input checked="" type="radio"/> NONE | YES | NO |
| 3. Are Custody seals on sample containers intact? | <input checked="" type="radio"/> NONE | YES | NO |
| 4. Is there a COC (Chain-of-Custody) present or other representative documents? | | <input checked="" type="radio"/> YES | NO |
| 5. Are the COC and bottle labels complete and legible? | | <input checked="" type="radio"/> YES | NO |
| 6. Is the COC in agreement with samples received? (IDs, dates, times, no. of samples, no. of containers, matrix, requested analyses, etc.) | | YES | <input checked="" type="radio"/> NO |
| 7. Were airbills / shipping documents present and/or removable? | DROP OFF: | <input checked="" type="radio"/> YES | NO |
| 8. Are all aqueous samples requiring preservation preserved correctly? (excluding volatiles) | N/A | YES | <input checked="" type="radio"/> NO |
| 9. Are all aqueous non-preserved samples pH 4-9? | N/A | <input checked="" type="radio"/> YES | NO |
| 10. Is there sufficient sample for the requested analyses? | | <input checked="" type="radio"/> YES | NO |
| 11. Were all samples placed in the proper containers for the requested analyses? | | <input checked="" type="radio"/> YES | NO |
| 12. Are all samples within holding times for the requested analyses? | | <input checked="" type="radio"/> YES | NO |
| 13. Were all sample containers received intact? (not broken or leaking, etc.) | | <input checked="" type="radio"/> YES | NO |
| 14. Are all samples requiring no headspace (VOC, GRO, Rx CN/S, radon), headspace free? Size of bubble: ___ < green pea ___ > green pea | <input checked="" type="radio"/> N/A | YES | NO |
| 15. Were samples checked for and free from the presence of residual chlorine? (Applicable when PM has indicated samples are from a chlorinated water source; note if field preservation with sodium thiosulfate was not observed.) | <input checked="" type="radio"/> N/A | YES | NO |
| 16. Were the samples shipped on ice? | | <input checked="" type="radio"/> YES | NO |
| 17. Were cooler temperatures measured at 0.1-6.0°C? | IR gun used*: #2 <input checked="" type="radio"/> #4 | RAD ONLY | <input checked="" type="radio"/> YES |
| Cooler #: <u>1</u> | | | |
| Temperature (°C): <u>1.8</u> | | | |
| No. of custody seals on cooler: <u>0</u> | | | |
| External µR/hr reading: <u>12</u> | | | |
| Background µR/hr reading: <u>12</u> | | | |
| Were external µR/hr readings ≤ two times background and within DOT acceptance criteria? <input checked="" type="radio"/> YES / NO / NA (If no, see Form 008.) | | | |

Additional Information: PROVIDE DETAILS BELOW FOR A NO RESPONSE TO ANY QUESTION ABOVE, EXCEPT #1 AND #16.

add on - sample # 2

1. See page 2

2. Bottle sent for tritium labelled as Gamma
Preserved bottle sent for Gamma labelled as tritium.

If applicable, was the client contacted? YES / NO / NA Contact: _____ Date/Time: _____

Project Manager Signature / Date: [Signature] 5/25/06

*IR Gun #2: Oakton, SN 29922500201-0066

*IR Gun #4: Oakton, SN 2372220101-0002

0605177
0506177
JUL
5/24/06

ORIGIN ID: GJTA (9) 263-7800
CANDY SMITH
CORDILLERAN COMPLIANCE SVCS. INC.
826 21 ROAD

Ship Date: 23MAY06
Actual Wgt: 25.0 LB MAN
System#: 144334/CAFE2285
Account 5 235727234

GRAND JUNCTION, CO 1505
UNITED STATES US

TO (BDD) 443-1511

PARAGON ANALYTICS
225 COMMERCE RIVE


FedEx
Express

FORT COLLINS, CO 80524

12
1.8



REF EB4243



Delivery Address
Air Code

BILL SENDER

PRIORITY OVRNIGHT

WED
Deliver By:

TRK# 7290 1771 5077 Form 0201

24MAY06

DEN AR

:80524 -C-LS

72 FTCA

Part # 156146-434 NRIT 9-05



000047

PARAGON ANALYTICS
Radiochemistry Data Package

Section 9

**ADDITIONAL
SUPPORTING
DOCUMENTATION**



000048

Liquid Scintillation Counter

Instrument Calibration

Initial Efficiency Calibration

Standards Traceability

LS6500 10mL geometry Tritium Background Determination

As noted in QASS 288299, the background count rate is determined from the average of the reagent blanks (not the method blank) from each batch.

Control limits for reagent blanks are established from 30 individual historical data points (10 batches). Limits are +/- 3 standard deviation from 30 individual historical data points. Individual reagent blanks and the average of reagent blanks from each batch are in control. Count Rate (CPM) is within the established control limits.

| Updated on 6/13/2006. MBC 6/13/06. | | | | | | | | | | |
|------------------------------------|--------------------|------------------|------------|-------|---------------------------|-------|--------|---------------------------|-------|--------|
| Sample ID | Count Duration (m) | Count Rate (CPM) | Total Cts. | Mean | Individual Reagent Blanks | | | Average of Reagent Blanks | | |
| | | | | | LCL | UCL | Pass ? | LCL | UCL | Pass ? |
| 3H060522-1CB1 | 100 | 10.67 | 1067.00 | | 10.06 | 12.64 | PASS | | | |
| 3H060522-1CB2 | 100 | 10.83 | 1083.00 | | 10.06 | 12.64 | PASS | | | |
| 3H060522-1CB3 | 100 | 10.98 | 1098.00 | 10.83 | 10.06 | 12.64 | PASS | 10.06 | 12.64 | PASS |

000050

QUALITY ASSURANCE SUMMARY SHEET

288299

PAI W.O. # / BATCH 411 3H - 10ml gen
 TEST 3H
 METHOD LSC
 SOP/REV (PREP) -
 SOP/REV (ANAL) was 5/17/05
7042 70427

Briefly document any QA or other problems or deviations associated with the analysis of samples. Problems could result from: log-in, color, odor, dilution, consistency, scheduling, equipment, or instrumentation, or may include documentation of minor deviations necessary due to unique DQO's or sample characteristics.

Per technical manager's instructions, the count rate from the reagent blanks (not the method blank) for the batch will be used for background determination. Data quality is not believed to be affected.

Note - this is a permanent procedural change that will be reflected in the next revision of SOP 704.

~~_____

_____~~

TECHNICIAN/ANALYST Leah Palko
 DEPARTMENT MANAGER Daniel C. Ben

DATE 5/17/05
 DATE 5/17/05

Tritium "Window 2" Control Limits (LS 6500)

The background count rate is determined from the average of the reagent blanks for the batch.

Window 2 control limits are established using the average count rate from the three reagent blanks associated with each prep batch +/- 3X the estimated poisson uncertainty.

Updated 6/13/2006 MBC

| Sample ID | Count Duration (min) | Average count Duration (min.) | Count Rate (CPM) | Batch Average Reagent Blank | Lower Control Limit | Upper Control Limit | PASS / FAIL |
|---------------|----------------------|-------------------------------|------------------|-----------------------------|---------------------|---------------------|-------------|
| 3H060522-1CB1 | 100 | | 41.07 | | | | |
| 3H060522-1CB2 | 100 | | 40.16 | | | | |
| 3H060522-1CB3 | 100 | 100 | 41.17 | 40.80 | 38.88 | 42.72 | PASS |

000052

LS6500 10mL geometry Tritium Background Determination

As noted in QASS 288299, the background count rate is determined from the average of the reagent blanks (not the method blank) from each batch.

Control limits for reagent blanks are established from 30 individual historical data points (10 batches). Limits are +/- 3 standard deviation from 30 individual historical data points. Individual reagent blanks and the average of reagent blanks from each batch are in control. Count Rate (CPM) is within the established control limits.

Updated on 6/13/2006. MBC 6/13/06.

| Sample ID | Count | | Total Cts. | Mean | Individual Reagent Blanks | | | Average of Reagent Blanks | | |
|---------------|--------------|------------|------------|-------|---------------------------|-------|--------|---------------------------|-------|--------|
| | Duration (m) | Rate (CPM) | | | LCL | UCL | Pass ? | LCL | UCL | Pass ? |
| 3H060606-2CB1 | 90 | 11.46 | 1031.40 | | 10.06 | 12.64 | PASS | | | |
| 3H060606-2CB2 | 90 | 11.02 | 991.80 | | 10.06 | 12.64 | PASS | | | |
| 3H060606-2CB3 | 90 | 11.12 | 1000.80 | 11.20 | 10.06 | 12.64 | PASS | 10.06 | 12.64 | PASS |

000053

Tritium "Window 2" Control Limits (LS 6500)

The background count rate is determined from the average of the reagent blanks for the batch.

Window 2 control limits are established using the average count rate from the three reagent blanks associated with each prep batch +/- 3X the estimated poisson uncertainty.

Updated 6/13/2006 MBC

| Sample ID | Count Duration (min) | Average count Duration (min.) | Count Rate (CPM) | Batch Average Reagent Blank | Lower Control Limit | Upper Control Limit |
|---------------|----------------------|-------------------------------|------------------|-----------------------------|---------------------|---------------------|
| 3H060606-2CB1 | 90 | | 40.96 | | | |
| 3H060606-2CB2 | 90 | | 41.18 | | | |
| 3H060606-2CB3 | 90 | 90 | 41.57 | 41.24 | 39.21 | 43.27 |

000054

Tritium Efficiency Calibration (10ml geometry)

10ml sample + 10ml Ultima Gold LLT

2/12/2006 Count Time (Min.) = 60

Beckman LS 6500

Standard used: 699.2613.19

94000.71 dpm/ml as of 9/3/1998
 1/2 life = 12.23 yrs.
 Current Activity = 61645.12 dpm/ml
 Volume = 0.20 ml
 Average Spike Activity = 12329.02 dpm

| Sample ID | WIND1 cpm | WIND2 cpm | %LUMEX | H# |
|-----------------|-----------|-----------|--------|----------------|
| 0516025-1 | 13.78 | 47.48 | 0.04 | 140.9 |
| 0516025-2 | 12.82 | 45.83 | 0.04 | 140.6 |
| 0516025-3 | 13.80 | 46.90 | 0.04 | 140.4 |
| 0516025-7 | 2620.52 | 48.05 | 0.00 | 140.4 |
| 0516025-8 | 2689.33 | 50.93 | 0.00 | 139.9 |
| 0516025-9 | 2711.89 | 48.38 | 0.00 | 140.5 |
| Average LCS= | 2673.91 | 46.74 | 0.02 | 140.5 averages |
| Average Bckgrd= | 13.47 | | | |

| Net cpm= | WIND2 * | %LUMEX | H# | |
|----------------------|----------------|----------------|---------------|-------------------|
| 2660.45 | 50.65 | 5.00 | 155.5 | UCL |
| /Known dpm= 12329.02 | 42.82 | 0.00 | 125.5 | LCL |
| Efficiency= 0.2158 | See Tech. Mgr. | See Tech. Mgr. | Std. Addition | Corrective Action |

*WIND2 control limits are based on average of blanks. CL used for calibration only.

Instrument Technician: Krista Bishop 2/15/06
 Signature & Date

Supervisory Review: Bence Hallen 2/15/06
 Signature & Date

ID# H3- 10 ML GEO.

12 FEB 2006 18:17

USER: 8

COMMENT:LS6500

PRESET TIME : 60.00
 DATA CALC : CPM H# :YES SAMPLE REPEATS: 1 PRINTER : STD
 COUNT BLANK : NO IC# : NO REPLICATES : 1 RS232 : OFF
 TWO PHASE : NO AQC : NO CYCLE REPEATS : 1 DISK :EDIT
 SCINTILLATOR: LIQUID LUMEX: NO LOW SAMPLE REJ: 0
 LOW LEVEL : YES HALF LIFE CORRECTION DATE: none

CHAN: 100.0 - 320.0 %ERROR: 2.00 FACTOR: 1.000000 BKG. SUB: 0
 CHAN: 380.0 - 990.0 %ERROR:20.00 FACTOR: 1.000000 BKG. SUB: 0

| SAM NO | POS | TIME MIN | H# | WIND1 | | WIND2 | | LUMEX % | ELAPSED TIME |
|-----------|------|-------------|-------|---------|--------|-------|--------|------------|-----------------|
| | | | | CPM | %ERROR | CPM | %ERROR | | |
| 1 | 29-1 | 60.00 | 140.9 | 13.78 | 6.95 | 47.48 | 3.75 | 0.04 | 61.03 |
| 2 | 29-2 | 60.00 | 140.6 | 12.82 | 7.21 | 45.83 | 3.81 | 0.04 | 122.13 |
| 3 | 29-3 | 60.00 | 140.4 | 13.80 | 6.95 | 46.90 | 3.77 | 0.04 | 183.25 |
| 4 | 29-4 | 3.85 | 140.4 | 2620.52 | 1.99 | 48.05 | 14.70 | 0.00 | 187.74 |
| 5 | 29-5 | 3.75 | 139.9 | 2689.33 | 1.99 | 50.93 | 14.47 | 0.00 | 192.14 |
| 6 | 29-6 | 3.70 | 140.5 | 2711.89 | 2.00 | 48.38 | 14.95 | 0.00 | 196.49 |
| 7 | 29-7 | 60.00 | 141.3 | 51.58 | 3.60 | 46.25 | 3.80 | 0.02 | 257.59 |

WMS
 2/15/06

000056

306214

LSC Run Log

Instrument ID: LS 6500

| Date | Sample ID | Count Time (min.) | Rack & Position | Test | User # | Batch ID | Position Check | Initials | Comments |
|---------|--------------|-------------------|-----------------|---------|--------|------------|----------------|----------|----------|
| 2/11/06 | 0601025-1D | 360 | 19 - 2 | N163 | 12 | N1060125-1 | UM3 | UM3 | NA |
| ↓ | N1060125-1MB | ↓ | ↓ - 2 | ↓ | ↓ | ↓ | ↓ | ↓ | ↓ |
| 2/12/06 | LCS | 360 | 19 - 2 | N163 | 12 | ↓ | UM3 | UM3 | NA |
| ↓ | DAILY QC | 10 | ↓ - 1-4 | ↓ | ↓ | ↓ | ↓ | ↓ | ↓ |
| 2/12/06 | 0516025-1 | 60 | 29 - 1 | H3-10mL | 0 | 0516025 | UM3 | UM3 | NA |
| ↓ | -2 | ↓ | ↓ - 2 | ↓ | ↓ | ↓ | ↓ | ↓ | ↓ |
| ↓ | -3 | ↓ | ↓ - 3 | ↓ | ↓ | ↓ | ↓ | ↓ | ↓ |
| ↓ | -7 | 3.85 | ↓ - 4 | ↓ | ↓ | ↓ | ↓ | ↓ | ↓ |
| ↓ | -8 | 3.75 | ↓ - 5 | ↓ | ↓ | ↓ | ↓ | ↓ | ↓ |
| ↓ | -9 | 3.70 | ↓ - 6 | ↓ | ↓ | ↓ | ↓ | ↓ | ↓ |
| ↓ | -13 | 60 | ↓ - 7 | ↓ | ↓ | ↓ | ↓ | ↓ | ↓ |
| 12 | | | | | | | | | |
| 13 | | | | | | | | | |
| 14 | | | | | | | | | |
| 15 | | | | | | | | | |
| 16 | | | | | | | | | |
| 17 | | | | | | | | | |
| 18 | | | | | | | | | |
| 19 | | | | | | | | | |
| 20 | | | | | | | | | |
| 21 | | | | | | | | | |
| 22 | | | | | | | | | |
| 23 | | | | | | | | | |
| 24 | | | | | | | | | |
| 25 | | | | | | | | | |
| 26 | | | | | | | | | |
| 27 | | | | | | | | | |
| 28 | | | | | | | | | |
| 29 | | | | | | | | | |
| 30 | | | | | | | | | |

UM3 2/14/06

UM3 2/14/06

Reviewed by / Date

306214

000057

Liquid Scintillation Benchsheet

Paragon Analytios, Inc.

Prep Batch: 18026H8.XLB

Method: WATER

File Name:

Prep Analyst: JJK
 Prep Date: 1/27/2006
 Prep SOP: 700RB

Non-Routine Pretreatment? Y / N
 Re-extraction? Y / N

| Lab Sample ID | Reok Pos | Reok | Milk | Report. Beq/g | Amplis Volume | Units | % Moisture | Water Added | Alquot Volume | Spikes Volume | Spikes Standard | Spine Pipet | Remarks |
|---------------|----------|------|-------|---------------|---------------|-------|------------|-------------|---------------|---------------|-----------------|-------------|---------|
| 0610026-1 | 28 | 1 | WATER | AR | 10 ML | 10 ML | 100 | 0 | 50 | | | | |
| 0610026-2 | 28 | 2 | WATER | AR | 10 ML | 10 ML | 100 | 0 | 50 | | | | |
| 0610026-3 | 28 | 3 | WATER | AR | 10 ML | 10 ML | 100 | 0 | 50 | | | | |
| 0610026-4 | | | WATER | AR | 5 ML | 5 ML | 100 | 0 | 50 | | | | |
| 0610026-5 | | | WATER | AR | 5 ML | 5 ML | 100 | 0 | 50 | | | | |
| 0610026-6 | 28 | 4 | WATER | AR | 10 ML | 10 ML | 100 | 0 | 48 | | 1888.2613.18 | RS-005 | |
| 0610026-7 | 28 | 5 | WATER | AR | 10 ML | 10 ML | 100 | 0 | 48 | | 1888.2613.18 | RS-005 | |
| 0610026-8 | 28 | 6 | WATER | AR | 10 ML | 10 ML | 100 | 0 | 48 | | 1888.2613.18 | RS-005 | |
| 0610026-9 | 28 | 7 | WATER | AR | 10 ML | 10 ML | 100 | 0 | 48 | | 1888.2613.18 | RS-005 | |
| 0610026-10 | | | WATER | AR | 5 ML | 5 ML | 100 | 0 | 48 | | 1888.2613.18 | RS-005 | |
| 0610026-11 | | | WATER | AR | 5 ML | 5 ML | 100 | 0 | 48 | | 1888.2613.18 | RS-005 | |
| 0610026-12 | 28 | 7 | WATER | AR | 10 ML | 10 ML | 100 | 0 | 48.5 | | 488.2613.80 | ST-002 | |
| 0610026-13 | | | WATER | AR | 5 ML | 5 ML | 100 | 0 | 48.5 | | 488.2613.80 | ST-002 | |
| 0610026-14 | | | WATER | AR | 5 ML | 5 ML | 100 | 0 | 48.5 | | 488.2613.80 | ST-002 | |

Spiked By: JJK
 Spine Witness: JJK

Counted by: LLT
 Scintillation Pipet: T-002
 Alquot Pipet: RS-007, RS-011

Reviewed By: JJK
 Received By: JJK

18026H8.XLB

18026H8.XLB

Liquid Sulfidation Benchsheet

Prep Analyst: JRIK
 Prep Date: 4/27/05
 Prep SOP: 700RP
 Non-Routine Pretreatment? Y / N
 Re-extraction? Y / N
 Matrix: WATER
 File Name: 16026H.XLS

| Lab Sample ID | Reack Pos | Reack | Matrix | Report Basis | Analyte Volume | Units | % Moisture | Water Added | Aliquot Volume | Splice Volume | Splice Standard | Splice Pipet | Remarks |
|---------------|-----------|-------|--------|--------------|----------------|-------|------------|-------------|----------------|---------------|-----------------|--------------|---------|
| 0516025-1 | | | WATER | AR | 10 ML | | 100 | 0 | 50 | | | | |
| 0516025-2 | | | WATER | AR | 10 ML | | 100 | 0 | 50 | | | | |
| 0516025-3 | | | WATER | AR | 10 ML | | 100 | 0 | 50 | | | | |
| 0516025-4 | 1 | | WATER | AR | 5 ML | | 100 | 0 | 50 | | | | |
| 0516025-5 | 2 | | WATER | AR | 5 ML | | 100 | 0 | 50 | | | | |
| 0516025-6 | 3 | | WATER | AR | 5 ML | | 100 | 0 | 50 | | | | |
| 0516025-7 | | | WATER | AR | 10 ML | | 100 | 0 | 48 | | 1699.2613.19 | RS-005 | |
| 0516025-8 | | | WATER | AR | 10 ML | | 100 | 0 | 48 | | 1699.2613.19 | RS-006 | |
| 0516025-9 | | | WATER | AR | 10 ML | | 100 | 0 | 48 | | 1699.2613.19 | RS-005 | |
| 0516025-10 | | | WATER | AR | 10 ML | | 100 | 0 | 48 | | 1699.2613.19 | RS-006 | |
| 0516025-11 | | | WATER | AR | 5 ML | | 100 | 0 | 48 | | 1699.2613.19 | RS-005 | |
| 0516025-12 | | | WATER | AR | 5 ML | | 100 | 0 | 48 | | 1699.2613.19 | RS-006 | |
| 0516025-13 | | | WATER | AR | 10 ML | | 100 | 0 | 48.8 | | 0.8 485.2613.80 | ST-002 | |
| 0516025-14 | | | WATER | AR | 5 ML | | 100 | 0 | 48.8 | | 0.6 485.2613.80 | BT-002 | |

Spiked By: JRIK
 Splice Witness: JRIK
 on 4/27/05
 evaluated By: JRIK
 on 4/27/05
 Retained By: JRIK
 on 4/27/05

DRAFT

16026H.XLS

TRITIUM RUN LOG

268468 Logbook/Page No.

SCP 700 Rev 9

Form 999005rL.doc (3/30/2003)

| Date | Workorder/ Sample Number | Column ID | Vial ID | Run No. | Comments | Technician's Initials |
|--------------------|-----------------------------|--------------|------------|------------|--------------------------------|--------------------------|
| 4/20/05 | 3H05042001 H.B | D | T5 | 1 | Batch 3H05042001 | JKA |
| L | 3H05042001 LCB | E | T11 | L | | L |
| 4/20/05 | 0516024-1,-4 | A | T5 | 1 | Den. H ₂ O 3H Calib | JKA |
| | -2,-5 | B | T2 | | | |
| | -3,-6 | C | T10 | | | |
| | 0516024-7,-10 | D | T10 | | | |
| | -8,-11 | E | T9 | | | |
| | -9,-12 | F | T3 | | | |
| | 0516024-13,-14 | A | 111 | 2 | | |
| 4/27/05 | 0516025-1,-4 | A | T1 | 1 | Den. H ₂ O 3H Calib | JKA |
| | -2,-5 | B | T5 | | | |
| | -3,-6 | C | T10 | | | |
| | 0516025-7,-10 | D | T11 | | | |
| | -8,-11 | E | 103 | | | |
| | -9,-12 | F | 111 | | | |
| | 0516025-13,-14 | A | 609 | 2 | | |
| 4/29/05 | 05050235 -1 | A | T1 | 1 | Batch: 3H05042901 | JKA |
| | -10 | B | T10 | | | |
| | -11 | C | T11 | | | |
| | 3H050429-1 LCB | D | 111 | | | |
| | 3H050429-1 LCB | E | 609 | | | |
| 5/15/05 | | | | | | |

Reviewed by / Date END 5/15/05

Paragon Analytics

QUALITY ASSURANCE SUMMARY SHEET

PAI W.O. #/BATCH LSC -
TEST LSC
METHOD LSC
SOP/REV (PREP) -
SOP/REV (ANAL) 704 r7

Briefly document any QA or other problems or deviations associated with the analysis of samples. Problems could result from: log-in, color, odor, dilution, consistency, scheduling, equipment, or instrumentation, or may include documentation of minor deviations necessary due to unique DQO's or sample characteristics.


Chm 9/23/05

Chm 9/23/05

As of 9/22/05, control limits for the "Window 2" count rate are established using the average count rate from the three reagent blanks associated with each prep batch +/- 3X the estimated poisson uncertainty.

Chm 9/23/05

Chm 9/23/05



TECHNICIAN/ANALYST *Chris M. King*

DATE 9/23/05

DEPARTMENT MANAGER *Brenda Kelley*

DATE 9/23/05

253979

FORM 302r6.doc (4/22/04)

000061

699.2613.19

34

Reference No. 2613

Prepare a working dilution of approximately 25,000 dpm/ml by diluting Std 699.2613.18 with DI water

dpm/ml by
Std ID: 699.2613.19
Description: 3H
Expiration: 8/1/06
Activity: 94000.71 dpm/ml
2s Uncertainty: 676.81 dpm/ml
Ref. Date: 9/2/03
Ref. Time: N/A
Prep Date: 7/29/04
Matrix/Comp: DI Water
Half Life (y): 1.23E+01
Prep by: JLK
Requires NCR for ICPT work

1) See Density on page 18 of log 2613

2) Transfer Std. 699.2613.18 to a 125 mL Amber glass bottle

Mass of bottle w/p.c. 194.7389g
Mass of bottle + Std 94.6785g
Net mass of Std 1.0714g

3) Add DI water to final dilution

Mass of bottle w/p.c. above 194.7389g
Mass of bottle Std. + DI water 172.4826g
Net mass of Std + DI water 77.7437g

4) Final Activity Calculations

$$\left(\frac{1.0714 \text{ g}}{77.7437 \text{ g}}\right) \left(\frac{94000.71 \text{ dpm}}{1.0714 \text{ g}}\right) \left(\frac{1.0714 \text{ g}}{1.0714 \text{ g}}\right) \left(\frac{1.0714 \text{ g}}{1.0714 \text{ g}}\right) \left(\frac{1.0714 \text{ g}}{1.0714 \text{ g}}\right) \left(\frac{1.0714 \text{ g}}{1.0714 \text{ g}}\right) \left(\frac{1.0714 \text{ g}}{1.0714 \text{ g}}\right)$$

Std ID: 699.2613.19

Description: 3H
Expiration: 7/28/04
Activity: 94000.71 dpm/ml
2s Uncertainty: 676.81 dpm/ml
Ref. Date: 9/3/98
Ref. Time: N/A
Prep Date: 7/28/03
Matrix/Comp: DI Water
Half Life (y): 1.23E+01

Chris Doyle 7/28/03

Std ID: 699.2613.19

Description: 3H
Expiration: 7/30/05
Activity: 94000.71 dpm/ml
2s Uncertainty: 676.81 dpm/ml
Ref. Date: 9/3/98
Ref. Time: N/A
Prep Date: 7/29/04
Matrix/Comp: DI Water
Half Life (y): 1.23E+01

JLK 9/1/04

Prepare a 1% dilution of Amprobe 6099 by diluting with DI water.

1) Determine the density of DI water

MASS of empty Class A volumetric flask 67.7550 g Bal #
 MASS of flask + 100 ml of DI water 100.3755 g J
 Net MASS of DI water 99.5779 g

2) Transfer Amprobe of amprobe 6099 into a 20 ml Amer glass Vort-vile

MASS of Vort-vile w/o lid 24.8883g Bal #
 MASS of opened ampoule before transfer 210.4130g
 MASS of opened ampoule after transfer 36.0281g
 Net mass of Str transferred 4.8851g

3) Add DI water to final dilution

MASS of Vort-vile w/ampoules 24.8883g Bal #
 MASS of Vort-vile, std + DI water 602.9642g
 Net mass of Str + DI water 38.5765g

4) Final Activity Calculation

$$\left(\frac{1.24 \times 10^4 \text{ KBq/g}}{1} \right) (4.8851 \text{ g}) \left(\frac{60000 \text{ Bq}}{1 \text{ MBq}} \right) \left(\frac{1000 \text{ Bq}}{1 \text{ MBq}} \right) (0.9958 \text{ g/ml})$$

38.5765g

24,787,412.57 Bq/ml

U.S. Department of Commerce
 National Institute of Standards
 and Technology
 BRM 4927F
 Hydrogen-3
 22 MBq in distilled water

CAUTION
 RADIOACTIVE



Continued on Page

Chad W. ...
 Signature

7/31/53
 Date

[Signature]
 Signature

7/31/53
 Date



National Institute of Standards & Technology

Certificate

PAID 8077
rec'd 5-07-03

Standard Reference Material 4927F Hydrogen-3 Radioactivity Standard

This Standard Reference Material (SRM) consists of radioactive hydrogen-3, as water, in 5 mL of distilled water. The solution is contained in a flame-sealed NIST borosilicate-glass ampoule. The SRM is intended for the calibration of beta-particle counting instruments and for the monitoring of radiochemical procedures.

Radiological Hazard

The SRM ampoule contains hydrogen-3 with a total activity of approximately 3.2 MBq. Hydrogen-3 decays by beta-particle emission. None of the beta particles escape from the SRM ampoule. During the decay process no photons are emitted. Approximate unshielded dose rates at several distances (as of the reference time) are given in note [a]*. There is no detectable external radiation. The SRM should be used only by persons qualified to handle radioactive material.

Chemical Hazard

The SRM ampoule contains only distilled water. There is no chemical hazard. If the ampoule is to be opened to transfer the solution, the recommended procedure is given on page 2.

Storage and Handling

The SRM should be stored and used at a temperature between 5 and 45 °C. The solution in an unopened ampoule should remain stable and homogeneous until at least September 2008.

The ampoule (or any subsequent container) should always be clearly marked as containing radioactive material. If the ampoule is transported it should be packed, marked, labeled, and shipped in accordance with the applicable national, international, and carrier regulations. The solution in the ampoule is a dangerous good (hazardous material) because of the radioactivity.

Preparation

This Standard Reference Material was prepared in the Physics Laboratory, Ionizing Radiation Division, Radioactivity Group, L.R. Karam, Group Leader. The overall technical direction and physical measurements leading to certification were provided by L.L. Lucas and M.P. Unterwieser of the Radioactivity Group.

The support aspects involved in the preparation, certification, and issuance of this SRM were coordinated through the Standard Reference Materials Program by J.W.L. Thomas.

Bert M. Coursey, Chief
Ionizing Radiation Division

Nancy M. Trahey, Chief
Standard Reference Materials Program

Gaithersburg, Maryland 20899
June 1999
Half-life and text revised October 2000

Recommended Procedure for Opening the SRM Ampoule

- 1) If the SRM solution is to be diluted, it is recommended that the diluting solution have a composition comparable to that of the SRM solution.
- 2) Wear eye protection, gloves, and protective clothing and work over a tray with absorbent paper in it. Work in a fume hood.
- 3) Shake the ampoule to wet all of the inside surface of the ampoule. Return the ampoule to the upright position.
- 4) Check that all of the liquid has drained out of the neck of the ampoule. If necessary, gently tap the neck to speed the process.
- 5) Holding the ampoule upright, score the narrowest part of the neck with a scribe or diamond pencil.
- 6) Lightly wet the scored line. This reduces the crack propagation velocity and makes for a cleaner break.
- 7) Hold the ampoule upright with a paper towel, a wiper, or a support jig. Position the scored line away from you. Using a paper towel or wiper to avoid contamination, snap off the top of the ampoule by pressing the narrowest part of the neck away from you while pulling the tip of the ampoule towards you.
- 8) Transfer the solution from the ampoule using a pycnometer or a pipet with dispenser handle. NEVER PIPETTE BY MOUTH.
- 9) Seal any unused SRM solution in a flame-sealed glass ampoule, if possible, to minimize the evaporation loss.

See also reference [4]*.

PROPERTIES OF SRM 4927E

Certified values

| | |
|---|--|
| Solution density | $(0.998 \pm 0.002) \text{ g}\cdot\text{mL}^{-1}$ at 20.0 °C [b]* |
| Radionuclide | Hydrogen-3 |
| Reference time | 1200 EST, 3 September 1998 |
| Massic activity of the solution [c] | 634.7 kBq·g ⁻¹ |
| Relative expanded uncertainty ($k=2$) | 0.72% [d] [e] |

Uncertified values

| | | | |
|--|--|--------------------------------------|------------------------------------|
| Physical Properties: | | | |
| Source description | Liquid in flame-sealed NIST borosilicate-glass ampoule | | |
| Ampoule specifications | Body outside diameter | (16.5 ± 0.5) mm | |
| | Wall thickness | (0.60 ± 0.04) mm | |
| | Radium content | Less than 2.5% | |
| | Lead-oxide content | Less than 0.02% | |
| | Other heavy elements | Trace quantities | |
| Solution mass | Approximately 5.0 g | | |
| Chemical Properties: | | | |
| Solution composition | Chemical Formula | Concentration (mol·L ⁻¹) | Mass Fraction (g·g ⁻¹) |
| | H ₂ O | 55 | 1.00 |
| | ³ H ₂ O | 6×10^{-7} | 1×10^{-6} |
| Radiological Properties: | | | |
| Radionuclidic impurities | None detected [f] | | |
| Half lives used | Hydrogen-3; (4500 ± 6) a [g] | | |
| Calibration method and measuring instrument(s) | 4πβ gas counting of SRM 4927E using the NIST length-compensated internal gas proportional counters and intercomparison of SRMs 4927E/4927F using two 4πβ liquid-scintillation counting systems [h] | | |

EVALUATION OF THE UNCERTAINTY OF THE MASSIC ACTIVITY [d]²

| Input Quantity x_i , the source of uncertainty (and individual uncertainty components where appropriate) | Method Used To Evaluate $u(x_i)$, the standard uncertainty of x_i (A) denotes evaluation by statistical methods (B) denotes evaluation by other methods | Relative Uncertainty Of Input Quantity, $u(x_i)/x_i$, (%) [I] | Relative Sensitivity Factor, $ dy/dx_i $ (x_i/y) [J] | Relative Uncertainty Of Output Quantity, $u(y)/y$, (%) [K] |
|--|---|---|---|--|
| Massic count rate of SRM 4927E, corrected for background and decay [h] | Standard deviation of the mean for 23 sets of gas counting measurements (A) | 0.18 | 1.0 | 0.18 |
| Gram-mole measurements | Estimated (B) | 0.20 | 1.0 | 0.20 |
| Live-time [p] | Estimated (B) | 0.10 | 1.0 | 0.10 |
| Extrapolation of count-rate-versus-energy to zero energy | Estimated (B) | 0.20 | 1.0 | 0.20 |
| Half life of ^{226}Ra | Standard uncertainty of the half life (A) | 0.18 [m] | 0.009 [n] | 0.002 |
| Liquid-scintillation intercomparison of SRM 4927F and SRM 4927E | Standard deviation of the mean for 7 sets of liquid-scintillation measurements (A) | 0.06 | 1.0 | 0.06 |
| Radioactive impurities | Limit of detection (B) [q] | 100 | 0.0005 | 0.05 |
| Relative Combined Standard Uncertainty of the Output Quantity, $u_c(y)/y$, (%) | | | | 0.36 |
| Coverage Factor, k | | | | $\times 2$ |
| Relative Expanded Uncertainty of the Output Quantity, $U(y)/y$, (%) | | | | 0.72 |

NOTES

[a] The Sievert is the SI unit for dose equivalent. See reference [1]. One μSv is equal to 0.1 μrem .
 Distance from Ampoule (cm): 1 30 100
 Approximate Dose Rate ($\mu\text{Sv/h}$): <0.1 (Not detectable)

[b] The stated uncertainty is two times the standard uncertainty.

[c] *Massic activity* is the preferred name for the quantity activity divided by the total mass of the sample. See reference [1].

[d] The reported value, y , of massic activity (activity per unit mass) at the reference time was not measured directly but was derived from measurements and calculations of other quantities. This can be expressed as $y = f(x_1, x_2, x_3, \dots, x_n)$, where f is a mathematical function derived from the assumed model of the measurement process.

The value, x_i , used for each input quantity i has a standard uncertainty, $u(x_i)$, that generates a corresponding uncertainty in y , $u_i(y) = |\partial y / \partial x_i| \cdot u(x_i)$, called a component of combined standard uncertainty of y .

The combined standard uncertainty of y , $u_c(y)$, is the positive square root of the sum of the squares of the components of combined standard uncertainty.

The combined standard uncertainty is multiplied by a coverage factor of $k = 2$ to obtain U , the expanded uncertainty of y .

Since it can be assumed that the possible estimated values of the massic activity are approximately normally distributed with approximate standard deviation $u_c(y)$, the unknown value of the massic activity is believed to lie in the interval $\pm U$ with a level of confidence of approximately 95 percent.

For further information on the expression of uncertainties, see references [2] and [3].

[e] The value of each standard uncertainty component, and hence the value of the expanded uncertainty itself, is a best estimate based upon all available information, but is only approximately known. That is to say, the "uncertainty of the uncertainty" is large and not well known. This is true for uncertainties evaluated by statistical methods (e.g., the relative standard deviation of the standard deviation of the mean for the massic response is approximately 50%) and for uncertainties evaluated by other methods (which could easily be over estimated or under estimated by substantial amounts). The unknown value of the expanded uncertainty is believed to lie in the interval $U/2$ to $2U$ (i.e., within a factor of 2 of the estimated value).

[f] The estimated limit of detection for radionuclides is $300 \text{ Bq} \cdot \text{g}^{-1}$.

[g] The stated uncertainty is the standard uncertainty. See reference [5].

[h] Extensive gas-counting measurements were made on the SRM 4927E solution during 1998 and 1999. The SRM 4927E solution was intercompared with the SRM 4927E solution using liquid-scintillation counting.

[i] Relative standard uncertainty of the input quantity x_i .

- [f] The relative change in the output quantity divided by the relative change in the input quantity x_i . If $|\partial y / \partial x_i| \cdot (x_i / y) = 1.0$, then a 1% change in x_i results in a 1% change in y . If $|\partial y / \partial x_i| \cdot (x_i / y) = 0.05$, then a 1% change in x_i results in a 0.05% change in y .
- [g] Relative component of combined standard uncertainty of output quantity, rounded to two significant figures or less. The relative component of combined standard uncertainty of y is given by $u_i(y) / y = |\partial y / \partial x_i| \cdot u(x_i) / y = |\partial y / \partial x_i| \cdot (x_i / y) \cdot u(x_i) / x_i$. The numerical values of $u(x_i) / x_i$, $|\partial y / \partial x_i| \cdot (x_i / y)$, and $u_i(y) / y$, all dimensionless quantities, are listed in columns 3, 4, and 5, respectively. Thus, the value in column 5 is equal to the value in column 4 multiplied by the value in column 3. The input quantities are independent, or very nearly so. Hence the covariances are zero or negligible.
- [h] The relative standard uncertainty of $\lambda \cdot t$ is determined by the relative standard uncertainty of λ (i.e., of the half-life). The relative standard uncertainty of t is negligible.
- [i] $|\partial y / \partial x_i| \cdot (x_i / y) = |\lambda \cdot t|$
- [j] The live time is determined by counting the pulses from a gated crystal-controlled oscillator.
- [k] The standard uncertainty for each undetected impurity that might reasonably be expected to be present is estimated to be equal to the estimated limit of detection for that impurity, i.e., $u(x_i) / x_i = 100\%$. $|\partial y / \partial x_i| \cdot (x_i / y) = \{(\text{response per Bq of impurity}) / (\text{response per Bq of } {}^{\text{H}}\text{-3})\} \cdot \{(\text{Bq of impurity}) / (\text{Bq of } {}^{\text{H}}\text{-3})\}$. Thus $u_i(y) / y$ is the relative change in y if the impurity were present with a massic activity equal to the estimated limit of detection.

REFERENCES

- [1] International Organization for Standardization (ISO), *ISO Standards Handbook - Quantities and Units*, 1993. Available from the American National Standards Institute, 11 West 42nd Street, New York, NY 10036, U.S.A. 1-212-642-4900.
- [2] International Organization for Standardization (ISO), *Guide to the Expression of Uncertainty in Measurement*, 1993. Available from the American National Standards Institute, 11 West 42nd Street, New York, NY 10036, U.S.A. 1-212-642-4900. (Listed under ISO miscellaneous publications as "ISO Guide to the Expression 1993".)
- [3] B. N. Taylor and C. E. Kuyatt, *Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Results*, NIST Technical Note 1297, 1994. Available from the Superintendent of Documents, U.S. Government Printing Office, Washington, DC 20407, U.S.A.
- [4] National Council on Radiation Protection and Measurements Report No. 58, *A Handbook of Radioactivity Measurements Procedures*, Second Edition, 1985. Available from the National Council on Radiation Protection and Measurements, 7910 Woodmont Avenue, Bethesda, MD 20814 U.S.A.
- [5] L.L. Lucas and M.P. Unterwiesing, *Comprehensive Review and Critical Evaluation of the Half-Life of Tritium*, J. Res. Natl. Inst. Stand. Technol. 105, 541-549 (2000).

DAILY INSTRUMENT PERFORMANCE CHECKS - LS6500

Daily IPCs consist of the following standards;

Efficiency Check

| Beckman Tritium Standard | Beckman C-14 Standard | PAI Reagent Blank |
|--------------------------|-----------------------|-------------------|
| Lot HNP0509 | Lot CNP3311 | 10/13/2004 |
| 105800 dpm | 101000 dpm | |
| 10/29/2003 REF | 10/29/2003 REF | |
| 10/29/2008 EXP | 10/29/2008 EXP | |

| Interim Control Limits | | | | | 1/25/2006 |
|------------------------|--------------------------------|------------------|----------------------|----------------------|-----------|
| | <u>Decay Corrected Tritium</u> | <u>Carbon-14</u> | <u>Reagent Blank</u> | <u>Quench Factor</u> | |
| UCL | 53731.57 | 73819.79 | 18.5 | 145.8 | |
| Mean Value | 48824.45 | 67108.90 | 14.9 | 132.5 | |
| LCL | 43962.20 | 60398.01 | 11.2 | 119.3 | |

r:\inst\sc\ls6500\prg\0805\sq6500.xls

| Date | Decay Corrected | | | C-14 CPM | Reagent | | Quench | | |
|-----------|-----------------|----------|----|----------|-------------|------|--------|-------|----|
| | H-3 CPM | H-CPM | | | H-3 Bkg CPM | H# | | | |
| 2/6/2006 | 43114.8 | 49022.92 | OK | 66888.5 | OK | 12.1 | OK | 138.6 | OK |
| 2/7/2006 | 43619.6 | 49604.56 | OK | 67084.8 | OK | 12.7 | OK | 134.3 | OK |
| 2/8/2006 | 43544.1 | 49526.35 | OK | 67772.1 | OK | 14.9 | OK | 133 | OK |
| 2/10/2006 | 43536.4 | 49532.90 | OK | 67085.9 | OK | 12.9 | OK | 134.6 | OK |
| 2/11/2006 | 42704.3 | 48593.70 | OK | 67500.5 | OK | 13.4 | OK | 134.5 | OK |
| 2/12/2006 | 43445.7 | 49444.99 | OK | 67272.7 | OK | 16.1 | OK | 134.4 | OK |
| 2/13/2006 | 44224.3 | 50338.88 | OK | 66942.8 | OK | 16.3 | OK | 135.3 | OK |

000070

H3 10 mL

Calibration Source Check

LS6500
Date: 2/12/2006
Nuclide: H3
Half Life: 1.223E+01 yr.

Calibration Check Source:

Spike Standard: 486.2613.60
Reference Date : 3/23/1998
Spiked DPM : 2852.9 dpm/mL
Spike Volume : 0.5 mL
Spiked into : 49.5 mL
Current Spk. Act. : 8.30 pCi/mL

Calibration Check Source Count

| Sample ID | Rack | Pos | Sam. Date | Cnt. Dur. | Anal. Vol. | GrsCPM | BkgCPM | Efficiency | Activity | Units | Chem. Yield |
|------------|------|-----|-----------|-----------|------------|--------|--------|------------|----------|--------|-------------|
| 0516025-13 | 29 | 7 | 4/27/2005 | 60 | 10 | 51.58 | 13.47 | 0.2158 | 8.32 | pCi/mL | 100% |

LCS Recovery: 100.3%
PAI Limits: 85-115%

R:\INSTL\SC\LS6500\PRG\CAL\VER

000071

PREPARE A WORKING DILUTION AT ~ 2500 dpm/ml OF 3H BY DILUTING STANDARD 486.1284.85 WITH DI WATER: (ACTIVITY OF 486.1284.85 = 948004.38 dpm/ml)

1.) DETERMINE DENSITY OF DI WATER:
 BAL 12 MASS OF EMPTY CLASS A 100 ml VOLUMETRIC FLASK: 62.6207g
 N/ MASS OF CLASS A VOLUMETRIC FLASK + 100 ml DI H₂O: 162.2711g
 NET MASS OF DI H₂O: 99.6504g
 $F = 0.9965 \text{ g/ml}$

2.) TRANSFER STANDARD 486.1284.85 TO A 500 ml amber BOTTLE:
 MASS OF BOTTLE w/ LID: 256.68g
 MASS OF BOTTLE + STD: 258.18g
 NET MASS OF STD: 1.50g

3.) ADD DI H₂O TO FINAL DILUTION:
 MASS OF BOTTLE w/ LID (FROM ABOVE): 256.68g
 MASS OF BOTTLE + STD + DI H₂O: 753.4g
 NET MASS OF STD + DI H₂O: 496.72g

4.) FINAL ACTIVITY CALCULATION:
 $(0.9965 \text{ g/ml}) \times (1.50 \text{ g}) \times (948004.38 \text{ dpm/ml}) = 2852.9 \text{ dpm/ml}$
 $496.72 \text{ g} = 2852.9 \text{ dpm/ml}$

| | | | |
|----------------------------------|-------------|----------------------------------|-------------|
| Std ID: 486.2613.60 | RG 10/5/04 | Std ID: 486.2613.60 | RG 10/5/04 |
| Description: 3H | | Description: 3H | |
| Expiration: 9/28/05 | | Expiration: 8/11/06 | |
| Activity: 2852.90 | dpm/ml | Activity: 2852.90 | dpm/ml |
| Uncertainty: 95.10 | dpm/ml | Uncertainty: 95.10 | dpm/ml |
| Ref. Date: 3/23/99 | | Ref. Date: 3/23/99 | |
| Ref Time: N/A | | Ref Time: N/A | |
| Prep Date: 9/27/04 | Prep by: AF | Prep Date: 9/27/04 | Prep by: AF |
| Matrix Comp: DI H ₂ O | | Matrix Comp: DI H ₂ O | |
| Half Life (Y): 1.24E+01 | | Half Life (Y): 1.24E+01 | |

Requires NCR for ICPT work

Continued on Page

Read and Understood By

Signed: Date: 10/5/04

Signed: Date: 10/5/04

| | | | | | |
|---------------------------------------|--------------------|-----------------|-------------------------|----------------|-------------------|
| <u>Sal</u> | <u>496</u> | <u>1282.24</u> | <u>100.000 dpm/disk</u> | <u>27.9711</u> | |
| | | | | | |
| | | | | | |
| <u>HS</u> | <u>496</u> | <u>1282.24</u> | <u>43.6011</u> | | |
| | | | | | |
| <u>Hand (dilution)</u> | | | <u>12.7170g</u> | | <u>RG 3/29/03</u> |
| <u>Std ID:</u> | <u>486.1284.85</u> | | | | |
| <u>Description:</u> | <u>H-3</u> | | | | |
| <u>Activity:</u> | <u>948004.38</u> | <u>dpm/ml</u> | | | |
| <u>2s Uncertainty:</u> | <u>4740.02</u> | <u>dpm/ml</u> | | | |
| <u>Ref. Date:</u> | <u>3/23/98</u> | | | | |
| <u>Ref Time:</u> | <u>na</u> | | | | |
| <u>Prep Date:</u> | <u>5/3/00</u> | <u>Prep by:</u> | <u>RTS</u> | | |
| <u>Expiration:</u> | <u>3/23/03</u> | | | | |
| <u>Matrix/Comp.</u> | <u>D1 H2O</u> | | | | |
| <u>Half Life (y):</u> | <u>1.23E+01</u> | | | | |
| <u>MAY REQUIRE NCR FOR IGPT WORK.</u> | | | | | |
| <u>Std ID:</u> | <u>486.1284.85</u> | | | | |

| | | | | | |
|---|---|--------------------|------------|--|--|
| <u>Description:</u> | <u>3H</u> | | | | |
| <u>Expiration:</u> | <u>7/20/05</u> | | | | |
| <u>Activity:</u> | <u>948004.38</u> | <u>dpm/ml</u> | | | |
| <u>2s Uncertainty:</u> | <u>4740.02</u> | <u>dpm/ml</u> | | | |
| <u>Ref. Date:</u> | <u>3/23/98</u> | | | | |
| <u>Ref Time:</u> | <u>N/A</u> | | | | |
| <u>Prep Date:</u> | <u>7/29/04</u> | <u>Prep by:</u> | <u>JLK</u> | | |
| <u>Matrix/Comp.</u> | <u>D1 H2O</u> | | | | |
| <u>Half Life (y):</u> | <u>1.23E+01</u> | | | | |
| <u>Description:</u> | <u>Tritium in Water Stock Solution</u> | | | | |
| <u>Activity:</u> | <u>427029</u> | <u>pCi/g or mL</u> | | | |
| <u>Uncertainty:</u> | <u>pCi/g or mL</u> | | | | |
| <u>Ref. Date:</u> | <u>3/23/98</u> | | | | |
| <u>Ref Time:</u> | <u>10:00 AM</u> | | | | |
| <u>Prep Date:</u> | <u>3/23/98</u> | <u>Prep by:</u> | <u>RTS</u> | | |
| <u>Expiration:</u> | <u>3/23/98</u> | | | | |
| <u>Matrix/Comp.</u> | <u>D1 Water</u> | | | | |
| <u>Short Lived -- Decay Correct with each use</u> | | | | | |
| <u>Handwritten:</u> | <u>Handwritten notes and signatures</u> | | | | |
| <u>Handwritten:</u> | <u>Handwritten notes and signatures</u> | | | | |
| <u>Handwritten:</u> | <u>Handwritten notes and signatures</u> | | | | |

Read and Understood By: RTS Date: 6/22/98

Signed: _____ Date: _____

PROJECT

Tritium Stock Solution 486-1284-B4

Notebook No. 183 E & O 4
Continued From Page

All solutions will be for 100% (Analytical: 5103-55543-207)
I need instructions for a 40 ml (17.4 ml) -
Certificate - 2302 not available, see 3/2/98 - (1) 1/1/03

Activity 51034 ml
Volume 50.34 g
Matrix DI

Std ID: 486-1284-B4

Std ID: 486-1284-B4

Description: 3H
Expiration: 7/30/05
Activity: 1014184.00 pCi/g or ml
2s Uncertainty: 50709.00 pCi/g or ml
Ref. Date: 3/23/98
Ref. Time: N/A
Prep Date: 7/29/04 Prep by: JLC
Matrix/Comp: DI Water
Half Life (y): 1.22E+01

Description: H-3
Activity: 1014184.00 pCi/g or ml
2s Uncertainty: 50709.00 pCi/g or ml
Ref. Date: 3/23/98
Ref. Time: 0.417
Prep Date: 3/30/98 Prep by: KTS
Expiration: 7/31/04
Matrix/Comp: DI Water
Half Life (y): 1.22E+01

Activity 51034 ml
Volume 50.34 g
Matrix DI
8/13/04

Std ID: 486-1284-B4

Description: Tritium in Water Stock Solution
Activity: 1014184 pCi/g or ml
Uncertainty: pCi/g or ml
Ref. Date: 3/23/98 3/23/98 7/2/02
Ref. Time: 10:00
Prep Date: 3/30/98 Prep by: KTS
Expiration: 3/29/03
Matrix/Comp: DI Water
Short Lived - Decay Correct with each use

Verified as verified on 7/2/02
Expiration 7/3/04 @ 7/2/02

Continued on Page

[Signature]
5/3/00
Signed

Read and Understood By
[Signature]
Signed

5/3/00
Date



ANALYTICS

1300 Seaboard Industrial Blvd.
Atlanta, Georgia 30318 - U.S.A.

Phone (404) 552-2577
Fax: (404) 552-2097

PAID 48
rec'd 3-24-98

CERTIFICATE OF CALIBRATION
Standard Radionuclide Source

55543-307

H-3 in Water 50 mL in Flame Sealed Vial

This standard radionuclide source was prepared using an aliquot measured gravimetrically from a master radionuclide solution standard. The master radionuclide solution standard was calibrated by the Department Des Applications Et De La Metrologie Des Rayonnements Ionisants (DAMRI), Paris, France, as Number 24057. The calibration was checked by liquid scintillation counting after source preparation.

ANALYTICS maintains traceability to the National Institute of Standards and Technology through Measurements Assurance Programs as described in USNRC Reg. Guide 4.15, Revision 1.

| | |
|--------------------|--------------------------|
| ISOTOPE: | H-3 |
| ACTIVITY (dps): | 1.889 E+06 |
| CALIBRATION DATE: | March 23, 1998 12:00 EST |
| HALF-LIFE: | 12.43 years |
| TOTAL UNCERTAINTY: | 5.0% |

50.34 grams of water.

P O NUMBER 23149, Item 1.

PREPARED BY:

M. D. Currie
M. D. Currie, Radiochemist

Q A APPROVED:

LS [Signature] 3/23/98

000075

Liquid Scintillation Counter

Quality Control Data

Daily Instrument Performance
Checks

000076

DAILY INSTRUMENT PERFORMANCE CHECKS - LS6500

Daily IPCs consist of the following standards;

Efficiency Check

| Beckman Tritium Standard | Beckman C-14 Standard | PAI Reagent Blank |
|---------------------------------|------------------------------|--------------------------|
| Lot HNP0509 | Lot CNP3311 | 10/13/2004 |
| 105800 dpm | 101000 dpm | |
| 10/29/2003 REF | 10/29/2003 REF | |
| 10/29/2008 EXP | 10/29/2008 EXP | |

Historical Control Limits

| | <u>Decay Corrected Tritium</u> | <u>Carbon-14</u> | <u>Reagent Blank</u> | <u>Quench Factor</u> | 3/4/2006 |
|-------------------|--------------------------------|------------------|----------------------|----------------------|----------|
| UCL | 50655.48 | 68696.60 | 18.9 | 138.8 | |
| Mean Value | 49177.70 | 67446.32 | 14.9 | 135.1 | |
| LCL | 47699.92 | 66196.05 | 11.0 | 131.4 | |

r:\inst\sc\ls6500\prg\0605\sq6500.xls

*Auto calibration performed, data used only to verify control limits. KB 4/07/06

| <u>Date</u> | <u>Decay Corrected</u> | | | <u>C-14 CPM</u> | | <u>Reagent <MAM</u> | | <u>Quench</u> | |
|-------------|------------------------|--------------|----|-----------------|-----|------------------------|--------------|---------------|----|
| | <u>H-3 CPM</u> | <u>H-CPM</u> | | | | <u>H-3 Bkg CPM</u> | <u>g CPM</u> | | |
| 6/10/2006 | 42856.3 | 49671.78 | OK | 66167.5 | Low | 12.9 | OK | 140.2 | OK |
| 6/11/2006 | 41557.2 | 48173.53 | OK | 66079.2 | Low | 17 | OK | 141.7 | OK |
| 6/12/2006 | 42498.0 | 49271.73 | OK | 66066.2 | Low | 13.8 | OK | 140.9 | OK |

000077

DAILY INSTRUMENT PERFORMANCE CHECKS - LS6500

Daily IPCs consist of the following standards;

Efficiency Check

| Beckman Tritium Standard | Beckman C-14 Standard | PAI Reagent Blank |
|--------------------------|-----------------------|-------------------|
| Lot HNP0509 | Lot CNP3311 | 10/13/2004 |
| 105800 dpm | 101000 dpm | |
| 10/29/2003 REF | 10/29/2003 REF | |
| 10/29/2008 EXP | 10/29/2008 EXP | |

Historical Control Limits

| | <u>Decay Corrected Tritium</u> | <u>Carbon-14</u> | <u>Reagent Blank</u> | <u>Quench Factor</u> | 3/4/2006 |
|-------------------|--------------------------------|------------------|----------------------|----------------------|----------|
| UCL | 50655.48 | 68696.60 | 18.9 | 138.8 | |
| Mean Value | 49177.70 | 67446.32 | 14.9 | 135.1 | |
| LCL | 47699.92 | 66196.05 | 11.0 | 131.4 | |

r:\inst\sc\ls6500\prg\0805\sq6500.xls

*Auto calibration performed, data used only to verify control limits. KB 4/07/06

| <u>Date</u> | <u>Decay Corrected</u> | | | <u>C-14 CPM</u> | | <u>Reagent <MAM</u> | | <u>Quench</u> | |
|-------------|------------------------|--------------|----|-----------------|----|------------------------|---------------|---------------|----|
| | <u>H-3 CPM</u> | <u>H-CPM</u> | | | | <u>H-3 Bkg CPM</u> | <u>cg CPM</u> | | |
| 6/20/2006 | 42742.5 | 49616.50 | OK | 66965.7 | OK | 17.5 | OK | 141.3 | OK |
| 6/22/2006 | 42547.6 | 49405.52 | OK | 67231 | OK | 15.7 | OK | 140 | OK |

000078