

November 29, 2006

Report to:

James Hix
Cordilleran Compliance Services
5550 Marshall Street
Arvada, CO 80002

Bill to:

James Hix
Cordilleran Compliance Services
5550 Marshall Street
Arvada, CO 80002

Project ID: ED04243

ACZ Project ID: L59449

James Hix:

Enclosed are the analytical results for sample(s) submitted to ACZ Laboratories, Inc. (ACZ) on October 17, 2006. This project has been assigned to ACZ's project number, L59449. Please reference this number in all future inquiries.

All analyses were performed according to ACZ's Quality Assurance Plan, version 11.0. The enclosed results relate only to the samples received under L59449. Each section of this report has been reviewed and approved by the appropriate Laboratory Supervisor, or a qualified substitute.

Except as noted, the test results for the methods and parameters listed on ACZ's current NELAC certificate letter (#ACZ) meet all requirements of NELAC.

This report shall be used or copied only in its entirety. ACZ is not responsible for the consequences arising from the use of a partial report.

All samples and sub-samples associated with this project will be disposed of after December 29, 2006. If the samples are determined to be hazardous, additional charges apply for disposal (typically less than \$10/sample). If you would like the samples to be held longer than ACZ's stated policy or to be returned, please contact your Project Manager or Customer Service Representative for further details and associated costs. ACZ retains analytical reports for five years.

If you have any questions or other needs, please contact your Project Manager.



Cordilleran Compliance Services

November 29, 2006

Project ID: ED04243

ACZ Project ID: L59449

Sample Receipt

ACZ Laboratories, Inc. (ACZ) received 3 ground water samples from Cordilleran Compliance Services on October 17, 2006. The samples were received in good condition. Upon receipt, the sample custodian removed the samples from the cooler, inspected the contents, and logged the samples into ACZ's computerized Laboratory Information Management System (LIMS). The samples were assigned ACZ LIMS project number L59449. The custodian verified the sample information entered into the computer against the chain of custody (COC) forms and sample bottle labels.

Holding Times

All analyses except those qualified with an ACZ 'H' flag were performed within EPA recommended holding times. The BTEX analyses were run outside of method holding time due to instrument issues.

Sample Analysis

These samples were analyzed for inorganic, organic and radiochemistry parameters. The individual methods are referenced on both, the ACZ invoice and the analytical reports. The extended qualifier reports may contain footnotes qualifying specific elements due to QC failures. In addition the following has been noted with this specific project:

1. The Gamma Scan and Tritium analyses were subcontracted to Paragon Analytical Labs in Ft. Collins, Colorado.
2. The BTEX analyses were qualified with the ACZ 'N1' flag due to high MTBE recovery on one of the quality control elements. There was insufficient sample remaining for reanalysis.

Cordilleran Compliance Services

Project ID: ED04243
 Sample ID: CW-W902

ACZ Sample ID: **L59449-01**
 Date Sampled: 10/16/06 09:40
 Date Received: 10/17/06
 Sample Matrix: Ground Water

Inorganic Prep

Parameter	EPA Method	Result	Qual	XQ	Units	MDL	PQL	Date	Analyst
Total Hot Plate Digestion	M200.2 ICP							10/20/06 13:42	erf

Metals Analysis

Parameter	EPA Method	Result	Qual	XQ	Units	MDL	PQL	Date	Analyst
Boron, total	M200.7 ICP	0.11			mg/L	0.01	0.05	10/25/06 19:45	msh
Calcium, total	M200.7 ICP	60.4			mg/L	0.2	1	10/25/06 19:45	msh
Iron, total	M200.7 ICP		U		mg/L	0.02	0.05	10/25/06 19:45	msh
Magnesium, total	M200.7 ICP	27.6			mg/L	0.2	1	10/25/06 19:45	msh
Manganese, total	M200.7 ICP		U		mg/L	0.005	0.03	10/25/06 19:45	msh
Potassium, total	M200.7 ICP	0.8	B		mg/L	0.3	1	10/25/06 19:45	msh
Selenium, total	M200.7 ICP		U		mg/L	0.04	0.2	10/26/06 18:31	gme
Sodium, total	M200.7 ICP	50.5			mg/L	0.3	1	10/25/06 19:45	msh

Wet Chemistry

Parameter	EPA Method	Result	Qual	XQ	Units	MDL	PQL	Date	Analyst
Alkalinity as CaCO3	SM2320B - Titration								
Bicarbonate as CaCO3		293			mg/L	2	20	10/30/06 0:00	ct
Carbonate as CaCO3			U		mg/L	2	20	10/30/06 0:00	ct
Hydroxide as CaCO3			U		mg/L	2	20	10/30/06 0:00	ct
Total Alkalinity		293			mg/L	2	20	10/30/06 0:00	ct
Bromide	M300.0 - Ion Chromatography		U	*	mg/L	0.1	0.5	10/26/06 16:29	nps
Chloride	M300.0 - Ion Chromatography	1.6	B		mg/L	0.5	3	10/26/06 16:29	nps
Fluoride	SM4500F-C	0.3	B	*	mg/L	0.1	0.5	10/26/06 11:09	ct
Lab Filtration	SM 3030 B			*				10/18/06 10:15	mls
Nitrate/Nitrite as N	M353.2 - H2SO4 preserved	0.35		*	mg/L	0.02	0.1	10/21/06 20:55	pjb
Nitrogen, ammonia	M350.1 - Automated Phenate		U	*	mg/L	0.05	0.5	10/25/06 14:19	nps
Phosphate	Calculation based on Ortho Phosphorus	0.09	B		mg/L	0.03	0.15	11/28/06 0:00	calc
Phosphorus, ortho dissolved	M365.1 - Automated Ascorbic Acid	0.03	B	*	mg/L	0.01	0.05	10/17/06 21:27	pjb
Residue, Filterable (TDS) @180C	M160.1 - Gravimetric	410			mg/L	10	20	10/23/06 16:31	ct
Sulfate	SM4500 SO4-D	60			mg/L	10	50	11/01/06 14:28	mhm/ct
Sulfide as S	376.2 - Methylene Blue		U	*	mg/L	0.02	0.1	10/19/06 17:11	ct

Cordilleran Compliance Services

Project ID: ED04243
 Sample ID: EG-SP902

ACZ Sample ID: **L59449-02**
 Date Sampled: 10/16/06 14:40
 Date Received: 10/17/06
 Sample Matrix: Ground Water

Inorganic Prep

Parameter	EPA Method	Result	Qual	XQ	Units	MDL	PQL	Date	Analyst
Total Hot Plate Digestion	M200.2 ICP							10/20/06 13:54	erf

Metals Analysis

Parameter	EPA Method	Result	Qual	XQ	Units	MDL	PQL	Date	Analyst
Boron, total	M200.7 ICP	0.27			mg/L	0.01	0.05	10/25/06 19:49	msh
Calcium, total	M200.7 ICP	67.0			mg/L	0.2	1	10/25/06 19:49	msh
Iron, total	M200.7 ICP	0.03	B		mg/L	0.02	0.05	10/25/06 19:49	msh
Magnesium, total	M200.7 ICP	30.9			mg/L	0.2	1	10/25/06 19:49	msh
Manganese, total	M200.7 ICP		U		mg/L	0.005	0.03	10/25/06 19:49	msh
Potassium, total	M200.7 ICP	0.5	B		mg/L	0.3	1	10/25/06 19:49	msh
Selenium, total	M200.7 ICP		U		mg/L	0.2	1	10/26/06 18:35	gme
Sodium, total	M200.7 ICP	56.8			mg/L	0.3	1	10/25/06 19:49	msh

Wet Chemistry

Parameter	EPA Method	Result	Qual	XQ	Units	MDL	PQL	Date	Analyst
Alkalinity as CaCO3	SM2320B - Titration								
Bicarbonate as CaCO3		320			mg/L	2	20	10/30/06 0:00	ct
Carbonate as CaCO3			U		mg/L	2	20	10/30/06 0:00	ct
Hydroxide as CaCO3			U		mg/L	2	20	10/30/06 0:00	ct
Total Alkalinity		320			mg/L	2	20	10/30/06 0:00	ct
Bromide	M300.0 - Ion Chromatography		U	*	mg/L	0.5	3	10/26/06 16:47	nps
Chloride	M300.0 - Ion Chromatography		U	*	mg/L	3	10	10/26/06 16:47	nps
Fluoride	SM4500F-C	0.5	B	*	mg/L	0.1	0.5	10/26/06 11:11	ct
Lab Filtration	SM 3030 B			*				10/18/06 10:17	mls
Nitrate/Nitrite as N	M353.2 - H2SO4 preserved	0.17		*	mg/L	0.02	0.1	10/21/06 21:00	pjb
Nitrogen, ammonia	M350.1 - Automated Phenate		U	*	mg/L	0.05	0.5	10/25/06 14:21	nps
Phosphate	Calculation based on Ortho Phosphorus	0.03	B		mg/L	0.03	0.15	11/28/06 0:00	calc
Phosphorus, ortho dissolved	M365.1 - Automated Ascorbic Acid	0.01	B	*	mg/L	0.01	0.05	10/17/06 21:29	pjb
Residue, Filterable (TDS) @180C	M160.1 - Gravimetric	430			mg/L	10	20	10/23/06 16:36	ct
Sulfate	SM4500 SO4-D	80			mg/L	10	50	11/01/06 14:32	mhm/ct
Sulfide as S	376.2 - Methylene Blue		U	*	mg/L	0.02	0.1	10/19/06 17:47	ct

Report Header Explanations

<i>Batch</i>	A distinct set of samples analyzed at a specific time
<i>Found</i>	Value of the QC Type of interest
<i>Limit</i>	Upper limit for RPD, in %.
<i>Lower</i>	Lower Recovery Limit, in % (except for LCSS, mg/Kg)
<i>MDL</i>	Method Detection Limit. Same as Minimum Reporting Limit. Allows for instrument and annual fluctuations.
<i>PCN/SCN</i>	A number assigned to reagents/standards to trace to the manufacturer's certificate of analysis
<i>PQL</i>	Practical Quantitation Limit, typically 5 times the MDL.
<i>QC</i>	True Value of the Control Sample or the amount added to the Spike
<i>Rec</i>	Amount of the true value or spike added recovered, in % (except for LCSS, mg/Kg)
<i>RPD</i>	Relative Percent Difference, calculation used for Duplicate QC Types
<i>Upper</i>	Upper Recovery Limit, in % (except for LCSS, mg/Kg)
<i>Sample</i>	Value of the Sample of interest

QC Sample Types

<i>AS</i>	Analytical Spike (Post Digestion)	<i>LCSWD</i>	Laboratory Control Sample - Water Duplicate
<i>ASD</i>	Analytical Spike (Post Digestion) Duplicate	<i>LFB</i>	Laboratory Fortified Blank
<i>CCB</i>	Continuing Calibration Blank	<i>LFM</i>	Laboratory Fortified Matrix
<i>CCV</i>	Continuing Calibration Verification standard	<i>LFMD</i>	Laboratory Fortified Matrix Duplicate
<i>DUP</i>	Sample Duplicate	<i>LRB</i>	Laboratory Reagent Blank
<i>ICB</i>	Initial Calibration Blank	<i>MS</i>	Matrix Spike
<i>ICV</i>	Initial Calibration Verification standard	<i>MSD</i>	Matrix Spike Duplicate
<i>ICSAB</i>	Inter-element Correction Standard - A plus B solutions	<i>PBS</i>	Prep Blank - Soil
<i>LCSS</i>	Laboratory Control Sample - Soil	<i>PBW</i>	Prep Blank - Water
<i>LCSSD</i>	Laboratory Control Sample - Soil Duplicate	<i>PQV</i>	Practical Quantitation Verification standard
<i>LCSW</i>	Laboratory Control Sample - Water	<i>SDL</i>	Serial Dilution

QC Sample Type Explanations

Blanks	Verifies that there is no or minimal contamination in the prep method or calibration procedure.
Control Samples	Verifies the accuracy of the method, including the prep procedure.
Duplicates	Verifies the precision of the instrument and/or method.
Spikes/Fortified Matrix	Determines sample matrix interferences, if any.
Standard	Verifies the validity of the calibration.

ACZ Qualifiers (Qual)

B	Analyte concentration detected at a value between MDL and PQL.
H	Analysis exceeded method hold time. pH is a field test with an immediate hold time.
R	Poor spike recovery accepted because the other spike in the set fell within the given limits.
T	High Relative Percent Difference (RPD) accepted because sample concentrations are less than 10x the MDL.
U	Analyte was analyzed for but not detected at the indicated MDL
V	High blank data accepted because sample concentration is 10 times higher than blank concentration
W	Poor recovery for Silver quality control is accepted because Silver often precipitates with Chloride.
X	Quality control sample is out of control.
Z	Poor spike recovery is accepted because sample concentration is four times greater than spike concentration.

Method References

(1)	EPA 600/4-83-020. Methods for Chemical Analysis of Water and Wastes, March 1983.
(2)	EPA 600/R-93-100. Methods for the Determination of Inorganic Substances in Environmental Samples, August 1993.
(3)	EPA 600/R-94-111. Methods for the Determination of Metals in Environmental Samples - Supplement I, May 1994.
(5)	EPA SW-846. Test Methods for Evaluating Solid Waste, Third Edition with Update III, December 1996.
(6)	Standard Methods for the Examination of Water and Wastewater, 19th edition, 1995.

Comments

(1)	QC results calculated from raw data. Results may vary slightly if the rounded values are used in the calculations.
(2)	Soil, Sludge, and Plant matrices for Inorganic analyses are reported on a dry weight basis.
(3)	Animal matrices for Inorganic analyses are reported on an "as received" basis.

Cordilleran Compliance Services

ACZ Project ID: **L59449**

ACZ ID	WORKNUM	PARAMETER	METHOD	QUAL	DESCRIPTION
L59449-01	WG215415	Bromide	M300.0 - Ion Chromatography	RA	Relative Percent Difference (RPD) was not used for data validation because the sample concentration is too low for accurate evaluation (< 10x MDL).
	WG215412	Fluoride	SM4500F-C	RA	Relative Percent Difference (RPD) was not used for data validation because the sample concentration is too low for accurate evaluation (< 10x MDL).
	WG215199	Nitrate/Nitrite as N	M353.2 - H2SO4 preserved	M1	Matrix spike recovery was high, the method control sample recovery was acceptable.
			M353.2 - H2SO4 preserved	RA	Relative Percent Difference (RPD) was not used for data validation because the sample concentration is too low for accurate evaluation (< 10x MDL).
	WG215368	Nitrogen, ammonia	M350.1 - Automated Phenate	RA	Relative Percent Difference (RPD) was not used for data validation because the sample concentration is too low for accurate evaluation (< 10x MDL).
	WG214983	Phosphorus, ortho dissolved	M365.1 - Automated Ascorbic Acid	RA	Relative Percent Difference (RPD) was not used for data validation because the sample concentration is too low for accurate evaluation (< 10x MDL).
			M365.1 - Automated Ascorbic Acid	ZU	Analysis date/time precedes filter date/time. A portion of sample was filtered and analyzed prior to the creation of a Filter workgroup.
WG215077	Sulfide as S	376.2 - Methylene Blue	RA	Relative Percent Difference (RPD) was not used for data validation because the sample concentration is too low for accurate evaluation (< 10x MDL).	
L59449-02	WG215415	Bromide	M300.0 - Ion Chromatography	DH	Sample required dilution due to high TDS and/or EC value.
			M300.0 - Ion Chromatography	RA	Relative Percent Difference (RPD) was not used for data validation because the sample concentration is too low for accurate evaluation (< 10x MDL).
	WG215412	Fluoride	M300.0 - Ion Chromatography	DH	Sample required dilution due to high TDS and/or EC value.
			SM4500F-C	RA	Relative Percent Difference (RPD) was not used for data validation because the sample concentration is too low for accurate evaluation (< 10x MDL).
	WG215199	Nitrate/Nitrite as N	M353.2 - H2SO4 preserved	M1	Matrix spike recovery was high, the method control sample recovery was acceptable.
			M353.2 - H2SO4 preserved	RA	Relative Percent Difference (RPD) was not used for data validation because the sample concentration is too low for accurate evaluation (< 10x MDL).
	WG215368	Nitrogen, ammonia	M350.1 - Automated Phenate	RA	Relative Percent Difference (RPD) was not used for data validation because the sample concentration is too low for accurate evaluation (< 10x MDL).
WG214983	Phosphorus, ortho dissolved	M365.1 - Automated Ascorbic Acid	RA	Relative Percent Difference (RPD) was not used for data validation because the sample concentration is too low for accurate evaluation (< 10x MDL).	
		M365.1 - Automated Ascorbic Acid	ZU	Analysis date/time precedes filter date/time. A portion of sample was filtered and analyzed prior to the creation of a Filter workgroup.	
WG215077	Sulfide as S	376.2 - Methylene Blue	RA	Relative Percent Difference (RPD) was not used for data validation because the sample concentration is too low for accurate evaluation (< 10x MDL).	

Cordilleran Compliance ServicesProject ID: ED04243
Sample ID: CW-W902ACZ Sample ID: **L59449-01**
Date Sampled: 10/16/06 9:40
Date Received: 10/17/06
Sample Matrix: Ground Water**BTEX with MTBE**Analysis Method: **M8021B GC/PID**
Extract Method:Workgroup: **WG215679**
Analyst: *ccp*
Extract Date:
Analysis Date: **10/31/06 11:20**

Compound	CAS	Result	QUAL	Dilution	XQ	Units	MDL	PQL
Benzene	71-43-2		UH	1	*	ug/L	0.2	0.5
Ethylbenzene	100-41-4		UH	1	*	ug/L	0.2	1
m p Xylene	1330 20 7		UH	1	*	ug/L	0.4	2
Methyl Tert Butyl Ether	1634-04-4		UH	1	*	ug/L	0.2	1
o Xylene	95-47-6		UH	1	*	ug/L	0.2	1
Toluene	108-88-3		UH	1	*	ug/L	0.2	1
Surrogate Recoveries	CAS	% Recovery		Dilution	XQ	Units	LCL	UCL
Bromofluorobenzene	460-00-4	94		1		%	83	117

Cordilleran Compliance ServicesProject ID: ED04243
Sample ID: CW-W902ACZ Sample ID: **L59449-01**
Date Sampled: 10/16/06 9:40
Date Received: 10/17/06
Sample Matrix: Ground Water**Methane**Analysis Method: **RSK 175 (GC/FID)**
Extract Method:Workgroup: **WG215622**
Analyst: *jj*
Extract Date:
Analysis Date: **10/30/06 17:21**

Compound	CAS	Result	QUAL	Dilution	XQ	Units	MDL	PQL
Methane	74-82-8		U	1	*	mg/L	0.002	0.002

Cordilleran Compliance ServicesProject ID: ED04243
Sample ID: EG-SP902ACZ Sample ID: **L59449-02**
Date Sampled: 10/16/06 14:40
Date Received: 10/17/06
Sample Matrix: Ground Water**BTEX with MTBE**Analysis Method: **M8021B GC/PID**
Extract Method:Workgroup: **WG215679**
Analyst: *ccp*
Extract Date:
Analysis Date: **10/31/06 9:55**

Compound	CAS	Result	QUAL	Dilution	XQ	Units	MDL	PQL
Benzene	71-43-2		UH	1	*	ug/L	0.2	0.5
Ethylbenzene	100-41-4		UH	1	*	ug/L	0.2	1
m p Xylene	1330 20 7		UH	1	*	ug/L	0.4	2
Methyl Tert Butyl Ether	1634-04-4		UH	1	*	ug/L	0.2	1
o Xylene	95-47-6		UH	1	*	ug/L	0.2	1
Toluene	108-88-3		UH	1	*	ug/L	0.2	1
Surrogate Recoveries	CAS	% Recovery		Dilution	XQ	Units	LCL	UCL
Bromofluorobenzene	460-00-4	95.9		1		%	83	117

Cordilleran Compliance ServicesProject ID: ED04243
Sample ID: EG-SP902ACZ Sample ID: **L59449-02**
Date Sampled: 10/16/06 14:40
Date Received: 10/17/06
Sample Matrix: Ground Water**Methane**Analysis Method: **RSK 175 (GC/FID)**
Extract Method:Workgroup: **WG215622**
Analyst: *jj*
Extract Date:
Analysis Date: **10/30/06 17:29**

Compound	CAS	Result	QUAL	Dilution	XQ	Units	MDL	PQL
Methane	74-82-8		U	1	*	mg/L	0.002	0.002

Cordilleran Compliance ServicesProject ID: ED04243
Sample ID: TB100906-04ACZ Sample ID: **L59449-03**
Date Sampled: 10/16/06 14:48
Date Received: 10/17/06
Sample Matrix: Ground Water**BTEX with MTBE**Analysis Method: **M8021B GC/PID**
Extract Method:Workgroup: **WG215679**
Analyst: *ccp*
Extract Date:
Analysis Date: **10/31/06 10:37**

Compound	CAS	Result	QUAL	Dilution	XQ	Units	MDL	PQL
Benzene	71-43-2		UH	1	*	ug/L	0.2	0.5
Ethylbenzene	100-41-4		UH	1	*	ug/L	0.2	1
m p Xylene	1330 20 7		UH	1	*	ug/L	0.4	2
Methyl Tert Butyl Ether	1634-04-4		UH	1	*	ug/L	0.2	1
o Xylene	95-47-6		UH	1	*	ug/L	0.2	1
Toluene	108-88-3	0.8	JH	1	*	ug/L	0.2	1
Surrogate Recoveries	CAS	% Recovery		Dilution	XQ	Units	LCL	UCL
Bromofluorobenzene	460-00-4	94.9		1		%	83	117

Report Header Explanations

<i>Batch</i>	A distinct set of samples analyzed at a specific time
<i>Found</i>	Value of the QC Type of interest
<i>Limit</i>	Upper limit for RPD, in %.
<i>Lower</i>	Lower Recovery Limit, in % (except for LCSS, mg/Kg)
<i>LCL</i>	Lower Control Limit
<i>MDL</i>	Method Detection Limit. Same as Minimum Reporting Limit. Allows for instrument and annual fluctuations.
<i>PCNI/SCN</i>	A number assigned to reagents/standards to trace to the manufacturer's certificate of analysis
<i>PQL</i>	Practical Quantitation Limit
<i>QC</i>	True Value of the Control Sample or the amount added to the Spike
<i>Rec</i>	Amount of the true value or spike added recovered, in % (except for LCSS, mg/Kg)
<i>RPD</i>	Relative Percent Difference, calculation used for Duplicate QC Types
<i>Upper</i>	Upper Recovery Limit, in % (except for LCSS, mg/Kg)
<i>UCL</i>	Upper Control Limit
<i>Sample</i>	Value of the Sample of interest

QC Sample Types

<i>SURR</i>	Surrogate	<i>LFM</i>	Laboratory Fortified Matrix
<i>INTS</i>	Internal Standard	<i>LFMD</i>	Laboratory Fortified Matrix Duplicate
<i>DUP</i>	Sample Duplicate	<i>LRB</i>	Laboratory Reagent Blank
<i>LCSS</i>	Laboratory Control Sample - Soil	<i>MS/MSD</i>	Matrix Spike/Matrix Spike Duplicate
<i>LCSW</i>	Laboratory Control Sample - Water	<i>PBS</i>	Prep Blank - Soil
<i>LFB</i>	Laboratory Fortified Blank	<i>PBW</i>	Prep Blank - Water

QC Sample Type Explanations

Blanks	Verifies that there is no or minimal contamination in the prep method procedure.
Control Samples	Verifies the accuracy of the method, including the prep procedure.
Duplicates	Verifies the precision of the instrument and/or method.
Spikes/Fortified Matrix	Determines sample matrix interferences, if any.

ACZ Qualifiers (Qual)

B	Analyte detected in daily blank
H	Analysis exceeded method hold time.
J	Analyte concentration detected at a value between MDL and PQL
R	Poor spike recovery accepted because the other spike in the set fell within the given limits.
T	High Relative Percent Difference (RPD) accepted because sample concentrations are less than 10x the MDL.
U	Analyte was analyzed for but not detected at the indicated MDL
V	High blank data accepted because sample concentration is 10 times higher than blank concentration
W	Poor recovery for Silver quality control is accepted because Silver often precipitates with Chloride.
X	Quality control sample is out of control.
Z	Poor spike recovery is accepted because sample concentration is four times greater than spike concentration.
P	Analyte concentration differs from second detector by more than 40%.
E	Analyte concentration is estimated due to result exceeding calibration range.
M	Analyte concentration is estimated due to matrix interferences.

Method References

- (1) EPA 600/4-83-020. Methods for Chemical Analysis of Water and Wastes, March 1983.
- (2) EPA 600/4-90/020. Methods for the Determination of Organic Compounds in Drinking Water (I), July 1990.
- (3) EPA 600/R-92/129. Methods for the Determination of Organic Compounds in Drinking Water (II), July 1990.
- (5) EPA SW-846. Test Methods for Evaluating Solid Waste, Third Edition with Update III, December, 1996.
- (6) Standard Methods for the Examination of Water and Wastewater, 19th edition, 1995.

Comments

- (1) QC results calculated from raw data. Results may vary slightly if the rounded values are used in the calculations.
- (2) Organic analyses are reported on an "as received" basis.

Cordilleran Compliance Services

ACZ Project ID: **L59449**

ACZ ID	WORKNUM	PARAMETER	METHOD	QUAL	DESCRIPTION		
L59449-01	WG215679	Benzene	M8021B GC/PID	H1	Sample analysis performed past holding time.		
		Ethylbenzene	M8021B GC/PID	H1	Sample analysis performed past holding time.		
		m p Xylene	M8021B GC/PID	H1	Sample analysis performed past holding time.		
		Methyl Tert Butyl Ether	M8021B GC/PID	C6	Sample RPD between the primary and confirmatory analysis exceeded 40%. Per EPA Method 8000B, the higher value was reported as there was no obvious chromatographic interference.		
			M8021B GC/PID	H1	Sample analysis performed past holding time.		
			M8021B GC/PID	M1	Matrix spike recovery was high, the method control sample recovery was acceptable.		
			M8021B GC/PID	N1	See Case Narrative.		
			M8021B GC/PID	V7	Calibration verification recovery was above the method control limit for this analyte, however the average % difference or % drift for all the analytes met method criteria.		
			o Xylene	M8021B GC/PID	H1	Sample analysis performed past holding time.	
			Toluene	M8021B GC/PID	H1	Sample analysis performed past holding time.	
		WG215622	*All Compounds*	RSK 175 (GC/FID)	RA	Relative Percent Difference (RPD) was not used for data validation because the sample concentration is too low for accurate evaluation (< 10x MDL).	
		L59449-02	WG215679	Benzene	M8021B GC/PID	H1	Sample analysis performed past holding time.
				Ethylbenzene	M8021B GC/PID	H1	Sample analysis performed past holding time.
m p Xylene	M8021B GC/PID			H1	Sample analysis performed past holding time.		
Methyl Tert Butyl Ether	M8021B GC/PID			C6	Sample RPD between the primary and confirmatory analysis exceeded 40%. Per EPA Method 8000B, the higher value was reported as there was no obvious chromatographic interference.		
	M8021B GC/PID			H1	Sample analysis performed past holding time.		
	M8021B GC/PID			M1	Matrix spike recovery was high, the method control sample recovery was acceptable.		
	M8021B GC/PID			N1	See Case Narrative.		
	M8021B GC/PID			V7	Calibration verification recovery was above the method control limit for this analyte, however the average % difference or % drift for all the analytes met method criteria.		
	o Xylene			M8021B GC/PID	H1	Sample analysis performed past holding time.	
	Toluene			M8021B GC/PID	H1	Sample analysis performed past holding time.	
L59449-03	WG215679			Benzene	M8021B GC/PID	H1	Sample analysis performed past holding time.
				Ethylbenzene	M8021B GC/PID	H1	Sample analysis performed past holding time.
				m p Xylene	M8021B GC/PID	H1	Sample analysis performed past holding time.
		Methyl Tert Butyl Ether	M8021B GC/PID	C6	Sample RPD between the primary and confirmatory analysis exceeded 40%. Per EPA Method 8000B, the higher value was reported as there was no obvious chromatographic interference.		
			M8021B GC/PID	H1	Sample analysis performed past holding time.		
			M8021B GC/PID	M1	Matrix spike recovery was high, the method control sample recovery was acceptable.		
			M8021B GC/PID	N1	See Case Narrative.		
			M8021B GC/PID	V7	Calibration verification recovery was above the method control limit for this analyte, however the average % difference or % drift for all the analytes met method criteria.		
			o Xylene	M8021B GC/PID	H1	Sample analysis performed past holding time.	
			Toluene	M8021B GC/PID	H1	Sample analysis performed past holding time.	

Cordilleran Compliance Services

Project ID: ED04243
Sample ID: CW-W902
Locator:

ACZ Sample ID: **L59449-01**
Date Sampled: 10/16/06 9:40
Date Received: 10/17/06
Sample Matrix:

Gamma Emitting Nuclides

Prep Method:

Parameter	Measure Date	Prep Date	Result	Error(+/-)	LLD	Units	XQ	Analyst
Gamma Emitting Nuclides			see attach- ments					n/a

Tritium in water

Prep Method:

Parameter	Measure Date	Prep Date	Result	Error(+/-)	LLD	Units	XQ	Analyst
Tritium in water			see attach- ments					n/a

Cordilleran Compliance Services

Project ID: ED04243
Sample ID: EG-SP902
Locator:

ACZ Sample ID: **L59449-02**
Date Sampled: 10/16/06 14:40
Date Received: 10/17/06
Sample Matrix:

Gamma Emitting Nuclides

Prep Method:

Parameter	Measure Date	Prep Date	Result	Error(+/-)	LLD	Units	XQ	Analyst
Gamma Emitting Nuclides			see attach- ments					n/a

Tritium in water

Prep Method:

Parameter	Measure Date	Prep Date	Result	Error(+/-)	LLD	Units	XQ	Analyst
Tritium in water			see attach- ments					n/a



Report Header Explanations

Table with 2 columns: Term and Definition. Includes terms like Batch, Error(+/-), Found, Limit, LCL, LLD, PCN/SCN, PQL, QC, Rec, RER, UCL, and Sample.

QC Sample Types

Table with 4 columns: Term, Description, Abbreviation, and Full Name. Includes DUP, LCSS, LCSW, MS/MSD, PBS, and PBW.

QC Sample Type Explanations

Table with 2 columns: Term and Explanation. Includes Blanks, Control Samples, Duplicates, and Matrix Spikes.

ACZ Qualifiers (Qual)

Table with 2 columns: Qualifier and Description. Includes H, R, T, U, V, X, and Z.

Method Prefix Reference

Table with 2 columns: Prefix and Reference. Includes M, SM, D, RP, and ESM.

Comments

- (1) Solid matrices are reported on a dry weight basis.
(2) Preparation method: "Method" indicates preparation defined in analytical method.
(3) QC results calculated from raw data. Results may vary slightly if the rounded values are used in the calculations.

Cordilleran Compliance Services

ACZ Project ID: **L59449**

ACZ ID	WORKNUM	PARAMETER	METHOD	QUAL	DESCRIPTION
--------	---------	-----------	--------	------	-------------

No extended qualifiers associated with this analysis

Cordilleran Compliance Services

ACZ Project ID: **L59449**

Gas Chromatography

The following parameters are not offered for certification or are not covered by NELAC certificate #ACZ.

Methane	RSK 175 (GC/FID)
---------	------------------

Wet Chemistry

The following parameters are not offered for certification or are not covered by NELAC certificate #ACZ.

Lab Filtration	SM 3030 B
Sulfide as S	376.2 - Methylene Blue

Cordilleran Compliance Services
 ED04243

ACZ Project ID: L59449
 Date Received: 10/17/2006
 Received By:
 Date Printed: 10/18/2006

Receipt Verification

	YES	NO	NA
1) Does this project require special handling procedures such as CLP protocol?			X
2) Are the custody seals on the cooler intact?	X		
3) Are the custody seals on the sample containers intact?			X
4) Is there a Chain of Custody or other directive shipping papers present?	X		
5) Is the Chain of Custody complete?	X		
6) Is the Chain of Custody in agreement with the samples received?	X		
7) Is there enough sample for all requested analyses?	X		
8) Are all samples within holding times for requested analyses?	X		
9) Were all sample containers received intact?	X		
10) Are the temperature blanks present?			X
11) Are the trip blanks (VOA and/or Cyanide) present?			X
12) Are samples requiring no headspace, headspace free?	X		
13) Do the samples that require a Foreign Soils Permit have one?			X

Exceptions: If you answered no to any of the above questions, please describe

Two trip blanks were received with this project. The project manager indicated to dispose of one trip blank and to log in the other. TB100906-05 was disposed of.

Contact (For any discrepancies, the client must be contacted)

N/A

Shipping Containers

Cooler Id	Temp (°C)	Rad (µR/hr)
1505	7.6	16

Client must contact ACZ Project Manager if analysis should not proceed for samples received outside of thermal preservation acceptance criteria.

Notes

Cordilleran Compliance Services
 ED04243

ACZ Project ID: L59449
 Date Received: 10/17/2006
 Received By:

Sample Container Preservation

SAMPLE	CLIENT ID	R < 2	G < 2	BK < 2	Y < 2	YG < 2	B < 2	O < 2	T > 12	N/A	RAD	ID
L59449-01	CW-W902	Y			Y				Y			<input type="checkbox"/>
L59449-02	EG-SP902	Y			Y				Y			<input type="checkbox"/>
L59449-03	TB100906-04									X		<input type="checkbox"/>

Sample Container Preservation Legend

Abbreviation	Description	Container Type	Preservative/Limits
R	Raw/Nitric	RED	pH must be < 2
B	Filtered/Sulfuric	BLUE	pH must be < 2
BK	Filtered/Nitric	BLACK	pH must be < 2
G	Filtered/Nitric	GREEN	pH must be < 2
O	Raw/Sulfuric	ORANGE	pH must be < 2
P	Raw/NaOH	PURPLE	pH must be > 12 *
T	Raw/NaOH Zinc Acetate	TAN	pH must be > 12
Y	Raw/Sulfuric	YELLOW	pH must be < 2
YG	Raw/Sulfuric	YELLOW GLASS	pH must be < 2
N/A	No preservative needed	Not applicable	
RAD	Gamma/Beta dose rate	Not applicable	must be < 250 µR/hr

* pH check performed by analyst prior to sample preparation

Sample IDs Reviewed By: _____



PARAGON ANALYTICS

225 Commerce Drive ♦ Fort Collins, CO 80524 ♦ (800) 443-1511 ♦ (970) 490-1511 ♦ FAX (970) 490-1522

November 27, 2006

Mr. Tony Antalek
ACZ Laboratories, Inc.
2773 Downhill Drive
Steamboat Springs, CO 80487

Re: Paragon Workorder: 06-10-164
Client Project Name: None Submitted
Client Project Number: L59449

Dear Mr. Antalek:

Two water samples were received from ACZ Laboratories, Inc. on October 20, 2006. The samples were scheduled for Tritium (pages 1-68) and Gamma Spectroscopy (pages 1-146) analyses.

The results for these analyses are contained in the enclosed reports.

Thank you for your confidence in Paragon Analytics. Should you have any questions, please call.

Sincerely,

Paragon Analytics
Lance Steere
Senior Project Manager

LRS/jb
Enclosure: Report

Paragon Analytics

Sample Number(s) Cross-Reference Table

Paragon OrderNum: 0610164

Client Name: ACZ Laboratories, Inc.

Client Project Name:

Client Project Number: L59449

Client PO Number:

Client Sample	Lab Sample Number	COC Number	Matrix	Date Collected	Time Collected
L59449-01	0610164-1		WATER	16-Oct-06	9:40
L59449-02	0610164-2		WATER	16-Oct-06	14:40

CONDITION OF SAMPLE UPON RECEIPT FORM

Paragon Analytics

Client: ACIZ
Project Manager: LS

Workorder No: 0610164
Initials: SL Date: 10-20-06

1. Does this project require any special handling in addition to standard Paragon procedures?		YES	<input checked="" type="radio"/> NO
2. Are custody seals on shipping containers intact?	NONE	<input checked="" type="radio"/> YES	NO
3. Are Custody seals on sample containers intact?	<input checked="" type="radio"/> NONE	YES	NO
4. Is there a COC (Chain-of-Custody) present or other representative documents?		<input checked="" type="radio"/> YES	NO
5. Are the COC and bottle labels complete and legible?		<input checked="" type="radio"/> YES	NO
6. Is the COC in agreement with samples received? (IDs, dates, times, no. of samples, no. of containers, matrix, requested analyses, etc.)		<input checked="" type="radio"/> YES	NO
7. Were airbills / shipping documents present and/or removable?	DROP OFF	<input checked="" type="radio"/> YES	NO
8. Are all aqueous samples requiring preservation preserved correctly? (excluding volatiles)	N/A	<input checked="" type="radio"/> YES	NO
9. Are all aqueous non-preserved samples pH 4-9?	<input checked="" type="radio"/> N/A	YES	NO
10. Is there sufficient sample for the requested analyses?		<input checked="" type="radio"/> YES	NO
11. Were all samples placed in the proper containers for the requested analyses?		<input checked="" type="radio"/> YES	NO
12. Are all samples within holding times for the requested analyses?		<input checked="" type="radio"/> YES	NO
13. Were all sample containers received intact? (not broken or leaking, etc.)		<input checked="" type="radio"/> YES	NO
14. Are all samples requiring no headspace (VOC, GRO, Rx CN/S, radon), headspace free? Size of bubble: _____ < green pea _____ > green pea	N/A	<input checked="" type="radio"/> YES	NO
15. Were samples checked for and free from the presence of residual chlorine? (Applicable when PM has indicated samples are from a chlorinated water source; note if field preservation with sodium thiosulfate was not observed.)	<input checked="" type="radio"/> N/A	YES	NO
16. Were the samples shipped on ice?		<input checked="" type="radio"/> YES	NO
17. Were cooler temperatures measured at 0.1-6.0°C? IR gun used*: <input checked="" type="radio"/> #2 #4		RAD ONLY	<input checked="" type="radio"/> YES NO
Cooler #: <u>1</u>			
Temperature (°C): <u>2.4</u>			
No. of custody seals on cooler: <u>0</u>			
External µR/hr reading: <u>12</u>			
Background µR/hr reading: <u>12</u>			
Were external µR/hr readings ≤ two times background and within DOT acceptance criteria? <input checked="" type="radio"/> YES / NO / NA (If no, see Form 008.)			

Additional Information: PROVIDE DETAILS BELOW FOR A NO RESPONSE TO ANY QUESTION ABOVE, EXCEPT #1 AND #16.

→ Sample #1 (L 59449-01) recieved only 2 40 ml Vials for 8260? (V1)
~~or H3 (if these vials are for H3, they are according to the label, Preserved with HCL → 1 Liter poly pres. w. HNO3)~~
 Sample #2 (L 59449-01) recieved only 1 40ml Vial for Voc on H3?
~~Pres with HCL~~
 40 ml Vials for H3 Not preserved. Limited Vol. for H3

Discussed per. w/ client. Enough, just
 If applicable, was the client contacted? YES / NO / NA Contact: [Signature] Date/Time: _____
 Project Manager Signature / Date: [Signature] 10/23/06

*IR Gun #2: Oakton, SN 29922500201-0066

*IR Gun #4: Oakton, SN 2372220101-0002

SL 10-20-06



Paragon Analytics

Radiochemistry Case Narrative

Tritium


ACZ Laboratories, Inc.

L59449

Paragon WO 0610164


1. This report consists of the analytical results and supporting documentation for two water samples received by Paragon on 10/20/06.
2. These samples were prepared according to Paragon Analytics procedures SOP700R9. Modifications were made to this procedure per QASS 313613.
3. The samples were analyzed for the presence of tritium according to Paragon Analytics procedure SOP704R8. The analyses were completed on 11/13/06.
4. The analysis results for these samples are reported in units of pCi/L. The samples were not filtered prior to analysis.
5. No anomalous situations were encountered during the preparation or analysis of these samples. All quality control criteria were met.

The data contained in the following report have been reviewed and approved by the personnel listed below. In addition, Paragon Analytics certifies that the analyses reported herein are true, complete and correct within the limits of the methods employed.



Emily Walter
Radiochemistry Instrument Technician

11/16/06
Date



Radiochemistry Final Data Review

11/16/06
Date

1

PARAGON ANALYTICS, INC.
Radiochemistry Data Package

Section 1

**SAMPLE RESULTS
SUMMARY**

000002

Tritium Analysis By Liquid Scintillation Sample Results Summary

Client Name: ACZ Laboratories, Inc. **Laboratory Name:** Paragon Analytics **Page:** 1 of 1
Client Project Name: **PAI Work Order:** 0610164 **Reported on:** Tuesday, November 14, 2006
Client Project Number: L59449 **1:24:30 PM**

Lab Sample ID	Client Sample ID	Sample Type	Nuclide	Result +/- 2 s TPU	MDC	Units	Matrix	Prep Batch	Date Analyzed	Flags
0610164-1	L59449-01	Sample	H-3	-70 +/- 190	330	pCi/l	WATER	3H061106-1	11/12/06	U
0610164-2	L59449-02	Sample	H-3	-100 +/- 190	330	pCi/l	WATER	3H061106-1	11/12/06	U

Comments:

Data Package ID: H30610164-1

Qualifiers/Flags:
 U - Result is less than the sample specific MDC.
 LT - Result is less than Requested MDC, greater than sample specific MDC.
 Y1 - Chemical Yield is in control at 100-110%. Quantitative Yield is assumed.
 Y2 - Chemical Yield outside default limits.
 M - The requested MDC was not met.
 M3 - The requested MDC was not met, but the reported activity is greater than the reported MDC.

Abbreviations:
 TPU - Total Propagated Uncertainty (see PAI SOP 743)
 MDC - Minimum Detectable Concentration (see PAI SOP 709)
 BDL - Below Detection Limit

000003

2

PARAGON ANALYTICS, INC.
Radiochemistry Data Package

Section 2

**QC RESULTS
SUMMARY**

000004

Tritium Analysis By Liquid Scintillation

PAI 704 Rev 8

Method Blank Results

Lab Name: Paragon Analytics

Work Order Number: 0610164

Client Name: ACZ Laboratories, Inc.

ClientProject ID: L59449

Lab ID: 3H061106-1MB

Sample Matrix: WATER

Prep SOP: PAI 700 Rev 9

Date Collected: 06-Nov-06

Date Prepared: 06-Nov-06

Date Analyzed: 13-Nov-06

Prep Batch: 3H061106-1

QCBatchID: 3H061106-1-1

Run ID: 3H061106-1A

Count Time: 60 minutes

Final Aliquot: 10.0 ml

Result Units: pCi/l

File Name: Manual Entry

CASNO	Target Nuclide	Result +/- 2 s TPU	MDC	Lab Qualifier
10028-17-8	H-3	30 +/- 200	330	U

Comments:

Qualifiers/Flags:

U - Result is less than the sample specific MDC.

Y1 - Chemical Yield is in control at 100-110%. Quantitative Yield is assumed.

Y2 - Chemical Yield outside default limits.

LT - Result is less than Requested MDC, greater than sample specific MDC.

Abbreviations:

TPU - Total Propagated Uncertainty (see PAI SOP 743)

MDC - Minimum Detectable Concentration (see PAI SOP 709)

BDL - Below Detection Limit

M - Requested MDC not met.

B - Analyte concentration greater than MDC.

B3 - Analyte concentration greater than MDC but less than Requested MDC.

Data Package ID: H30610164-1

Tritium Analysis By Liquid Scintillation

PAI 704 Rev 8

Laboratory Control Sample(s)

Lab Name: Paragon Analytics

Work Order Number: 0610164

Client Name: ACZ Laboratories, Inc.

ClientProject ID: L59449

Lab ID: 3H061106-1LCS	Sample Matrix: WATER Prep SOP: PAI 700 Rev 9 Date Collected: 06-Nov-06 Date Prepared: 06-Nov-06 Date Analyzed: 13-Nov-06	Prep Batch: 3H061106-1 QCBatchID: 3H061106-1-1 Run ID: 3H061106-1A Count Time: 60 minutes	Final Aliquot: 9.90 ml Result Units: pCi/l File Name: Manual Entry
-----------------------	--	--	--

CASNO	Target Nuclide	Results +/- 2s TPU	MDC	Spike Added	% Rec	Control Limits	Lab Qualifier
10028-17-8	H-3	10200 +/- 1600	300	10700	95.3	85 - 115	P

Comments:

Qualifiers/Flags:

- U - Result is less than the sample specific MDC.
- LT - Result is less than Requested MDC, greater than sample specific MDC.
- Y1 - Chemical Yield is in control at 100-110%. Quantitative Yield is assumed.
- Y2 - Chemical Yield outside default limits.
- L - LCS Recovery below lower control limit.
- H - LCS Recovery above upper control limit.
- P - LCS Recovery within control limits.
- M - The requested MDC was not met.
- M3 - The requested MDC was not met, but thereported activity is greater than the reported MDC.

Abbreviations:

- TPU - Total Propagated Uncertainty (see PAI SOP 743)
- MDC - Minimum Detectable Concentration (see PAI SOP 709)

Data Package ID: H30610164-1

PARAGON ANALYTICS, INC.
Radiochemistry Data Package

3

Section 3

**INDIVIDUAL
SAMPLE RESULTS**

000007

Tritium Analysis By Liquid Scintillation

PAI 704 Rev 8
Sample Results

Lab Name: Paragon Analytics
Work Order Number: 0610164
Client Name: ACZ Laboratories, Inc.
ClientProject ID: L59449

Field ID: L59449-01 Lab ID: 0610164-1	Sample Matrix: WATER Prep SOP: PAI 700 Rev 9 Date Collected: 16-Oct-06 Date Prepared: 06-Nov-06 Date Analyzed: 12-Nov-06	Prep Batch: 3H061106-1 QCBatchID: 3H061106-1-1 Run ID: 3H061106-1A Count Time: 60 minutes Report Basis: Unfiltered	Final Aliquot: 10.0 ml Prep Basis: Unfiltered Moisture(%): 100.000 Result Units: pCi/l File Name: Manual Entry
--	--	--	--

CASNO	Target Nuclide	Result +/- 2 s TPU	MDC	Lab Qualifier
10028-17-8	H-3	-70 +/- 190	330	U

Comments:

Qualifiers/Flags:

- U - Result is less than the sample specific MDC.
- Y1 - Chemical Yield is in control at 100-110%. Quantitative Yield is assumed.
- Y2 - Chemical Yield outside default limits.
- LT - Result is less than Requested MDC, greater than sample specific MDC.
- M3 - The requested MDC was not met, but the reported activity is greater than the reported MDC.
- M - The requested MDC was not met.

Abbreviations:

- TPU - Total Propagated Uncertainty (see PAI SOP 743)
- MDC - Minimum Detectable Concentration (see PAI SOP 709)
- BDL - Below Detection Limit

Data Package ID: H30610164-1

Tritium Analysis By Liquid Scintillation

PAI 704 Rev 8
Sample Results

Lab Name: Paragon Analytics
Work Order Number: 0610164
Client Name: ACZ Laboratories, Inc.
ClientProject ID: L59449

Field ID: L59449-02 Lab ID: 0610164-2	Sample Matrix: WATER Prep SOP: PAI 700 Rev 9 Date Collected: 16-Oct-06 Date Prepared: 06-Nov-06 Date Analyzed: 12-Nov-06	Prep Batch: 3H061106-1 QCBatchID: 3H061106-1-1 Run ID: 3H061106-1A Count Time: 60 minutes Report Basis: Unfiltered	Final Aliquot: 10.0 ml Prep Basis: Unfiltered Moisture(%): 100.000 Result Units: pCi/l File Name: Manual Entry
--	--	--	--

CASNO	Target Nuclide	Result +/- 2 s TPU	MDC	Lab Qualifier
10028-17-8	H-3	-100 +/- 190	330	U

Comments:

Qualifiers/Flags:

- U - Result is less than the sample specific MDC.
- Y1 - Chemical Yield is in control at 100-110%. Quantitative Yield is assumed.
- Y2 - Chemical Yield outside default limits.
- LT - Result is less than Requested MDC, greater than sample specific MDC.
- M3 - The requested MDC was not met, but the reported activity is greater than the reported MDC.
- M - The requested MDC was not met.

Abbreviations:

- TPU - Total Propagated Uncertainty (see PAI SOP 743)
- MDC - Minimum Detectable Concentration (see PAI SOP 709)
- BDL - Below Detection Limit

Data Package ID: H30610164-1

PARAGON ANALYTICS, INC.
Radiochemistry Data Package

Section 4

4

RAW DATA

000010

Tritium Analysis By Liquid Scintillation Raw Data Report

Laboratory Name: Paragon Analytics
 PAI Work Order: 0610164

Prep SOP: PAI 700
 Analytical SOP: PAI 704

Reported on: Tuesday, November 14, 2006
 1:02:49 PM

Sample ID QC Type	Nuclide Type	Sample Date/Time	Prep Batch QCBatchID	Ingrwth Date /Time	Quench Factor %Lum	Matrix %Moist.	Samp Alq Analy Alq	Inst ID DetID	AnRunID File Name	Count Date/Time	GrossCPM BkgCPM	BaseEff ProgEff	CntDur(min) Yield	Activity +/- 2 s TPU	MDC DeclEv	ReportUnits ReportBasis	DER RPD	%Spk. Recov Flags
0610164-1	H-3	10/16/06	3H061106-1	NA	138.4	WATER	43 ml	L56000	3H061106-1A	11/12/06	7.720	23.86%	60	-70	330	pCi/l	NA	U
SMP	Trg. Analyte	9:40:00 AM	3H061106-1-1	NA	0.11	100	10 ml	27-06	Manual Entry	12:54 PM	8.080	NA	NA	190	NA	Unfiltered	NA	U
0610164-2	H-3	10/16/06	3H061106-1	NA	138.9	WATER	43 ml	L56000	3H061106-1A	11/12/06	7.550	23.86%	60	-100	330	pCi/l	NA	U
SMP	Trg. Analyte	2:40:00 PM	3H061106-1-1	NA	0.11	100	10 ml	27-06	Manual Entry	1:56 PM	8.080	NA	NA	190	NA	Unfiltered	NA	U
3H061106-1	H-3	11/6/06	3H061106-1	NA	138.9	WATER	50 ml	L56000	3H061106-1A	11/13/06	8.230	23.86%	60	30	330	pCi/l	NA	U
MB	Trg. Analyte	10:29:27 AM	3H061106-1-1	NA	0.14	NA	10 ml	27-07	Manual Entry	11:14 AM	8.080	NA	NA	200	NA	Unfiltered	NA	U
3H061106-1	H-3	11/6/06	3H061106-1	NA	138.2	WATER	50 ml	L56000	3H061106-1A	11/13/06	61.700	23.86%	60	10200	300	pCi/l	NA	95.3
LCS	Trg. Analyte	10:29:27 AM	3H061106-1-1	NA	0.07	NA	9.9 ml	27-08	Manual Entry	12:15 PM	8.080	NA	NA	1600	NA	Unfiltered	NA	P

Comments:

Data Package ID: H30610164-1

Qualifiers/Flags:

- U - Result is less than the sample specific MDC.
- Y1 - Chemical Yield is in control at 100-110%. Quantitative yield is assumed.
- Y2 - Chemical Yield outside default limits.
- W - DER is greater than Warning Limit of 1.42
- D - DER is greater than Control Limit of 2.13
- + - Duplicate RPD not within limits
- LT - Result is less than Request MDC, greater than sample specific MDC
- * - Analyte Basis is 'As Received' while the Report Basis is 'Dry Weight'
- # - Analyte Basis is 'Dry Weight' while the Report Basis is 'As Received'.

Notes:

- 1) The Tracer results are not yield corrected (i.e. activity measured not activity added).
- 2) Where sample time is not available, 12:00 PM (Mountain) is used for decay correction.

Abbreviations:

- TR - Tracer
- TA - Target Analyte
- TPU - Total Propagated Uncertainty (see PAI SOP 743)
- MDC - Minimum Detectable Concentration (see PAI SOP 709)
- DER - Duplicate Error Ratio
- BDL - Below Detection Limit

11 NOV 2006 20:37

ID: 3H: 10-ML GEO

USER: 8

COMMENT: L56000

PRESET TIME : 60.00
 DATA CALC : CPM H# : YES SAMPLE REPEATS: 1 PRINTER : STD
 COUNT BLANK : NO IC# : NO REPLICATES : 1 RS232 : OFF
 TWO PHASE : NO AGC : NO CYCLE REPEATS : 1
 SCINTILLATOR: LIQUID LUMEX: NO LOW SAMPLE REJ: 0
 LOW LEVEL : YES HALF LIFE CORRECTION DATE: none

CHAN: 50.0 - 250.0 %ERROR: 1.75 FACTOR: 1.000000 BKG. SUB: 0
 CHAN: 450.0 - 900.0 %ERROR: 20.00 FACTOR: 1.000000 BKG. SUB: 0

ALPHA-BETA DISCRIMINATION: NO

SAM NO	POS	TIME MIN	H#	WIND1		WIND2		LUMEX %	ELAPSED TIME
				CPM	%ERROR	CPM	%ERROR		
1	37-1	60.00	139.5	8.67	8.77	28.93	4.80	0.14	61.02
2	37-2	60.00	139.1	8.37	8.93	29.68	4.74	0.14	122.15
3	37-3	60.00	139.0	8.02	9.12	29.82	4.73	0.13	183.27
4	37-4	60.00	138.5	7.52	9.42	29.38	4.76	0.13	244.40
5	37-5	60.00	139.2	7.68	9.31	28.33	4.85	0.13	305.54
6	37-6	60.00	139.3	8.18	9.03	29.78	4.73	0.12	366.67
7	37-7	60.00	138.8	7.70	9.30	29.27	4.77	0.11	427.77
8	37-8	60.00	138.5	7.73	9.28	30.20	4.70	0.11	488.86
9	37-9	60.00	139.5	8.80	8.70	30.23	4.70	0.11	550.01
10	37-10	60.00	139.5	8.08	9.08	29.82	4.73	0.11	611.13
11	37-11	60.00	138.7	7.53	9.41	29.82	4.73	0.11	672.24
12	37-12	60.00	139.4	8.25	8.99	28.62	4.83	0.11	733.39
13	27-1	60.00	139.6	8.48	8.86	28.73	4.82	0.12	794.63
14	27-2	60.00	138.8	8.18	9.03	29.52	4.75	0.11	855.71
15	27-3	60.00	138.9	7.38	9.50	28.72	4.82	0.11	916.81
16	27-4	60.00	138.6	8.43	8.89	28.72	4.82	0.11	977.91
17	27-5	60.00	138.4	7.72	9.29	28.82	4.81	0.11	1039.01
18	27-6	60.00	138.9	7.55	9.40	27.93	4.89	0.11	1100.09
19	27-7	SAMPLE TERMINATED:							

OK
 W 11/14/06

000012

13 NOV 2006 11:14

ID: 34: 10-ML GEO

USER: 8 COMMENT: LS6000

PRESET TIME : 60.00
 DATA CALC : CPM H# : YES SAMPLE REPEATS: 1 PRINTER : STD
 COUNT BLANK : NO IC# : NO REPLICATES : 1 RS232 : OFF
 TWO PHASE : NO ADC : NO CYCLE REPEATS : 1
 SCINTILLATOR: LIQUID LUMEX: NO LOW SAMPLE REJ: 0
 LOW LEVEL : YES HALF LIFE CORRECTION DATE: none

CHAN: 50.0 - 250.0 %ERROR: 1.75 FACTOR: 1.000000 BKG. SUB: 0
 CHAN: 450.0 - 900.0 %ERROR: 20.00 FACTOR: 1.000000 BKG. SUB: 0

ALPHA-BETA DISCRIMINATION: NO

SAM NO	POS	TIME MIN	H#	WIND1		WIND2		LUMEX %	ELAPSED TIME
				CPM	%ERROR	CPM	%ERROR		
MISSING SAMPLE									
7	27-7	60.00	138.9	8.23	9.00	30.28	4.69	0.14	61.14
8	27-8	60.00	138.2	61.70	3.29	31.25	4.62	0.07	122.25
9	27-9	60.00	137.6	8.03	9.11	29.68	4.74	0.16	183.38

OK
11/14/06

LSC Run Log 310861 Instrument ID: LS 6000

Date	Sample ID	CountTime (min.)	Rack & Position	Test	User #	Batch ID	Position Check	Initials	Comments
11/11/06	34061106-1CB1	60	37 - 1	U3-10mL	8	34061106-1	MC	MC	NA
	0610133-1		22						
	2		23						
	3		24						
	4		25						
	5		26						
	6		27						
	7		28						
	8		29						
	9		30						
	10		31						
	11		32						
	12		33						
	13		34						
	14		35						
	15		36						
	16		37						
	17		38						
	18		39						
	19		40						
	20		41						
	21		42						
	22		43						
	23		44						
	24		45						
	25		46						
	26		47						
	27		48						
	28		49						
	29		50						
	30		51						

Reviewed by / Date GAN 11/14/06

TRITIUM RUN LOG

297984

Logbook/Page No.

SOP 700 Rev 9

Form 1306r1.doc (3/30/2003)

Date	Workorder/ Sample Number	Column ID	Flask ID	Run No.	Comments	Technician's Initials
11/6/06	0410133 -4	E	602	1	batch 3H06/106-1	OBC
	↓ -5	F	T2	↓		
	0410133 -6	A	T12	2		
	↓ -7	B	T5			
	↓ -8	C	103			
	0410141 -1	D	101			
	↓ -2	E	T8			
	↓ -3	F	600			
	↓ -4	A	T8	3		
	↓ -5	B	103			
	0410164 -1	C	602			
	↓ -2	D	603			
	0611014 -2	E	600			
	↓ -3 Dup	F	101	↓		
	↓ -4	A	T7	4		
	↓ -4 MS	B	T9			
	3H06/106-1 MB	C	T2			
	↓ CCS	D	T3	↓		
<div style="border: 1px solid black; width: 100px; height: 100px; transform: rotate(45deg); margin: 0 auto;"></div>						
				OBC		
				11/7/06		

PARAGON ANALYTICS, INC.
Radiochemistry Data Package

Section 5

QUALITY ASSURANCE
SUMMARY REPORTS

5

000016

QUALITY ASSURANCE SUMMARY SHEET

PAI W.O. # / BATCH FOR ALL REDUCED
TEST VOLUME
METHOD SAMPLES
SOP/REV (PREP) - PREP
SOP/REV (ANAL) _____

Briefly document any QA or other problems or deviations associated with the analysis of samples. Problems could result from: log-in, color, odor, dilution, consistency, scheduling, equipment, or instrumentation, or may include documentation of minor deviations necessary due to unique DQO's or sample characteristics.

DUE TO LIMITED SAMPLE VOLUME, A
REDUCED ALIQUOT WAS USED TO PREPARE
THESE SAMPLES.

[A large diagonal line is drawn across the remaining text area, indicating that no further notes are present.]

TECHNICIAN/ANALYST *[Signature]*

DATE 5/19/06

DEPARTMENT MANAGER *[Signature]*

DATE 5/19/06

PARAGON ANALYTICS, INC.
Radiochemistry Data Package

Section 6

**LABORATORY
BENCH SHEETS**

6

000018

Radiochemistry Instrument Worksheet

Prep Batch: 3H061106-1

Paragon Analytics

Prep Procedure: H3 10 - mL

Analytical QASS / NCR? Y (N) **NA**

Prep Num	LabID	QC Type	Init Alq	Fin Alq	Units	Cnt 1 File/Inst	Cnt 1 Rack-Pos	Cnt 1 Pos Chk By	Cnt 2 File/Inst	Cnt 2 Rack-Pos	Cnt 2 Pos Chk By	Cnt 3 File/Inst	Cnt 3 Rack-Pos	Cnt 3 Pos Chk By	Notes
1	0610133-1	SMP	50	10	ml	150000 39-2 MC	3								 <div style="text-align: right; margin-right: 50px;">3H061106-1</div>
1	0610133-2	SMP	50	10	ml		4								
1	0610133-3	SMP	50	10	ml		5								
1	0610133-3	DUP	50	10	ml		6								
1	0610133-4	SMP	50	10	ml		7								
1	0610133-5	SMP	50	10	ml		8								
1	0610133-6	SMP	50	10	ml		9								
1	0610133-7	SMP	50	10	ml		10								
1	0610133-8	SMP	50	10	ml		21								
1	0610141-1	SMP	50	10	ml		22								
1	0610141-2	SMP	50	10	ml		29-1								
1	0610141-3	SMP	50	10	ml		2								
1	0610141-4	SMP	50	10	ml		3								
1	0610141-5	SMP	50	10	ml		4								
1	0610164-1	SMP	43	10	ml		5								
1	0610164-2	SMP	43	10	ml		6								
1	061014-2	SMP	50	10	ml		7								
1	0611014-2	DUP	50	10	ml										
1	0611014-4	SMP	50	10	ml										
1	0611014-4	MS	50	9.900990	ml										
1	3H061106-1CB1	MB	50	10	ml		39-1 MC								
1	3H061106-1CB2	MB	50	10	ml		11								
1	3H061106-1CB3	MB	50	10	ml		27-9								
1	3H061106-1	MB	50	10	ml		27-7								
1	3H061106-1	LCS	50	9.900990	ml		18								

00001

Radiochemistry Instrument Worksheet

Prep Batch: 3H061106-1

Paragon Analytics

Prep Procedure: **H3**

Analytical QASS / NCR? Y **(N) NA**

Prep Num	LabID	QC Type	Init Alq	Fin Alq	Units	Cnt 1 File/Inst	Cnt 1 Rack-Pos	Cnt 1 Pos Chk By	Cnt 2 File/Inst	Cnt 2 Rack-Pos	Cnt 2 Pos Chk By	Cnt 3 File/Inst	Cnt 3 Rack-Pos	Cnt 3 Pos Chk By	Notes

Spike Solution Information

Soln. #	Nuclide	SolnID	Prep Conc	Units	Prep Date	Aliquot	Units	Pipet ID
S1	H-3	486.3020.23	2,383.682	DPM/ml	11/06/06	0.5	ml	ST-002































000020

Radiochemistry Instrument Worksheet

Prep Batch: 3H061106-1

Paragon Analytics

Sample Barcodes

0610164-1		0610141-1	
0610133-1		0610133-2	
0611014-2		0610164-2	
0610141-2		0610133-3	
0610141-3		0610133-4	
0611014-4		0610141-4	
0610133-5		0610141-5	
0610133-6		0610133-7	
0610133-8		0611014-2DUP	
0610133-3DUP		3H061106-1LCS	
3H061106-1LCS		3H061106-1CB1MB	
3H061106-1CB2MB		3H061106-1CB3MB	
3H061106-1MB		3H061106-1CB3MB	
3H061106-1CB2MB		3H061106-1CB1MB	
3H061106-1MB		0611014-4MS	

000021

Radiochemistry Prep Worksheet

Prep Batch: 3H061106-1

Paragon Analytics

Prep Procedure: H3

Reviewed By: DBC *DB* Review Date: 11/7/2006

Non-Routine Pre-Treatment? Y N Batch: *MLA*

Prep QASS / NCR? Y N *11/6/06*

Prep SOP: PAI 700 Rev: 9
 Prep SOP: NONE
 Matrix Class: liquid

Prep Analyst: Derek B. Caduff
 Prep Date: 11/6/2006
 Prep Dept: RS

Balance:
 Balance:

Cocktail: UG LLT
 Cocktail Pipet: T-002
 Aliquot Pipet: RS-007

Samp Num	Prep Num	LabID	QC Type	Dish No.	Init Alq ml	Fin Alq ml	Prep Basis	Water Added(ml)	Moisture(%)	Analysis Vol.(ml)	Standards	Prep Notes
1	1	0610133-1	SMP		50	10	Unfiltered	0	100	10		
2	1	0610133-2	SMP		50	10	Unfiltered	0	100	10		
3	1	0610133-3	SMP		50	10	Unfiltered	0	100	10		
4	1	0610133-3	DUP		50	10	Unfiltered	0	100	10		
5	1	0610133-4	SMP		50	10	Unfiltered	0	100	10		
6	1	0610133-5	SMP		50	10	Unfiltered	0	100	10		
7	1	0610133-6	SMP		50	10	Unfiltered	0	100	10		
8	1	0610133-7	SMP		50	10	Unfiltered	0	100	10		
9	1	0610133-8	SMP		50	10	Unfiltered	0	100	10		
10	1	0610141-1	SMP		50	10	Unfiltered	0	100	10		
11	1	0610141-2	SMP		50	10	Unfiltered	0	100	10		
12	1	0610141-3	SMP		50	10	Unfiltered	0	100	10		
13	1	0610141-4	SMP		50	10	Unfiltered	0	100	10		
14	1	0610141-5	SMP		50	10	Unfiltered	0	100	10		
15	1	0610164-1	SMP		43	10	Unfiltered	0	100	10		
16	1	0610164-2	SMP		43	10	Unfiltered	0	100	10		
17	1	061014-2	SMP		50	10	Unfiltered	0	100	10		
18	1	061014-2	DUP		50	10	Unfiltered	0	100	10		
19	1	061014-4	SMP		50	10	Unfiltered	0	100	10		
20	1	061014-4	MS		50	9.900990	Unfiltered	0	100	10	S1	
21	1	3H061106-1CB1	MB		50	10	Unfiltered	0	100	10		
22	1	3H061106-1CB2	MB		50	10	Unfiltered	0	100	10		
23	1	3H061106-1CB3	MB		50	10	Unfiltered	0	100	10		
24	1	3H061106-1	MB		50	10	Unfiltered	0	100	10		
25	1	3H061106-1	LCS		50	9.900990	Unfiltered	0	100	10	S1	

000022

Comments:
 UG LLT LOT #97-060701

Radiochemistry Prep Worksheet

Prep Batch: 3H061106-1

Paragon Analytics

Prep Procedure: H3

Reviewed By: DBC *DBC* Review Date: 11/7/2006

Non-Routine Pre-Treatment? Y / N Batch: *N/A* Re-Prep? Y / N Batch: *N/A* Prep QASS/NCR? Y / N *317613*

Prep SOP: PAI 700 Rev: 9
 Prep SOP: NONE
 Matrix Class: liquid

Prep Analyst: Derek B. Caduff
 Prep Date: 11/6/2006
 Prep Dept: RS

Balance: _____
 Balance: _____
 Cocktail: UG LLLT
 Cocktail Pipet: T-002
 Aliquot Pipet: RS-007

Samp Num	Prep Num	LabID	QC Type	Dish No.	Init Alq ml	Fin Alq ml	Prep Basis	Water Added(ml)	Moisture(%)	Analysis Vol.(ml)	Standards	Prep Notes

Spiked By: Derek B. Caduff Date: 11/6/2006

Witnessed By: Jeff Kujawa Date: 11/6/2006

Spike Solution Information

Soln #	Nuclide	SolnID	Prep Conc	Units	Prep Date	Aliquot Units	Pipet ID
S1	H-3	486.3020.23	2,383.682	DPM/ml	11/06/06	0.5 ml	ST-002

000029

Comments
 UG LLLT LOT #97-060701

Radiochemistry Prep Worksheet

Prep Batch: 3H061106-1

Paragon Analytics

Prep Batch Not Validated!!!

Prep Procedure: H3

Reviewed By:

Review Date:

Non-Routine Pre-Treatment? Y / N

Re-Prep? Y / N

Batch: *NA*

Prep QASS / NCR? N

313613

Prep SOP: PAI 700 Rev: 9

Prep Analyst: Derek B. Caduff *DK*

Balance:

Cocktail: UG LLT

Prep SOP: NONE

Prep Date: 11/6/2006

Balance:

Cocktail Pipet: T-002

Matrix Class: liquid

Prep Dept: RS

Aliquot Pipet: RS-007

Samp Num	Prep Num	LabID	QC Type	Dish No.	Init Alq sample	Fin Alq sample	Prep Basis	Water Added(ml)	Moisture(%)	Analysis Vol.(ml)	Standards	Prep Notes
1	1	0610133-1	SMP		50	10	Unfiltered	0	100	10		
2	1	0610133-2	SMP		50	10	Unfiltered	0	100	10		
3	1	0610133-3	SMP		50	10	Unfiltered	0	100	10		
4	1	0610133-3	DUP		50	10	Unfiltered	0	100	10		
5	1	0610133-4	SMP		50	10	Unfiltered	0	100	10		
6	1	0610133-5	SMP		50	10	Unfiltered	0	100	10		
7	1	0610133-6	SMP		50	10	Unfiltered	0	100	10		
8	1	0610133-7	SMP		50	10	Unfiltered	0	100	10		
9	1	0610133-8	SMP		50	10	Unfiltered	0	100	10		
10	1	0610141-1	SMP		50	10	Unfiltered	0	100	10		
11	1	0610141-2	SMP		50	10	Unfiltered	0	100	10		
12	1	0610141-3	SMP		50	10	Unfiltered	0	100	10		
13	1	0610141-4	SMP		50	10	Unfiltered	0	100	10		
14	1	0610141-5	SMP		50	10	Unfiltered	0	100	10		
15	1	0610164-1	SMP		50	10	Unfiltered	0	100	10		
16	1	0610164-2	SMP		50	10	Unfiltered	0	100	10		
17	1	0610142	SMP		50	10	Unfiltered	0	100	10		
18	1	0610142	DUP		50	10	Unfiltered	0	100	10		
19	1	0610144	SMP		50	10	Unfiltered	0	100	10		
20	1	0610144	MS		50	9.900990	Unfiltered	0	100	10	S1	
21	1	3H061106-1CB1	MB		50	10	Unfiltered	0	100	10		
22	1	3H061106-1CB2	MB		50	10	Unfiltered	0	100	10		
23	1	3H061106-1CB3	MB		50	10	Unfiltered	0	100	10		
24	1	3H061106-1	MB		50	10	Unfiltered	0	100	10		
25	1	3H061106-1	LCS		50	9.900990	Unfiltered	0	100	10	S1	

000024

Comments

UG LLT LOT #97-060701

Radiochemistry Prep Worksheet

Prep Batch: 3H061106-1

Paragon Analytics

Prep Procedure: H3

Prep Batch Not Validated!!!

Reviewed By:

Review Date:

Non-Routine Pre-Treatment? Y N

Batch: N/A

Prep QASS / NCR? Y N

Prep SOP: PAI 700 Rev: 9

Prep Analyst: Derek B. Caduff *DK*

Cocktail: UG LLT

Prep SOP: NONE

Prep Date: 11/6/2006

Cocktail Pipet: T-002

Matrix Class: liquid

Prep Dept: RS

Aliquot Pipet: RS-007

Samp Num	Prep Num	LabID	QC Type	Dish No.	Init Alq sample	Fin Alq sample	Prep Basis	Water Added(ml)	Moisture(%)	Analysis Vol.(ml)	Standards	Prep Notes

Spiked By: Derek B. Caduff *DK* Date: 11/6/06

Witnessed By: *DK* Date: 11/6/06

Soln #	Nuclide	SolnID	Prep Conc	Units	Prep Date	Aliquot Units	Pipet ID
S1	H-3	486.3020.23	2,383.682	DFM/ml	11/06/06	0.5 ml	ST-002

Ex 4/05/07

00000

Comments
UG LLT LOT #97-060701

SAMPLE CONDITION FORM (LIQUID)

ANALYST: *DBC*

ANALYSIS DATE: *11/6/06*

METHOD: *3H*

WORK ORDER	SAMPLE ID	SAMPLE CONDITION		
		pH	Color	Remarks
<i>0610133</i>	<i>1</i>	<i>~7</i>	<i>clear</i>	<i>N/A</i>
↓	<i>2</i>			
	<i>3</i>			
	<i>4</i>			
	<i>5</i>			
	<i>6</i>			
	<i>7</i>			
	<i>8</i>			
<i>0610141</i>	<i>1</i>			
↓	<i>2</i>			
	<i>3</i>			
	<i>4</i>			
	<i>5</i>			
<i>0610164</i>	<i>1</i>			
↓	<i>2</i>			
<i>0611014</i>	<i>2</i>			
↓	<i>4</i>			
<i>DBC</i> <i>11/6/06</i>				

PARAGON ANALYTICS, INC.
Radiochemistry Data Package

Section 7

**STANDARDS
TRACEABILITY
DOCUMENTS**



000027

PREPARE A WORKING DILUTION OF 486.1284.85 AT ≈ 2500 dpm/ml

1. ^{231.106} TRANSFER 486.1284.85 TO AN AMBER GLASS BOTTLE
 MASS OF BOTTLE w/o LID: 259.20g ^{BAL26}
 MASS OF BOTTLE + STD: 261.25g \downarrow
 NET MASS OF STD: 2.05g

2. ^{231.106} DILUTE TO 500 ml WITH DI WATER (SEE BELOW FOR ρ DETERMINATION)
 MASS OF BOTTLE + STD + H₂O: 760.0g ^{BAL26}
 MASS OF BOTTLE (from above): 259.20g \downarrow
 NET MASS OF DILUTED STD: 500.8g

3. DETERMINE DENSITY OF DI WATER
 MASS OF 100 ml VOLUMETRIC FLASK: 56.4421g ^{BAL12}
 MASS OF FLASK + H₂O: 156.1980g \downarrow
 NET MASS OF H₂O: 99.7559g
 $\rho = 0.997559$ g/ml

4. FINAL ACTIVITY CALCULATION

$$\frac{(486.1284.85 \text{ dpm/ml})}{500.8 \text{ g}} \times (0.9976 \text{ g/ml}) \times (2.05 \text{ g}) \times (2.22 \text{ dpm/g}) = 3871.3 \frac{\text{dpm}}{\text{ml}}$$
^{364/17/06}


Strid ID: 486.3020.23

Description: H-3
 Expiration: 4/15/07
 Activity: 3871.30 dpm/mL
 2s Uncertainty: 193.57 dpm/mL
 Ref. Date: 3/23/98
 Ref Time: N/A
 Prep Date: 3/24/06 Prep by: AF
 Matrix/Comp: DI H2O
 Half Life (y): 1.24E+01


Reverification Log		
Analysis Date	Initials	Expiration Date

Continued on Page

Read and Understood By


Signature

3/24/06
Date


Signature

4/17/06
Date

Tritium Spike Solution 486 1284.85

Continued From Page _____

Vol # 486 1284.84 was diluted to form a soln approx 100,000 dpm/mL as follows:

VOA tare 27.9711 g

VOA + HS (486 1284.84) 42.6011 g

Final (dil w/ DI H₂O) 62.7170 g

RG 9/16/05

Std ID: 486.1284.85

Description: 3H

Expiration: 8/11/06

Activity: 948004.38 dpm/mL

2s Uncertainty: 47400.22 dpm/mL

Ref. Date: 3/23/98

Ref Time: N/A

Prep Date: 5/3/00

Matrix/Comp. DI Water

Half Life (y): 1.23E+01

rep by: RTS

RG 9/16/05

dil factor

42.6011 g

- 27.9711 g

14.63 g STD

x 10.141876 pCi/g

14.837521 pCi

(62.7170 g - 27.9711 g)

427,029.4 pCi/g

RG 8/17/04

Std ID: 486.1284.85

Std ID: 486.1284.85

Description: H-3

Activity: 948004.38 dpm/mL

2s Uncertainty: 4740.02 dpm/mL

Ref. Date: 3/23/98

Ref Time: na

Prep Date: 5/3/00

Expiration: 3/25/03

Matrix/Comp. DI H₂O

Half Life (y): 1.23E+01

Prep by: RTS

RG 3/25/04

RG 3/28/03

RG 3/28/03

MAY REQUIRE NCR FOR ICPT WORK.

Description: 3H

Expiration: 7/30/05

Activity: 948004.38 dpm/mL

2s Uncertainty: 4740.02 dpm/mL

Ref. Date: 3/23/98

Ref Time: N/A

Prep Date: 7/29/04

Matrix/Comp. DI H₂O

Half Life (y): 1.23E+01

Prep by: JLK

RG 8/17/04

Std ID: 486.1284.85

Description: Tritium in Water Stock Solution

Activity: 427029 pCi/g or mL

Uncertainty: pCi/g or mL

Ref. Date: ~~3/30/98~~ 3/23/98 1-7-02

Ref Time: 10:00 AM 3/30/98

Prep Date: ~~3/30/98~~ 3/23/98

Expiration: ~~3/29/03~~ 3/23/98

Matrix/Comp. DI Water 3/29/03

Short Lived - Decay Correct with each use

Prep by: RTS

CEM 10/4/02

CEM 12/4/02

Verified 5/27/98 JUS

Standard re-verified 3/25/03. RG 3/28/03

Expires 3/25/04

Standard re-verified. New exp date: 7/30/05 RG 8/17/04

Continued on Page _____

Read and Understood By

Signature: [Signature] Date: 5/3/00

Signature: [Signature] Date: 5/22/98

PROJECT

Tritium Stock Solution 486-1284.84

Notebook No. 1101404
Continued From Page

PAI solution number 486-1284.84 (Analytical 51355543-BA7)

was transferred to a 40 ml TGA vial.

Certificate of Analysis not available on 3/30/98 - will obtain same.

Activity 51.054 uCi
Volume 50.34 grams
Matrix H₂O

RG 8/13/04
Std ID: 486.1284.84

Std ID: 486.1284.84

RG 8/13/04

Description: 3H
Expiration: 7/30/05
Activity: 1014184.00 pCi/g or ml
2s Uncertainty: 50709.00 pCi/g or ml
Ref. Date: 3/23/98
Ref Time: N/A
Prep Date: 7/29/04 Prep by: JLK
Matrix/Comp. DI Water
Half Life (y): 1.23E+01

Description: H-3
Activity: 1014184.00 pCi/g or ml
2s Uncertainty: 50709.00 pCi/g or ml
Ref. Date: 3/23/98
Ref Time: 0.417
Prep Date: 3/30/98 Prep by: RTS
Expiration: 7/3/04
Matrix/Comp. DI Water
Half Life (y): 1.24E+01

RG 8/13/04

RG 8/13/04

Std ID: 486.1284.84

Standard re-verified on 7/30/04. Expires 7/30/05.
RG 8/13/04

Description: Tritium in Water Stock Solution
Activity: 1014184 pCi/g or mL
Uncertainty: pCi/g or mL
Ref. Date: ~~3/23/98~~ 3/23/98 7-7-02
Ref Time: 10:00
Prep Date 3/30/98 Prep by: RTS
Expiration 3/29/03
Matrix/Comp. DI Water
Short Lived - Decay Correct with each use

Verified 5/21/98

Standard re-verified on 7/3/04
Expires 7/3/04

Continued on Page

[Signature]
Signed 5/5/00

Date

Read and Understood By

[Signature]

Signed

5/20/99
Date

000030



ANALYTICS

1380 Seaboard Industrial Blvd.
Atlanta, Georgia 30318 · U.S.A.

Phone (404) 352-8677
Fax (404) 352-2837

*PAID 48.
rec'd 3-24-98*

CERTIFICATE OF CALIBRATION

Standard Radionuclide Source

55543-307

H-3 in Water 50 mL in Flame Sealed Vial

This standard radionuclide source was prepared using an aliquot measured gravimetrically from a master radionuclide solution standard. The master radionuclide solution standard was calibrated by the Department Des Applications Et De La Metrologie Des Rayonnements Ionisants (DAMRI), Paris, France, as Number 24057. The calibration was checked by liquid scintillation counting after source preparation.

ANALYTICS maintains traceability to the National Institute of Standards and Technology through Measurements Assurance Programs as described in USNRC Reg. Guide 4.15, Revision 1.

ISOTOPE:	H-3
ACTIVITY (dps):	1.889 E+06
CALIBRATION DATE:	March 23, 1998 12:00 EST.
HALF-LIFE:	12.43 years
TOTAL UNCERTAINTY:	5.0%

50.34 grams of water.

P O NUMBER 21143, Item 1

PREPARED BY: *M. D. Currie*
M. D. Currie, Radiochemist

Q A APPROVED: *LCM 3/23/98*

PARAGON ANALYTICS, INC.
Radiochemistry Data Package

Section 8

CHAIN OF CUSTODY

8

000032

Paragon Analytics

Sample Number(s) Cross-Reference Table

Paragon OrderNum: 0610164

Client Name: ACZ Laboratories, Inc.

Client Project Name:

Client Project Number: L59449

Client PO Number:

Client Sample	Lab Sample Number	COC Number	Matrix	Date Collected	Time Collected
L59449-01	0610164-1		WATER	16-Oct-06	9:40
L59449-02	0610164-2		WATER	16-Oct-06	14:40

0610164

ACZ Laboratories, Inc.

CHAIN of CUSTODY

2773 Downhill Drive Steamboat Springs, CO 80487 (800) 334-5493

Report to:

Name: Tony Antalek
 Company: ACZ LABS, INC.
 E-mail: TAntalek@acz.com

Address: 2773 DOWNHILL DR
 STEAMBOAT SPGS, CO 80487
 Telephone: (970) 879-6590

Copy of Report to:

Name: N/A
 Company:

E-mail:
 Telephone:

Invoice to:

Name: YVONNE BAKER
 Company: ACZ LAS, INC.
 E-mail: yb@acz.com

Address: 2773 DOWNHILL DR
 STEAMBOAT SPGS, CO 80487
 Telephone: (970) 870-6590

If sample(s) received past holding time (HT), or if insufficient HT remains to complete analysis before expiration, shall ACZ proceed with requested short HT analyses? YES
 If "NO" then ACZ will contact client for further instruction. If neither "YES" nor "NO" is indicated, ACZ will proceed with the requested analyses, even if HT is expired, and data will be qualified. NO

PROJECT INFORMATION

ANALYSFS REQUESTED (attach list or use quote number)

Quote #:	Project/PO #:	Shipping Co.:	Tracking #:	Reporting state for compliance testing:	Are any samples NRC licensable material?	SAMPLE IDENTIFICATION	DATE:TIME	Matrix	# of Containers	GAMMA SCAN	H3								
1	L59449-01					L59449-01	10/16/2006 09:40	GW	3	<	<								
2	L59449-02					L59449-02	10/16/2006 14:40	GW	32	<	<								

Matrix SW (Surface Water) - GW (Ground Water) - WW (Waste Water) - DW (Drinking Water) - SL (Sludge) - SO (Soil) - OL (Oil) - Other (Specify)

REMARKS
 SUBCONTRACTED - Paragon Analytical, Attn: Lance Steere, 225 Commerce Drive, Ft. Collins, CO 80524 (970) 490-1511 - PLEASE SHIP UPS GROUND SERVICE
 SEND ONE EACH RRAD & TWO EACH VLUP PER SAMPLE

RELINQUISHED BY:	DATE:TIME	RECEIVED BY:	DATE:TIME	Page
T. ANTALEK	10/19/2006 1600	<i>Ann Zapp</i>	10-20-06 1110	Of

000034

CONDITION OF SAMPLE UPON RECEIPT FORM

Paragon Analytics

Client: ACZ
 Project Manager: LS

Workorder No: 0610164
 Initials: SL Date: 10-20-06

1. Does this project require any special handling in addition to standard Paragon procedures?	NONE	<input checked="" type="radio"/> YES	<input checked="" type="radio"/> NO
2. Are custody seals on shipping containers intact?	<input checked="" type="radio"/> NONE	<input checked="" type="radio"/> YES	<input checked="" type="radio"/> NO
3. Are Custody seals on sample containers intact?		<input checked="" type="radio"/> YES	<input checked="" type="radio"/> NO
4. Is there a COC (Chain-of-Custody) present or other representative documents?		<input checked="" type="radio"/> YES	<input checked="" type="radio"/> NO
5. Are the COC and bottle labels complete and legible?		<input checked="" type="radio"/> YES	<input checked="" type="radio"/> NO
6. Is the COC in agreement with samples received? (IDs, dates, times, no. of samples, no. of containers, matrix, requested analyses, etc.)		<input checked="" type="radio"/> YES	<input checked="" type="radio"/> NO
7. Were airbills / shipping documents present and/or removable?	DROP OFF	<input checked="" type="radio"/> YES	<input checked="" type="radio"/> NO
8. Are all aqueous samples requiring preservation preserved correctly? (excluding volatiles)	N/A	<input checked="" type="radio"/> YES	<input checked="" type="radio"/> NO
9. Are all aqueous non-preserved samples pH 4-9?	<input checked="" type="radio"/> N/A	<input checked="" type="radio"/> YES	<input checked="" type="radio"/> NO
10. Is there sufficient sample for the requested analyses?		<input checked="" type="radio"/> YES	<input checked="" type="radio"/> NO
11. Were all samples placed in the proper containers for the requested analyses?		<input checked="" type="radio"/> YES	<input checked="" type="radio"/> NO
12. Are all samples within holding times for the requested analyses?		<input checked="" type="radio"/> YES	<input checked="" type="radio"/> NO
13. Were all sample containers received intact? (not broken or leaking, etc.)		<input checked="" type="radio"/> YES	<input checked="" type="radio"/> NO
14. Are all samples requiring no headspace (VOC, GRO, Rx CN/S, radon), headspace free? Size of bubble: ___ < green pea ___ > green pea	N/A	<input checked="" type="radio"/> YES	<input checked="" type="radio"/> NO
15. Were samples checked for and free from the presence of residual chlorine? (Applicable when PM has indicated samples are from a chlorinated water source; note if field preservation with sodium thiosulfate was not observed.)	<input checked="" type="radio"/> N/A	<input checked="" type="radio"/> YES	<input checked="" type="radio"/> NO
16. Were the samples shipped on ice?		<input checked="" type="radio"/> YES	<input checked="" type="radio"/> NO
17. Were cooler temperatures measured at 0.1-6.0°C?	IR gun used*: <input checked="" type="radio"/> #2 #4	<input checked="" type="radio"/> YES	<input checked="" type="radio"/> NO
Cooler #: <u>1</u>			
Temperature (°C): <u>2.4</u>			
No. of custody seals on cooler: <u>0</u>			
External µR/hr reading: <u>12</u>			
Background µR/hr reading: <u>12</u>			
Were external µR/hr readings ≤ two times background and within DOT acceptance criteria? <input checked="" type="radio"/> YES / NO / NA (If no, see Form 008.)			

Additional Information: PROVIDE DETAILS BELOW FOR A NO RESPONSE TO ANY QUESTION ABOVE, EXCEPT #1 AND #16.

→ Sample #1 (L 59449-01) recieved only 2 40 ml Vials for ~~R260?~~ (u) or H3 (if these vials are for H3, they are according to the label, Preserved with HCL → 1 Liter poly pres. w. HND3)
 Sample #2 (L 59449-01) recieved only 1 40ml Vial for ~~Voc~~ or H3?
 ↓
 Pres with HCL

40 ml Vials for H3 Not preserved. Limited Vol. for H3
 → Discussed per. w/ client. Enough, just

If applicable, was the client contacted? YES / NO / NA Contact: [Signature] Date/Time: _____

Project Manager Signature / Date: [Signature] 10/20/06

*IR Gun #2: Oakton, SN 29922500201-0066 *IR Gun #4: Oakton, SN 2372220101-0002


2010-20-06

0610164

SHIPPING DEPARTMENT
(970)879-6590
ACZ LABORATORIES INC
2773 DOWNHILL DRIVE
STEAMBOAT SPRING CO 80487

15 LBS 1 OF 1

SHIP LANCE STEERE
(970)490-1511
TO: PARAGON ANALYTICAL LABS
225 COMMERCE DRIVE
FT COLLINS CO 80524




CO 805 0-01



2.4

UPS GROUND

TRACKING #: 1Z 810 130 03 9407 9509



BILLING: P/P

Unauthorized copying or reuse of any part of this page is prohibited. For more information, contact the Copyright Clearance Center, Inc., 222 Rosewood Drive, Danvers, MA 01923. Copyright © 2000 by UPS. All rights reserved. Printed in the USA. 0000036

DWS ID Lab

1110



1Z8101300394079509

000036

PARAGON ANALYTICS
Radiochemistry Data Package

Section 9

ADDITIONAL
SUPPORTING
DOCUMENTATION

9

000037

Liquid Scintillation Counter

Instrument Calibration

Initial Efficiency Calibration

Standards Traceability

009038

LS6000 10-mL geometry Tritium Background Determination

Interim control limits are established from the initial calibration for the geometry of interest. Limits are +/- 3 standard deviations from the initial unquenched calibration blank data. Once enough historical data is acquired, new historical limits are set as follows: Control limits for reagent blanks are established from 30 individual historical data points (10 batches). Limits are +/- 3 standard deviations from 30 individual historical data points. Individual reagent blanks and the average of reagent blanks from each batch are in control if the Count Rate (CPM) is within the established control limits.

CURRENTLY UNDER HISTORICAL LIMITS!

Updated on 9/25/2006. MBC

CT DATE	#	Sample ID	Count Duration (m)	Count Rate (CPM)	Total Cts.	Mean	Individual Reagent Blanks			Average of Reagent Blanks		
							LCL	UCL	Pass?	LCL	UCL	Pass?
11/11/06	7	3H061106-1CB1	60	8.67	520.20		4.81	11.03	PASS	4.81	11.03	PASS?
11/12/06	8	3H061106-1CB2	60	7.53	451.80	8.08	4.81	11.03	PASS	4.81	11.03	PASS
11/13/06	9	3H061106-1CB3	60	8.03	481.80		4.81	11.03	PASS	4.81	11.03	PASS

000039

10-mL Tritium "Window 2" Control Limits (LS 6000)

The background count rate is determined from the average of the reagent blanks for the batch.

Window 2 control limits are established using the average count rate from the three reagent blanks associated with each prep batch +/- 3X the estimated poisson uncertainty.

UPDATED 9/26/2006 MBC

CT DATE	Sample ID	Count Duration (min.)	Average count Duration (min.)	Count Rate (CPM)	Batch Average Reagent Blank	Lower Control Limit	Upper Control Limit	PASS/ FAIL
11/11/06	3H061106-1CB1	60		28.93				
11/12/06	3H061106-1CB2	60	60	29.82	29.48	27.37	31.58	PASS
11/13/06	3H061106-1CB3	60	60	29.68				

000040

10-mL Tritium Efficiency Calibration LS6000

10 mL sample + 10 mL Ultima Gold LLT
 9/18/2006
 LS6000


Standard used: 699.2613.19

94000.71 dpm/ml as of 9/3/1998
 1/2 life = 12.23 yrs.
 current activity = 59594.72 dpm/ml
 volume = 0.1961 ml
 Average Spike Activity = 11685.24 dpm

Sample ID	WIND1 cpm	WIND2 cpm	%LUMEX	SQP(E)	
0619009-S4	2859.570	27.830	0.00	141.40	
0619009-S5	2748.960	35.620	0.00	141.10	
0619009-S6	2781.280	30.210	0.00	140.70	
0619009-B4	8.050	29.050	0.17	140.80	
0619009-B5	8.770	30.400	0.17	141.00	
0619009-B6	7.700	29.180	0.19	139.30	
average LCS=	2796.603	29.543	0.09	140.72	averages
average bkg=	8.173				

net cpm=	WIND2 cpm	%LUMEX	SQP(E)	
2788.43	31.65	5.00	155.72	UCL
/known dpm= 11685.24	27.44	0.00	126.72	LCL
efficiency= 0.2386	See Tech. Mgr.	See Tech. Mgr.	Std. Addition	Corrective Action

Instrument Technician:  9/20/06
 Signature & Date

Supervisory Review:  9/25/06
 Signature & Date

³H Efficiency Calibration Verification / Method Blank Verification 09/18/06

Calibration Source Check

LS 6000
 Analysis Date: 9/18/2006
 Nuclide: ³H
 Half Life: 12.23

Calibration Check Source:

Spike Standard: 486.3020.23
 Reference Date: 3/23/1998
 Spiked DPM/mL: 3871.30
 Spike Volume: 0.5
 Spiked into: 50.0
 Current Spk. Act.: 10.82

Calibration Check Source Count

Sample ID	Position	Prep Date	Cnt. Dur.	Anal. Vol.	GrsCPM	BkgCPM	Efficiency	Activity	Units	Chem. Yield	LCS Recovery	Pass/Fail
0620022-ICV4	47-7	8/21/2006	60	9.90099	61.430	8.17	0.2386	10.20	pCi/mL	100%	94.2%	PASS
0620022-ICV5	47-8	8/21/2006	60	9.90099	59.720	8.17	0.2386	9.87	pCi/mL	100%	91.2%	PASS
0620022-ICV6	47-9	8/21/2006	60	9.90099	60.250	8.17	0.2386	9.97	pCi/mL	100%	92.1%	PASS

1 σ IU 0.056
 1 σ PU 0.051

Method Blank Check Source Count

Sample ID	Rack	Prep Date	Cnt. Dur.	Anal. Vol.	GrsCPM	BkgCPM	Efficiency	Chem. Yield	k (denom.)	activity	MDC	Pass/Fail	Units	2s CU	1 σ IU	1 σ PU	2s TPU
0620022-ICB4	47-10	8/21/2006	60	10	7.32	8.17	0.2386	100%	5.275	-0.1618	0.32	PASS	pCi/mL	0.193	-0.01812	-0.01650	0.194
0620022-ICB5	47-11	8/21/2006	60	10	7.6	8.17	0.2386	100%	5.275	-0.1087	0.32	PASS	pCi/mL	0.194	-0.01217	-0.01109	0.195
0620022-ICB6	47-12	8/21/2006	60	10	7.4	8.17	0.2386	100%	5.275	-0.1466	0.32	PASS	pCi/mL	0.193	-0.01642	-0.01495	0.194

000042

ID: 3H: 10-ML GEO.

18 SEP 2006 21:13

USER: 4

COMMENT: LS6000

PRESET TIME : 60.00

DATA CALC : CPM H# : YES SAMPLE REPEATS: 1 PRINTER : STD

COUNT BLANK : NO IC# : NO REPLICATES : 1 RS232 : OFF

TWO PHASE : NO ABC : NO CYCLE REPEATS : 1

SCINTILLATOR: LIQUID LUMEX: NO LOW SAMPLE REJ: 0

LOW LEVEL : YES HALF LIFE CORRECTION DATE: none

CHAN: 50.0 - 250.0 %ERROR: 1.75 FACTOR: 1.000000 BKS. SUB: 0

CHAN: 450.0 - 900.0 %ERROR: 20.00 FACTOR: 1.000000 BKS. SUB: 0

ALPHA-BETA DISCRIMINATION: NO

SAM NO	POS	TIME MIN	H#	WIND1		WIND2		LUMEX %	ELAPSED TIME
				CPM	%ERROR	CPM	%ERROR		
1	47-1	4.60	141.4	2859.57	1.74	27.83	17.68	0.00	5.15
2	47-2	4.80	141.1	2748.96	1.74	35.62	15.29	0.00	10.63
3	47-3	4.70	140.7	2781.28	1.75	30.21	16.78	0.00	15.99
4	47-4	60.00	140.8	8.05	9.10	29.05	4.79	0.17	77.11
5	47-5	60.00	141.0	8.77	8.72	30.40	4.68	0.17	138.22
6	47-6	60.00	139.3	7.70	9.30	29.18	4.78	0.19	199.36
7	47-7	60.00	138.5	61.43	3.29	27.65	4.91	0.08	260.47
8	47-8	60.00	138.0	59.72	3.34	27.70	4.91	0.07	321.61
9	47-9	60.00	138.0	60.25	3.33	28.00	4.88	0.07	382.71
10	47-10	60.00	140.6	7.32	9.55	29.00	4.79	0.13	443.85
11	47-11	60.00	138.8	7.60	9.37	27.37	4.94	0.13	504.98
12	47-12	60.00	138.7	7.40	9.49	29.22	4.78	0.12	566.13

MC
9/19/06

000043

Radiochemistry Instrument Worksheet

Prep Batch: H0607125

Paragon Analytics

Analytical QASS / NCR? Y / N

Prep Procedure: H3 SINGLE-POINT CALIBRATION (5ML/10ML)

Prep Num	LabID	QC Type	Init Aliq	Fin Aliq	Units	Cnt 1 File/Inst	Cnt 1 Rack-Pos	Cnt 1 Chk By	Cnt 2 File/Inst	Cnt 2 Rack-Pos	Cnt 2 Chk By	Cnt 3 File/Inst	Cnt 3 Rack-Pos	Cnt 3 Chk By	Notes
1	0619009-B1	MB	50	5	ml										
1	0619009-B2	MB	50	5	ml										
1	0619009-B3	MB	50	5	ml										
1	0619009-B4	MB	50	10	ml										
1	0619009-B5	MB	50	10	ml										
1	0619009-B6	MB	50	10	ml										
1	0619009-S1	LCS	50	4.901961	ml										
1	0619009-S2	LCS	50	4.901961	ml										
1	0619009-S3	LCS	50	4.901961	ml										
1	0619009-S4	LCS	50	9.803922	ml										
1	0619009-S5	LCS	50	9.803922	ml										
1	0619009-S6	LCS	50	9.803922	ml										

Soln #	Nuclide	ScintID	Prep Conc	Units	Prep Date	Aliquot	Units	Pipet ID
S1	H-3	699.2613.19	60,437.988	DPM/ml	07/12/06	1	ml	RS-006

Sample Barcodes

0619009-S6LCS		0619009-S5LCS	
0619009-S4LCS		0619009-S3LCS	
0619009-S2LCS		0619009-S1LCS	
0619009-B6MB		0619009-B5MB	
0619009-B4MB		0619009-B3MB	
0619009-B2MB		0619009-B1MB	

Radiochemistry Prep Worksheet

Prep Batch: 3106071216

Paragon Analytics

Prep Procedure: H3 Reviewed By: JRK/RL Review Date: 7/13/2006

Non-Routine Pre-Treatment? Y / N Batch: NA Re-Prep? Y / N Prep QASS / NCR? Y / N Cocktall: UG LLT
 Prep Analyst: Jeff Kujawa Balance: Cocktall Pipet: T-002
 Prep Date: 7/12/2006 Balance: Aliquot Pipet: Prep Dept: RS

Stamp	Prep Num	LabID	QC Type	Dish No.	Init Alq ml	Fin Alq ml	Prep Basis	Water Added (ml)	Moisture (%)	Analysis Vol. (ml)	Standards	Prep Notes
	1	0619009-B1	MB		50	5	Unfiltered	0	100	5		
	2	0619009-B2	MB		50	5	Unfiltered	0	100	5		
	3	0619009-B3	MB		50	5	Unfiltered	0	100	5		
	4	0619009-B4	MB		50	10	Unfiltered	0	100	10		
	5	0619009-B5	MB		50	10	Unfiltered	0	100	10		
	6	0619009-B6	MB		50	10	Unfiltered	0	100	10		
	7	0619009-S1	LCS		50	4.901961	Unfiltered	0	100	5	S1	
	8	0619009-S2	LCS		50	4.901961	Unfiltered	0	100	5	S1	
	9	0619009-S3	LCS		50	4.901961	Unfiltered	0	100	5	S1	
	10	0619009-S4	LCS		50	9.803922	Unfiltered	0	100	10	S1	
	11	0619009-S5	LCS		50	9.803922	Unfiltered	0	100	10	S1	
	12	0619009-S6	LCS		50	9.803922	Unfiltered	0	100	10	S1	

Spiked By: Jeff Kujawa Date: 7/12/2006
 Witnessed By: Kimberly Arnett Date: 7/12/2006

Spiked Solution Information

Soln #	Nuclide	SolnID	Prep Conc	Units	Prep Date	Aliquot	Units	Pipet ID
S1	H-3	699.2613.19	60,437.998	CPM/ml	07/12/06	1	ml	RS-006

Comments: UG LLT LOT# 97-060101
 Page 1 of 1 H3 Bench Sheet Supersedes: 21KPC 15.78
 Date Printed: 7/13/2006 10:06 LIMS Version: 5.357A

000046

Radiochemistry Prep Worksheet

Prep Batch# 3H06071235

Paragon Analytics

Prep Procedure: H3
 Re-Prep? Y / N Batch: _____
 Re-Prep? Y / N Batch: _____
 Reviewed By: _____
 Review Date: _____

Prep Batch Not Validated!!!

Non-Routine Pre-Treatment? Y / N Batch: _____
 Prep Analyst: Jeff Kujawa *JK*
 Prep Date: 7/12/2006
 Prep Dept: RS
 Balance: _____
 Cocktail: UG LLT
 Cocktail Pipet: T-002
 Aliquot Pipet: _____

Samp Num	Prep Num	LabID	QC Type	Dish No.	Init Aliq ml	Flt Aliq ml	Prep Basis	Water Added(ml)	Moisture(%)	Analysis Vol.(ml)	Standards	Prep Notes
1	1	0619009-B1	MB		50	5	Unfiltered	0	100	5		
2	1	0619009-B2	MB		50	5	Unfiltered	0	100	5		
3	1	0619009-B3	MB		50	5	Unfiltered	0	100	5		
4	1	0619009-B4	MB		50	10	Unfiltered	0	100	10		
5	1	0619009-B5	MB		50	10	Unfiltered	0	100	10		
6	1	0619009-B6	MB		50	10	Unfiltered	0	100	10		
7	1	0619009-S1	LCS		50	4.901961	Unfiltered	0	100	5	S1	
8	1	0619009-S2	LCS		50	4.901961	Unfiltered	0	100	5	S1	
9	1	0619009-S3	LCS		50	4.901961	Unfiltered	0	100	5	S1	
10	1	0619009-S4	LCS		50	9.603922	Unfiltered	0	100	10	S1	
11	1	0619009-S5	LCS		50	9.603922	Unfiltered	0	100	10	S1	
12	1	0619009-S6	LCS		50	9.603922	Unfiltered	0	100	10	S1	

Spiked By: *JK* Date: 7/12/06
 Witnessed By: *JK* Date: 7/12/06

Soln #	Nuclide	SolnID	Prep Conc	Units	Prep Date	Aliquot Units	Pipet ID
S1	H-3	699.2613.19	60,437.998	DFM/ml	07/12/06	1 ml	RS-006

exp. 6/11/06

0000

Comments
 UG LLT LOT# 97-0607101

Supersedes: *KA*

TRITIUM RUN LOG

297950

Logbook/Page No.

SOP 700 Rev 9

Form 1306r1.doc (3/30/2003)

Date	Workorder/ Sample Number	Column ID	Flask ID	Run No.	Comments	Technician's Initials
7/11/06	0607032 -BMS	B	T4	1	Batch 3H060711-1	VBA
	-16	C	T3			
	-6ms	D	604			
	-16ms	F	608			
	3H060711-1 -MB	A	T11	2		
	-CS	B	101			
7/11/06	3H060711-1 LCSD	C	608	2	Batch 3H060711-1	VBA
7/12/06	0619009 -S1, S4	A	103	1	Batch 3H060712-5	JLK
	-S2, S5	B	600			
	-S3, S6	C	609			
	-D1, B4	D	T9			
	-B2, B5	F	T10			
	-B3, B6	A	T12	2		
<div style="border: 1px solid black; width: 100px; height: 100px; transform: rotate(45deg); margin: 0 auto; display: flex; align-items: center; justify-content: center;"> JLK 7/13/06 </div>						

Reviewed by/Date JLK 7/13/06

000048

Radiochemistry Instrument Worksheet

Prep Batch: 3H06082 ES

Paragon Analytics

Prep Procedure: H3 SINGLE POINT ICB's/ICV's (5ML(10ML)) Analytical QASS / NCR? Y / N

Prep Num	LabID	QC Type	Init Alq	Fin Alq	Units	Cnt 1 File/Inst	Cnt 1 Rack-Pos	Cnt 1 Chk By	Cnt 2 File/Inst	Cnt 2 Rack-Pos	Cnt 2 Chk By	Cnt 3 File/Inst	Cnt 3 Rack-Pos	Cnt 3 Chk By	Notes
1	0620022-ICB1	MB	50	5	ml										
1	0620022-ICB2	MB	50	5	ml										
1	0620022-ICB3	MB	50	5	ml										
1	0620022-ICB4	MB	50	10	ml		LS-6000	47-10							MC
1	0620022-ICB5	MB	50	10	ml			11							
1	0620022-ICB6	MB	50	10	ml			12							
1	0620022-ICV1	LCS	50	4.950495	ml										
1	0620022-ICV2	LCS	50	4.950495	ml										
1	0620022-ICV3	LCS	50	4.950495	ml										
1	0620022-ICV4	LCS	50	9.900990	ml		LS-6000	47-7							MC
1	0620022-ICV5	LCS	50	9.900990	ml			18							
1	0620022-ICV6	LCS	50	9.900990	ml			19							

Spike Solution Information

Soln #	Nuclide	SolnID	Prep Conc	Units	Prep Date	Aliquot Units	Pipet ID
S1	H-3	486.3020.23	2,412.044	DPM/ml	08/21/06	0.5	ST-002

Sample Barcodes

0620022-ICV6LCS		0620022-ICV5LCS	
0620022-ICV4LCS		0620022-ICV3LCS	
0620022-ICV2LCS		0620022-ICV1LCS	
0620022-ICB6MB		0620022-ICB5MB	
0620022-ICB4MB		0620022-ICB3MB	
0620022-ICB2MB		0620022-ICB1MB	

Radiochemistry Prep Worksheet

Prep Batch: 3H060821-3

Paragon Analytics

Prep Procedure: H3

Reviewed By: JRK / *huc* Review Date: 8/21/2006

Non-Routine Pre-Treatment? Y / N / Batch: *LA* Re-Prep? Y / N / Prep QASS / NCR? Y / N / *LA*

Prep Analyst: Jeff Kujawa Balance: Cocktali: UG LLT
 Prep Date: 8/21/2006 Balance: Cocktail Pipet: T-002
 Prep Dept: RS Aliquot Pipet:

Sample Num	Prep Num	LabID	QC Type	Dish No.	Init Aliq ml	Fin Aliq ml	Prep Basis	Water Added(ml)	Moisture(%)	Analysis Vol.(ml)	Standards	Prep Notes
1	1	0620022-ICB1	MB		50	5	Unfiltered	0	100	5		<i>LA</i>
2	1	0620022-ICB2	MB		50	5	Unfiltered	0	100	5		
3	1	0620022-ICB3	MB		50	5	Unfiltered	0	100	5		
4	1	0620022-ICB4	MB		50	10	Unfiltered	0	100	10		
5	1	0620022-ICB5	MB		50	10	Unfiltered	0	100	10		
6	1	0620022-ICB6	MB		50	10	Unfiltered	0	100	10		
7	1	0620022-ICV1	LCS		50	4.900495	Unfiltered	0	100	5	S1	
8	1	0620022-ICV2	LCS		50	4.950495	Unfiltered	0	100	5	S1	
9	1	0620022-ICV3	LCS		50	4.950495	Unfiltered	0	100	5	S1	
10	1	0620022-ICV4	LCS		50	9.900990	Unfiltered	0	100	10	S1	
11	1	0620022-ICV6	LCS		50	9.900990	Unfiltered	0	100	10	S1	
12	1	0620022-ICV6	LCS		50	9.900990	Unfiltered	0	100	10	S1	

Spiked By: Jeff Kujawa Date: 8/21/2006

Witnessed By: Derek B. Caduff Date: 8/21/2006

Soln #	Nuclide	SolnID	Prep Conc	Units	Prep Date	Aliquot	Units	Flask ID
S1	H-3	486.3020.23	2,412.044	DPM/ml	08/21/06	0.5	ml	ST-002

Comments
UG LLT LOT# 97-060101

Radiochemistry Prep Worksheet

Prep Batch: 3H0608213

Paragon Analytics

Prep Procedure: H3

Prep Batch Not Validated!!!

Reviewed By: _____

Review Date: _____

Non-Routine Pre-Treatment? Y / N Batch: _____ Re-Prep? Y / N Batch: _____ Prep QASS / NCR? Y / N _____

Prep SOP: PAI 700 Rev: 9

Prep Analyst: Jeff Kujawa *JK*

Prep Date: 8/21/2006

Prep Dept: RS

Cocktail: UG LLT

Cocktail Pipet: T-002

Aliquot Pipet: _____

Matrix Class: liquid

Stamp	Prep Num	LabID	QC Type	Dish No.	Init Aliq ml	Fin Aliq ml	Prep Basis	Water Added(ml)	Moisture(%)	Analysis Vol.(ml)	Standards	Prep Notes
	1	0620022-ICB1	MB	50	5	5	Unfiltered	0	100	5		
	2	0620022-ICB2	MB	50	5	5	Unfiltered	0	100	5		
	3	0620022-ICB3	MB	50	5	5	Unfiltered	0	100	5		
	4	0620022-ICB4	MB	50	10	10	Unfiltered	0	100	10		
	5	0620022-ICB5	MB	50	10	10	Unfiltered	0	100	10		
	6	0620022-ICB6	MB	50	10	10	Unfiltered	0	100	10		
	7	0620022-ICV1	LCS	50	4.950495	4.950495	Unfiltered	0	100	5	S1	
	8	0620022-ICV3	LCS	50	4.950495	4.950495	Unfiltered	0	100	5	S1	
	9	0620022-ICV4	LCS	50	9.900990	9.900990	Unfiltered	0	100	10	S1	
	10	0620022-ICV5	LCS	50	9.900990	9.900990	Unfiltered	0	100	10	S1	
	11	0620022-ICV6	LCS	50	9.900990	9.900990	Unfiltered	0	100	10	S1	
	12	0620022-ICV6	LCS	50	9.900990	9.900990	Unfiltered	0	100	10	S1	

Spiked By: *JK* Date: 8/21/06
 Witnessed By: *DBL* Date: 8/21/06

Spike Solution Information							
Soln #	Nuclide	SolnID	Prep Conc	Units	Prep Date	Aliquot Units	Pipet ID
S1	H-3	488.3020.23	2.412.044	DFM/ml	08/21/06	0.5	ST-002

EXP: 9/13/07

0000

Comments
 UG LLT LOT# 97-060101

Page 1 of 1 H3 Bench Sheet
 Date Printed: 8/21/2006 14:09

Paragon Analytics
 LIMS Version: 5.415A

Supersedes: *kt*

250000

Paragon Analyti

TRITIUM RUN LOG

297959

Logbook/Page No.

SOP 700 Rev 9

Form 1306r1.doc (3/30/2003)

Technician's Initials

Comments

Run No.

Flask ID

Column ID

Sample Number

Date

8/21/04	0607066-6/MS	B	604	2	Batch 3/H0608D1-1	DRB
---------	--------------	---	-----	---	-------------------	-----

	-16	C	111			
--	-----	---	-----	--	--	--

	-11	D	T3			
--	-----	---	----	--	--	--

	3/H0608D1-1/MS	E	102			
--	----------------	---	-----	--	--	--

	-1255	F	101			
--	-------	---	-----	--	--	--

8/11/04	0607066-6/MS	A	T2	3	Batch 3/H0608D1-3	MR
---------	--------------	---	----	---	-------------------	----

	-ECS3	B	T4			
--	-------	---	----	--	--	--

	-ECS3	C	600			
--	-------	---	-----	--	--	--

	-ECS3	D	609			
--	-------	---	-----	--	--	--

	-ECS3	E	101			
--	-------	---	-----	--	--	--

	-ECS3	F	111			
--	-------	---	-----	--	--	--

--	--	--	--	--	--	--

Reviewed by / Date

8/21/04

MR

MR
8/21/04

Prepare a working dilution at approx 75,000 dpm/ml by diluting std 699.2613.18 with DI water.

Prepared by RG 9/16/05
Std ID: 699.2613.19
Description: 3H
Expiration: 8/11/06
Activity: 94000.71 dpm/mL
2s Uncertainty: 676.81 dpm/mL
Ref. Date: 9/3/98
Ref Time: NIA 9/24/08
Prep Date: 7/29/04 rep by: JLK
Matrix/Comp. DI Water 8/12/06
Half Life (y): 1.23E+01
Requires NCR for ICPT work. RG 9/16/05

1) See Density on page 18 of logbook 2613

2) Transfer std. 699.2613.18 to a 125 mL Amber glass bottle

Mass of bottle w/o lid 94.7389g Bal #12
Mass of bottle + std 96.6785g
Net mass of std. 1.9396g

3) Add DI water to final dilution

Reverification Log
Analysis Date Initials Expiration Date
9/24/06 RG 9/24/07

Mass of bottle w/o lid from above 94.7389g Bal #12
Mass of bottle, std, + DI water 192.4026g
Net mass of std, + DI water 97.6637g

4) Final Activity Calculation

$$(634.7 \text{ kBq/g}) (4.8151 \text{ g}) (60 \text{ dpm/Bq}) (1000 \text{ Bq/kBq}) (0.9958 \text{ g/mL}) (1.9396 \text{ g})$$

$$(38.5765 \text{ g}) (97.6637 \text{ g})$$

$$= 94,000 \pm 71 \text{ DPM/mL}$$

Std ID: 699.2613.19

Std ID: 699.2613.19

Description: 3H
Expiration: 7/28/04
Activity: 94000.71 dpm/mL

* NCR for ICPT work SD 8/11/04

2s Uncertainty: 676.81 dpm/mL
Ref. Date: 9/3/98
Ref Time: NIA
Prep Date: 7/28/03 Prep by: JRK
Matrix/Comp. DI Water
Half Life (y): 1.23E+01

Description: 3H
Expiration: 7/30/05
Activity: 94000.71 dpm/mL

2s Uncertainty: 676.81 dpm/mL
Ref. Date: 9/3/98
Ref Time: NIA 9/24/06
Prep Date: 7/29/04 7/23/05 Prep by: JLK
Matrix/Comp. DI Water
Half Life (y): 1.23E+01

Signed: *Chad Wayne* SD 8/11/04 Date: 7/28/03

Signed: *[Signature]* Date: 8/11/03

Prepare a 1st dilution of Ampoule 699 by diluting with DI water.

1) Determine the density of DI water

Mass of empty Class A ^{100 mL} volumetric flask	107.7956 g	Bal #
Mass of flask + 100 mL of DI water	107.3755 g	
Net mass of DI water	99.5799 g	

$$\rho = 0.9958 \text{ g/mL}$$

2) Transfer contents of ampoule 699 into a 40 mL Amber glass VOA vial

Mass of VOA vial w/o lid	24.3883g	Bal #
Mass of opened ampoule before transfer	40.4132g	
Mass of opened ampoule after transfer	36.0281g	
Net mass of Std transferred	4.8151g	

3) Add DI water to final dilution

Mass of VOA vial ^{w/ lid} from above	24.3883g	Bal #
Mass of VOA vial, std, + DI water	62.9648g	
Net mass of Std + DI water	38.5765g	

4) Final Activity Calculation

$$(634.7 \text{ kBq/g}) (4.8151 \text{ g}) \left(\frac{600 \text{ dpm}}{37 \text{ kBq}} \right) \left(\frac{1000 \text{ Bq}}{1 \text{ kBq}} \right) (0.9958 \text{ g/mL})$$

$$38.5765 \text{ g}$$

$$= 14,733,412.57 \frac{\text{dpm}}{\text{mL}}$$

U.S. Department of Commerce
National Institute of Standards
and Technology
SRM 4927F
Hydrogen-3
<4 MBq in distilled water

CAUTION
RADIOACTIVE



Continued on Page

Chad Wajsb
Signed

7/28/03
Date

has read / understood By
[Signature]
Signed

7/31/03
Date



National Institute of Standards & Technology

Certificate

PAF ID 0699
recd 5-09-03

Standard Reference Material 4927F Hydrogen-3 Radioactivity Standard

This Standard Reference Material (SRM) consists of radioactive hydrogen-3, as water, in 5 mL of distilled water. The solution is contained in a flame-sealed NIST borosilicate-glass ampoule. The SRM is intended for the calibration of beta-particle counting instruments and for the monitoring of radiochemical procedures.

Radiological Hazard

The SRM ampoule contains hydrogen-3 with a total activity of approximately 3.2 MBq. Hydrogen-3 decays by beta-particle emission. None of the beta particles escape from the SRM ampoule. During the decay process no photons are emitted. Approximate unshielded dose rates at several distances (as of the reference time) are given in note [a]*. There is no detectable external radiation. The SRM should be used only by persons qualified to handle radioactive material.

Chemical Hazard

The SRM ampoule contains only distilled water. There is no chemical hazard. If the ampoule is to be opened to transfer the solution, the recommended procedure is given on page 2.

Storage and Handling

The SRM should be stored and used at a temperature between 5 and 65 °C. The solution in an unopened ampoule should remain stable and homogeneous until at least September 2008.

The ampoule (or any subsequent container) should always be clearly marked as containing radioactive material. If the ampoule is transported it should be packed, marked, labeled, and shipped in accordance with the applicable national, international, and carrier regulations. The solution in the ampoule is a dangerous good (hazardous material) because of the radioactivity.

Preparation

This Standard Reference Material was prepared in the Physics Laboratory, Ionizing Radiation Division, Radioactivity Group, L.R. Karam, Group Leader. The overall technical direction and physical measurements leading to certification were provided by L.L. Lucas and M.P. Unterweger of the Radioactivity Group.

The support aspects involved in the preparation, certification, and issuance of this SRM were coordinated through the Standard Reference Materials Program by J.W.L. Thomas.

Bert M. Coursey, Chief
Ionizing Radiation Division

Nancy M. Trahey, Chief
Standard Reference Materials Program

Gaithersburg, Maryland 20899
June 1999
Half-life and text revised October 2000

Recommended Procedure for Opening the SRM Ampoule

- 1) If the SRM solution is to be diluted, it is recommended that the diluting solution have a composition comparable to that of the SRM solution.
- 2) Wear eye protection, gloves, and protective clothing and work over a tray with absorbent paper in it. Work in a fume hood.
- 3) Shake the ampoule to wet all of the inside surface of the ampoule. Return the ampoule to the upright position.
- 4) Check that all of the liquid has drained out of the neck of the ampoule. If necessary, gently tap the neck to speed the process.
- 5) Holding the ampoule upright, score the narrowest part of the neck with a scribe or diamond pencil.
- 6) Lightly wet the scored line. This reduces the crack propagation velocity and makes for a cleaner break.
- 7) Hold the ampoule upright with a paper towel, a wiper, or a support jig. Position the scored line away from you. Using a paper towel or wiper to avoid contamination, snap off the top of the ampoule by pressing the narrowest part of the neck away from you while pulling the tip of the ampoule towards you.
- 8) Transfer the solution from the ampoule using a pycnometer or a pipet with dispenser handle. NEVER PIPETTE BY MOUTH.
- 9) Seal any unused SRM solution in a flame-sealed glass ampoule, if possible, to minimize the evaporation loss.

See also reference [4]*.

PROPERTIES OF SRM 4927F

Certified values

Solution density	$(0.998 \pm 0.002) \text{ g} \cdot \text{mL}^{-1}$ at 20.0 °C [b]*
Radionuclide	Hydrogen-3
Reference time	1200 EST, 3 September 1998
Massic activity of the solution [c]	$634.7 \text{ kBq} \cdot \text{g}^{-1}$
Relative expanded uncertainty ($k=2$)	0.72% [d] [e]

Uncertified values

Physical Properties:			
Source description	Liquid in flame-sealed NIST borosilicate-glass ampoule		
Ampoule specifications	Body outside diameter	(16.5 ± 0.5) mm	
	Wall Thickness	(0.60 ± 0.04) mm	
	Barium content	Less than 2.5%	
	Lead-oxide content	Less than 0.02%	
	Other heavy elements	Trace quantities	
Solution mass	Approximately 5.0 g		
Chemical Properties:			
Solution composition	Chemical Formula	Concentration (mol·L ⁻¹)	Mass Fraction (g·g ⁻¹)
	H ₂ O ³ HHO	55 6×10^{-7}	1.00 1×10^{-8}
Radiological Properties:			
Radionuclidic impurities	None detected [f]		
Half lives used	Hydrogen-3: (4500 ± 8) d [g]		
Calibration method and measuring instrument(s)	4πβ gas counting of SRM 4927E using the NIST length-compensated internal gas proportional counters and intercomparison of SRMs 4927E/4927F using two 4πβ liquid-scintillation counting systems [h]		

EVALUATION OF THE UNCERTAINTY OF THE MASSIC ACTIVITY [d]*

Input Quantity x_i , the source of uncertainty (and individual uncertainty components where appropriate)	Method Used To Evaluate $u(x_i)$, the standard uncertainty of x_i (A) denotes evaluation by statistical methods (B) denotes evaluation by other methods	Relative Uncertainty Of Input Quantity, $u(x_i)/x_i$, (%) [j]	Relative Sensitivity Factor, $ \partial y/\partial x_i \cdot$ (x_i/y) [j]	Relative Uncertainty Of Output Quantity, $u_i(y)/y$, (%) [k]
Massic count rate of SRM 4927E, corrected for background and decay [h]	Standard deviation of the mean for 23 sets of gas counting measurements (A)	0.18	1.0	0.18
Gram-mole measurements	Estimated (B)	0.20	1.0	0.20
Live-time [p]	Estimated (B)	0.10	1.0	0.10
Extrapolation of count-rate-versus-energy to zero energy	Estimated (B)	0.20	1.0	0.20
Half life of H-3	Standard uncertainty of the half life (A)	0.18 [m]	0.009 [n]	0.002
Liquid-scintillation intercomparison of SRM 4927F and SRM 4927E	Standard deviation of the mean for 7 sets of liquid-scintillation measurements (A)	0.06	1.0	0.06
Radionuclidic impurities	Limit of detection (B) [q]	100.	0.0005	0.05
Relative Combined Standard Uncertainty of the Output Quantity, $u_c(y)/y$, (%)				0.36
Coverage Factor, k				<u>x 2</u>
Relative Expanded Uncertainty of the Output Quantity, U/y , (%)				0.72

NOTES

- [a] The Sievert is the SI unit for dose equivalent. See reference [1]. One μSv is equal to 0.1 mrem.
Distance from Ampoule (cm): 1 30 100
Approximate Dose Rate ($\mu\text{Sv/h}$): <0.1 (Not detectable)
- [b] The stated uncertainty is two times the standard uncertainty.
- [c] Massic activity is the preferred name for the quantity activity divided by the total mass of the sample. See reference [1].
- [d] The reported value, y , of massic activity (activity per unit mass) at the reference time was not measured directly but was derived from measurements and calculations of other quantities. This can be expressed as $y = f(x_1, x_2, x_3, \dots, x_n)$, where f is a mathematical function derived from the assumed model of the measurement process.
- The value, x_i , used for each input quantity i has a standard uncertainty, $u(x_i)$, that generates a corresponding uncertainty in y , $u_i(y) \equiv |\partial y / \partial x_i| \cdot u(x_i)$, called a component of combined standard uncertainty of y .
- The combined standard uncertainty of y , $u_c(y)$, is the positive square root of the sum of the squares of the components of combined standard uncertainty.
- The combined standard uncertainty is multiplied by a coverage factor of $k = 2$ to obtain U , the expanded uncertainty of y .
- Since it can be assumed that the possible estimated values of the massic activity are approximately normally distributed with approximate standard deviation $u_c(y)$, the unknown value of the massic activity is believed to lie in the interval $y \pm U$ with a level of confidence of approximately 95 percent.
- For further information on the expression of uncertainties, see references [2] and [3].
- [e] The value of each standard uncertainty component, and hence the value of the expanded uncertainty itself, is a best estimate based upon all available information, but is only approximately known. That is to say, the "uncertainty of the uncertainty" is large and not well known. This is true for uncertainties evaluated by statistical methods (e.g., the relative standard deviation of the standard deviation of the mean for the massic response is approximately 50%) and for uncertainties evaluated by other methods (which could easily be over estimated or under estimated by substantial amounts). The unknown value of the expanded uncertainty is believed to lie in the interval $U/2$ to $2U$ (i.e., within a factor of 2 of the estimated value).
- [f] The estimated limit of detection for radionuclidic impurities is $300 \text{ Bq} \cdot \text{g}^{-1}$.
- [g] The stated uncertainty is the standard uncertainty. See reference [5].
- [h] Extensive gas-counting measurements were made on the SRM 4927E solution during 1998 and 1999. The SRM 4927F solution was intercompared with the SRM 4927E solution using liquid-scintillation counting.
- [i] Relative standard uncertainty of the input quantity x_i .

- [j] The relative change in the output quantity y divided by the relative change in the input quantity x_i . If $|\partial y/\partial x_i| \cdot (x_i/y) = 1.0$, then a 1% change in x_i results in a 1% change in y . If $|\partial y/\partial x_i| \cdot (x_i/y) = 0.05$, then a 1% change in x_i results in a 0.05% change in y .
- [k] Relative component of combined standard uncertainty of output quantity y , rounded to two significant figures or less. The relative component of combined standard uncertainty of y is given by $u_c(y)/y = |\partial y/\partial x_i| \cdot u(x_i)/y = |\partial y/\partial x_i| \cdot (x_i/y) \cdot u(x_i)/x_i$. The numerical values of $u(x_i)/x_i$, $|\partial y/\partial x_i| \cdot (x_i/y)$, and $u_c(y)/y$, all dimensionless quantities, are listed in columns 3, 4, and 5, respectively. Thus, the value in column 5 is equal to the value in column 4 multiplied by the value in column 3. The input quantities are independent, or very nearly so. Hence the covariances are zero or negligible.
- [m] The relative standard uncertainty of $\lambda \cdot t$ is determined by the relative standard uncertainty of λ (i.e., of the half life). The relative standard uncertainty of t is negligible.
- [n] $|\partial y/\partial x_i| \cdot (x_i/y) = |\lambda \cdot t|$
- [p] The live time is determined by counting the pulses from a gated crystal-controlled oscillator.
- [q] The standard uncertainty for each undetected impurity that might reasonably be expected to be present is estimated to be equal to the estimated limit of detection for that impurity, i.e. $u(x_i)/x_i = 100\%$. $|\partial y/\partial x_i| \cdot (x_i/y) = \{(\text{response per Bq of impurity})/(\text{response per Bq of H-3})\} \cdot \{(\text{Bq of impurity})/(\text{Bq of H-3})\}$. Thus $u_c(y)/y$ is the relative change in y if the impurity were present with a massic activity equal to the estimated limit of detection.

REFERENCES

- [1] International Organization for Standardization (ISO), *ISO Standards Handbook - Quantities and Units*, 1993. Available from the American National Standards Institute, 11 West 42nd Street, New York, NY 10036, U.S.A. 1-212-642-4900.
- [2] International Organization for Standardization (ISO), *Guide to the Expression of Uncertainty in Measurement*, 1993. Available from the American National Standards Institute, 11 West 42nd Street, New York, NY 10036, U.S.A. 1-212-642-4900. (Listed under ISO miscellaneous publications as "ISO Guide to the Expression 1993".)
- [3] B. N. Taylor and C. E. Kuyatt, *Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Results*, NIST Technical Note 1297, 1994. Available from the Superintendent of Documents, U.S. Government Printing Office, Washington, DC 20407, U.S.A.
- [4] National Council on Radiation Protection and Measurements Report No. 58, *A Handbook of Radioactivity Measurements Procedures*, Second Edition, 1985. Available from the National Council on Radiation Protection and Measurements, 7910 Woodmont Avenue, Bethesda, MD 20814 U.S.A.
- [5] L.L. Lucas and M.P. Unterweger, *Comprehensive Review and Critical Evaluation of the Half-Life of Tritium*, J. Res. Natl. Inst. Stand. Technol. 105, 541-549 (2000).

PREPARE A WORKING DILUTION OF 486.1284.85 AT ≈ 2500 dpm/ml

1. ^{2.51.06} TRANSFER 486.1284.85 TO AN AMBER GLASS BOTTLE
 MASS OF BOTTLE W/O LID: 259.20g BALZG
 MASS OF BOTTLE + STD: 261.25g
 NET MASS OF STD: 2.05g

2. ^{2.21.06} DILUTE TO 500ml WITH DI WATER (SEE BELOW FOR ρ DETERMINATION)
 MASS OF BOTTLE + STD + H₂O: 760.0g BALZG
 MASS OF BOTTLE (from above): 259.20g
 NET MASS OF DILUTED STD: 500.8g

3. DETERMINE DENSITY OF DI WATER
 MASS OF 100ml VOLUMETRIC FLASK: 56.4421g BALZG
 MASS OF FLASK + H₂O: 156.1980g
 NET MASS OF H₂O: 99.7559g
 $\rho = 0.997559$ g/ml

4. FINAL ACTIVITY CALCULATION

$$\left(\frac{486.1284.85 \text{ dpm}}{500.8 \text{ g}} \right) \left(0.9976 \text{ g/ml} \right) \left(2.05 \text{ g} \right) \left(2.22 \text{ dpm/g} \right) = 3871.3 \frac{\text{dpm}}{\text{ml}}$$

Std ID: 486.3020.23

Description: H-3

Expiration: 4/15/07
 Activity: 3871.30 dpm/mL

2s Uncertainty: 193.57 dpm/mL

Ref. Date: 3/23/98

Ref Time: N/A

Prep Date: 3/24/06 Prep by: AF


Matrix/Comp. DI H2O

Half Life (y): 1.24E+01


Reverification Log		
Analysis Date	Initials	Expiration Date

Continued on Page

Read and Understood By



3/24/06



4/17/06

Signed

Date

Signed

Date

Tritium Spike Solution 486 1284.85

Vol # 486 1284.84 was diluted to from a sub approx
100,000 dpm/ml as follows:

Wt taken 27.9711g
Wt + H₂O (486 1284.84) 42.6011g
Final (dil w/ DI H₂O) 62.7170g

PG 9/16/05
Std ID: 486.1284.85
Description: 3H
Expiration: 8/11/06
Activity: 948004.38 dpm/mL
2s Uncertainty: 47400.22 dpm/mL
Ref. Date: 3/23/98
Ref Time: N/A
Prep Date: 5/3/00
Matrix/Comp. DI Water
Half Life (y): 1.23E+01
Prep by: RTS
PG 9/16/05

dil factor
42.6011g
- 27.9711g
14.63g STD
x 1.0141876 pCi/g
14.837521 pCi
= (62.7170g - 27.9711g)
427,029.4 pCi/g
PG 8/17/04
Std ID: 486.1284.85

PG 3/28/03
Std ID: 486.1284.85
Description: H-3
Activity: 948004.38 dpm/ml
2s Uncertainty: 4740.02 dpm/ml
Ref. Date: 3/23/98
Ref Time: na
Prep Date: 5/3/00
Expiration: 3/25/03
Matrix/Comp. DI H₂O
Half Life (y): 1.23E+01
Prep by: RTS
PG 3/28/03
PG 3/28/03

MAY REQUIRE NCR FOR ICPT WORK.
PG 3/28/03

Description: 3H
Expiration: 7/30/05
Activity: 948004.38 dpm/mL
2s Uncertainty: 4740.02 dpm/mL
Ref. Date: 3/23/98
Ref Time: N/A
Prep Date: 7/29/04
Matrix/Comp. DI H₂O
Half Life (y): 1.23E+01
Prep by: JLK
PG 8/17/04

Std ID: 486.1284.85
Description: Tritium in Water Stock Solution
Activity: 427029 pCi/g or mL
Uncertainty: pCi/g or mL
Ref. Date: ~~3/30/98~~ 3/23/98
Ref Time: 10:00 AM
Prep Date: 3/30/98
Expiration: ~~3/29/03~~ 3/23/98
Matrix/Comp. DI Water
Short Lived - Decay Correct with each use
Prep by: RTS
PG 3/28/03

Verified 5/27/98 JUS
Standard re-verified 3/25/03 PG 3/28/03
Expires 3/25/04
Standard re-verified. New exp date: 7/30/05 PG 8/17/04

Read and Understood By
Signature: JUS Date: 5/22/98
Signature: _____ Date: _____
Signature: _____ Date: _____

Tritium Stock Solution 486.1284.84 - 12.84.84

PAI solution number 486.1284.84 (Analytics: JRS:55543-327)

1 was transferred to a 40 ml TDA vial.

Certificate - Biotek not available on 3/30/98 - will obtain tomorrow.

Activity 51.054 uCi
Volume 50.34 g/ml
Matrix H₂O

RG 8/13/04
Std ID: 486.1284.84

Std ID: 486.1284.84

RG 8/13/04

Description: 3H
Expiration: 7/30/05
Activity: 1014184.00 pCi/g or ml
2s Uncertainty: 50709.00 pCi/g or ml
Ref. Date: 3/23/98
Ref Time: N/A
Prep Date: 7/29/04 Prep by: JLK
Matrix/Comp. DI Water
Half Life (y): 1.23E+01
RG 8/13/04

RG 8/13/04

Description: H-3
Activity: 1014184.00 pCi/g or ml
2s Uncertainty: 50709.00 pCi/g or ml
Ref. Date: 3/23/98
Ref Time: 0.417
Prep Date: 3/30/98 Prep by: RTS
Expiration: 7/3/04
Matrix/Comp. DI Water
Half Life (y): 1.24E+01

RG 8/13/04

Standard re-verified on 7/30/04. Expires 7/30/05.
RG 8/13/04

Std ID: 486.1284.84

Description: Tritium in Water Stock Solution
Activity: 1014184 pCi/g or mL
Uncertainty: pCi/g or mL
Ref. Date: ~~3/23/98~~ 3/30/98 7-7-02
Ref Time: 10:00
Prep Date 3/30/98 Prep by: RTS
Expiration 3/29/03
Matrix/Comp. DI Water
Short Lived - Decay Correct with each use

Verified 5/27/98

Standard re-verified on 7/3/04. Expires 7/3/04. RG 7/7/03

Continued on Page

[Signature]
Signed

5/3/00
Date

Read and Understood By

[Signature]
Signed

5/30/99
Date



ANALYTIX

1380 Seaboard Industrial Blvd.
Atlanta, Georgia 30318 U.S.A.

Phone (404) 352-8577
Fax (404) 352-2837

*PAJ ID 486
rec'd 3-24-98*

CERTIFICATE OF CALIBRATION

Standard Radionuclide Source

55543-307

H-3 in Water 50 mL in Flame Sealed Vial

This standard radionuclide source was prepared using an aliquot measured gravimetrically from a master radionuclide solution standard. The master radionuclide solution standard was calibrated by the Department Des Applications Et De La Metrologie Des Rayonnements Ionisants (DAMRI), Paris, France, as Number 24057. The calibration was checked by liquid scintillation counting after source preparation.

ANALYTIX maintains traceability to the National Institute of Standards and Technology through Measurements Assurance Programs as described in USNRC Reg. Guide 4.15, Revision 1.

ISOTOPE:	H-3
ACTIVITY (dps):	1.889 E+06
CALIBRATION DATE:	March 23, 1998 12:00 EST.
HALF-LIFE:	12.43 years
TOTAL UNCERTAINTY:	5.0%

50.34 grams of water.

P O NUMBER 21143, Item 1

PREPARED BY: *M. D. Currie*
M. D. Currie, Radiochemist

Q A APPROVED: *LCM w/ 3/23/98*

DAILY INSTRUMENT PERFORMANCE CHECKS - LS6000 (LL OFF, LUMEX OFF)

Daily IPCs consist of the following standards;

Efficiency Check -

Beckman Tritium Standard

Lot HNZ0202

101900.00 dpm

2/17/2005 REF

2/17/2010 EXP

Beckman C-14 Standard

Lot CNZ3112

98500.0 dpm

2/17/2005 REF

2/17/2010 EXP

PAI Reagent Blank

RB 5/7/03 LLT # 97-3021

The quench factor is measured on the Reagent Blank

Interim Control Limits as of 9/20/06 MBC

	<u>Decay Corrected Tritium</u>	<u>Carbon-14</u>	<u>Reagent Blank</u>	<u>Quench Factor</u>
UCL	75824.8	84792.9	38.8	123.1
Mean Value	68931.6	77084.4	33.3	111.9
LCL	62038.5	69376.0	27.9	100.7

<u>Date</u>	<u>Decay Corrected</u>			<u>Reagent</u>			<u>Quench</u>		
	<u>H-3 CPM</u>	<u>H-CPM</u>	<u>PASS?</u>	<u>C-14 CPM</u>	<u>PASS?</u>	<u>H-3 Bkg CPM</u>	<u>PASS?</u>	<u>H#</u>	<u>PASS?</u>
9/14/2006	62988.60	68831.32	OK	76980.9	OK	35.4	OK	112.4	OK
9/18/2006	63005.30	68892.14	OK	77074.3	OK	31.9	OK	111.7	OK
9/18/2006	62900.30	68777.33	OK	77019.8	OK	30.6	OK	111.8	OK
9/18/2006	62990.50	68875.96	OK	77113.1	OK	33.3	OK	112.1	OK
9/18/2006	62957.30	68839.66	OK	77024.2	OK	32.9	OK	111.7	OK
9/18/2006	63016.50	68904.39	OK	77085.6	OK	31.4	OK	112.0	OK
9/18/2006	63036.60	68926.37	OK	77121.0	OK	33.0	OK	112.2	OK
9/18/2006	63090.50	68985.30	OK	77089.7	OK	32.7	OK	111.9	OK
9/18/2006	63225.40	69132.81	OK	77013.2	OK	36.7	OK	111.7	OK
9/18/2006	62960.90	68843.59	OK	77216.4	OK	32.2	OK	112.1	OK
9/18/2006	62982.00	68866.67	OK	77049.7	OK	34.8	OK	111.5	OK
9/19/2006	62941.50	68833.02	OK	77117.7	OK	33.4	OK	111.9	OK

Liquid Scintillation Counter

Quality Control Data

Daily Instrument Performance
Checks

DAILY INSTRUMENT PERFORMANCE CHECKS - LS6000 (LL OFF, LUMEX OFF)

Daily IPCs consist of the following standards;

Efficiency Check -

Beckman Tritium Standard

Lot HNZ0202

101900.00 dpm
2/17/05 REF
2/17/10 EXP

Beckman C-14 Standard

Lot CNZ3112

98500.0 dpm
2/17/05 REF
2/17/10 EXP

PAI Reagent Blank

RB 5/7/03 LLT # 97-3021

The quench factor is measured on the Reagent Blank

Historical Control Limits as of 11/16/06 MBC

	<u>Decay Corrected Tritium</u>	<u>Carbon-14</u>	<u>Reagent Blank</u>	<u>Quench Factor</u>
UCL	70286.2	78615.0	38.2	113.1
Mean Value	68908.0	77073.5	32.9	112.0
LCL	67529.9	75532.0	27.7	111.0

<u>Date</u>	<u>Decay Corrected</u>			<u>Reagent</u>			<u>Quench</u>		
	<u>H-3 CPM</u>	<u>H-CPM</u>	<u>PASS?</u>	<u>C-14 CPM</u>	<u>PASS?</u>	<u>H-3 Bkg CPM</u>	<u>PASS?</u>	<u>H#</u>	<u>PASS?</u>
11/10/06	62523.30	68927.35	OK	77231.4	OK	31.8	OK	112.7	OK
11/11/06	62353.50	68750.79	OK	77020.0	OK	35.4	OK	112.1	OK
11/11/06	62382.10	68782.32	OK	77132.1	OK	31.3	OK	112.1	OK
11/13/06	62170.30	68569.98	OK	76985.7	OK	32.1	OK	112.1	OK
11/13/06	62332.40	68748.77	OK	76963.7	OK	32.7	OK	112.6	OK
11/14/06	62204.80	68618.64	OK	77096.3	OK	33.9	OK	112.0	OK

DAILY CHECK LL ON ⁹⁹Tc SOURCE- LS6000

⁹⁹ Tc standard		spike	known activity	
03/18/99	REF		488.15	dpm/mL
03/13/07	EXP		2440.75	dpm

HISTORICAL Control Limit as of 10/18/2006

	<u>blank</u>	<u>spike</u>
UCL	24.98	567.20
Mean Value	19.10	516.59
LCL	14.76	477.89

Date	Blank count rate	Pass ?	Spiked count rate	Pass ?
11/10/06	19	OK	529.5	OK
11/11/06	18.8	OK	530.1	OK
11/11/06	20.2	OK	520.6	OK
11/13/06	22.7	OK	549.4	OK
11/13/06	18	OK	536.1	OK
11/14/06	19.8	OK	547.4	OK



Paragon Analytics

Radiochemistry Case Narrative Gamma Spectroscopy


ACZ Laboratories, Inc.

L59449


Paragon Work Order 0610164

1. This report consists of analysis results for two water samples received by Paragon Analytics on 10/20/2006. The analysis results for these samples are reported in units of pCi/L. The samples were not filtered prior to analysis.
2. These samples were prepared according to Paragon Analytics procedure PA SOP739R8.
3. The samples were analyzed for the presence of gamma emitting radionuclides according to Paragon Analytics procedure PA SOP713R9. The analyses were completed on 11/5/2006.
4. Paragon Analytics has found there to be a significant low bias to ^{214}Pb and ^{214}Bi results when using a mixed nuclide gamma source for efficiency calibrations. The magnitude of this bias has been determined to be approximately 32% for ^{214}Bi , and 23% for ^{214}Pb . Therefore, any reported results for ^{214}Pb and ^{214}Bi are flagged with a "J" qualifier, indicating the activity values to be an estimated value. Results are reported without further qualification.
5. Technical considerations made in the creation of the gamma spectroscopy library used in this analysis are detailed in the document "Technical Comments Regarding Gamma Spectroscopy Libraries" found in Section 4.
6. There are cases where the magnitude of negative activity is greater than the 2σ TPU. Under typical conditions, where background data is normally distributed and analyzed by paired observations, this event is likely to occur at least 2.5% of the time. Review of the data does not indicate a problem with the instrument or reporting systems and results are reported without further qualification.
7. No problems were encountered with either the client samples or the associated quality control samples. All quality control criteria were met.


The data contained in the following report have been reviewed and approved by the personnel listed below. In addition, Paragon Analytics certifies that the analyses reported herein are true, complete and correct within the limits of the methods employed.




Radiochemistry Instrument Technician



Date



Radiochemistry Final Data Review



Date

Paragon Analytics

A Division of DataChem Laboratories, Inc.

000001

PARAGON ANALYTICS
Radiochemistry Data Package

Section 1

**SAMPLE RESULTS
SUMMARY**

**Due to the nature of gamma spectroscopy
data a summary report is not provided.**

**Please refer to the individual sample results
in Section 3.**

2

PARAGON ANALYTICS
Radiochemistry Data Package

Section 2

**QC RESULTS
SUMMARY**

000004

Gamma Spectroscopy Results

PAI 713 Rev 9

Method Blank Results

Lab Name: Paragon Analytics

Work Order Number: 0610164

Client Name: ACZ Laboratories, Inc.

ClientProject ID: L59449

Lab ID: GS061025-4MB

Library: FANP

Sample Matrix: WATER

Prep SOP: PAI 739 Rev 8

Date Collected: 25-Oct-06

Date Prepared: 25-Oct-06

Date Analyzed: 02-Nov-06

Prep Batch: GS061025-4

QCBatchID: GS061025-4-2

Run ID: GS061025-4A

Count Time: 400 minutes

Final Aliquot: 1000 ml

Result Units: pCi/l

File Name: 062634d06

CASNO	Target Nuclide	Result +/- 2 s TPU	MDC	Lab Qualifier
14331-83-0	Ac-228	-10 +/- 30	51	U
14391-76-5	Ag-110m	2.7 +/- 4.5	7.5	U
14682-66-7	Al-26	1.6 +/- 5.5	9.6	U
14596-10-2	Am-241	-12 +/- 50	86	U
13966-02-4	Be-7	-54 +/- 36	66	U
14913-49-6	Bi-212	48 +/- 66	110	U
14733-03-0	Bi-214	8 +/- 15	25	U,J
13982-30-4	Ce-139	-0.4 +/- 3.6	6.1	U
14762-78-8	Ce-144	-12 +/- 25	42	U
14093-03-9	Co-56	1.6 +/- 7.5	12.9	U
13981-50-5	Co-57	-1.8 +/- 3.5	5.9	U
13981-38-9	Co-58	1.5 +/- 4.5	7.6	U
10198-40-0	Co-60	5.8 +/- 5.5	8.8	U
14392-02-0	Cr-51	9 +/- 32	54	U
13967-70-9	Cs-134	-3.2 +/- 4.7	8.2	U

Comments:

Qualifiers/Flags:

U - Result is less than the sample specific MDC or less than the associated TPU

Y1 - Chemical Yield is in control at 100-110%. Quantitative Yield is assumed.

Y2 - Chemical Yield outside default limits.

LT - Result is less than Requested MDC, greater than sample specific MDC.

SQ - Spectral quality prevents accurate quantitation.

SI - Nuclide identification and/or quantitation is tentative.

TI - Nuclide identification is tentative.

R - Nuclide has exceeded 8 half-lives.

M - Requested MDC not met.

B - Analyte concentration greater than MDC.

B3 - Analyte concentration greater than MDC but less than Requested MDC.

Abbreviations:

TPU - Total Propagated Uncertainty (see PAI SOP 743)

MDC - Minimum Detectable Concentration (see PAI SOP 709)

BDL - Below Detection Limit

Data Package ID: GSW0610164-1

Gamma Spectroscopy Results

PAI 713 Rev 9

Method Blank Results

Lab Name: Paragon Analytics

Work Order Number: 0610164

Client Name: ACZ Laboratories, Inc.

ClientProject ID: L59449

Lab ID: GS061025-4MB

Library: FANP

Sample Matrix: WATER
Prep SOP: PAI 739 Rev 8
Date Collected: 25-Oct-06
Date Prepared: 25-Oct-06
Date Analyzed: 02-Nov-06

Prep Batch: GS061025-4
QCBatchID: GS061025-4-2
Run ID: GS061025-4A
Count Time: 400 minutes

Final Aliquot: 1000 ml
Result Units: pCi/l
File Name: 062634d06

CASNO	Target Nuclide	Result +/- 2 s TPU	MDC	Lab Qualifier
10045-97-3	Cs-137	-3.8 +/- 5.0	8.9	U
14683-23-9	Eu-152	1 +/- 24	42	U
15585-10-1	Eu-154	-15 +/- 25	46	U
14391-16-3	Eu-155	24 +/- 16	26	U
14596-12-4	Fe-59	10.3 +/- 8.5	13.4	U
10043-66-0	I-131	-0.9 +/- 4.6	7.9	U
13966-00-2	K-40	30 +/- 100	180	U
13966-31-9	Mn-54	-1.2 +/- 4.8	8.5	U
13966-32-0	Na-22	2.7 +/- 4.6	7.7	U
14681-63-1	Nb-94	-2.7 +/- 4.8	8.5	U
13967-76-5	Nb-95	2.3 +/- 4.4	7.3	U
15100-28-4	Pa-234m	-180 +/- 770	1380	U
15092-94-1	Pb-212	2.1 +/- 7.3	12.2	U
15067-28-4	Pb-214	7 +/- 13	22	U,J
13967-48-1	Ru-106	-7 +/- 45	78	U

Comments:

Qualifiers/Flags:

U - Result is less than the sample specific MDC or less than the associated TPU
Y1 - Chemical Yield is in control at 100-110%. Quantitative Yield is assumed.
Y2 - Chemical Yield outside default limits.
LT - Result is less than Requested MDC, greater than sample specific MDC.
SQ - Spectral quality prevents accurate quantitation.
SI - Nuclide identification and/or quantitation is tentative.
TI - Nuclide identification is tentative.
R - Nuclide has exceeded 8 half-lives.
M - Requested MDC not met.
B - Analyte concentration greater than MDC.
B3 - Analyte concentration greater than MDC but less than Requested MDC.

Abbreviations:

TPU - Total Propagated Uncertainty (see PAI SOP 743)
MDC - Minimum Detectable Concentration (see PAI SOP 709)
BDL - Below Detection Limit

Data Package ID: GSW0610164-1

000006

Gamma Spectroscopy Results

PAI 713 Rev 9

Method Blank Results

Lab Name: Paragon Analytics
Work Order Number: 0610164
Client Name: ACZ Laboratories, Inc.
ClientProject ID: L59449

Lab ID: GS061025-4MB

Library: FANP

Sample Matrix: WATER
Prep SOP: PAI 739 Rev 8
Date Collected: 25-Oct-06
Date Prepared: 25-Oct-06
Date Analyzed: 02-Nov-06

Prep Batch: GS061025-4
QC Batch ID: GS061025-4-2
Run ID: GS061025-4A
Count Time: 400 minutes

Final Aliquot: 1000 ml
Result Units: pCi/l
File Name: 062634d06

CASNO	Target Nuclide	Result +/- 2 s TPU	MDC	Lab Qualifier
14683-10-4	Sb-124	-12.5 +/- 5.4	9.8	U
14234-35-6	Sb-125	0 +/- 12	21	U
13967-63-0	Sc-46	-0.1 +/- 4.6	8.1	U
15623-47-9	Th-227	-16 +/- 25	43	U
15065-10-8	Th-234	-20 +/- 110	190	U
14913-50-9	Tl-208	-3.9 +/- 6.6	11.4	U
15117-96-1	U-235	-15 +/- 25	44	U
13982-39-3	Zn-65	-3.1 +/- 9.9	17.8	U

Comments:

Qualifiers/Flags:

U - Result is less than the sample specific MDC or less than the associated TPU
Y1 - Chemical Yield is in control at 100-110%. Quantitative Yield is assumed.
Y2 - Chemical Yield outside default limits.
LT - Result is less than Requested MDC, greater than sample specific MDC.
SQ - Spectral quality prevents accurate quantitation.
SI - Nuclide identification and/or quantitation is tentative.
TI - Nuclide identification is tentative.
R - Nuclide has exceeded 8 half-lives.
M - Requested MDC not met.
B - Analyte concentration greater than MDC.
B3 - Analyte concentration greater than MDC but less than Requested MDC.

Abbreviations:

TPU - Total Propagated Uncertainty (see PAI SOP 743)
MDC - Minimum Detectable Concentration (see PAI SOP 709)
BDL - Below Detection Limit

Data Package ID: GSW0610164-1

000007

Gamma Spectroscopy Results

PAI 713 Rev 9

Laboratory Control Sample(s)

Lab Name: Paragon Analytics
Work Order Number: 0610164
Client Name: ACZ Laboratories, Inc.
ClientProject ID: L59449

Lab ID: GS061025-4LCS	Sample Matrix: WATER Prep SOP: PAI 739 Rev 8 Date Collected: 25-Oct-06 Date Prepared: 25-Oct-06 Date Analyzed: 03-Nov-06	Prep Batch: GS061025-4 QCBatchID: GS061025-4-2 Run ID: GS061025-4A Count Time: 30 minutes	Final Aliquot: 1000 ml Result Units: pCi/l File Name: 062543d02
-----------------------	--	--	---

CASNO	Target Nuclide	Results +/- 2s TPU	MDC	Spike Added	% Rec	Control Limits	Lab Qualifier
14596-10-2	Am-241	97000 +/- 12000	2000	97600	99.7	85 - 115	P
10198-40-0	Co-60	41800 +/- 4900	100	42000	99.5	85 - 115	P
10045-97-3	Cs-137	38200 +/- 4500	200	37000	103	85 - 115	P,M3

Comments:

Qualifiers/Flags:

U - Result is less than the sample specific MDC or less than the associated TPU
LT - Result is less than Requested MDC, greater than sample specific MDC.
Y1 - Chemical Yield is in control at 100-110%. Quantitative Yield is assumed.
Y2 - Chemical Yield outside default limits.
L - LCS Recovery below lower control limit.
H - LCS Recovery above upper control limit.
P - LCS Recovery within control limits.
M - The requested MDC was not met.
M3 - The requested MDC was not met, but thereported activity is greater than the reported MDC.

Abbreviations:

TPU - Total Propagated Uncertainty (see PAI SOP 743)
MDC - Minimum Detectable Concentration (see PAI SOP 709)

SQ - Spectral quality prevents accurate quantitation.
SI - Nuclide identification and/or quantitation is tentative.
TI - Nuclide identification is tentative.
R - Nuclide has exceeded 8 halfives.

Data Package ID: GSW0610164-1

000008

PARAGON ANALYTICS
Radiochemistry Data Package

3

Section 3

**INDIVIDUAL
SAMPLE RESULTS**

000009

Gamma Spectroscopy Results

PAI 713 Rev 9
Sample Results

Lab Name: Paragon Analytics
Work Order Number: 0610164
Client Name: ACZ Laboratories, Inc.
ClientProject ID: L59449

Field ID: L59449-01 Lab ID: 0610164-1 Library: FANP	Sample Matrix: WATER Prep SOP: PAI 739 Rev 8 Date Collected: 16-Oct-06 Date Prepared: 25-Oct-06 Date Analyzed: 05-Nov-06	Prep Batch: GS061025-4 QCBatchID: GS061025-4-2 Run ID: GS061025-4A Count Time: 300 minutes Report Basis: Unfiltered	Final Aliquot: 975 ml Prep Basis: Unfiltered Moisture(%): NA Result Units: pCi/l File Name: 063147d08
---	--	---	---

CASNO	Target Nuclide	Result +/- 2 s TPU	MDC	Lab Qualifier
14331-83-0	Ac-228	7 +/- 22	38	U
14391-76-5	Ag-110m	0.7 +/- 4.3	7.5	U
14682-66-7	Al-26	-0.2 +/- 5.5	9.9	U
14596-10-2	Am-241	0.5 +/- 5.2	8.9	U
13966-02-4	Be-7	26 +/- 40	66	U
14913-49-6	Bi-212	26 +/- 84	142	U
14733-03-0	Bi-214	2 +/- 14	23	U,J
13982-30-4	Ce-139	-0.4 +/- 2.8	4.9	U
14762-78-8	Ce-144	-3 +/- 18	31	U
14093-03-9	Co-56	-9.9 +/- 9.0	17.1	U
13981-50-5	Co-57	1.3 +/- 2.0	3.4	U
13981-38-9	Co-58	-2.4 +/- 5.1	9.2	U
10198-40-0	Co-60	-2.3 +/- 5.2	9.6	U
14392-02-0	Cr-51	-35 +/- 48	85	U
13967-70-9	Cs-134	-0.7 +/- 4.1	7.2	U

Comments:

Qualifiers/Flags:

U - Result is less than the sample specific MDC or less than the associated TPU
Y1 - Chemical Yield is in control at 100-110%. Quantitative Yield is assumed.
Y2 - Chemical Yield outside default limits.
LT - Result is less than Requested MDC, greater than sample specific MDC.
M3 - The requested MDC was not met, but the reported activity is greater than the reported MDC.
M - The requested MDC was not met.

SQ - Spectral quality prevents accurate quantitation.
SI - Nuclide identification and/or quantitation is tentative.
TI - Nuclide identification is tentative.
R - Nuclide has exceeded 8 half-lives.
G - Sample density differs by more than 15% of LCS density.

Abbreviations:

TPU - Total Propagated Uncertainty (see PAI SOP 743)
MDC - Minimum Detectable Concentration (see PAI SOP 709)
BDL - Below Detection Limit

Data Package ID: GSW0610164-1

Gamma Spectroscopy Results

PAI 713 Rev 9
Sample Results

Lab Name: Paragon Analytics
Work Order Number: 0610164
Client Name: ACZ Laboratories, Inc.
ClientProject ID: L59449

Field ID: L59449-01
Lab ID: 0610164-1

Library: FANP

Sample Matrix: WATER
Prep SOP: PAI 739 Rev 8
Date Collected: 16-Oct-06
Date Prepared: 25-Oct-06
Date Analyzed: 05-Nov-06

Prep Batch: GS061025-4
QC Batch ID: GS061025-4-2
Run ID: GS061025-4A
Count Time: 300 minutes
Report Basis: Unfiltered

Final Aliquot: 975 ml
Prep Basis: Unfiltered
Moisture(%): NA
Result Units: pCi/l
File Name: 063147d08

CASNO	Target Nuclide	Result +/- 2 s TPU	MDC	Lab Qualifier
10045-97-3	Cs-137	-1.4 +/- 4.7	8.3	U
14683-23-9	Eu-152	3 +/- 23	40	U
15585-10-1	Eu-154	15 +/- 26	44	U
14391-16-3	Eu-155	-2.8 +/- 7.4	12.9	U
14596-12-4	Fe-59	5 +/- 12	20	U
10043-66-0	I-131	5 +/- 19	33	U
13966-00-2	K-40	-29 +/- 78	138	U
13966-31-9	Mn-54	-1.3 +/- 4.4	7.8	U
13966-32-0	Na-22	2.1 +/- 4.8	8.2	U
14681-63-1	Nb-94	-3.0 +/- 4.3	7.8	U
13967-76-5	Nb-95	-3.2 +/- 4.6	8.5	U
15100-28-4	Pa-234m	-90 +/- 780	1380	U
15092-94-1	Pb-212	-2.8 +/- 7.2	12.4	U
15067-28-4	Pb-214	1 +/- 12	20	U,J
13967-48-1	Ru-106	-4 +/- 42	73	U

Comments:

Qualifiers/Flags:

U - Result is less than the sample specific MDC or less than the associated TPU
Y1 - Chemical Yield is in control at 100-110%. Quantitative Yield is assumed.
Y2 - Chemical Yield outside default limits.
LT - Result is less than Requested MDC, greater than sample specific MDC.
M3 - The requested MDC was not met, but the reported activity is greater than the reported MDC.
M - The requested MDC was not met.

SQ - Spectral quality prevents accurate quantitation.
S1 - Nuclide identification and/or quantitation is tentative.
TI - Nuclide identification is tentative.
R - Nuclide has exceeded 8 half-lives.
G - Sample density differs by more than 15% of LCS density.

Abbreviations:

TPU - Total Propagated Uncertainty (see PAI SOP 743)
MDC - Minimum Detectable Concentration (see PAI SOP 709)
BDL - Below Detection Limit

Data Package ID: GSW0610164-1

Gamma Spectroscopy Results

PAI 713 Rev 9
Sample Results

Lab Name: Paragon Analytics
Work Order Number: 0610164
Client Name: ACZ Laboratories, Inc.
ClientProject ID: L59449

Field ID: L59449-01 Lab ID: 0610164-1	Sample Matrix: WATER Prep SOP: PAI 739 Rev 8 Date Collected: 16-Oct-06 Date Prepared: 25-Oct-06 Date Analyzed: 05-Nov-06	Prep Batch: GS061025-4 QCBatchID: GS061025-4-2 Run ID: GS061025-4A Count Time: 300 minutes Report Basis: Unfiltered	Final Aliquot: 975 ml Prep Basis: Unfiltered Moisture(%): NA Result Units: pCi/l File Name: 063147d08
--	--	---	---

CASNO	Target Nuclide	Result +/- 2 s TPU	MDC	Lab Qualifier
14683-10-4	Sb-124	-1.8 +/- 5.0	8.9	U
14234-35-6	Sb-125	0.9 +/- 9.0	16.3	U
13967-63-0	Sc-46	1.8 +/- 4.7	8.1	U
15623-47-9	Th-227	-2 +/- 18	31	U
15065-10-8	Th-234	-29 +/- 57	97	U
14913-50-9	Tl-208	2.4 +/- 6.4	10.8	U
15117-96-1	U-235	5 +/- 17	29	U
13982-39-3	Zn-65	-2 +/- 11	19	U

Comments:

Qualifiers/Flags:

U - Result is less than the sample specific MDC or less than the associated TPU
Y1 - Chemical Yield is in control at 100-110%. Quantitative Yield is assumed.
Y2 - Chemical Yield outside default limits.
LT - Result is less than Requested MDC, greater than sample specific MDC.
M3 - The requested MDC was not met, but the reported activity is greater than the reported MDC.
M - The requested MDC was not met.

SQ - Spectral quality prevents accurate quantitation.
SI - Nuclide identification and/or quantitation is tentative.
TI - Nuclide identification is tentative.
R - Nuclide has exceeded 8 half-lives.
G - Sample density differs by more than 15% of LCS density.

Abbreviations:

TPU - Total Propagated Uncertainty (see PAI SOP 743)
MDC - Minimum Detectable Concentration (see PAI SOP 709)
BDL - Below Detection Limit

Data Package ID: GSW0610164-1

Gamma Spectroscopy Results

PAI 713 Rev 9
Sample Results

Lab Name: Paragon Analytics
Work Order Number: 0610164
Client Name: ACZ Laboratories, Inc.
ClientProject ID: L59449

Field ID: L59449-02
Lab ID: 0610164-2

Library: FANP

Sample Matrix: WATER
Prep SOP: PAI 739 Rev 8
Date Collected: 16-Oct-06
Date Prepared: 25-Oct-06
Date Analyzed: 05-Nov-06

Prep Batch: GS061025-4
QC Batch ID: GS061025-4-2
Run ID: GS061025-4A
Count Time: 300 minutes
Report Basis: Unfiltered

Final Aliquot: 990 ml
Prep Basis: Unfiltered
Moisture(%): NA
Result Units: pCi/l
File Name: 062554d02

CASNO	Target Nuclide	Result +/- 2 s TPU	MDC	Lab Qualifier
14331-83-0	Ac-228	1 +/- 22	38	U
14391-76-5	Ag-110m	-3.8 +/- 4.8	8.7	U
14682-66-7	Al-26	3.4 +/- 5.6	9.4	U
14596-10-2	Am-241	16 +/- 30	50	U
13966-02-4	Be-7	21 +/- 42	70	U
14913-49-6	Bi-212	58 +/- 62	100	U
14733-03-0	Bi-214	6 +/- 16	26	U,J
13982-30-4	Ce-139	-0.9 +/- 4.9	8.3	U
14762-78-8	Ce-144	-6 +/- 22	38	U
14093-03-9	Co-56	7.2 +/- 10	17.2	U
13981-50-5	Co-57	-0.7 +/- 2.9	5.1	U
13981-38-9	Co-58	-2.9 +/- 5.9	10.5	U
10198-40-0	Co-60	0.5 +/- 5.7	10.0	U
14392-02-0	Cr-51	-26 +/- 47	84	U
13967-70-9	Cs-134	-7.7 +/- 4.9	9.1	U

Comments:

Qualifiers/Flags:

U - Result is less than the sample specific MDC or less than the associated TPU
Y1 - Chemical Yield is in control at 100-110%. Quantitative Yield is assumed.
Y2 - Chemical Yield outside default limits.
LT - Result is less than Requested MDC, greater than sample specific MDC.
M3 - The requested MDC was not met, but the reported activity is greater than the reported MDC.
M - The requested MDC was not met.

SQ - Spectral quality prevents accurate quantitation.
SI - Nuclide identification and/or quantitation is tentative.
TI - Nuclide identification is tentative.
R - Nuclide has exceeded 8 half-lives.
G - Sample density differs by more than 15% of LCS density.

Abbreviations:

TPU - Total Propagated Uncertainty (see PAI SOP 743)
MDC - Minimum Detectable Concentration (see PAI SOP 709)
BDL - Below Detection Limit

Data Package ID: GSW0610164-1

Gamma Spectroscopy Results

PAI 713 Rev 9
Sample Results

Lab Name: Paragon Analytics
Work Order Number: 0610164
Client Name: ACZ Laboratories, Inc.
ClientProject ID: L59449

Field ID: L59449-02	Sample Matrix: WATER	Prep Batch: GS061025-4	Final Aliquot: 990 ml
Lab ID: 0610164-2	Prep SOP: PAI 739 Rev 8	QCBatchID: GS061025-4-2	Prep Basis: Unfiltered
Library: FANP	Date Collected: 16-Oct-06	Run ID: GS061025-4A	Moisture(%): NA
	Date Prepared: 25-Oct-06	Count Time: 300 minutes	Result Units: pCi/l
	Date Analyzed: 05-Nov-06	Report Basis: Unfiltered	File Name: 062554d02

CASNO	Target Nuclide	Result +/- 2 s TPU	MDC	Lab Qualifier
10045-97-3	Cs-137	4.3 +/- 5.0	8.2	U
14683-23-9	Eu-152	-4 +/- 26	47	U
15585-10-1	Eu-154	12 +/- 27	45	U
14391-16-3	Eu-155	4 +/- 11	19	U
14596-12-4	Fe-59	2 +/- 12	21	U
10043-66-0	I-131	0 +/- 22	38	U
13966-00-2	K-40	-9 +/- 85	147	U
13966-31-9	Mn-54	-0.9 +/- 5.0	8.9	U
13966-32-0	Na-22	-7.0 +/- 5.6	10.7	U
14681-63-1	Nb-94	-1.9 +/- 4.9	8.7	U
13967-76-5	Nb-95	1.7 +/- 5.5	9.4	U
15100-28-4	Pa-234m	-410 +/- 840	1520	U
15092-94-1	Pb-212	1.4 +/- 10	17.3	U
15067-28-4	Pb-214	3 +/- 13	21	U,J
13967-48-1	Ru-106	-7 +/- 45	79	U

Comments:

Qualifiers/Flags:

U - Result is less than the sample specific MDC or less than the associated TPU
Y1 - Chemical Yield is in control at 100-110%. Quantitative Yield is assumed.
Y2 - Chemical Yield outside default limits.
LT - Result is less than Requested MDC, greater than sample specific MDC.
M3 - The requested MDC was not met, but the reported activity is greater than the reported MDC.
M - The requested MDC was not met.

SQ - Spectral quality prevents accurate quantitation.

SI - Nuclide identification and/or quantitation is tentative.

TI - Nuclide identification is tentative.

R - Nuclide has exceeded 8 halfives.

G - Sample density differs by more than 15% of LCS density.

Abbreviations:

TPU - Total Propagated Uncertainty (see PAI SOP 743)

MDC - Minimum Detectable Concentration (see PAI SOP 709)

BDL - Below Detection Limit

Data Package ID: GSW0610164-1

Gamma Spectroscopy Results

PAI 713 Rev 9
Sample Results

Lab Name: Paragon Analytics
Work Order Number: 0610164
Client Name: ACZ Laboratories, Inc.
ClientProject ID: L59449

Field ID: L59449-02
Lab ID: 0610164-2

Library: FANP

Sample Matrix: WATER
Prep SOP: PAI 739 Rev 8
Date Collected: 16-Oct-06
Date Prepared: 25-Oct-06
Date Analyzed: 05-Nov-06

Prep Batch: GS061025-4
QCBatchID: GS061025-4-2
Run ID: GS061025-4A
Count Time: 300 minutes
Report Basis: Unfiltered

Final Aliquot: 990 ml
Prep Basis: Unfiltered
Moisture(%): NA
Result Units: pCi/l
File Name: 062554d02

CASNO	Target Nuclide	Result +/- 2 s TPU	MDC	Lab Qualifier
14683-10-4	Sb-124	-9.0 +/- 6.4	11.6	U
14234-35-6	Sb-125	5 +/- 11	20	U
13967-63-0	Sc-46	0.4 +/- 5.4	9.4	U
15623-47-9	Th-227	-8 +/- 22	38	U
15065-10-8	Th-234	-2 +/- 98	164	U
14913-50-9	Tl-208	-1.0 +/- 7.1	12.5	U
15117-96-1	U-235	15 +/- 17	27	U
13982-39-3	Zn-65	-10 +/- 11	20	U

Comments:

Qualifiers/Flags:

U - Result is less than the sample specific MDC or less than the associated TPU
Y1 - Chemical Yield is in control at 100-110%. Quantitative Yield is assumed.
Y2 - Chemical Yield outside default limits.
LT - Result is less than Requested MDC, greater than sample specific MDC.
M3 - The requested MDC was not met, but the reported activity is greater than the reported MDC.
M - The requested MDC was not met.

SQ - Spectral quality prevents accurate quantitation.
SI - Nuclide identification and/or quantitation is tentative.
TI - Nuclide identification is tentative.
R - Nuclide has exceeded 8 half-lives.
G - Sample density differs by more than 15% of LCS density.

Abbreviations:

TPU - Total Propagated Uncertainty (see PAI SOP 743)
MDC - Minimum Detectable Concentration (see PAI SOP 709)
BDL - Below Detection Limit

Data Package ID: GSW0610164-1

PARAGON ANALYTICS
Radiochemistry Data Package

Section 4

4

RAW DATA

000016

 SEEKER G A M M A A N A L Y S I S R E S U L T S PS Version 1.8.4

Paragon Analytics, Div. of DataChem Lab
 GammaScan

Geo 1 / Water

Sample ID: 0610164-1 GS061025-4

```

-----
Sampling Start:   10/16/2006 12:00:00 | Counting Start:   11/05/2006 09:27:20
Sampling Stop:   10/16/2006 12:00:00 | Decay Time. . . . . 4.77E+002 Hrs
Buildup Time. . . . . 0.00E+000 Hrs | Live Time . . . . . 18000 Sec
Sample Size . . . . . 9.75E-001 L | Real Time . . . . . 18036 Sec
Collection Efficiency . . . . . 1.0000 | Spc. File . . . . . 063147D08.SPC
-----
  
```

Detector #: 8 (Detector 8)

Energy(keV) = -0.56 + 0.500*Ch + 0.00E+00*Ch^2 + 0.00E+00*Ch^3 11/05/2006

FWHM(keV) = 0.72 + 0.007*En + 8.11E-04*En^2 + 0.00E+00*En^3 08/17/2006

Where En = Sqrt(Energy in keV)

Search Sensitivity: 1.00 | Sigma Multiplier: 2.00 | Search Start/End: 80/4000

PEAK SEARCH RESULTS

PK. #	ENERGY (keV)	ADDRESS CHANNEL	NET/MDA COUNTS	UN-CERTAINTY	C.L. COUNTS	BKG COUNTS	FWHM (keV)	FLAG
1	46.45	93.96	138	56	42	322	0.90	a
2	53.48	108.00	13	26	21	120	0.41	a NET< CL
3	63.28	127.58	211	60	43	348	0.93	a
4	66.20	133.43	289	74	54	464	1.26	b
5	74.27	149.55	90	68	54	460	1.24	a
6	77.02	155.05	74	49	38	288	0.82	b
7	92.59	186.17	190	58	42	322	0.91	a
8	119.14	239.24	25	31	24	146	0.45	a
9	139.67	280.26	145	48	35	240	0.74	a
10	175.15	351.18	52	51	40	277	1.07	a
11	185.82	372.50	109	55	42	305	1.11	a
12	198.36	397.56	148	44	30	182	0.81	a
13	238.59	477.96	75	35	25	143	0.72	a
14	295.20	591.10	19	26	20	97	0.54	a NET< CL
15	352.24	705.10	73	38	28	141	0.93	a
16	511.01	1022.41	461	65	40	242	1.95	a
17	558.48	1117.26	123	35	22	95	1.15	a
18	596.34	1192.94	68	40	30	148	1.59	a
19	609.33	1218.89	58	33	25	119	1.12	a
20	802.76	1605.47	41	24	17	64	1.18	a
21	911.04	1821.87	28	21	15	48	1.25	a
22	927.70	1855.18	25	27	20	70	2.02	a
23	1460.05	2919.10	40	22	15	45	1.57	a

 SEEKER BACKGROUND SUBTRACT RESULTS Vers. 2.2.1

Paragon Analytics, Div. of DataChem Lab
 GammaScan

Background File: DET081103.BKG (061103-8 WEEKLY BKG)

Bkg.File Detector #: 8

=====

BACKGROUND SUBTRACT RESULTS

=====

PK#	ENERGY (keV)	OLD NET COUNTS	OLD UN- CERTAINTY	OLD CR.LEVEL	NEW NET COUNTS	NEW UN- CERTAINTY	NEW CR.LEVEL	FLAG
1	46.45	138	56	42	7	72	59	NET<CL
3	63.28	211	60	43	43	80	65	NET<CL
4	66.20	289	74	54	216	92	72	
5	74.27	90	68	54	26	84	68	NET<CL
6	77.02	74	49	38	8	69	57	NET<CL
7	92.59	190	58	42	-42	82	69	NET<CL
9	139.67	145	48	35	90	61	47	
11	185.82	109	55	42	1	73	60	NET<CL
12	198.36	148	44	30	83	63	50	
13	238.59	75	35	25	-22	55	46	NET<CL
14	295.20	19	26	20	-17	48	40	NET<CL
15	352.24	73	38	28	6	57	47	NET<CL
16	511.01	461	65	40	-49	101	84	NET<CL
17	558.48	123	35	22	35	51	41	NET<CL
19	609.33	58	33	25	6	53	44	NET<CL
20	802.76	41	24	17	-6	40	33	NET<CL
21	911.04	28	21	15	11	33	27	NET<CL
23	1460.05	40	22	15	-12	33	28	NET<CL

Paragon Analytics, Div. of DataChem Lab
GammaScan

Geo 1 / Water

Sample ID: 0610164-1 GS061025-4

Sampling Start: 10/16/2006 12:00:00 | Counting Start: 11/05/2006 09:27:20
Sampling Stop: 10/16/2006 12:00:00 | Decay Time: 4.77e+002 Hrs
Buildup Time: 0.00e+000 Hrs | Live Time: 18000 Sec
Sample Size: 9.75e-001 L | Real Time: 18036 Sec
Collection Efficiency: 1.0000 | Spectrum File: 063147D08.SPC
Cr. Level Confidence Interval: 95 % | Det. Limit Confidence Interval: 95 %

Detector #: 8 (Detector 8)

Efficiency File: (D08)(Sh01).EFF (Geo 1 Eff Cal)

Eff.=1/[6.36E-01*En^-1.19E+00 + 1.20E+02*En^9.04E-01] 09/11/2006

Library File: .FANP.LIB (FANP (Fiss. Act. and Nat. Products))

MEASURED or MDA CONCENTRATIONS

Table with columns: Nuclide, ENERGY (keV), N (T), Concentration (pCi/L), MDA, Critical Level, Halflife (hrs). Rows include Th-234, Sb-125, Pb-212, Pb-214, Bi-214, Ac-228, K-40, Am-241, Eu-155, Co-57, Ce-144, U-235, Ce-139, Th-227, Cr-51, I-131, Be-7, Tl-208, Sb-124, Cs-134, Ru-106, Ag-110M, Cs-137.

MEASURED or MDA CONCENTRATIONS

Nuclide	ENERGY E (keV)	N T	Concentration		MDA	Critical Level	Halflife (hrs)
			(pCi/L)			
Nb-94	702.50	N-3.01E+00	+ -	4.30E+00	7.76E+00	3.69E+00	1.78E+08
Bi-212	727.17	N 2.58E+01	+ -	8.43E+01	1.42E+02	6.82E+01	1.67E+04
Nb-95	765.82	N-3.20E+00	+ -	4.60E+00	8.46E+00	3.98E+00	1.54E+03
Co-58	810.75	N-2.45E+00	+ -	5.09E+00	9.17E+00	4.33E+00	1.70E+03
Mn-54	834.81	N-1.32E+00	+ -	4.38E+00	7.81E+00	3.68E+00	7.49E+03
Sc-46	889.26	N 1.77E+00	+ -	4.74E+00	8.09E+00	3.78E+00	2.01E+03
Pa-234m	1001.03	N-8.81E+01	+ -	7.82E+02	1.38E+03	6.49E+02	3.92E+13
Eu-154	1004.80	N 1.53E+01	+ -	2.63E+01	4.40E+01	2.06E+01	7.45E+04
Fe-59	1099.22	N 4.94E+00	+ -	1.15E+01	1.95E+01	9.11E+00	1.08E+03
Zn-65	1115.52	N-2.35E+00	+ -	1.06E+01	1.88E+01	8.84E+00	5.85E+03
Co-56	1238.28	N-9.93E+00	+ -	8.88E+00	1.71E+01	8.00E+00	1.86E+03
Na-22	1274.54	N 2.10E+00	+ -	4.79E+00	8.17E+00	3.77E+00	2.28E+04
Co-60	1332.51	N-2.33E+00	+ -	5.22E+00	9.58E+00	4.47E+00	4.62E+04
Eu-152	1408.08	N 3.41E+00	+ -	2.26E+01	3.99E+01	1.83E+01	1.17E+05
Al-26	1808.65	N-2.37E-01	+ -	5.49E+00	9.93E+00	4.53E+00	6.31E+09

MEASURED TOTAL: 1.05E+02 +- 3.05E+02 pCi/L

UNKNOWN, SUM or ESCAPE PEAKS

PK. #	ENERGY (keV)	ADDRESS CHANNEL	NET COUNTS	UN- CERTAINTY	C.L. COUNTS	BKG COUNTS	FWHM (keV)	FLAG
1	46.45	93.96	7	72	59	322	0.90	Deleted
2	53.48	108.00	13	26	21	120	0.41	Deleted
3	63.28	127.58	43	80	65	348	0.93	Deleted
4	66.20	133.43	216	92	72	464	1.26	Unknown
5	74.27	149.55	26	84	68	460	1.24	Deleted
6	77.02	155.05	8	69	57	288	0.82	Deleted
8	119.14	239.24	25	31	24	146	0.45	Unknown
9	139.67	280.26	90	61	47	240	0.74	Unknown
11	185.82	372.50	1	73	60	305	1.11	Deleted
12	198.36	397.56	83	63	50	182	0.81	Unknown
14	295.20	591.10	-17	48	40	97	0.54	Deleted
16	511.01	1022.41	-49	101	84	242	1.95	Deleted
17	558.48	1117.26	35	51	41	95	1.15	Deleted
18	596.34	1192.94	68	40	30	148	1.59	Unknown
20	802.76	1605.47	-6	40	33	64	1.18	Deleted
22	927.70	1855.18	25	27	20	70	2.02	Unknown

c:\SEEKER\BIN\063147d08.res Analysis Results Saved.

 SEEKER G A M M A A N A L Y S I S R E S U L T S PS Version 1.8.4

Paragon Analytics, Div. of DataChem Lab
 GammaScan

Geo 1 / Water

Sample ID: 0610164-2 GS061025-4

```
-----
Sampling Start: 10/16/2006 12:00:00 | Counting Start: 11/05/2006 09:27:05
Sampling Stop: 10/16/2006 12:00:00 | Decay Time. . . . . 4.77E+002 Hrs
Buildup Time. . . . . 0.00E+000 Hrs | Live Time . . . . . 18000 Sec
Sample Size . . . . . 9.90E-001 L | Real Time . . . . . 18007 Sec
Collection Efficiency . . . . . 1.0000 | Spc. File . . . . . 062554D02.SPC
-----
```

```
-----
Detector #: 2 (Detector 2)
Energy(keV) = -0.77 + 0.501*Ch + 2.77E-08*Ch^2 + 0.00E+00*Ch^3 11/05/2006
FWHM(keV) = 0.69 + 0.006*En + 1.18E-03*En^2 + 0.00E+00*En^3 05/16/2006
Where En = Sqrt(Energy in keV)
-----
```

```
-----
Search Sensitivity: 1.00 | Sigma Multiplier: 2.00 | Search Start/End: 80/4000
-----
```

PEAK SEARCH RESULTS

```
=====
```

PK. #	ENERGY (keV)	ADDRESS CHANNEL	NET/MDA COUNTS	UN- CERTAINTY	C.L. COUNTS	BKG COUNTS	FWHM (keV)	FLAG
1	66.36	134.10	116	52	39	299	0.83 a	
2	74.81	151.00	13	32	26	165	0.42 a	NET< CL
3	84.80	170.96	30	42	34	228	0.75 a	NET< CL
4	92.63	186.60	127	51	38	268	0.92 a	
5	139.97	281.17	94	38	27	165	0.62 a	
6	143.88	289.00	31	35	27	165	0.66 b	
7	174.78	350.72	39	35	27	164	0.70 a	
8	185.72	372.59	84	33	22	123	0.50 a	
9	198.23	397.58	140	52	38	247	1.09 a	
10	233.03	467.10	22	32	25	142	0.58 a	NET< CL
11	238.53	478.09	90	40	29	173	0.87 a	
12	278.67	558.28	27	33	26	132	0.76 a	
13	295.00	590.91	40	36	28	153	0.90 a	
14	351.88	704.54	77	36	26	132	1.11 a	
15	511.00	1022.42	420	66	42	240	2.27 a	Wide Pk
16	558.50	1117.31	106	35	24	103	1.25 a	
17	595.75	1191.72	36	31	24	122	1.07 a	
18	609.51	1219.20	74	37	27	132	1.61 a	
19	803.03	1605.79	42	31	23	95	1.80 a	
20	911.21	1821.88	26	18	13	39	0.97 a	
21	1461.30	2920.65	50	24	16	41	2.55 a	

```
=====
```

 SEEKER BACKGROUND SUBTRACT RESULTS Vers. 2.2.1

Paragon Analytics, Div. of DataChem Lab
 GammaScan

Background File: DET021103.BKG (061103-2 WEEKLY BKG)

Bkg.File Detector #: 2

=====

BACKGROUND SUBTRACT RESULTS

=====

PK#	ENERGY (keV)	OLD NET COUNTS	OLD UN- CERTAINTY	OLD CR.LEVEL	NEW NET COUNTS	NEW UN- CERTAINTY	NEW CR.LEVEL	FLAG
1	66.36	116	52	39	28	72	59	NET<CL
2	74.81	13	32	26	-94	103	87	NET<CL
4	92.63	127	51	38	-1	72	60	NET<CL
5	139.97	94	38	27	59	48	37	
8	185.72	84	33	22	-18	69	57	NET<CL
9	198.23	140	52	38	81	72	57	
11	238.53	90	40	29	9	66	54	NET<CL
13	295.00	40	36	28	7	46	38	NET<CL
14	351.88	77	36	26	13	53	43	NET<CL
15	511.00	420	66	42	-81	99	82	NET<CL
16	558.50	106	35	24	54	48	38	
18	609.51	74	37	27	21	53	43	NET<CL
19	803.03	42	31	23	2	42	35	NET<CL
20	911.21	26	18	13	1	30	25	NET<CL
21	1461.30	50	24	16	-3	34	28	NET<CL

Paragon Analytics, Div. of DataChem Lab
GammaScan

Geo 1 / Water

Sample ID: 0610164-2 GS061025-4

```

-----
Sampling Start: 10/16/2006 12:00:00 | Counting Start: 11/05/2006 09:27:05
Sampling Stop: 10/16/2006 12:00:00 | Decay Time . . . . . 4.77e+002 Hrs
Buildup Time . . . . . 0.00e+000 Hrs | Live Time . . . . . 18000 Sec
Sample Size . . . . . 9.90e-001 L | Real Time . . . . . 18007 Sec
Collection Efficiency . . . . . 1.0000 | Spectrum File . . . . . 062554D02.SPC
Cr. Level Confidence Interval: 95 % | Det. Limit Confidence Interval: 95 %
-----
    
```

Detector #: 2 (Detector 2)

Efficiency File: (D02)(Sh01).EFF (Geo 1 Eff Cal)
Eff.=1/[3.02E-03*En^-3.85E+00 + 1.33E+02*En^7.85E-01] 05/18/2006

Library File:FANP.LIB (FANP (Fiss. Act. and Nat. Products))

MEASURED or MDA CONCENTRATIONS

Nuclide	ENERGY E (keV)	N T	Concentration (pCi/L)	MDA	Critical Level	Halflife (hrs)
Th-234	92.50	N	1.68E+00 +- 9.77E+01	1.64E+02	8.04E+01	3.92E+13
U-235	143.76	N	1.47E+01 +- 1.66E+01	2.71E+01	1.29E+01	6.17E+12
Sb-125	Average:x	N	5.27E+00 +- 1.07E+01	2.43E+04
	176.29	N	3.22E+01 +- 2.90E+01	4.68E+01	2.23E+01	2.43E+04
	427.95	N	1.05E+00 +- 1.15E+01	1.97E+01	9.38E+00	2.43E+04
Pb-212	238.63	N	1.39E+00 +- 1.03E+01	1.73E+01	8.45E+00	1.67E+04
Tl-208	Average:x	N	9.67E-01 +- 7.15E+00	1.67E+04
	277.36	N	3.25E+01 +- 3.92E+01	6.46E+01	3.07E+01	1.67E+04
	583.14	N	2.12E+00 +- 7.27E+00	1.25E+01	6.05E+00	1.67E+04
Pb-214	351.99	N	3.10E+00 +- 1.26E+01	2.12E+01	1.03E+01	1.40E+07
Bi-214	609.32	N	6.23E+00 +- 1.58E+01	2.65E+01	1.28E+01	1.40E+07
Ac-228	911.07	N	5.92E-01 +- 2.18E+01	3.78E+01	1.79E+01	5.04E+04
K-40	1460.75	N	8.57E+00 +- 8.47E+01	1.47E+02	7.01E+01	1.12E+13
Am-241	59.54	N	1.64E+01 +- 2.97E+01	4.95E+01	2.38E+01	3.80E+06
Eu-155	105.31	N	3.85E+00 +- 1.13E+01	1.91E+01	9.14E+00	4.35E+04
Co-57	122.07	N	7.25E-01 +- 2.94E+00	5.06E+00	2.44E+00	6.48E+03
Ce-144	133.53	N	6.40E+00 +- 2.17E+01	3.76E+01	1.81E+01	6.82E+03
Ce-139	165.85	N	8.55E-01 +- 4.87E+00	8.25E+00	4.03E+00	3.30E+03
Th-227	236.00	N	7.52E+00 +- 2.20E+01	3.82E+01B	1.84E+01	1.90E+05
Cr-51	320.07	N	2.56E+01 +- 4.72E+01	8.37E+01	4.00E+01	6.65E+02
I-131	364.48	N	0.00E+00 +- 2.21E+01	3.81E+01	1.82E+01	1.93E+02
Be-7	477.56	N	2.08E+01 +- 4.17E+01	7.00E+01	3.31E+01	1.28E+03
Sb-124	602.71	N	8.98E+00 +- 6.28E+00	1.16E+01B	5.56E+00	1.44E+03
Cs-134	604.66	N	7.74E+00 +- 4.84E+00	9.06E+00B	4.34E+00	1.81E+04
Ru-106	621.84	N	7.34E+00 +- 4.52E+01	7.91E+01	3.76E+01	8.84E+03

MEASURED or MDA CONCENTRATIONS

Nuclide	ENERGY E (keV)	N T	Concentration (pCi/L)	MDA	Critical Level	Halflife (hrs)
Ag-110M	657.75	N-3.78E+00	+ - 4.78E+00	8.72E+00	4.14E+00	6.00E+03
Cs-137	661.62	N 4.28E+00	+ - 5.02E+00	8.23E+00	3.88E+00	2.64E+05
Nb-94	702.50	N-1.87E+00	+ - 4.91E+00	8.69E+00	4.13E+00	1.78E+08
Bi-212	727.17	N 5.83E+01	+ - 6.19E+01	1.01E+02	4.70E+01	1.67E+04
Nb-95	765.82	N 1.70E+00	+ - 5.50E+00	9.39E+00	4.42E+00	1.54E+03
Co-58	810.75	N-2.94E+00	+ - 5.85E+00	1.05E+01	4.98E+00	1.70E+03
Mn-54	834.81	N-8.68E-01	+ - 5.04E+00	8.89E+00	4.20E+00	7.49E+03
Sc-46	889.26	N 4.33E-01	+ - 5.41E+00	9.43E+00	4.42E+00	2.01E+03
Pa-234m	1001.03	N-4.10E+02	+ - 8.38E+02	1.53E+03	7.16E+02	3.92E+13
Eu-154	1004.80	N 1.16E+01	+ - 2.66E+01	4.51E+01	2.10E+01	7.45E+04
Fe-59	1099.22	N 1.57E+00	+ - 1.23E+01	2.14E+01	9.99E+00	1.08E+03
Zn-65	1115.52	N-1.00E+01	+ - 1.07E+01	2.02E+01	9.49E+00	5.85E+03
Co-56	1238.28	N 7.23E+00	+ - 1.04E+01	1.72E+01	8.02E+00	1.86E+03
Na-22	1274.54	N-7.05E+00	+ - 5.49E+00	1.07E+01	5.01E+00	2.28E+04
Co-60	1332.51	N 5.08E-01	+ - 5.70E+00	9.99E+00	4.65E+00	4.62E+04
Eu-152	1408.08	N-3.77E+00	+ - 2.63E+01	4.72E+01	2.19E+01	1.17E+05
Al-26	1808.65	N 3.43E+00	+ - 5.61E+00	9.43E+00	4.28E+00	6.31E+09

MEASURED TOTAL: 1.61E+02 +- 3.09E+02 pCi/L

UNKNOWN, SUM or ESCAPE PEAKS

PK. #	ENERGY (keV)	ADDRESS CHANNEL	NET COUNTS	UN-CERTAINTY	C.L. COUNTS	BKG COUNTS	FWHM (keV)	FLAG
1	66.36	134.10	28	72	59	299	0.83	Deleted
2	74.81	151.00	-94	103	87	165	0.42	Deleted
3	84.80	170.96	30	42	34	228	0.75	Deleted
5	139.97	281.17	59	48	37	165	0.62	Unknown
8	185.72	372.59	-18	69	57	123	0.50	Deleted
9	198.23	397.58	81	72	57	247	1.09	Unknown
10	233.03	467.10	22	32	25	142	0.58	Deleted
13	295.00	590.91	7	46	38	153	0.90	Deleted
15	511.00	1022.42	-81	99	82	240	2.27	Deleted
16	558.50	1117.31	54	48	38	103	1.25	Unknown
17	595.75	1191.72	36	31	24	122	1.07	Unknown
19	803.03	1605.79	2	42	35	95	1.80	Deleted

c:\SEEKER\BIN\062554d02.res Analysis Results Saved.

 SEEKER G A M M A A N A L Y S I S R E S U L T S PS Version 1.8.4

Paragon Analytics, Div. of DataChem Lab
 GammaScan

Geo 1 / Water

Sample ID: GS061025-4MB GS061025-4

```
-----
Sampling Start:   11/02/2006 22:00:00 | Counting Start:   11/02/2006 22:36:31
Sampling Stop:   11/02/2006 22:00:00 | Decay Time. . . . . 6.09E-001 Hrs
Buildup Time. . . . . 0.00E+000 Hrs | Live Time . . . . . 24000 Sec
Sample Size . . . . . 1.00E+000 L | Real Time . . . . . 24010 Sec
Collection Efficiency . . . . . 1.0000 | Spc. File . . . . . 062634D06.SPC
-----
```

Detector #: 6 (Detector 6)

$$\text{Energy (keV)} = -0.65 + 0.501 \cdot \text{Ch} - 2.97\text{E-}08 \cdot \text{Ch}^2 + 0.00\text{E+}00 \cdot \text{Ch}^3 \quad 11/02/2006$$

$$\text{FWHM (keV)} = 1.19 + -0.002 \cdot \text{En} + 7.29\text{E-}04 \cdot \text{En}^2 + 0.00\text{E+}00 \cdot \text{En}^3 \quad 07/24/2006$$

Where En = Sqrt(Energy in keV)

Search Sensitivity: 1.00 | Sigma Multiplier: 2.00 | Search Start/End: 80/4000

=====

PEAK SEARCH RESULTS

=====

PK. #	ENERGY (keV)	ADDRESS CHANNEL	NET/MDA COUNTS	UN- CERTAINTY	C.L. COUNTS	BKG COUNTS	FWHM (keV)	FLAG
1	66.38	133.81	162	63	47	471	0.94	a
2	139.89	280.55	113	55	42	364	0.95	a
3	159.00	318.70	29	41	32	256	0.64	a NET< CL
4	185.92	372.45	53	39	30	220	0.65	a
5	198.44	397.44	191	54	38	310	0.95	a
6	237.62	475.64	31	84	68	657	2.15	a NET< CL Wide Pk
7	352.05	704.08	90	44	32	207	1.09	a
8	500.25	999.94	49	39	30	176	1.49	a
9	511.23	1021.88	570	76	48	320	2.77	a Wide Pk
10	558.39	1116.02	155	39	25	129	1.31	a
11	596.05	1191.22	99	33	22	112	0.91	a
12	609.20	1217.47	78	34	24	125	1.16	a
13	803.10	1604.59	25	29	23	96	1.68	a
14	1460.95	2918.12	142	31	16	46	2.02	a
15	1763.65	3522.60	21	16	11	25	1.66	a

 SEEKER B A C K G R O U N D S U B T R A C T R E S U L T S Vers. 2.2.1

Paragon Analytics, Div. of DataChem Lab
 GammaScan

Background File: DET061027.BKG (061027-6 WEEKLY BKG)

Bkg.File Detector #: 6

=====

BACKGROUND SUBTRACT RESULTS

=====

PK#	ENERGY (keV)	OLD NET COUNTS	OLD UN- CERTAINTY	OLD CR.LEVEL	NEW NET COUNTS	NEW UN- CERTAINTY	NEW CR.LEVEL	FLAG
1	66.38	162	63	47	112	86	69	
2	139.89	113	55	42	53	80	65	NET<CL
4	185.92	53	39	30	14	59	48	NET<CL
5	198.44	191	54	38	118	80	63	
6	237.62	31	84	68	-12	100	82	NET<CL
7	352.05	90	44	32	36	66	54	NET<CL
9	511.23	570	76	48	74	117	95	NET<CL
10	558.39	155	39	25	105	55	42	
11	596.05	99	33	22	68	57	45	
12	609.20	78	34	24	30	59	48	NET<CL
13	803.10	25	29	23	-21	44	37	NET<CL
14	1460.95	142	31	16	12	47	38	NET<CL
15	1763.65	21	16	11	2	30	25	NET<CL

MEASURED or MDA CONCENTRATIONS

Nuclide	ENERGY E (keV)	N T	Concentration (pCi/L)	MDA	Critical Level	Halflife (hrs)
Nb-95	765.82	N	2.28E+00 +- 4.35E+00	7.30E+00	3.46E+00	1.54E+03
Co-58	810.75	N	1.46E+00 +- 4.50E+00	7.64E+00	3.62E+00	1.70E+03
Mn-54	834.81	N	1.24E+00 +- 4.82E+00	8.47E+00	4.03E+00	7.49E+03
Sc-46	889.26	N	1.45E-01 +- 4.62E+00	8.05E+00	3.81E+00	2.01E+03
Ac-228	911.07	N	9.96E+00 +- 2.96E+01	5.11E+01	2.47E+01	5.04E+04
Pa-234m	1001.03	N	1.79E+02 +- 7.74E+02	1.38E+03	6.48E+02	3.92E+13
Eu-154	1004.80	N	1.55E+01 +- 2.52E+01	4.62E+01	2.17E+01	7.45E+04
Fe-59	1099.22	N	1.03E+01 +- 8.43E+00	1.33E+01	6.21E+00	1.08E+03
Zn-65	1115.52	N	3.09E+00 +- 9.94E+00	1.78E+01	8.37E+00	5.85E+03
Co-56	1238.28	N	1.58E+00 +- 7.49E+00	1.30E+01	6.05E+00	1.86E+03
Na-22	1274.54	N	2.69E+00 +- 4.61E+00	7.74E+00	3.58E+00	2.28E+04
Co-60	1332.51	N	5.79E+00 +- 5.45E+00	8.75E+00	4.07E+00	4.62E+04
Eu-152	1408.08	N	1.03E+00 +- 2.38E+01	4.21E+01	1.95E+01	1.17E+05
Al-26	1808.65	N	1.60E+00 +- 5.52E+00	9.59E+00	4.40E+00	6.31E+09

MEASURED TOTAL: 1.52E+02 +- 3.23E+02 pCi/L

UNKNOWN, SUM or ESCAPE PEAKS

PK. #	ENERGY (keV)	ADDRESS CHANNEL	NET COUNTS	UN-CERTAINTY	C.L. COUNTS	BKG COUNTS	FWHM (keV)	FLAG
1	66.38	133.81	112	86	69	471	0.94	Unknown
2	139.89	280.55	53	80	65	364	0.95	Deleted
3	159.00	318.70	29	41	32	256	0.64	Deleted
4	185.92	372.45	14	59	48	220	0.65	Deleted
5	198.44	397.44	118	80	63	310	0.95	Unknown
6	237.62	475.64	-12	100	82	657	2.15	Deleted
8	500.25	999.94	49	39	30	176	1.49	Unknown
9	511.23	1021.88	74	117	95	320	2.77	Deleted
10	558.39	1116.02	105	55	42	129	1.31	Unknown
11	596.05	1191.22	68	57	45	112	0.91	Unknown
13	803.10	1604.59	-21	44	37	96	1.68	Deleted
15	1763.65	3522.60	2	30	25	25	1.66	Deleted

c:\SEEKER\BIN\062634d06.res Analysis Results Saved.

SEEKER CALIBRATION RESULTS Version 2.0.4

Sample ID: DAILY CHECK
Stds. Match Tolerance: 2.00 keV

Detector Number: 06 Calibration Date: 11/16/2006 12:59:26

Energy(keV) = -0.69 + 0.501*Ch +-9.15e-09*Ch^2 + 0.00e+00*Ch^3

Table with 5 columns: Pk.#, Measured Centroid, Calculated Energy, Energy (keV), % Difference. Contains 4 rows of peak data.

Calibration Results Saved.

Paragon Analytics, Div. of DataChem Lab
 GammaScan

Geo 1 / Water

Sample ID: GS061025-4LCS GS061025-4

```

-----
Sampling Start: 11/03/2006 07:00:00 | Counting Start: 11/03/2006 07:20:43
Sampling Stop: 11/03/2006 07:00:00 | Decay Time. . . . . 3.45E-001 Hrs
Buildup Time. . . . . 0.00E+000 Hrs | Live Time . . . . . 1800 Sec
Sample Size . . . . . 1.00E+000 L | Real Time . . . . . 1825 Sec
Collection Efficiency . . . . . 1.0000 | Spc. File . . . . . .062543D02.SPC
-----
  
```

Detector #: 2 (Detector 2)

Energy (keV) = -0.77 + 0.500*Ch + 4.01E-08*Ch^2 + 0.00E+00*Ch^3 11/03/2006

FWHM(keV) = 0.69 + 0.006*En + 1.18E-03*En^2 + 0.00E+00*En^3 05/16/2006

Where En = Sqrt(Energy in keV)

Search Sensitivity: 1.00 | Sigma Multiplier: 2.00 | Search Start/End: 80/4000

=====

PEAK SEARCH RESULTS

=====

PK. #	ENERGY (keV)	ADDRESS CHANNEL	NET/MDA COUNTS	UN-CERTAINTY	C.L. COUNTS	BKG COUNTS	FWHM (keV)	FLAG
1	50.98	103.45	210	196	159	4320	1.18 a	
2	58.37	118.21	713	286	231	6570	1.91 a	Wide Pk
3	59.55	120.56	13622	268	108	2329	0.85 b	
4	70.41	142.26	96	100	81	1453	0.73 a	
5	87.98	177.38	10504	233	91	1659	0.87 a	
6	98.92	199.25	48	70	56	780	0.44 a	NET< CL
7	99.47	200.36	21	84	68	1039	0.61 b	NET< CL
8	103.05	207.52	87	71	56	780	0.53 c	
9	122.08	245.53	2326	138	82	1343	0.86 a	
10	136.61	274.58	346	131	103	1827	1.17 a	
11	166.07	333.46	142	99	79	1261	0.86 a	
12	186.81	374.92	74	118	96	1690	0.90 a	NET< CL
13	317.55	636.20	32	88	72	1099	0.93 a	NET< CL
14	353.32	707.68	40	65	52	670	0.61 a	NET< CL
15	476.30	953.46	81	62	49	583	0.69 a	
16	511.10	1023.01	61	104	84	1170	1.58 a	NET< CL
17	630.53	1261.68	41	57	46	481	1.05 a	NET< CL
18	644.13	1288.86	49	60	48	524	1.04 a	
19	661.74	1324.03	22572	312	68	852	1.63 a	
20	1173.40	2346.41	18534	280	54	501	2.33 a	
21	1332.64	2664.55	16729	261	26	109	2.52 a	HiResid

062543D02.SPC Analyzed by

SEEKER B A C K G R O U N D S U B T R A C T R E S U L T S Vers. 2.2.1

Paragon Analytics, Div. of DataChem Lab
GammaScan

Background File: DET021027.BKG (061027-2 WEEKLY BKG)

Bkg.File Detector #: 2

=====

BACKGROUND SUBTRACT RESULTS

=====

PK#	ENERGY (keV)	OLD NET COUNTS	OLD UN- CERTAINTY	OLD CR.LEVEL	NEW NET COUNTS	NEW UN- CERTAINTY	NEW CR.LEVEL	FLAG
5	87.98	10504	233	91	10504	233	91	
12	186.81	74	118	96	67	118	96	NET<CL
14	353.32	40	65	52	34	65	52	NET<CL
16	511.10	61	104	84	11	104	85	NET<CL

Paragon Analytics, Div. of DataChem Lab
 GammaScan

Geo 1 / Water

Sample ID: GS061025-4LCS GS061025-4

```

-----
Sampling Start: 11/03/2006 07:00:00 | Counting Start: 11/03/2006 07:20:43
Sampling Stop: 11/03/2006 07:00:00 | Decay Time. . . . . 3.45e-001 Hrs
Buildup Time. . . . . 0.00e+000 Hrs | Live Time . . . . . 1800 Sec
Sample Size . . . . . 1.00e+000 L | Real Time . . . . . 1825 Sec
Collection Efficiency . . . . . 1.0000 | Spectrum File . . . . . 062543D02.SPC
Cr. Level Confidence Interval: 95 % | Det. Limit Confidence Interval: 95 %
-----
  
```

Detector #: 2 (Detector 2)

Efficiency File: (D02)(Sh01).EFF (Geo 1 Eff Cal)
 Eff.=1/[3.02E-03*En^-3.85E+00 + 1.33E+02*En^7.85E-01] 05/18/2006

Library File:ANALYTICAL.LIB (Analytical)

MEASURED or MDA CONCENTRATIONS

Nuclide	ENERGY (keV)	N T	Concentration (pCi/L)	MDA	Critical Level	Halflife (hrs)
Am-241	59.54	9.74E+04	+ - 1.91E+03	1.56E+03	7.69E+02	3.79E+06
Cd-109	88.02	2.38E+05	+ - 5.27E+03	4.17E+03	2.06E+03	1.11E+04
Co-57	122.07	1.44E+03	+ - 8.58E+01	1.03E+02	5.06E+01	6.50E+03
Ce-139	165.85	9.42E+01	+ - 6.56E+01	1.06E+02	5.24E+01	3.30E+03
Cs-137	661.62	3.82E+04	+ - 5.27E+02	2.34E+02	1.15E+02	2.64E+05
Co-60	Average:x	4.18E+04	+ - 4.54E+02	4.62E+04
	1173.21	4.19E+04	+ - 6.33E+02	2.51E+02	1.22E+02	4.62E+04
	1332.48	4.18E+04	+ - 6.51E+02	1.35E+02	6.43E+01	4.62E+04
Hg-203	279.18	MDA	1.71E+02	8.40E+01	1.12E+03
Sn-113	391.68	MDA	2.51E+02	1.24E+02	2.76E+03
Y-88	898.02	MDA	3.31E+02	1.63E+02	2.56E+03

MEASURED TOTAL: 4.17E+05 +- 8.32E+03 pCi/L

UNKNOWN, SUM or ESCAPE PEAKS

PK. #	ENERGY (keV)	ADDRESS CHANNEL	NET COUNTS	UN-CERTAINTY	C.L. COUNTS	BKG COUNTS	FWHM (keV)	FLAG
1	50.98	103.45	210	196	159	4320	1.18	Unknown
2	58.37	118.21	713	286	231	6570	1.91	Unknown
4	70.41	142.26	96	100	81	1453	0.73	Unknown
6	98.92	199.25	48	70	56	780	0.44	Deleted
7	99.47	200.36	21	84	68	1039	0.61	Deleted

=====

UNKNOWN,SUM or ESCAPE PEAKS

=====

PK. #	ENERGY (keV)	ADDRESS CHANNEL	NET COUNTS	UN- CERTAINTY	C.L. COUNTS	BKG COUNTS	FWHM (keV)	FLAG
8	103.05	207.52	87	71	56	780	0.53	Unknown
10	136.61	274.58	346	131	104	1827	1.17	Unknown
12	186.81	374.92	67	118	96	1690	0.90	Deleted
13	317.55	636.20	32	88	72	1099	0.93	Deleted
14	353.32	707.68	34	65	52	670	0.61	Deleted
15	476.30	953.46	81	62	49	583	0.69	Unknown
16	511.10	1023.01	11	104	85	1170	1.58	Deleted
17	630.53	1261.68	41	57	46	481	1.05	Deleted
18	644.13	1288.86	49	60	48	524	1.04	Unknown

c:\SEEKER\BIN\062543d02.res Analysis Results Saved.

Gamma Spectrometer Run Log

Date: 11/2/06

Reviewed By/Date: JP 11/3/06

Sample ID	Ver1	Det. No.	Geo2	Count Dur. (min.) ³	Start Time	Analyst	File ID/Comments	Saved?
0610189-1	JP	2	13	30	8:11	JP	062535D02.SPC	JP
↓ -2	JP	9	↓	↓	↓	↓	061306D09.SPC	JP
0610197-2	JP	6	13	↓	8:33	JP	062627D06.SPC	JP
0610189-3	JP	2	13	30	9:07	JP	062536D02.SPC	JP
↓ -1D	JP	9	↓	↓	↓	↓	061307D09.SPC	JP
0610197-3	JP	6	↓	↓	↓	↓	062628D06.SPC	JP
0613011-8(799)	JP	8	18	30	9:24	JP	063124D08.SPC	JP
0610189-4	JP	2	13	30	9:45	JP	062537D02.SPC	JP
GS061027-2MB	JP	9	↓	↓	↓	↓	061308D09.SPC	JP
GS061027-1MB	JP	6	↓	↓	↓	↓	062629D06.SPC	JP
0610074-1	JP	8	1	210	9:57	JP	063126D08.SPC	JP
0610197-2D	JP	9	13	30	10:35	JP	061309D09.SPC	JP
GS061027-2LCS	JP	2	↓	30	↓	↓	062538D02.SPC	JP
GS061027-1LCS	JP	6	13	30	11:06	JP	062630D06.SPC	JP
0610161-1	JP	9	1	330	11:09	JP	061310D09.SPC	JP
↓ -4	JP	2	↓	↓	↓	↓	062539D02.SPC	JP
↓ -6	JP	6	↓	315	11:37	JP	062632D06.SPC	JP
0610161-9	JP	8	1	200	13:34	JP	063127D08.SPC	JP
↓ -1D	JP	2	↓	315	16:57	JP	062540D02.SPC	JP
0610179-1	JP	9	↓	330	16:57	↓	061311D09.SPC	JP
↓ -2	JP	8	↓	300	↓	↓	063128D08.SPC	JP
↓ -4	JP	6	1	315	17:05	JP	062633D06.SPC	JP
0610179-3	JP	9	1	400	22:36	JP	061312D09.SPC	JP
GS061025-4MB	JP	6	↓	↓	↓	↓	062634D06.SPC	JP

JP 11/2/06

- Analyst will verify the position, detector, and geometry when the sample is removed from the detector.
- Calibration geometry.
- Count duration.

KEY:

- * sample was counted on a puck
- ↑ sample was counted with air flow arrow pointing up
- ↓ sample was counted with air flow arrow pointing down

Gamma Spectrometer Run Log

Date: 11/3/06

Reviewed By/Date: JP 11/4/06

Sample ID	Ver ¹	Det. No.	Geo ²	Count Dur. (min.) ³	Start Time	Analyst	File ID/Comments	Saved?
G5061025-4LCS	JP	2	1	30	7:21	JP	06254300Z.SPC	JP
0610179-1D	JP	8	1	180	7:28	JP	063132D08.SPC	JP
0610141-1	Q	9	1	300	7:46	JP	061315D09.SPC	Q
0611025-1	SI	2	7*	60	10:23	JP	062544D0Z.SPC	SI
↓ -9	SI	6	↓		↓	↓	062636D06.SPC	SI
0611025-17	SI	8	7*		10:33	JP	063134D08.SPC 063133D08.SPC	SI
0611025-9D	Q	2	7*		12:09	SI	062545D0Z.SPC	Q
0611025-26	↓	6	7*		12:11	SI	062637D06.SPC	Q
0611025-27	↓	8	↓		12:11	SI	063135D08.SPC	Q
0611025-28	Q	2	↓		13:22	Q	062547D0Z.SPC	SI
0611025-30	Q	6	7*	30	14:58	Q	062638D06.SPC 062638D06.SPC	SI
0611025-31	Q	8	↓	30	14:58	Q	063138D08.SPC	SI
G5061103-1MB	SI	6	7*	↓	14:58	SI	062640D06.SPC	SI
G5061103-1LCS	SI	8	↓	30	15:48	Q	063141D08.SPC	Q
0611025-29	SI	2	7*	60	14:58	Q	062549D0Z.SPC	SI
0611025-29	SI	2	7*	60	14:52	SI	062549D0Z.SPC	SI
↓ -30	Q	6	7*	60	13:22	Q	062638D06.SPC	Q
↓ -31	↓	8	↓	↓	↓	↓	063138D08.SPC	Q
061103-2	JP	2	NA	1000	17:04	Q	062549D0Z.SPC	JP
↓ -6	JP	6	↓	↓	↓	↓	062641D06.SPC	JP
↓ -8	JP	8	↓	↓	↓	↓	063142D08.SPC	JP
↓ -9	JP	9	↓	↓	↓	↓	061316D09.SPC	JP

Q mislabeled

JP 11/4/06

- Analyst will verify the position, detector, and geometry when the sample is removed from the detector.
- Calibration geometry.
- Count duration.

KEY:

- * sample was counted on a puck
- ↑ sample was counted with air flow arrow pointing up
- ↓ sample was counted with air flow arrow pointing down

Gamma Spectrometer Run Log

Date: 11/5/06

Reviewed By/Date: J 11/2/06

Sample ID	Ver ¹	Det. No.	Geo ²	Count Dur. (min.) ³	Start Time	Analyst	File ID/Comments	Saved?
0610164-1	MC	8	1	300	9:27	JP	063147D08.SPC	MC
↓ -2	↓	2	↓	↓	↓	↓	062554D07.SPC	↓
GS061019-1MB	MC	6	1	335	↓	↓	062648D06.SPC	MC
0610045-4	MC	9	↓	335 ⁰	9:45	JP	061322D09.SPC	MC
0608143-17	Q	6	1	1000	16:07	MC	062649D06.SPC	Q
↓ -18	↓	8	↓	↓	↓	↓	063148D08.SPC	Q
GS061024-1MB	↓	9	↓	↓	↓	↓	061323D09.SPC	Q
0613011-2	↓	2	18	1000	↓	MC	062555D02.SPC	Q
<div style="border: 1px solid black; padding: 5px; display: inline-block;"> MC 11/5/06 </div>								

MC
11/5/06

- 1 Analyst will verify the position, detector, and geometry when the sample is removed from the detector.
- 2 Calibration geometry.
- 3 Count duration.

KEY: * sample was counted on a puck
 ↑ sample was counted with air flow arrow pointing up
 ↓ sample was counted with air flow arrow pointing down

OCT. TIME = 330
MIN.

318723 B

MC
11/5/06



Technical Comments Regarding Analysis using the FANP Gamma Spectroscopic Library

Analysis using the FANP (Fission, Activation, and Natural Products) library is limited to the list of gamma emitting radionuclides specified by Paragon Analyticals. Paragon Analyticals specifies all values assigned to the nuclides in this library. In cases where multiple gamma emissions are used to quantify activity, the most abundant emission is used for quantification in the absence of any supporting gamma emissions. It should be noted that the current software program used for gamma spectroscopic analysis is limited to a +/- 2.0 keV photo-peak resolution tolerance. Thus, any gamma emissions occurring within the same +/- 2.0 keV range will suffer interference, consequently preventing accurate quantification. Nuclide specific information regarding analysis using the FANP library is as follows:

Nuclide: ^{228}Ac Energy: various Photon Abundance: various

All activity values for ^{228}Ac are calculated using the half-life, $t_{1/2}=5.75$ years, of the long-lived ^{228}Ra parent. It is assumed that secular equilibrium is achieved between the ^{228}Ra parent and the ^{228}Ac progeny.

Nuclide: ^{212}Bi , ^{212}Pb , ^{208}Tl Energy: various Photon Abundance: various

All activity values for ^{212}Bi , ^{212}Pb , and ^{208}Tl are calculated using the half-life, $t_{1/2}=1.91$ years, of the long-lived ^{228}Th parent. It is assumed that secular equilibrium is achieved between the ^{228}Th parent and the ^{212}Bi , ^{212}Pb , ^{208}Tl progeny.

Nuclide: ^{214}Bi , ^{214}Pb Energy: various Photon Abundance: various

All activity values for ^{214}Bi and ^{214}Pb are calculated using the half-life, $t_{1/2}=1600$ years, of the long-lived ^{226}Ra parent. It is assumed that secular equilibrium is achieved between the ^{226}Ra parent and the ^{214}Bi and ^{214}Pb progeny.

Nuclide: ^{56}Co Energy: 1175.13 keV Photon Abundance: 0.0228

This emission for this nuclide suffers from possible resolution interference due to the ^{60}Co gamma emission occurring at 1173.23 keV (0.9997, abundance). Therefore, this emission will be used as an identifier only and not in the activity calculations for this nuclide.

Nuclides: ^{57}Co Energy: 122.07 Photon Abundance: 0.8560

The most abundant gamma emission specified for quantification of this nuclide suffers from possible resolution interference due to the ^{152}Eu gamma emission occurring at 121.78 keV (0.2050, abundance). Therefore, a possibility of a high bias to the ^{57}Co results may occur in the presence of elevated ^{152}Eu activity.

Nuclide: ^{134}Cs Energy: 604.66 Photon Abundance: 0.9762

The most abundant gamma emission specified for quantification of this nuclide suffers from possible resolution interference due to the ^{124}Sb gamma emission occurring at 602.71 keV (0.9826, abundance). Therefore, a possibility of a high bias to the ^{134}Cs results may occur in the presence of elevated ^{124}Sb activity.

Other gamma emissions used for quantification of this nuclide suffer from possible resolution interference due to multiple gamma emissions of ^{228}Ac . Therefore, a possible high bias to the ^{134}Cs activity results may occur in the presence of elevated ^{228}Ac activity.

Nuclide: ^{137}Cs Energy: 661.62 keV Photon Abundance: 0.8512

Cesium-137 does not emit any gamma photons useful for quantification. However, it can be assumed to be in secular equilibrium with the short-lived $^{137\text{m}}\text{Ba}$ daughter product. Therefore, the activity for ^{137}Cs is determined from the 661.62 keV gamma emission of the $^{137\text{m}}\text{Ba}$ daughter product. The calculated gamma photon abundance used in the library is the product of the 0.8998 abundance of the 661.62 keV $^{137\text{m}}\text{Ba}$ photon and the 0.946 branching ratio between ^{137}Ba and $^{137\text{m}}\text{Ba}$.

Nuclide: ^{155}Eu Energy: 105.31 Photon Abundance: 0.2120

The only gamma emission useful for quantification of this nuclide suffers from possible resolution interference due to the ^{235}U gamma emission occurring at 105 keV (0.0210, abundance). Therefore, a possibility of a high bias to the ^{155}Eu results may occur in the presence of elevated ^{235}U activity.

Nuclide: ^{40}K Energy: 1460.75 Photon Abundance: 0.1100

The only gamma emission useful for quantification of this nuclide suffers from possible resolution interference due to the ^{228}Ac gamma emission occurring at 1459.2 keV (0.0104, abundance). Therefore, a possibility of a high bias to the ^{40}K results may occur in the presence of elevated ^{228}Ac activity.

Nuclide: ^{54}Mn Energy: 834.81 Photon Abundance: 0.9997

The only gamma emission useful for quantification of this nuclide suffers from possible resolution interference due to the ^{228}Ac gamma emission occurring at 835.6 keV (0.0182, abundance). Therefore, a possibility of a high bias to the ^{54}Mn results may occur in the presence of elevated ^{228}Ac activity.

000038

Nuclide: ^{95}Nb Energy: 765.82 Photon Abundance: 0.9999

All activity values for ^{95}Nb are calculated using the half-life, $t_{1/2}=64.02$ days, of the ^{95}Zr parent. It is assumed that a transient equilibrium is achieved between the ^{95}Zr parent and the ^{95}Nb progeny.

The only gamma emission useful for quantification of this nuclide suffers from possible resolution interference due to the $^{234\text{m}}\text{Pa}$ gamma emission occurring at 766.6 keV (0.0020, abundance). Therefore, a possibility of a high bias to the ^{95}Nb results may occur in the presence of elevated $^{234\text{m}}\text{Pa}$ activity.

Nuclide: $^{234\text{m}}\text{Pa}$ Energy: 1001.03 Photon Abundance: 0.0059

All activity values for $^{234\text{m}}\text{Pa}$ are calculated using the half-life, $t_{1/2}=4.468\text{E}+09$ yrs, of the long-lived ^{238}U parent. It is assumed that secular equilibrium is achieved between the ^{238}U parent and the $^{234\text{m}}\text{Pa}$ progeny.

Nuclide: ^{106}Ru Energy: various Photon Abundance: various

Ru-106 does not emit any gamma photons. Therefore, all activity values for ^{106}Ru are calculated using the gamma emissions of the short-lived ^{106}Rh daughter. The half-life, $t_{1/2}=368.2$ days, of the ^{106}Ru parent is used in the activity calculations. It is assumed that a secular equilibrium is achieved between the ^{106}Ru parent and the ^{106}Rh progeny.

Nuclide: ^{124}Sb Energy: 602.71 Photon Abundance: 0.9826

The most abundant gamma emission specified for quantification of this nuclide suffers from possible resolution interference due to the ^{134}Cs gamma emission occurring at 604.66 keV (0.9762, abundance). Therefore, a possibility of a high bias to the ^{124}Sb results may occur in the presence of elevated ^{134}Cs activity.

Nuclide: ^{125}Sb Energy: 600.77 Photon Abundance: 0.1786

The gamma emission specified for quantification of this nuclide that occurs at 600.77 keV suffers from possible resolution interference due to the ^{124}Sb gamma emission occurring at 602.71 keV (0.9826, abundance). Therefore, this photo-peak will be used as an identifier only and not in the activity calculations for this nuclide.

Nuclide: ^{227}Th Energy: 236.00 Photon Abundance: 0.1230

All activity values for ^{227}Th are calculated using the half-life, $t_{1/2}=21.7$ yrs, of the long-lived ^{227}Ac parent. It is assumed that secular equilibrium is achieved between the ^{227}Ac parent and the ^{227}Th progeny.

Nuclide: ^{234}Th

Energy: 92.50

Photon Abundance: 0.0553

The 92.50 keV photo-peak used in this library for Th-234 quantification is actually two separate photo-peaks, occurring at 92.4 keV and 92.8 keV. The current software used for gamma spectroscopic analysis cannot resolve two photo-peaks that occur within the 2-keV resolution tolerance. Therefore, these two photopeaks are observed as a single photo-peak. Therefore, the average of the two photo-peak energies is used in this library. Also, the sum of the two photo-peak abundances, 0.0553, is used in the activity calculations for this observed 'single' photo-peak.

All activity values for ^{234}Th are calculated using the half-life, $t_{1/2}=4.468\text{E}+09$ yrs, of the long-lived ^{238}U parent. It is assumed that secular equilibrium is achieved between the ^{238}U parent and the ^{234}Th progeny.

Nuclide: ^{235}U

Energy: 185.70


Photon Abundance: 0.5720

Quantifying ^{235}U activity using the 185.70 keV photo-peak is vulnerable to a significant high bias due to interference from gamma emissions from ^{226}Ra occurring at 186.21 keV (0.0328, abundance). Therefore, this emission will be used as an identifier only and not in the activity calculations for this nuclide.


Gamma Spectroscopist

Radiochemistry Instrumentation Laboratory

2-23-05
Date


Radiochemistry Manager

2/23/05
Date

000040

OK JB
 2/23/05

Pk. #	Energy (keV)	Isotope Name	2ndary Pk #	Type	Gamma Fraction	Halflife
24	338.40	Ac-228	61	QUANT	0.1127	5.7500E+00 yrs
61	911.07	Ac-228	63	NET	0.2580	5.7500E+00 yrs
63	968.90	Ac-228	24	QUANT	0.1580	5.7500E+00 yrs
41	657.75	Ag-110M	43	NET	0.9314	2.4990E+02 dys
43	677.71	Ag-110M	46	QUANT	0.1054	2.4990E+02 dys
46	706.67	Ag-110M	49	QUANT	0.1646	2.4990E+02 dys
49	763.93	Ag-110M	59	QUANT	0.2198	2.4990E+02 dys
59	884.67	Ag-110M	62	QUANT	0.7163	2.4990E+02 dys
62	937.48	Ag-110M	84	QUANT	0.3375	2.4990E+02 dys
84	1384.27	Ag-110M	87	QUANT	0.2394	2.4990E+02 dys
87	1505.00	Ag-110M	41	QUANT	0.1289	2.4990E+02 dys
90	1808.65	Al-26	0	NET	0.9973	7.2000E+05 yrs
1	59.54	Am-241	0	NET	0.3590	4.3310E+02 yrs
30	477.56	Be-7	0	NET	0.1052	5.3440E+01 dys
48	727.17	Bi-212	0	NET	0.0658	1.9100E+00 yrs
38	609.32	Bi-214	73	NET	0.4609	1.6000E+03 yrs
73	1120.28	Bi-214	38	QUANT	0.1510	1.6000E+03 yrs
11	165.85	Ce-139	0	NET	0.8035	1.3766E+02 dys
7	133.53	Ce-144	44	NET	0.1109	2.8414E+02 dys
44	696.49	Ce-144	7	QUANT	0.0130	2.8414E+02 dys
56	846.81	Co-56	67	QUANT	0.9999	7.7300E+01 dys
67	1037.83	Co-56	76	QUANT	0.1400	7.7300E+01 dys
76	1175.13	Co-56	77	ID	0.0228	7.7300E+01 dys
77	1238.28	Co-56	81	NET	0.6760	7.7300E+01 dys
81	1360.22	Co-56	89	QUANT	0.0429	7.7300E+01 dys
89	1771.49	Co-56	56	QUANT	0.1570	7.7300E+01 dys
6	122.07	Co-57	8	NET	0.8560	2.7000E+02 dys
8	136.43	Co-57	6	QUANT	0.1068	2.7000E+02 dys
54	810.75	Co-58	0	NET	0.9945	7.0780E+01 dys
75	1173.23	Co-60	80	QUANT	0.9997	5.2721E+00 yrs
80	1332.51	Co-60	75	NET	0.9998	5.2721E+00 yrs
23	320.07	Cr-51	0	NET	0.1000	2.7700E+01 dys
31	563.26	Cs-134	32	QUANT	0.0835	2.0623E+00 yrs
32	569.29	Cs-134	37	QUANT	0.1538	2.0623E+00 yrs
37	604.66	Cs-134	52	NET	0.9762	2.0623E+00 yrs
52	795.76	Cs-134	53	QUANT	0.8553	2.0623E+00 yrs
53	801.84	Cs-134	74	QUANT	0.0869	2.0623E+00 yrs
74	1167.86	Cs-134	82	QUANT	0.0180	2.0623E+00 yrs
82	1365.13	Cs-134	31	QUANT	0.0304	2.0623E+00 yrs
42	661.62	Cs-137	0	NET	0.8512	3.0104E+01 yrs
25	344.30	Eu-152	51	QUANT	0.2650	1.3330E+01 yrs
51	778.90	Eu-152	69	QUANT	0.1294	1.3330E+01 yrs
69	1085.80	Eu-152	71	QUANT	0.1021	1.3330E+01 yrs
71	1112.07	Eu-152	85	QUANT	0.1364	1.3330E+01 yrs
85	1408.08	Eu-152	25	NET	0.2100	1.3330E+01 yrs
18	248.04	Eu-154	34	QUANT	0.0660	8.5019E+00 yrs
34	591.70	Eu-154	58	QUANT	0.0460	8.5019E+00 yrs
58	873.20	Eu-154	64	QUANT	0.1227	8.5019E+00 yrs
64	996.30	Eu-154	66	QUANT	0.1030	8.5019E+00 yrs
66	1004.80	Eu-154	18	NET	0.1801	8.5019E+00 yrs
4	105.31	Eu-155	0	NET	0.2120	4.9600E+00 yrs

Pk. #	Energy (keV)	Isotope Name	2ndary Pk #	Type	Gamma Fraction	Halflife
14	192.34	Fe-59	70	QUANT	0.0308	4.5100E+01 dys
70	1099.22	Fe-59	79	NET	0.5650	4.5100E+01 dys
79	1291.56	Fe-59	14	QUANT	0.4320	4.5100E+01 dys
20	284.29	I-131	27	QUANT	0.0614	8.0405E+00 dys
27	364.48	I-131	20	NET	0.8170	8.0405E+00 dys
86	1460.75	K-40	0	NET	0.1100	1.2800E+09 yrs
55	834.81	Mn-54	0	NET	0.9997	3.1220E+02 dys
78	1274.54	Na-22	0	NET	0.9994	2.6000E+00 yrs
45	702.50	Nb-94	0	NET	0.9790	2.0300E+04 yrs
50	765.82	Nb-95	0	NET	0.9999	6.4020E+01 dys
65	1001.03	Pa-234m	0	NET	0.0059	4.4680E+09 yrs
5	115.18	Pb-212	17	QUANT	0.0059	1.9100E+00 yrs
17	238.63	Pb-212	22	NET	0.4330	1.9100E+00 yrs
22	300.09	Pb-212	5	QUANT	0.0327	1.9100E+00 yrs
21	295.22	Pb-214	26	QUANT	0.1920	1.6000E+03 yrs
26	351.99	Pb-214	21	NET	0.3710	1.6000E+03 yrs
39	621.84	Ru-106	68	NET	0.0981	3.6820E+02 dys
68	1050.47	Ru-106	39	QUANT	0.0173	3.6820E+02 dys
36	602.71	Sb-124	40	NET	0.9826	6.0200E+01 dys
40	645.84	Sb-124	47	QUANT	0.0745	6.0200E+01 dys
47	713.82	Sb-124	83	QUANT	0.0238	6.0200E+01 dys
83	1368.21	Sb-124	88	QUANT	0.0251	6.0200E+01 dys
88	1691.04	Sb-124	36	QUANT	0.4779	6.0200E+01 dys
12	176.29	Sb-125	28	QUANT	0.0682	2.7702E+00 yrs
28	427.95	Sb-125	29	NET	0.3000	2.7702E+00 yrs
29	463.51	Sb-125	35	QUANT	0.1049	2.7702E+00 yrs
35	600.77	Sb-125	12	ID	0.1786	2.7702E+00 yrs
60	889.26	Sc-46	0	NET	0.9998	8.3850E+01 dys
16	236.00	Th-227	0	NET	0.1230	2.1700E+01 yrs
2	63.29	Th-234	3	QUANT	0.0390	4.4680E+09 yrs
3	92.50	Th-234	2	NET	0.0553	4.4680E+09 yrs
19	277.36	Tl-208	33	QUANT	0.0631	1.9100E+00 yrs
33	583.14	Tl-208	57	NET	0.8450	1.9100E+00 yrs
57	860.47	Tl-208	19	QUANT	0.1242	1.9100E+00 yrs
9	143.76	U-235	10	NET	0.1096	7.0379E+08 yrs
10	163.35	U-235	13	QUANT	0.0508	7.0379E+08 yrs
13	185.72	U-235	15	ID	0.5720	7.0379E+08 yrs
15	205.31	U-235	9	QUANT	0.0501	7.0379E+08 yrs
72	1115.52	Zn-65	0	NET	0.5060	2.4380E+02 dys

PARAGON ANALYTICS
Radiochemistry Data Package

Section 5

**QUALITY ASSURANCE
SUMMARY REPORTS**

5

000043

**No NON-CONFORMANCE REPORTS or
QUALITY ASSURANCE SUMMARY SHEETS
are included in this data package.**

000044

PARAGON ANALYTICS
Radiochemistry Data Package

Section 6

**LABORATORY
BENCH SHEETS**

6

000045

Radiochemistry Instrument Worksheet

Prep Batch: GS061025-4

Paragon Analytics

Prep Procedure: **GAMMASCAN**

Analytical QASS / NCR? Y / (N) *N/A*

Prep Num	Lab ID Collection Date	QC Type	Init Alq	Fin Alq	Units Geo.	Cnt 1 File Cnt Dur (min)	Cnt 1 Inst/Det	Cnt 1 Count Date	Cnt 2 File Cnt Dur (min)	Cnt 2 Inst/Det	Cnt 2 Count Date	Cnt 3 File Cnt Dur (min)	Cnt 3 Inst/Det	Cnt 3 Count Date	Notes
1	0610141-1 10/17/06 08:45	SMP	1000	1000	ml 01										
1	0610141-2	SMP	1000	1000	ml 01										
1	0610141-3 10/17/06 09:30	SMP	1000	1000	ml 01										
1	0610141-4 10/17/06 10:08	SMP	1000	1000	ml 01										
1	0610141-5 10/17/06 10:47	SMP	1000	1000	ml 01										
1	0610161-1 10/19/06 09:50	SMP	1000	1000	ml 01										
1	0610161-1 10/19/06 09:50	DUP	1000	1000	ml 01										
1	0610161-4 10/19/06 09:15	SMP	930	930	ml 01										
1	0610161-6 10/19/06 09:27	SMP	1000	1000	ml 01										
1	0610161-9 10/19/06 09:00	SMP	1000	1000	ml 01										
1	0610164-1 10/16/06 09:40	SMP	975	975	ml 01										
1	0610164-2 10/16/06 14:40	SMP	990	990	ml 01										
1	0610178-1 10/23/06 08:40	SMP	1000	1000	ml 01										
1	0610178-1 10/23/06 08:40	DUP	1000	1000	ml 01										
1	0610179-1 10/19/06 13:45	SMP	1000	1000	ml 01										
1	0610179-1 10/19/06 13:45	DUP	1000	1000	ml 01										
1	0610179-2 10/19/06 13:35	SMP	1000	1000	ml 01										
1	0610179-3 10/19/06 13:25	SMP	970	970	ml 01										
1	0610179-4 10/19/06 14:05	SMP	1000	1000	ml 01										
1	GS061025-4A 10/25/06 14:16	MB	1000	1000	ml 01										
1	GS061025-4B 10/25/06 14:16	MB	1000	1000	ml 01										
1	GS061025-4C 10/25/06 14:16	MB	1000	1000	ml 01										
1	GS061025-4 10/25/06 14:16	MB	1000	1000	ml 01										
1	GS061025-4 10/25/06 14:16	LCS	1000	1000	ml 01										

See Appendix

000048

Radiochemistry Instrument Worksheet

Prep Batch: GS061025-4

Paragon Analytics

Prep Procedure: **GAMMASCAN**

Analytical QASS / NCR? Y / N *JA*

Prep Num	Lab ID	Collection Date	QC Type	Init Alq	Fin Alq	Units Geo.	Cnt 1		Cnt 2		Cnt 3		Notes
							File	Count Date	File	Count Date	File	Count Date	
							Cnt 1 File	Cnt 1 Count Date	Cnt 2 File	Cnt 2 Count Date	Cnt 3 File	Cnt 3 Count Date	
							(min)		(min)		(min)		

Spike Solution Information						
Soln #	Nuclide	SolnID	Prep Conc	Units	Prep Date	Aliquot Units
S1	AM-241	718	216,780	DPM/ml	10/25/06	1000 ml
S1	CO-60	718	93,625	DPM/ml	10/25/06	1000 ml
S1	CS-137	718	82,285	DPM/ml	10/25/06	1000 ml




























000047

Radiochemistry Instrument Worksheet

Prep Batch: GS061025-4

Paragon Analytics

Sample Barcodes

0610161-1		0610178-1	
0610164-1		0610179-1	
0610141-1		0610164-2	
0610179-2		0610141-2	
0610141-3		0610179-3	
0610141-4		0610161-4	
0610179-4		0610141-5	
0610161-6		0610161-9	
0610179-1DUP		0610161-1DUP	
0610178-1DUP		GS061025-4LCS	
GS061025-4LCS		GS061025-4LCS	
GS061025-4LCS		GS061025-4BMB	
GS061025-4MB		GS061025-4AMB	
GS061025-4CMB			

000048

11/5 *CAR

Radiochemistry Instrument Worksheet

Prep Batch: GS061025-4

Paragon Analytics

Prep Procedure: GAMMASCAN

Analytical QASS / NCR? Y / (N) *DA*

Prep Num	Lab ID Collection Date	QC Type	Init Alq	Fin Alq	Units Geo.	Cnt 1 File Cnt Dur (min)	Cnt 1 Inst/Det	Count Date	Cnt 2 File Cnt Dur (min)	Cnt 2 Inst/Det	Count Date	Cnt 3 File Cnt Dur (min)	Cnt 3 Inst/Det	Count Date	Notes
1	0610141-1	SMP	1000	1000	ml 0.1	300	9	11/3/06							FANP
1	0610141-2	SMP	1000	1000	ml 0.1	251	2	11/4/06							RG 11/22/06
1	0610141-3	SMP	1000	1000	ml 0.1	265	6								
1	0610141-4	SMP	1000	1000	ml 0.1	251	8								
1	0610141-5	SMP	1000	1000	ml 0.1		9								
1	0610161-1	SMP	1000	1000	ml 0.1	330	9	11/2/06							
1	0610161-1	DUP	1000	1000	ml 0.1	315	2								CT DUP
1	0610161-4	SMP	930	930	ml 0.1	330	2								RG 11/22/06
1	0610161-6	SMP	1000	1000	ml 0.1	315	6								
1	0610161-9	SMP	1000	1000	ml 0.1	200	8								
1	0610164-1	SMP	975	975	ml 0.1	306	8	11/5/06							RG 11/22/06
1	0610164-2	SMP	990	990	ml 0.1		2								
1	0610178-1	SMP	1000	1000	ml 0.1	1000	2	11/6/06	1000	8	11/9/06				TIM
1	0610178-1	DUP	1000	1000	ml 0.1	1000	6	11/7/06							CT DUP
1	0610179-1	SMP	1000	1000	ml 0.1	330	9	11/2/06							
1	0610179-1	DUP	1000	1000	ml 0.1	180	8	11/3/06							CT DUP
1	0610179-2	SMP	1000	1000	ml 0.1	300	8	11/2/06							
1	0610179-3	SMP	970	970	ml 0.1	400	9								
1	0610179-4	SMP	1000	1000	ml 0.1	315	6								
1	GS061025-4A	MB	1000	1000	ml 0.1	600	8	11/6/06							RG 11/22/06
1	GS061025-4B	MB	1000	1000	ml 0.1	400	6	11/2/06							
1	GS061025-4C	MB	1000	1000	ml 0.1										
1	GS061025-4	MB	1000	1000	ml 0.1										
1	GS061025-4	LCS	1000	1000	ml 0.1	30	2	11/3/06							

* PRINT SPECTRA RECOUNT - DET 2 on on 11/7/06

000049

Radiochemistry Instrument Worksheet

Prep Batch: GS061025-4

Paragon Analytics

Prep Procedure: **GAMMASCAN**

Analytical QASS / NCR? Y / N

Prep Num	Lab ID	Collection Date	QC Type	Init Alq	Fin Alq	Units Geo.	Cnt 1		Cnt 2		Cnt 3		Notes
							Inst/Det	Count Date	Inst/Det	Count Date	Inst/Det	Count Date	
							Cnt 1 File Cnt Dur (min)	Cnt 2 File Cnt Dur (min)	Cnt 3 File Cnt Dur (min)	Cnt 3 Inst/Det	Cnt 3 Count Date		

Spike Solution Information						
Soln #	Nuclide	Source	Prep Conc	Units	Prep Date	Aliquot Units
S1	AM-241	798	225.652	DPM/ml	10/25/06	1000 ml
S1	CO-60	798	125.472	DPM/ml	10/25/06	1000 ml
S1	CS-137	798	86.704	DPM/ml	10/25/06	1000 ml

798/21/06




























000050

Radiochemistry Instrument Worksheet

Prep Batch: GS061025-4

Paragon Analytics

Sample Barcodes

0610161-1		0610178-1	
0610164-1		0610179-1	
0610141-1		0610164-2	
0610179-2		0610141-2	
0610141-3		0610179-3	
0610141-4		0610161-4	
0610179-4		0610141-5	
0610161-6		0610161-9	
0610179-1DUP		0610161-1DUP	
0610178-1DUP		GS061025-4LCS	
GS061025-4LCS		GS061025-4LCS	
GS061025-4LCS		GS061025-4BMB	
GS061025-4MB		GS061025-4AMB	
GS061025-4CMB			

000051

Radiochemistry Prep Worksheet

Prep Batch: GS061025-4

Paragon Analytics

Prep Procedure: GAMMASCAN

Reviewed By: atf *[Signature]* Review Date: 11/18/2006

Non-Routine Pre-Treatment? Y / N Batch: *[Signature]*

Prep OASS / NCR? Y / N

Prep SOP: PAI 739 Rev: 8

Prep SOP: NONE

Matrix Class: liquid

Prep Analyst: Krystal A. Brown

Prep Date: 10/25/2006

Prep Dept: GM

Balance:

Balance:

Samp Num	Prep Num	LabID	QC Type	Dish No.	Init Aliq ml	Fin Aliq ml	Prep Basis	Geometry	Standards	Prep Notes
1	1	0610141-1	SMP		1000	1000	Unfiltered	01		
2	1	0610141-2	SMP		1000	1000	Unfiltered	01		
3	1	0610141-3	SMP		1000	1000	Unfiltered	01		
4	1	0610141-4	SMP		1000	1000	Unfiltered	01		
5	1	0610141-5	SMP		1000	1000	Unfiltered	01		
6	1	0610161-1	SMP		1000	1000	Filtered	01		
7	1	0610161-1	DUP		1000	1000	Filtered	01		Count dup due to insufficient volume
8	1	0610161-4	SMP		930	930	Unfiltered	01		Brought up to 1000 ml with DI water
9	1	0610161-6	SMP		1000	1000	Unfiltered	01		
10	1	0610161-9	SMP		1000	1000	Unfiltered	01		
11	1	0610164-1	SMP		975	975	Unfiltered	01		Brought up to 1000 ml with DI water
12	1	0610164-2	SMP		990	990	Unfiltered	01		Brought up to 1000 ml with DI water
13	1	0610178-1	SMP		1000	1000	Unfiltered	01		
14	1	0610178-1	DUP		1000	1000	Unfiltered	01		Count dup due to insufficient volume
15	1	0610179-1	SMP		1000	1000	Filtered	01		
16	1	0610179-1	DUP		1000	1000	Filtered	01		Count Dup due to insufficient volume
17	1	0610179-2	SMP		1000	1000	Unfiltered	01		
18	1	0610179-3	SMP		970	970	Unfiltered	01		Brought up to 1000 ml with DI water
19	1	0610179-4	SMP		1000	1000	Unfiltered	01		
20	1	GS061025-4A	MB		1000	1000	Unfiltered	01		
21	1	GS061025-4B	MB		1000	1000	Unfiltered	01		
22	1	GS061025-4C	MB		1000	1000	Unfiltered	01		
23	1	GS061025-4	MB		1000	1000	Unfiltered	01		
24	1	GS061025-4	LCS		1000	1000	Unfiltered	01	S1	

Comments

000052

Radiochemistry Prep Worksheet

Prep Batch: GS061025-4

Paragon Analytics

Prep Procedure: GAMMASCAN

Reviewed By: att [Signature] Review Date: 11/18/2006

Non-Routine Pre-Treatment? Y / N / Batch: NA Re-Prep? Y / N / Prep QASS / NCR? / N /

Prep SOP: PAI 739 Rev: 8
 Prep SOP: NONE
 Matrix Class: liquid

Prep Analyst: Krystal A. Brown
 Prep Date: 10/25/2006
 Prep Dept: GM

Balance:
 Balance:

Samp Num	Prep Num	LabID	QC Type	Dish No.	Init Alq ml	Fin Alq ml	Prep Basis	Geometry	Standards	Prep Notes

Spiked By: N/A Date: N/A
 Witnessed By: N/A Date: N/A

Spike Solution Information

Solin #	Nuclide	SolinID	Prep Conc	Units	Prep Date	Aliquot	Units	Pipet ID
S1	AM-241	718	216.780	DPM/ml	10/25/06	1000	ml	
S1	CO-60	718	93.625	DPM/ml	10/25/06	1000	ml	
S1	CS-137	718	82.285	DPM/ml	10/25/06	1000	ml	

000053

Comments

Supersedes: 10/27/06 9:36

Radiochemistry Prep Worksheet

Prep Batch: GS061025-4

11/20/06
Review Date: 11/20/2006

Reviewed By: kab *KAB*

Prep *QASS/NCR* *Y* *N* Batch: *318B15*

Paragon Analytics

Prep Procedure: GAMMASCAN

Non-Routine Pre-Treatment? *Y* / *N* Batch: *N/A*

Prep SOP: PAI 739 Rev: 8

Prep SOP: NONE

Matrix Class: liquid

Prep Analyst: Krystal A. Brown

Prep Date: 10/25/2006

Prep Dept: GM

Balance:

Balance:

Sample Num	Prep Num	LabID	QC Type	Dish No.	Init Alq ml	Fin Alq ml	Prep Basis	Geometry	Standards	Prep Notes
1	1	0610141-1	SMP	1000	1000	1000	Unfiltered	01		
2	1	0610141-2	SMP	1000	1000	1000	Unfiltered	01		
3	1	0610141-3	SMP	1000	1000	1000	Unfiltered	01		
4	1	0610141-4	SMP	1000	1000	1000	Unfiltered	01		
5	1	0610141-5	SMP	1000	1000	1000	Unfiltered	01		
6	1	0610161-1	SMP	1000	1000	1000	Filtered	01		
7	1	0610161-1	DUP	1000	1000	1000	Filtered	01		Count dup due to insufficient volume
8	1	0610161-4	SMP	930	930	930	Unfiltered	01		Brought up to 1000 ml with DI water
9	1	0610161-6	SMP	1000	1000	1000	Unfiltered	01		
10	1	0610161-9	SMP	1000	1000	1000	Unfiltered	01		
11	1	0610164-1	SMP	975	975	975	Unfiltered	01		Brought up to 1000 ml with DI water
12	1	0610164-2	SMP	990	990	990	Unfiltered	01		Brought up to 1000 ml with DI water
13	1	0610178-1	SMP	1000	1000	1000	Unfiltered	01		
14	1	0610178-1	DUP	1000	1000	1000	Unfiltered	01		Count dup due to insufficient volume
15	1	0610179-1	SMP	1000	1000	1000	Unfiltered	01		
16	1	0610179-1	DUP	1000	1000	1000	Unfiltered	01		Count Dup due to insufficient volume
17	1	0610179-2	SMP	1000	1000	1000	Unfiltered	01		
18	1	0610179-3	SMP	970	970	970	Unfiltered	01		Brought up to 1000 ml with DI water
19	1	0610179-4	SMP	1000	1000	1000	Unfiltered	01		
20	1	GS061025-4A	MB	1000	1000	1000	Unfiltered	01		
21	1	GS061025-4B	MB	1000	1000	1000	Unfiltered	01		
22	1	GS061025-4C	MB	1000	1000	1000	Unfiltered	01		
23	1	GS061025-4	MB	1000	1000	1000	Unfiltered	01		
24	1	GS061025-4	LCS	1000	1000	1000	Unfiltered	01	S1	

000054

Comments

Radiochemistry Prep Worksheet

Prep Batch: GS061025-4

Paragon Analytics

Prep Procedure: GAMMASCAN

Reviewed By: kab *KAB*

Review Date: 11/20/2006

Non-Routine Pre-Treatment? Y (N) Batch: *NA*

Prep QASS (NCR?) Y (N) *318815*

Prep SOP: PAI 739 Rev: 8

Prep Analyst: Krystal A. Brown

Prep Date: 10/25/2006

Prep Dept: GM

Balance:

Balance:

Prep Notes

Standards

Geometry

Prep Basis

Fin Alq

Init Alq

Dish No.

QC Type

LabID

Samp Num

Prep Num

Date: N/A

Spiked By: N/A

Date: N/A

Witnessed By: N/A

Spike Solution Information						
Soln #	Nuclide	SolnID	Prep Conc	Units	Prep Date	Aliquot Units
S1	AM-241	718	216.780	DPM/ml	10/25/06	1000 ml
S1	CO-60	718	93.625	DPM/ml	10/25/06	1000 ml
S1	CS-137	718	82.285	DPM/ml	10/25/06	1000 ml

Comments

000055

Radiochemistry Prep Worksheet

Prep Batch: GS061025-4

Paragon Analytics

Prep Procedure: **GAMMASCAN**

Reviewed By: kab *KAB*

Review Date: 10/27/2006

Non-Routine Pre-Treatment? Y N Batch: *N/A* Re-Prep? Y N Prep QASS / NCR? Y N *N/A*

Prep SOP: PAI 739 Rev: 8

Prep Analyst: Krystal A. Brown

Balance:

Prep Date: 10/25/2006

Balance:

Matrix Class: liquid

Prep Dept: GM

Stamp	Prep Num	LabID	QC Type	Dish No.	Init Alq ml	Fin Alq ml	Prep Basis	Geometry	Standards	Prep Notes
	1	0610141-1	SMP	1000	1000	1000	Unfiltered	01		
	2	0610141-2	SMP	1000	1000	1000	Unfiltered	01		
	3	0610141-3	SMP	1000	1000	1000	Unfiltered	01		
	4	0610141-4	SMP	1000	1000	1000	Unfiltered	01		
	5	0610141-5	SMP	1000	1000	1000	Unfiltered	01		
	6	0610161-1	SMP	1000	1000	1000	Unfiltered	01		
	7	0610161-1	DUP	1000	1000	1000	Unfiltered	01		
	8	0610161-4	SMP	930	930	930	Unfiltered	01		
	9	0610161-6	SMP	1000	1000	1000	Unfiltered	01		
	10	0610161-9	SMP	1000	1000	1000	Unfiltered	01		
	11	0610164-1	SMP	975	975	975	Unfiltered	01		
	12	0610164-2	SMP	990	990	990	Unfiltered	01		
	13	0610178-1	SMP	1000	1000	1000	Unfiltered	01		
	14	0610178-1	DUP	1000	1000	1000	Unfiltered	01		
	15	0610179-1	SMP	1000	1000	1000	Unfiltered	01		
	16	0610179-1	DUP	1000	1000	1000	Unfiltered	01		
	17	0610179-2	SMP	1000	1000	1000	Unfiltered	01		
	18	0610179-3	SMP	970	970	970	Unfiltered	01		
	19	0610179-4	SMP	1000	1000	1000	Unfiltered	01		
	20	GS061025-4A	MB	1000	1000	1000	Unfiltered	01		
	21	GS061025-4B	MB	1000	1000	1000	Unfiltered	01		
	22	GS061025-4C	MB	1000	1000	1000	Unfiltered	01		
	23	GS061025-4	MB	1000	1000	1000	Unfiltered	01		
	24	GS061025-4	LCS	1000	1000	1000	Unfiltered	01	S1	

000058

Comments

Radiochemistry Prep Worksheet

Prep Batch: GS061025-4

Paragon Analytics

Prep Procedure: GAMMASCAN

Reviewed By: kab Review Date: 10/27/2006

Non-Routine Pre-Treatment? Y / N Batch: Re-Prep? Y / N Batch: Prep GASS / NCR? Y / N

Prep SOP: PAI 739 Rev: 8
 Prep SOP: NONE
 Matrix Class: liquid

Prep Analyst: Krystal A. Brown Balance:
 Prep Date: 10/25/2006 Balance:
 Prep Dept: GM

Samp Num	Prep Num	LabID	QC Type	Dish No.	Init Alq ml	Fin Alq ml	Prep Basis	Geometry	Standards	Prep Notes

Spiked By: N/A Date: N/A

Witnessed By: N/A Date: N/A

Soln #	Nuclide	SolnID	Prep Conc	Units	Prep Date	Aliquot	Units	Pipet ID
S1	AM-241	798	225.652	DPM/ml	10/25/06	1000	ml	
S1	CO-60	798	125.472	DPM/ml	10/25/06	1000	ml	
S1	CS-137	798	86.704	DPM/ml	10/25/06	1000	ml	

000057

Comments

Radiochemistry Prep Worksheet

Prep Batch: GS061025-4

Paragon Analytics

Prep Procedure: **GAMMASCAN**

Prep Batch Not Validated!!!

Reviewed By: _____ Review Date: _____

Non-Routine Pre-Treatment? Y / N Batch: _____ Re-Prep? Y / N Batch: _____ Prep QASS / NCR? Y / N _____

Prep SOP: PAI 739 Rev: 8 Prep Analyst: Krystal A. Brown *KAB* Balance: _____

Prep SOP: NONE Prep Date: 10/25/2006 Balance: _____

Matrix Class: liquid Prep Dept: GM

Sampl Num	Prep Num	LabID	QC Type	Dish No.	Init Alq ml	Fin Alq ml	Prep Basis	Geometry	Standards	Prep Notes
1	1	0610141-1	SMP		1000	1000	Unfiltered	01		
2	1	0610141-2	SMP		1000	1000	Unfiltered	01		
3	1	0610141-3	SMP		1000	1000	Unfiltered	01		
4	1	0610141-4	SMP		1000	1000	Unfiltered	01		
5	1	0610141-5	SMP		1000	1000	Unfiltered	01		
6	1	0610161-1	SMP		1000	1000	Unfiltered	01		
7	1	0610161-1	DUP		1000	1000	Unfiltered	01		
8	1	0610161-4	SMP		1000	1000	Unfiltered	01		
9	1	0610161-6	SMP		1000	1000	Unfiltered	01		
10	1	0610161-9	SMP		1000	1000	Unfiltered	01		
11	1	0610164-1	SMP		1000	1000	Unfiltered	01		
12	1	0610164-2	SMP		1000	1000	Unfiltered	01		
13	1	0610178-1	SMP		1000	1000	Unfiltered	01		
14	1	0610179-1	SMP		1000	1000	Unfiltered	01		
15	1	0610179-1	DUP		1000	1000	Unfiltered	01		
16	1	0610179-2	SMP		1000	1000	Unfiltered	01		
17	1	0610179-3	SMP		1000	1000	Unfiltered	01		
18	1	0610179-4	SMP		1000	1000	Unfiltered	01		
19	1	GS061025-4A	MB		1000	1000	Unfiltered	01		
20	1	GS061025-4B	MB		1000	1000	Unfiltered	01		
21	1	GS061025-4C	MB		1000	1000	Unfiltered	01		
22	1	GS061025-4	MB		1000	1000	Unfiltered	01		
23	1	GS061025-4	LCS		1000	1000	Unfiltered	01	S1	

C6 10 178-1 Dup* *KAB* 10/27/06

000058

Radiochemistry Prep Worksheet

Prep Batch: GS061025-4

Paragon Analytics

Prep Procedure: **GAMMASCAN**

Prep Batch Not Validated!!!

Reviewed By:

Review Date:

Non-Routine Pre-Treatment? Y / N Batch: _____ Re-Prep? Y / N Batch: _____ Prep QASS / NCR? Y / N _____

Prep SOP: PAI 739 Rev: 8

Prep Analyst: Krystal A. Brown *KAB*

Balance:

Prep Date: 10/25/2006

Balance:

Matrix Class: liquid

Prep Dept: GM

Samp Num	Prep Num	LabID	QC Type	Dish No.	Init Aliq ml	Fin Aliq ml	Prep Basis	Geometry	Standards	Prep Notes

Spiked By: N/A Date: N/A

Witnessed By: _____ Date: _____

Spike Solution Information

Soln #	Nuclide	SolnID	Prep Conc	Units	Prep Date	Aliquot	Units	Pipet ID
S1	AM-241	798	225.652	DPM/ml	10/25/06	1000	ml	
S1	CO-60	798	125.472	DPM/ml	10/25/06	1000	ml	
S1	CS-137	798	86.704	DPM/ml	10/25/06	1000	ml	

000059

SAMPLE CONDITION FORM (LIQUID)

ANALYST: KAB

ANALYSIS DATE: 10/25/06

METHOD: Prep

WORK ORDER	SAMPLE ID	SAMPLE CONDITION		
		pH	Color	Remarks
0610161	1	7.2	Clear	Filtered
↓	4			Unfiltered
	6			
	9			
0610164	1			
↓	2			
0610178	1			
0610179	1			Filtered
↓	2			Unfiltered
	3			
	4			
0610141	1			
↓	2			
	3			
	4			
↓	5			
KAB 10/25/06				

PARAGON ANALYTICS
Radiochemistry Data Package

Section 7

**STANDARDS
TRACEABILITY
DOCUMENTS**

7

000061

CERTIFICATE OF CALIBRATION

Standard Radionuclide Source

66352-307

PAID 07/18
rec'd 8-20-03

1.0 Liter Solid in 138G GA-MA Beaker

This standard radionuclide source was prepared using aliquots measured gravimetrically from master radionuclide solution sources. The Am-241 was calibrated by 4 pi alpha liquid scintillation counting. All other radionuclides were calibrated using a germanium gamma spectrometer system. Calibration and purity were checked using a germanium gamma spectrometer system. At the time of calibration no interfering gamma-ray emitting impurities were detected. The gamma-ray emission rates for the most intense gamma-ray lines are given. Analytix maintains traceability to the National Institute of Standards and Technology through a Measurements Assurance Program as described in USNRC Regulatory Guide 4.15, Rev. 1, February, 1979.

US Patent 4,430,258; UK Patent GB2,149,194B; CA Patent 1,196,776.
Density of solid matrix 1.15 g/cc.

Calibration date: July 1, 2003 12:00 EST

ISOTOPE	GAMMA-RAY ENERGY	HALF-LIFE		GAMMA-RAYS PER SECOND	TOTAL UNCERTAINTY %
Am-241	59.5	432	y	1304	3.0
Cd-109	88	462.6	d	1862	3.3
Co-57	122	271.79	d	1032	2.8
Ce-139	166	137.6	d	1419	2.8
Hg-203	279	46.61	d	3194	2.7
Sn-113	392	115.1	d	1960	2.6
Cs-137	662	30.07	y	1260	3.0
Y-88	898	106.6	d	5060	2.6
Co-60	1173	5.2714	y	2402	2.7
Co-60	1332	5.2714	y	2427	2.6
Y-88	1836	106.6	d	5287	2.6

P O NUMBER EW060303, Item 2

SOURCE PREPARED BY:

M. D. Currie, Radiochemist

Q A APPROVED:

W.M. Putz 8-1-03

This standard will expire one year after the calibration date.

000062

ANALYTICS

1380 Seaboard Industrial Blvd.
Atlanta, Georgia 30318-U.S.A.

Phone (404) 352-8677
Fax (404) 352-2837
email: analytics@mindspring.com
www.analyticsinc.com

ANALYSIS OF UNCERTAINTY FOR MIXED GAMMA STANDARDS BATCH 113

CALIBRATION DATE: JULY 1, 2003 12:00 EST

ISOTOPE	ENERGY (keV)	CALIBRATION METHOD ¹	STATISTICS ²	CALIBRATION ²	PEAK FITTING ²	GEOMETRY ²	IMPURITIES ²	WEIGHTING ²	COMBINED STANDARD UNCERTAINTY ¹	RELATIVE EXPANDED UNCERTAINTY ² (K=2)
Cd-109	88	HPGe	0.16	1.1	0.88	0.8	0	0.2	1.64	3.3
Co-57	122	HPGe	0.23	0.93	0.71	0.7	0	0.2	1.40	2.8
Ce-139	166	HPGe	0.17	1.0	0.58	0.7	0	0.2	1.38	2.8
Hg-203	279	HPGe	0.11	1.1	0.34	0.7	0	0.2	1.37	2.7
Sn-113	392	HPGe	0.21	1.0	0.35	0.7	0	0.2	1.30	2.6
Cs-137	662	HPGe	0.36	1.1	0.60	0.7	0	0.2	1.49	3.0
Y-88	898	HPGe	0.19	1.0	0.33	0.7	0	0.2	1.29	2.6
Co-60	1173	HPGe	0.31	0.97	0.45	0.7	0	0.2	1.33	2.7
Co-60	1332	HPGe	0.33	0.93	0.48	0.7	0	0.2	1.32	2.6
Y-88	1836	HPGe	0.24	1.0	0.35	0.7	0	0.2	1.31	2.6

OPTIONAL ADDITIONAL ISOTOPES

Pb-210	46.5	4NLS	0.33	1.1	0	0.9	0.30	0.2	1.50	3.0
Am-241	59.5	4NLS	0.33	1.1	0	0.9	0.30	0.2	1.50	3.0
Sr-85	514	IC	0.30	1.1	0	0.7	0.17	0.2	1.36	2.7
Cs-134	605	IC	0.30	1.0	0	0.8	0.17	0.2	1.34	2.7
Cs-134	796	IC	0.30	1.0	0	0.8	0.17	0.2	1.34	2.7
Mn-54	835	IC	0.30	1.0	0	0.8	0.17	0.2	1.34	2.7
Zn-65	1116	IC	0.30	1.0	0	0.8	0.17	0.2	1.34	2.7

¹Calibration Methods:

4NLS (4 pi Liquid Scintillation Counting)
HPGe (High Purity Germanium Gamma Ray Spectrometer)
IC (Gamma Ray Ionization Chamber)

²As Percent (%)

No interfering gamma emitting impurities were detected during calibration. Depending on the resolution and energy dispersion (keV/channel) of the measuring system, the following spectral conflicts may occur: (1) between the 88 keV gamma-ray and the X-rays emitted in the decay of Hg-203, (2) between the 1333 keV gamma-ray and the 1325 keV single escape peak from the 1836 keV gamma-ray.

000063

PARAGON ANALYTICS
Radiochemistry Data Package

Section 8

CHAIN OF CUSTODY

8

000064

Paragon Analytics

Sample Number(s) Cross-Reference Table

Paragon OrderNum: 0610164

Client Name: ACZ Laboratories, Inc.

Client Project Name:

Client Project Number: L59449

Client PO Number:

Client Sample	Lab Sample Number	COC Number	Matrix	Date Collected	Time Collected
L59449-01	0610164-1		WATER	16-Oct-06	9:40
L59449-02	0610164-2		WATER	16-Oct-06	14:40

CONDITION OF SAMPLE UPON RECEIPT FORM

Paragon Analytics

Client: ACIZ
Project Manager: LS

Workorder No: 0610164
Initials: SL Date: 10-20-06

1. Does this project require any special handling in addition to standard Paragon procedures?		YES	<input checked="" type="radio"/> NO
2. Are custody seals on shipping containers intact?	NONE	<input checked="" type="radio"/> YES	NO
3. Are Custody seals on sample containers intact?	<input checked="" type="radio"/> NONE	YES	NO
4. Is there a COC (Chain-of-Custody) present or other representative documents?		<input checked="" type="radio"/> YES	NO
5. Are the COC and bottle labels complete and legible?		<input checked="" type="radio"/> YES	NO
6. Is the COC in agreement with samples received? (IDs, dates, times, no. of samples, no. of containers, matrix, requested analyses, etc.)		<input checked="" type="radio"/> YES	NO
7. Were airbills / shipping documents present and/or removable?	DROP OFF	<input checked="" type="radio"/> YES	NO
8. Are all aqueous samples requiring preservation preserved correctly? (excluding volatiles)	N/A	<input checked="" type="radio"/> YES	NO
9. Are all aqueous non-preserved samples pH 4-9?	<input checked="" type="radio"/> N/A	YES	NO
10. Is there sufficient sample for the requested analyses?		<input checked="" type="radio"/> YES	NO
11. Were all samples placed in the proper containers for the requested analyses?		<input checked="" type="radio"/> YES	NO
12. Are all samples within holding times for the requested analyses?		<input checked="" type="radio"/> YES	NO
13. Were all sample containers received intact? (not broken or leaking, etc.)		<input checked="" type="radio"/> YES	NO
14. Are all samples requiring no headspace (VOC, GRO, Rx CN/S, radon), headspace free? Size of bubble: ___ < green pea ___ > green pea	N/A	<input checked="" type="radio"/> YES	NO
15. Were samples checked for and free from the presence of residual chlorine? (Applicable when PM has indicated samples are from a chlorinated water source; note if field preservation with sodium thiosulfate was not observed.)	<input checked="" type="radio"/> N/A	YES	NO
16. Were the samples shipped on ice?		<input checked="" type="radio"/> YES	NO
17. Were cooler temperatures measured at 0.1-6.0°C?	IR gun used*: <input checked="" type="radio"/> #2 #4	RAD ONLY	<input checked="" type="radio"/> YES NO
Cooler #: <u>1</u>			
Temperature (°C): <u>2.4</u>			
No. of custody seals on cooler: <u>0</u>			
External µR/hr reading: <u>12</u>			
Background µR/hr reading: <u>12</u>			
Were external µR/hr readings ≤ two times background and within DOT acceptance criteria? <input checked="" type="radio"/> YES / NO / NA (If no, see Form 008.)			

Additional Information: PROVIDE DETAILS BELOW FOR A NO RESPONSE TO ANY QUESTION ABOVE, EXCEPT #1 AND #16.

→ Sample #1 (L 59449-01) recieved only 2 40 ml Vials for ~~8260?~~ (H3)
~~or~~ H3 (if these vials are for H3, they are according to the label, Preserved with HCL → 1 Liter poly pres. w. HNO3)
 Sample #2 (L 59449-01) recieved only 1 40ml Vial for ~~VOC~~ or H3?
 ↓
~~Pres with HCL~~
 40 ml Vials for H3 Not preserved. Limited Vol. for H3

If applicable, was the client contacted? YES / NO / NA Contact: [Signature] Date/Time: 10/20/06

Project Manager Signature / Date: [Signature] / 10/20/06


*IR Gun #2: Oakton, SN 29922500201-0066 *IR Gun #4: Oakton, SN 2372220101-0002

0610164

SHIPPING DEPARTMENT
(970)879-6590
ACZ LABORATORIES INC
2773 DOWNHILL DRIVE
STEAMBOAT SPRIN CO 80487

15 LBS 1 OF 1

SHIP LANCE STEERE
(970)490-1511
TO: PARAGON ANALYTICAL LABS
225 COMMERCE DRIVE
FT COLLINS CO 80524



CO 805 0-01



2.4

UPS GROUND

TRACKING #: 1Z 810 130 03 9407 9509



BILLING: P/P

International Shipping Notice: Consignee and addressee must be advised of the applicable country and office. Since this country is addressed by the Consignee of the shipment of certain items, please be responsible for the applicable country and office. This notice is for informational purposes only. It is not intended to constitute an offer of insurance or any other financial product. For more information, please contact your local UPS office. © 2008 United Parcel Service of America, Inc.

DWS ID Lab

1110



1Z8101300394079509

000068

PARAGON ANALYTICS
Radiochemistry Data Package

Section 9

**ADDITIONAL
SUPPORTING
DOCUMENTATION**

9

000069

Gamma Spectroscopy

Initial Calibration
Standards Traceability

000070

ANALYTICS



RSO#817
Rec'd 3/10/06
DWS

1380 Seaboard Industrial Blvd.
Atlanta, Georgia 30318 • U.S.A.

Phone (404) 352-8677
Fax (404) 352-2837

CERTIFICATE OF CALIBRATION

Standard Radionuclide Source

72391-307

Sand in 16 Ounce PP MRP Jar

This standard radionuclide source was prepared using aliquots measured gravimetrically from master radionuclide solution sources. The Am-241 was calibrated by 4 pi alpha liquid scintillation counting. All other radionuclides were calibrated using a germanium gamma spectrometer system. Calibration and purity were checked using a germanium gamma spectrometer system. At the time of calibration no interfering gamma-ray emitting impurities were detected. The gamma-ray emission rates for the most intense gamma-ray lines are given. Analytix maintains traceability to the National Institute of Standards and Technology through a Measurements Assurance Program as described in USNRC Regulatory Guide 4.15, Rev. 1, February, 1979.

Calibration date: January 1, 2006 12:00 EST

ISOTOPE	GAMMA-RAY ENERGY	HALF-LIFE		GAMMA-RAYS PER SECOND	TOTAL UNCERTAINTY %
Am-241	59.5	432	y	1329	3.0
Cd-109	88	462.6	d	1875	3.3
Co-57	122	271.79	d	996.0	3.0
Ce-139	166	137.6	d	1400	2.8
Hg-203	279	46.61	d	3215	2.7
Sn-113	392	115.1	d	1939	2.6
Cs-137	662	30.07	y	1252	3.0
Y-88	898	106.6	d	4692	2.6
Co-60	1173	5.2714	y	2378	2.7
Co-60	1332	5.2714	y	2399	2.6
Y-88	1836	106.6	d	4875	2.6

500 grams/290 mL of customer supplied sand.
P O NUMBER 71239, Rev., 2/3/06 REL., Item 3

SOURCE PREPARED BY: M. Taskaeva
M. Taskaeva, Radiochemist

Q A APPROVED: [Signature] 03-07-2006

This standard will expire one year after the calibration date.

000071

 SEEKER G A M M A A N A L Y S I S R E S U L T S PS Version 1.8.4

Paragon Analytics, Div. of DataChem Lab
 GammaScan

Geo 13 / Solid

Sample ID: 0613003-2 FWHM CAL (817)

```
-----
Sampling Start: 01/01/2006 12:00:00 | Counting Start: 05/16/2006 20:43:28
Sampling Stop: 01/01/2006 12:00:00 | Decay Time. . . . . 3.25E+003 Hrs
Buildup Time. . . . . 0.00E+000 Hrs | Live Time . . . . . 2700 Sec
Sample Size . . . . . 5.00E+002 g | Real Time . . . . . 2811 Sec
Collection Efficiency . . . . . 1.0000 | Spc. File . . . . . 061202D02.SPC
-----
```

Detector #: 2 (Detector 2)

Energy (keV) = -0.77 + 0.501*Ch + 3.48E-08*Ch^2 + 0.00E+00*Ch^3 05/16/2006
 FWHM (keV) = 0.65 + 0.009*En + 1.05E-03*En^2 + 0.00E+00*En^3 03/27/2006
 Where En = Sqrt(Energy in keV)

 Search Sensitivity: 1.00 | Sigma Multiplier: 2.00 | Search Start/End: 80/4000

=====

PEAK SEARCH RESULTS

=====

PK. #	ENERGY (keV)	ADDRESS CHANNEL	NET/MDA COUNTS	UN-CERTAINTY	C.L. COUNTS	BKG COUNTS	FWHM (keV)	FLAG
1	59.52	120.43	15714	364	217	9518	0.81	a
2	70.81	142.98	389	247	201	8923	0.68	a
3	72.86	147.07	779	250	201	8923	0.72	b
4	82.54	166.41	2090	451	363	19473	1.54	a HiResid Wide Pk
5	85.06	171.44	4401	687	554	32455	2.48	b HiResid
6	87.96	177.24	72878	609	232	10818	0.83	c HiResid
7	122.09	245.41	46605	523	243	10903	0.92	a
8	136.52	274.22	6040	319	229	9690	0.96	a
9	165.85	332.82	48502	510	212	8297	0.97	a
10	255.14	511.16	1549	248	194	6400	1.12	a
11	279.19	559.20	22194	369	179	5488	1.11	a
12	391.73	783.97	34450	415	153	4302	1.27	a
13	511.02	1022.22	730	230	184	4986	2.04	a
14	565.63	1131.29	61	96	78	1506	0.73	a NET< CL
15	661.72	1323.19	33826	408	145	3865	1.62	a
16	813.63	1626.56	707	144	110	2366	1.44	a HiResid
17	821.66	1642.61	222	190	155	3681	2.27	b HiResid
18	898.12	1795.31	39719	432	136	3430	1.93	a
19	1173.34	2344.89	38827	414	104	1854	2.26	a
20	1325.10	2647.92	1131	168	126	1863	4.33	a Wide Pk
21	1332.56	2662.82	35672	390	81	1075	2.50	b
22	1836.13	3668.26	24519	319	48	326	3.10	a HiResid

061201D02.SPC Analyzed by

SEEKER CALIBRATION RESULTS Version 2.0.4

Sample ID: DAILY CHECK

Stds. Match Tolerance: 2.00 keV

Detector Number: 02 Calibration Date. . . 05/16/2006 20:14:57

Energy(keV) = -0.77 + 0.501*Ch + 3.48e-08*Ch^2 + 0.00e+00*Ch^3

Pk. #	Measured Centroid	Calculated Energy	Energy (keV)	% Difference
1	120.40	59.51	59.50	0.01
2	1322.99	661.62	661.64	-0.00
3	2344.74	1173.26	1173.21	0.00
4	2662.59	1332.45	1332.48	-0.00

Calibration Results Saved.

061202D02.SPC Analyzed by

SEEKER CALIBRATION RESULTS Version 2.0.4

Sample ID: 0613003-2 FWHM CAL (817)

Stds. Match Tolerance: 2.00 keV

Detector Number: 02 Calibration Date. . . 05/16/2006 20:43:28

FWHM(keV) = $0.69 + 0.006*En + 1.18e-03*En^2 + 0.00e+00*En^3$
(Where En = SQR(Energy in keV))

Pk. #	Energy (keV)	Measured FWHM(keV)	% Diff.	Calculated FWHM(keV)	% Diff.	Prev.Calc. FWHM(keV)
1	59.50	0.808	-0.03	0.808	-2.65	0.787
2	88.04	0.833	2.18	0.851	-2.33	0.832
3	122.06	0.922	-2.25	0.901	-2.11	0.883
4	165.85	0.972	-0.89	0.964	-1.97	0.945
5	279.00	1.112	0.75	1.120	-1.95	1.099
6	391.68	1.269	0.18	1.271	-2.13	1.245
7	661.64	1.622	0.17	1.625	-2.73	1.581
8	898.02	1.929	-0.04	1.928	-3.25	1.868
9	1173.21	2.264	0.59	2.278	-3.79	2.195
10	1332.48	2.500	-0.85	2.479	-4.07	2.382
11	1836.01	3.104	0.18	3.109	-4.81	2.967

Calibration Results Saved.

 SEEKER G A M M A A N A L Y S I S R E S U L T S PS Version 1.8.4

Paragon Analytics, Div. of DataChem Lab
 GammaScan

Geo 13 / Solid

Sample ID: 0613003-6 FWHM CAL (817)

OK DUB 7/25/06

 Sampling Start: 01/01/2006 12:00:00 | Counting Start: 07/24/2006 15:42:28
 Sampling Stop: 01/01/2006 12:00:00 | Decay Time. 4.90E+003 Hrs
 Buildup Time. 0.00E+000 Hrs | Live Time 3600 Sec
 Sample Size 5.00E+002 g | Real Time 3727 Sec
 Collection Efficiency 1.0000 | Spc. File 061950D06.SPC

Detector #: 6 (Detector 6)

Energy(keV) = -0.68 + 0.501*Ch +-1.46E-08*Ch^2 + 0.00E+00*Ch^3 07/24/2006

FWHM(keV) = 0.72 + 0.012*En + 5.28E-04*En^2 + 0.00E+00*En^3 09/15/2005

Where En = Sqrt(Energy in keV)

 Search Sensitivity: 1.00 | Sigma Multiplier: 2.00 | Search Start/End: 80/4000

=====

PEAK SEARCH RESULTS

=====

PK. #	ENERGY (keV)	ADDRESS CHANNEL	NET/MDA COUNTS	UN-CERTAINTY	C.L. COUNTS	BKG COUNTS	FWHM (keV)	FLAG
1	59.49	120.17	10241	363	248	10458	1.21	a HiResid
2	72.99	147.12	281	261	213	9168	0.82	a
3	82.47	166.05	1593	452	366	18532	1.56	a HiResid Wide Pk
4	84.96	171.03	4655	673	542	29651	2.63	b HiResid
5	87.99	177.07	60307	616	306	14825	1.31	c HiResid
6	122.08	245.14	40603	536	290	13353	1.23	a
7	136.56	274.05	5061	326	241	9910	1.08	a
8	165.87	332.59	37403	494	253	10113	1.28	a
9	255.28	511.13	854	238	190	6148	1.07	a
10	279.26	559.02	8775	304	197	6128	1.30	a
11	391.78	783.71	25487	380	170	5004	1.43	a
12	491.58	983.00	5	103	85	1760	0.63	a NET< CL
13	510.52	1020.83	581	243	196	5393	2.09	a Wide Pk
14	511.88	1023.55	238	181	147	3734	1.50	b
15	512.86	1025.50	52	101	82	1660	0.63	c NET< CL
16	575.83	1151.25	104	109	88	1769	0.84	a
17	661.69	1322.70	37171	427	151	3724	1.64	a
18	814.03	1626.93	437	188	151	3368	1.92	a
19	898.14	1794.91	28356	378	141	3477	1.80	a
20	1173.31	2344.44	42262	428	99	1654	1.99	a
21	1325.43	2648.25	611	122	92	1250	2.93	a HiResid Wide Pk
22	1332.55	2662.48	38425	402	71	902	2.09	b HiResid

=====

PEAK SEARCH RESULTS

=====

PK. #	ENERGY (keV)	ADDRESS CHANNEL	NET/MDA COUNTS	UN- CERTAINTY	C.L. COUNTS	BKG COUNTS	FWHM (keV)	FLAG
23	1570.93	3138.57	14	55	45	492	1.02	a NET< CL
24	1835.94	3667.86	17334	267	37	221	2.42	a HiResid

061950D06.SPC Analyzed by

SEEKER CALIBRATION RESULTS Version 2.0.4

Sample ID: 0613003-6 FWHM CAL (817)
Stds. Match Tolerance: 2.00 keV

Detector Number: 06 Calibration Date. . . 07/24/2006 15:42:28

FWHM(keV) = $1.19 + -0.002*En + 7.29e-04*En^2 + 0.00e+00*En^3$
(Where En = SQR(Energy in keV))

Pk. #	Energy (keV)	Measured FWHM(keV)	% Diff.	Calculated FWHM(keV)	% Diff.	Prev.Calc. FWHM(keV)
1	59.50	1.206	1.04	1.219	-43.87	0.847
2	88.04	1.313	-6.18	1.236	-40.09	0.883
3	122.06	1.231	2.13	1.258	-36.61	0.921
4	165.85	1.276	0.77	1.286	-33.13	0.966
5	279.00	1.299	4.50	1.361	-26.93	1.072
6	391.68	1.432	0.33	1.437	-22.90	1.169
7	661.64	1.642	-1.27	1.621	-17.22	1.383
8	898.02	1.803	-1.01	1.785	-14.48	1.559
9	1173.21	1.994	-0.85	1.977	-12.56	1.756
10	1332.48	2.091	-0.11	2.088	-11.82	1.867
11	1836.01	2.423	0.78	2.443	-10.52	2.210

Calibration Results Saved.

OK
JJB
7/25/06

062429D08.SPC Analyzed by CW

SEEKER G A M M A A N A L Y S I S R E S U L T S PS Version 1.8.4

Paragon Analytics, Div. of DataChem Lab
GammaScan

Geo 13 / Solid

Sample ID: 0613003-8 FWHM CAL (817)

Sampling Start: 01/01/2006 12:00:00 | Counting Start: 08/17/2006 10:06:09
Sampling Stop: 01/01/2006 12:00:00 | Decay Time 5.47E+003 Hrs
Buildup Time 0.00E+000 Hrs | Live Time 1800 Sec
Sample Size 5.00E+002 g | Real Time 1908 Sec
Collection Efficiency 1.0000 | Spc. File 062429D08.SPC

Detector #: 8 (Detector 8)
Energy(keV) = -0.37 + 0.500*Ch + -4.29E-08*Ch^2 + 6.41E-11*Ch^3 08/17/2006
FWHM(keV) = 0.68 + 0.014*En + 5.09E-04*En^2 + 0.00E+00*En^3 12/14/2005
Where En = Sqrt(Energy in keV)

Search Sensitivity: 1.00 | Sigma Multiplier: 2.00 | Search Start/End: 80/4000

=====

PEAK SEARCH RESULTS

=====

PK. #	ENERGY (keV)	ADDRESS CHANNEL	NET/MDA COUNTS	UN- CERTAINTY	C.L. COUNTS	BKG COUNTS	FWHM (keV)	FLAG
1	49.65	100.03	2072	410	329	15977	1.52	a Wide Pk
2	59.55	119.83	58326	557	228	10450	0.83	a
3	66.82	134.36	3680	574	462	22523	2.49	a HiResid Wide Pk
4	72.39	145.50	2223	506	409	19520	2.14	b HiResid
5	74.85	150.42	294	203	165	6006	0.60	c HiResid
6	85.43	171.58	1866	630	514	26576	2.56	a Wide Pk
7	88.08	176.87	67069	566	187	7029	0.87	b
8	122.07	244.84	28275	377	141	3985	0.89	a
9	136.50	273.70	3465	211	144	3820	0.92	a
10	165.88	332.44	19459	322	133	3271	0.91	a
11	199.17	399.02	337	174	140	3351	1.21	a
12	255.20	511.05	560	126	96	1873	0.87	a
13	279.25	559.14	3487	184	116	2276	1.07	a
14	310.01	620.64	57	105	86	1476	0.75	a NET< CL
15	391.74	784.07	11961	249	99	1661	1.18	a
16	511.45	1023.40	302	125	99	1690	1.55	a
17	661.78	1323.90	19881	303	92	1467	1.45	a
18	714.54	1429.33	56	93	76	1057	1.39	a NET< CL
19	814.00	1628.10	204	95	74	1020	1.27	a
20	898.10	1796.12	12812	252	92	1557	1.62	a
21	1173.41	2345.92	22012	307	64	697	1.92	a
22	1325.55	2649.54	412	84	61	484	3.54	a Wide Pk
23	1332.70	2663.80	20223	289	40	286	2.07	b

062429D08.SPC Analyzed by

=====

PEAK SEARCH RESULTS

=====

PK. #	ENERGY (keV)	ADDRESS CHANNEL	NET/MDA COUNTS	UN- CERTAINTY	C.L. COUNTS	BKG COUNTS	FWHM (keV)	FLAG
24	1836.98	3668.79	7861	179	21	75	2.50	a

062429D08.SPC Analyzed by

SEEKER CALIBRATION RESULTS Version 2.0.4

Sample ID: 0613003-8 FWHM CAL (817)
Stds. Match Tolerance: 2.00 keV

Detector Number: 08 Calibration Date. . . 08/17/2006 10:06:09

FWHM(keV) = $0.72 + 0.007*En + 8.11e-04*En^2 + 0.00e+00*En^3$
(Where En = SQR(Energy in keV))

Pk. #	Energy (keV)	Measured FWHM(keV)	% Diff.	Calculated FWHM(keV)	% Diff.	Prev.Calc. FWHM(keV)
1	59.50	0.834	-1.06	0.825	-0.66	0.820
2	88.04	0.872	-1.42	0.860	-0.28	0.857
3	122.06	0.894	0.55	0.899	-0.13	0.898
4	165.85	0.914	3.44	0.947	-0.18	0.945
5	279.00	1.073	-0.77	1.065	-0.88	1.055
6	391.68	1.183	-0.51	1.177	-1.89	1.155
7	661.64	1.453	-1.16	1.436	-4.52	1.374
8	898.02	1.624	2.00	1.657	-6.69	1.553
9	1173.21	1.922	-0.64	1.909	-8.97	1.752
10	1332.48	2.069	-0.75	2.054	-10.16	1.864
11	1836.01	2.496	0.36	2.505	-13.46	2.208

Calibration Results Saved.

ANALYTICS

RSO # 798
Rec'd 8/5/05
DJS

1380 Seaboard Industrial Blvd.
Atlanta, Georgia 30318 - U.S.A.

Phone (404) 352-8677
Fax (404) 352-2837

CERTIFICATE OF CALIBRATION

Standard Radionuclide Source

71035A-307

1.0 Liter Solid in 138G GA-MA Beaker

This standard radionuclide source was prepared using aliquots measured gravimetrically from master radionuclide solution sources. The Am-241 was calibrated by 4 pi alpha liquid scintillation counting. All other radionuclides were calibrated using a germanium gamma spectrometer system. Calibration and purity were checked using a germanium gamma spectrometer system. At the time of calibration no interfering gamma-ray emitting impurities were detected. The gamma-ray emission rates for the most intense gamma-ray lines are given. Analytics maintains traceability to the National Institute of Standards and Technology through a Measurements Assurance Program as described in USNRC Regulatory Guide 4.15, Rev. 1, February, 1979.

US Patent 4,430,258; UK Patent GB2,149,194B; CA Patent 1,196,776.
Density of solid matrix 1.15 g/cc.

Calibration date: July 1, 2005 12:00 EST

ISOTOPE	GAMMA-RAY ENERGY	HALF-LIFE		GAMMA-RAYS PER SECOND	TOTAL UNCERTAINTY %
Am-241	59.5	432	y	1353	3.0
Cd-109	88	462.6	d	1912	3.3
Co-57	122	271.79	d	1048	3.0
Ce-139	166	137.6	d	1426	2.8
Hg-203	279	46.61	d	3260	2.7
Sn-113	392	115.1	d	2038	2.6
Cs-137	662	30.07	y	1268	3.0
Y-88	898	106.6	d	5008	2.6
Co-60	1173	5.2714	y	2475	2.7
Co-60	1332	5.2714	y	2499	2.6
Y-88	1836	106.6	d	5224	2.6

P O NUMBER 71239, Rel. 5/20/05, Item 1

SOURCE PREPARED BY: M. D. Currie
M. D. Currie, Radiochemist

Q A APPROVED: M. D. Currie 8-1-05

This standard will expire one year after the calibration date.

000081

CERTIFICATE OF CALIBRATION
Standard Radionuclide Source

73487-307

RSO# 824 Rec'd 8/29/06 GUS

1.0 Solid in 138G GA-MA Beaker

This standard radionuclide source was prepared using aliquots measured gravimetrically from master radionuclide solution sources. The Am-241 was calibrated by 4 pi alpha liquid scintillation counting. All other radionuclides were calibrated using a germanium gamma spectrometer system. Calibration and purity were checked using a germanium gamma spectrometer system. At the time of calibration no interfering gamma-ray emitting impurities were detected. The gamma-ray emission rates for the most intense gamma-ray lines are given. Analytisc maintains traceability to the National Institute of Standards and Technology through a Measurements Assurance Program as described in USNRC Regulatory Guide 4.15, Rev. 1, February, 1979.

US Patent 4,430,258; UK Patent GB2,149,194B; CA Patent 1,196,776.
Density of solid matrix 1.15 g/cc.

Calibration date: July 1, 2006 12:00 EST

ISOTOPE	GAMMA-RAY ENERGY	HALF-LIFE		GAMMA-RAYS PER SECOND	TOTAL UNCERTAINTY %
Am-241	59.5	432	Y	1323	3.0
Cd-109	88	462.6	d	1872	3.3
Co-57	122	271.79	d	984.9	3.0
Ce-139	166	137.6	d	1391	2.8
Hg-203	279	46.61	d	3088	2.7
Sn-113	392	115.1	d	1971	2.6
Cs-137	662	30.07	Y	1256	3.0
Y-88	898	106.6	d	4857	2.6
Co-60	1173	5.2714	Y	2377	2.7
Co-60	1332	5.2714	Y	2374	2.6
Y-88	1836	106.6	d	5084	2.6

P O NUMBER 71239, Rel. 7/31/06, Item 1

SOURCE PREPARED BY: *M. Taskaeva*
M. Taskaeva, Radiochemist

Q A APPROVED: *UM [Signature]* 8-24-06

This standard will expire one year after the calibration date.

 SEEKER G A M M A A N A L Y S I S R E S U L T S P S Version 1.8.4

Paragon Analytics, Div. of DataChem Lab
 GammaScan

Geo 1 / Water

Sample ID: 0613004-2 GEO 1 EFF CAL (798)

```

-----
Sampling Start: 07/01/2005 12:00:00 | Counting Start: 05/18/2006 09:12:40
Sampling Stop: 07/01/2005 12:00:00 | Decay Time. . . . . 7.70E+003 Hrs
Buildup Time. . . . . 0.00E+000 Hrs | Live Time . . . . . 3600 Sec
Sample Size . . . . . 1.00E+000 L | Real Time . . . . . 3704 Sec
Collection Efficiency . . . . . 1.0000 | Spc. File . . . . . 061207D02.SPC
-----
  
```

Detector #: 2 (Detector 2)

Energy(keV) = -0.85 + 0.501*Ch + 0.00E+00*Ch^2 + 0.00E+00*Ch^3 05/18/2006

FWHM(keV) = 0.69 + 0.006*En + 1.18E-03*En^2 + 0.00E+00*En^3 05/16/2006

Where En = Sqrt(Energy in keV)

 Search Sensitivity: 1.00 | Sigma Multiplier: 2.00 | Search Start/End: 80/4000

=====

PEAK SEARCH RESULTS

=====

PK. #	ENERGY (keV)	ADDRESS CHANNEL	NET/MDA COUNTS	UN- CERTAINTY	C.L. COUNTS	BKG COUNTS	FWHM (keV)	FLAG
1	59.47	120.46	27378	422	215	9358	0.80	a HiResid
2	69.30	140.11	727	471	385	20558	1.59	a Wide Pk
3	87.94	177.34	81294	632	224	10141	0.85	a
4	122.05	245.44	45606	480	179	6492	0.89	a
5	136.50	274.32	5577	250	165	5456	0.80	a HiResid
6	165.83	332.89	28936	404	180	5965	0.94	a
7	255.13	511.23	862	223	177	5336	1.07	a
8	279.18	559.26	2006	201	148	4041	1.01	a
9	310.08	620.98	191	209	171	4619	1.34	a
10	337.03	674.80	112	133	108	2661	0.76	a
11	391.75	784.10	17389	317	144	3832	1.23	a
12	511.32	1022.89	485	226	182	4669	2.05	a
13	513.40	1027.05	52	109	89	1796	0.87	b NET< CL
14	661.77	1323.35	47362	462	127	2989	1.57	a
15	813.72	1626.83	267	140	112	2433	1.43	a
16	898.19	1795.52	18141	315	134	3467	1.86	a
17	1115.40	2229.34	132	119	96	1955	1.55	a
18	1173.43	2345.22	52613	474	97	1671	2.20	a
19	1332.63	2663.19	47485	443	66	726	2.38	a HiResid
20	1836.11	3668.71	10531	210	37	198	2.97	a HiResid

061207D02.SPC Analyzed by

SEEKER BACKGROUND SUBTRACT RESULTS Version 1.8.2

Paragon Analytics, Div. of DataChem Lab
GammaScan

Background File: DET020517.BKG (060517-2 WEEKLY BKGD)

Bkg.File Detector #: 2

=====

BACKGROUND SUBTRACT RESULTS

=====

PK#	ENERGY (keV)	OLD NET COUNTS	OLD UN- CERTAINTY	OLD CR.LEVEL	NEW NET COUNTS	NEW UN- CERTAINTY	NEW CR.LEVEL	FLAG
1	59.47	27378	422	215	27376	422	216	
2	69.30	727	471	385	725	471	385	
3	87.94	81294	632	224	81288	632	224	
4	122.05	45606	480	179	45605	480	180	
10	337.03	112	133	108	109	133	108	
12	511.32	485	226	182	388	226	183	

061207D02.SPC Analyzed by

SEEKER CALIBRATION RESULTS Version 2.0.4

Sample ID: 0613004-2 GEO 1 EFF CAL (798)
Stds. Match Tolerance: 2.00 keV

Detector Number: 02 Calibration Date. . . 05/18/2006 09:12:40
Geometry File (D02)(Sh01).EFF ID. Geo 1 Eff Cal
Amount of Std. in Calib. Source: 1.000000 gm

Eff = 1 / [3.02e-03*En^-3.85e+00 + 1.33e+02*En^ 7.85e-01]
(Where En = Energy in MeV) (Exponential)

Pk. #	Energy (keV)	Measured Efficiency	% Difference	Calculated Efficiency	% Difference	Prev.Calc. Efficiency
1	59.50	5.63e-03	3.56	5.84e-03	0.32	5.86e-03
2	88.04	1.91e-02	-3.95	1.84e-02	-0.41	1.83e-02
3	122.06	2.74e-02	3.19	2.83e-02	0.09	2.83e-02
4	165.85	2.84e-02	-0.48	2.82e-02	0.63	2.84e-02
5	279.00	2.02e-02	0.64	2.04e-02	0.79	2.05e-02
6	391.68	1.64e-02	-4.30	1.57e-02	0.68	1.58e-02
7	661.64	1.06e-02	-1.63	1.04e-02	0.46	1.05e-02
8	898.02	8.10e-03	1.14	8.20e-03	0.33	8.22e-03
9	1173.21	6.63e-03	0.25	6.64e-03	0.21	6.66e-03
10	1332.48	5.92e-03	1.46	6.01e-03	0.15	6.02e-03
11	1836.01	4.51e-03	3.48	4.67e-03	0.01	4.68e-03

Calibration Results Saved.

Geometry 01 Calibration Verification: Gamma Mixed Nuclide Source											
Std. #	718	Detector	2	REF DATE :	7/1/2003	count date	5/18/2006				
FROM CALIBRATION CERTIFICATE				FROM ANALYTICS.LIB				EXPECTED ACTIVITY			
Isotope	KeV	Half Life(y)	Gamma Fraction:	Gamma Fraction:	Mass of Standard	DPS	pCi/L	Activity	Recovery	Pass/Fail	# of half-lives expired
Am-241	59.9	432.0000	1304	0.3590	L 1	3632.3	98170.6	98900	101%	Pass	0.01
Cd-109	88	1.2666	1862	0.0361		51578.9	1394025.6	1440000	103%	Pass	2.27
Co-57	122	0.7441	1032	0.8551		1206.9	32618.3	31900	98%	Pass	3.87
Ce-139	166	0.3768	1419	0.8035		1766.0	47730.4	>5 h-lives	>5 h-lives	>5 h-lives	7.64
Hg-203	279	0.1276	3194	0.7730		4132.0	111674.4	>5 h-lives	>5 h-lives	>5 h-lives	22.58
Sn-113	392	0.3151	1960	0.6490		3020.0	81622.5	>5 h-lives	>5 h-lives	>5 h-lives	9.14
Cs-137	662	30.0000	1260	0.8512		1480.3	40007.1	41100	103%	Pass	0.10
Y-88	898	0.2919	5060	0.9340		5417.6	146420.5	>5 h-lives	>5 h-lives	>5 h-lives	9.87
Co-60	1173	5.2714	2402	1.0000		2402.0	64918.9	64300	99%	Pass	0.55
Co-60	1332	5.2714	2427	1.0000		2427.0	65594.6	64300	98%	Pass	0.55
Y-88	1836	0.2919	5287	0.9938		5320.0	143783.3	>5 h-lives	>5 h-lives	>5 h-lives	9.87

000086

061208D02.SPC Analyzed by *W*

SEEKER G A M M A A N A L Y S I S R E S U L T S PS Version 1.8.4

Paragon Analytics, Div. of DataChem Lab
GammaScan

Geo 1 / Water

Sample ID: 0613004-2A GEO 1 CAL VER (718)

```

-----
Sampling Start:   07/01/2003 12:00:00 | Counting Start:   05/18/2006 10:32:24
Sampling Stop:   07/01/2003 12:00:00 | Decay Time. . . . . 2.52E+004 Hrs
Buildup Time. . . . . 0.00E+000 Hrs | Live Time . . . . . 1800 Sec
Sample Size . . . . . 1.00E+000 L | Real Time . . . . . 1829 Sec
Collection Efficiency . . . . . 1.0000 | Spc. File . . . . . 061208D02.SPC
-----

```

Detector #: 2 (Detector 2)

```

Energy(keV) = -0.85 + 0.501*Ch + 0.00E+00*Ch^2 + 0.00E+00*Ch^3 05/18/2006
FWHM(keV)   = 0.69 + 0.006*En + 1.18E-03*En^2 + 0.00E+00*En^3 05/16/2006
Where En = Sqrt(Energy in keV)
-----

```

Search Sensitivity: 1.00 | Sigma Multiplier: 2.00 | Search Start/End: 80/4000

PEAK SEARCH RESULTS

PK. #	ENERGY (keV)	ADDRESS CHANNEL	NET/MDA COUNTS	UN-CERTAINTY	C.L. COUNTS	BKG COUNTS	FWHM (keV)	FLAG
1	42.76	87.09	92	115	93	1925	0.71	a NET< CL
2	59.48	120.49	13766	274	116	2693	0.83	a
3	66.05	133.61	95	87	69	1189	0.41	a
4	87.93	177.31	13237	261	101	2052	0.83	a
5	122.06	245.46	3486	157	85	1471	0.83	a
6	136.40	274.10	473	125	96	1720	0.95	a
7	165.84	332.91	358	119	92	1577	1.00	a
8	175.42	352.04	76	98	79	1262	0.83	a NET< CL
9	194.02	389.18	93	102	83	1382	0.80	a
10	310.09	621.00	7	75	61	834	0.58	a NET< CL
11	318.89	638.57	69	79	64	918	0.74	a
12	343.65	688.02	40	87	71	1070	1.04	a NET< CL
13	353.24	707.17	73	87	70	1046	0.99	a
14	367.64	735.93	45	64	52	657	0.62	a NET< CL
15	391.82	784.22	151	107	85	1350	1.34	a
16	661.73	1323.28	22741	313	70	894	1.58	a
17	762.31	1524.16	53	99	81	1050	2.21	a NET< CL
18	1173.35	2345.07	19174	284	53	498	2.06	a HiResid
19	1332.58	2663.08	17943	270	28	133	2.39	a
20	1836.22	3668.92	31	17	11	21	1.91	a

061208D02.SPC Analyzed by

SEEKER B A C K G R O U N D S U B T R A C T R E S U L T S Vers. 2.2.1

Paragon Analytics, Div. of DataChem Lab
GammaScan

Background File: DET020517.BKG (060517-2 WEEKLY BKGD)

Bkg.File Detector #: 2

=====

BACKGROUND SUBTRACT RESULTS

=====

PK#	ENERGY (keV)	OLD NET COUNTS	OLD UN- CERTAINTY	OLD CR.LEVEL	NEW NET COUNTS	NEW UN- CERTAINTY	NEW CR.LEVEL	FLAG
2	59.48	13766	274	116	13765	274	116	
3	66.05	95	87	69	89	87	70	
4	87.93	13237	261	101	13234	261	101	
5	122.06	3486	157	85	3486	157	85	
13	353.24	73	87	70	68	87	71	NET<CL

Paragon Analytics, Div. of DataChem Lab
 GammaScan

Geo 1 / Water

Sample ID: 0613004-2A GEO 1 CAL VER (718)

```
-----
Sampling Start: 07/01/2003 12:00:00 | Counting Start: 05/18/2006 10:32:24
Sampling Stop: 07/01/2003 12:00:00 | Decay Time. . . . . 2.52e+004 Hrs
Buildup Time. . . . . 0.00e+000 Hrs | Live Time . . . . . 1800 Sec
Sample Size . . . . . 1.00e+000 L | Real Time . . . . . 1829 Sec
Collection Efficiency . . . . . 1.0000 | Spectrum File . . . . . 061208D02.SPC
Cr. Level Confidence Interval: 95 % | Det. Limit Confidence Interval: 95 %
-----
```

Detector #: 2 (Detector 2)

Efficiency File: (D02)(Sh01).EFF (Geo 1 Eff Cal)
 Eff.=1/[3.02E-03*En^-3.85E+00 + 1.33E+02*En^7.85E-01] 05/18/2006

Library File:ANALYTICAL.LIB (Analytical)

MEASURED or MDA CONCENTRATIONS

```
=====
```

Nuclide	ENERGY E (keV)	N T	Concentration		MDA	Critical Level	Halflife (hrs)
			(pCi/L)			
Am-241	59.54	9.89E+04	+ -	1.96E+03	1.68E+03	8.31E+02	3.79E+06
Cd-109	88.02	1.44E+06	+ -	2.84E+04	2.23E+04	1.10E+04	1.11E+04
Co-57	122.07	3.19E+04	+ -	1.44E+03	1.59E+03	7.81E+02	6.50E+03
Ce-139	165.85	4.73E+04	+ -	1.56E+04	2.47E+04	1.22E+04	3.30E+03
Cs-137	661.62	4.11E+04	+ -	5.66E+02	2.56E+02	1.26E+02	2.64E+05
Co-60	Average:x	6.43E+04	+ -	6.80E+02	4.62E+04
	1173.21	6.33E+04	+ -	9.39E+02	3.59E+02	1.75E+02	4.62E+04
	1332.48	6.55E+04	+ -	9.85E+02	2.17E+02	1.04E+02	4.62E+04
Hg-203	279.18	MDA	1.09E+09	5.37E+08	1.12E+03
Sn-113	391.68	MDA	8.95E+04r	4.36E+04	2.76E+03
Y-88	898.02	MDA	3.20E+05	1.58E+05	2.56E+03

MEASURED TOTAL: 1.73E+06 + - 4.87E+04 pCi/L

UNKNOWN, SUM or ESCAPE PEAKS

```
=====
```

PK. #	ENERGY (keV)	ADDRESS CHANNEL	NET COUNTS	UN- CERTAINTY	C.L. COUNTS	BKG COUNTS	FWHM (keV)	FLAG
1	42.76	87.09	92	115	93	1925	0.71	Deleted
3	66.05	133.61	89	87	70	1189	0.41	Unknown
6	136.40	274.10	473	125	96	1720	0.95	Unknown
8	175.42	352.04	76	98	79	1262	0.83	Deleted
9	194.02	389.18	93	102	83	1382	0.80	Unknown

061208D02.SPC Analyzed by

=====

UNKNOWN, SUM or ESCAPE PEAKS

=====

PK. #	ENERGY (keV)	ADDRESS CHANNEL	NET COUNTS	UN- CERTAINTY	C.L. COUNTS	BKG COUNTS	FWHM (keV)	FLAG
10	310.09	621.00	7	75	61	834	0.58	Deleted
11	318.89	638.57	69	79	64	918	0.74	Unknown
12	343.65	688.02	40	87	71	1070	1.04	Deleted
13	353.24	707.17	68	87	71	1046	0.99	Deleted
14	367.64	735.93	45	64	52	657	0.62	Deleted
15	391.82	784.22	151	107	85	1350	1.34	Unknown
17	762.31	1524.16	53	99	81	1050	2.21	Deleted
20	1836.22	3668.92	31	17	11	21	1.91	Unknown

000090

062299D06.SPC Analyzed by *cu*

SEEKER G A M M A A N A L Y S I S R E S U L T S PS Version 1.8.4

Paragon Analytics, Div. of DataChem Lab
GammaScan

Geo 1 / Water

Sample ID: 0613004-6 GEO 1 EFF CAL (824)

```

-----
Sampling Start: 07/01/2006 12:00:00 | Counting Start: 09/11/2006 10:06:11
Sampling Stop: 07/01/2006 12:00:00 | Decay Time . . . . . 1.73E+003 Hrs
Buildup Time . . . . . 0.00E+000 Hrs | Live Time . . . . . 3600 Sec
Sample Size . . . . . 1.00E+000 L | Real Time . . . . . 3816 Sec
Collection Efficiency . . . . . 1.0000 | Spc. File . . . . . 062299D06.SPC
-----

```

Detector #: 6 (Detector 6)

Energy(keV) = -0.69 + 0.501*Ch + -1.21E-08*Ch^2 + 0.00E+00*Ch^3 09/11/2006

FWHM(keV) = 1.19 + -0.002*En + 7.29E-04*En^2 + 0.00E+00*En^3 07/24/2006

Where En = Sqrt(Energy in keV)

Search Sensitivity: 1.00 | Sigma Multiplier: 2.00 | Search Start/End: 80/4000

=====

PEAK SEARCH RESULTS

=====

PK. #	ENERGY (keV)	ADDRESS CHANNEL	NET/MDA COUNTS	UN- CERTAINTY	C.L. COUNTS	BKG COUNTS	FWHM (keV)	FLAG
1	59.52	120.20	17321	432	282	16764	0.97	a
2	68.57	138.27	869	573	469	34236	1.81	a
3	70.84	142.82	2134	456	367	24899	1.33	b
4	72.87	146.86	2689	335	262	15562	0.90	c
5	82.43	165.95	3002	571	461	33017	1.82	a Wide Pk
6	85.57	172.22	9260	1014	819	66033	3.56	b
7	88.01	177.10	89635	696	292	18009	1.00	c
8	122.08	245.12	67543	624	284	17005	1.04	a
9	136.50	273.92	9248	419	306	18487	1.07	a
10	165.87	332.56	85526	661	253	13482	1.04	a HiResid Wide Pk
11	169.73	340.25	7	895	736	51681	3.84	b NET< CL HiResid
12	255.20	510.90	3206	298	227	10112	1.21	a
13	279.25	558.93	70487	588	208	8536	1.18	a
14	291.15	582.70	135	161	131	4242	0.68	a
15	391.79	783.63	64255	552	181	6023	1.27	a HiResid
16	486.62	972.99	90	162	132	3675	0.94	a NET< CL
17	511.23	1022.12	1613	274	216	7476	2.15	a
18	661.70	1322.56	40791	452	167	5398	1.50	a
19	814.06	1626.80	1148	160	120	2938	1.38	a
20	898.14	1794.69	74717	574	144	3827	1.67	a HiResid
21	1173.29	2344.11	46124	448	106	2151	1.85	a
22	1325.16	2647.37	1580	178	131	2472	3.03	a

=====

PEAK SEARCH RESULTS

=====

PK. #	ENERGY (keV)	ADDRESS CHANNEL	NET/MDA COUNTS	UN- CERTAINTY	C.L. COUNTS	BKG COUNTS	FWHM (keV)	FLAG
23	1332.56	2662.15	41933	425	92	1561	1.98	b
24	1835.95	3667.36	43750	422	46	363	2.29	a HiResid

062299D06.SPC Analyzed by

SEEKER BACKGROUND SUBTRACT RESULTS Version 1.8.2

Paragon Analytics, Div. of DataChem Lab
GammaScan

Background File: DET060908.BKG (060908-6 WEEKLY BKGD)

Bkg.File Detector #: 6

=====

BACKGROUND SUBTRACT RESULTS

=====

PK#	ENERGY (keV)	OLD NET COUNTS	OLD UN- CERTAINTY	OLD CR.LEVEL	NEW NET COUNTS	NEW UN- CERTAINTY	NEW CR.LEVEL	FLAG
17	511.23	1613	274	216	1535	275	216	

062299D06.SPC Analyzed by

SEEKER CALIBRATION RESULTS Version 2.0.4

Sample ID: 0613004-6 GEO 1 EFF CAL (824)
Stds. Match Tolerance: 2.00 keV

Detector Number: 06 Calibration Date. . . 09/11/2006 10:06:11
Geometry File (D06)(Sh01).EFF ID. Geo 1 Eff Cal
Amount of Std. in Calib. Source: 1.000000 gm

Crossover: 180.00 keV

Below Crossover Efficiency Fit:

$$\text{Eff} = 10^{[-2.87e+01 + 2.52e+01*En + -5.85e+00*En^2 + 0.00e+00*En^3]}$$

(Where En = LOG(Energy in keV)) (Polynomial)

Above Knee Efficiency Fit:

$$\text{Eff} = 10^{[-1.40e+00 + 1.00e+00*En + -6.55e-01*En^2 + 7.72e-02*En^3]}$$

(Where En = LOG(Energy in keV)) (Polynomial)

Pk. #	Energy (keV)	Measured Efficiency	% Difference	Calculated Efficiency	% Difference	Prev.Calc. Efficiency
1	59.50	3.64e-03	1.22	3.68e-03	-0.85	3.65e-03
2	88.04	1.48e-02	-4.48	1.42e-02	1.50	1.44e-02
3	122.06	2.29e-02	4.81	2.40e-02	2.25	2.46e-02
4	165.85	2.45e-02	-1.80	2.41e-02	1.96	2.46e-02
5	279.00	1.85e-02	0.02	1.85e-02	1.23	1.87e-02
6	391.68	1.40e-02	0.07	1.40e-02	1.78	1.42e-02
7	661.64	9.06e-03	-0.92	8.98e-03	1.35	9.10e-03
8	898.02	6.82e-03	1.55	6.93e-03	1.11	7.01e-03
9	1173.21	5.53e-03	0.06	5.53e-03	1.33	5.61e-03
10	1332.48	5.04e-03	-1.14	4.98e-03	1.66	5.06e-03
11	1836.01	3.82e-03	0.34	3.83e-03	3.40	3.97e-03

Calibration Results Saved.

OK
9/11/06

Geometry 01 Calibration Verification: Gamma Mixed Nuclide Source												
Std. #	798	Defector	6									
NEW SOURCE:	798			REF DATE:	7/1/2005							
FROM CALIBRATION CERTIFICATE			FROM ANALYTICAL LIB			EXPECTED ACTIVITY			count date			9/11/2006
Isotope	KeV	Half Life(y)	Gammas/Sec.	Gamma Fraction:	Mass of Standard		DPS	pCi/g	Activity	Recovery	Pass/Fail	# of half-lives expired
Am-241	59.9	432.0000	1353	0.3590	1	L	3768.8	101859.5	97000	95%	Pass	0.00
Cd-109	88	1.2666	1912	0.0361			52964.0	1431459.2	1490000	104%	Pass	0.94
Co-57	122	0.7441	1048	0.8551			1225.6	33124.0	32000	97%	Pass	1.61
Ce-139	166	0.3768	1426	0.8035			1774.7	47965.8	48500	101%	Pass	3.17
Hg-203	279	0.1276	3260	0.7730			4217.3	113982.0	NA	>5 h-lives	>5 h-lives	9.38
Sn-113	392	0.3151	2038	0.6490			3140.2	84870.7	88200	104%	Pass	3.80
Cs-137	662	30.0000	1268	0.8512			1489.7	40261.1	41000	102%	Pass	0.04
Y-88	898	0.2919	5008	0.9340			5361.9	144915.8	146000	101%	Pass	4.10
Co-60	1173	5.2714	2475	1.0000			2475.0	66891.9	68600	103%	Pass	0.23
Co-60	1332	5.2714	2499	1.0000			2499.0	67540.5	68100	101%	Pass	0.23
Y-88	1836	0.2919	5224	0.9938			5256.6	142070.0	144000	101%	Pass	4.10

000095

062302D06.SPC Analyzed by CW

SEEKER G A M M A A N A L Y S I S R E S U L T S PS Version 1.8.4

Paragon Analytics, Div. of DataChem Lab
GammaScan

Geo 1 / Water

Sample ID: 0613004-6A GEO 1 CAL VER (798)

```

-----
Sampling Start: 07/01/2005 12:00:00 | Counting Start: 09/11/2006 11:24:12
Sampling Stop: 07/01/2005 12:00:00 | Decay Time. . . . . 1.05E+004 Hrs
Buildup Time. . . . . 0.00E+000 Hrs | Live Time . . . . . 5400 Sec
Sample Size . . . . . 1.00E+000 L | Real Time . . . . . 5543 Sec
Collection Efficiency . . . . . 1.0000 | Spc. File . . . . . 062302D06.SPC
-----

```

Detector #: 6 (Detector 6)

Energy(keV) = -0.69 + 0.501*Ch +-1.21E-08*Ch^2 + 0.00E+00*Ch^3 09/11/2006

FWHM(keV) = 1.19 + -0.002*En + 7.29E-04*En^2 + 0.00E+00*En^3 07/24/2006

Where En = Sqrt(Energy in keV)

Search Sensitivity: 1.00 | Sigma Multiplier: 2.00 | Search Start/End: 80/4000

=====

PEAK SEARCH RESULTS

=====

PK. #	ENERGY (keV)	ADDRESS CHANNEL	NET/MDA COUNTS	UN- CERTAINTY	C.L. COUNTS	BKG COUNTS	FWHM (keV)	FLAG
1	59.48	120.14	25644	420	224	10565	0.93	a HiResid
2	86.98	175.03	3233	643	521	33393	2.65	a Wide Pk
3	88.01	177.10	79514	632	236	11715	1.03	b
4	122.07	245.11	43091	497	224	9901	1.08	a
5	136.45	273.81	5438	271	187	7422	1.00	a
6	165.85	332.52	20801	379	202	8020	1.08	a
7	255.21	510.93	358	180	145	4784	0.82	a
8	279.26	558.95	327	168	135	4143	0.77	a
9	391.77	783.60	11517	287	157	4568	1.31	a
10	432.05	864.01	65	158	129	3513	1.00	a NET< CL
11	510.34	1020.35	452	205	165	4781	1.82	a
12	512.17	1024.00	224	133	106	2608	1.05	b
13	611.66	1222.65	131	130	105	2407	1.22	a
14	661.68	1322.53	60926	518	128	3184	1.50	a
15	813.20	1625.08	362	282	230	6496	3.23	a Wide Pk
16	815.26	1629.18	-99	97	81	1624	0.86	b NET< CL
17	836.24	1671.08	248	176	143	3578	1.88	a
18	898.13	1794.65	11035	274	145	3901	1.63	a
19	1173.29	2344.11	64793	522	94	1687	1.87	a HiResid
20	1324.43	2645.91	169	88	69	780	2.54	a
21	1332.56	2662.14	57861	487	64	749	1.96	a
22	1460.94	2918.51	46	52	41	325	1.76	a
23	1835.94	3667.35	6407	163	25	111	2.29	a

062302D06.SPC Analyzed by

SEEKER B A C K G R O U N D S U B T R A C T R E S U L T S Vers. 2.2.1

Paragon Analytics, Div. of DataChem Lab
GammaScan

Background File: DET060908.BKG (060908-6 WEEKLY BKGD)

Bkg.File Detector #: 6

=====

BACKGROUND SUBTRACT RESULTS

=====

PK#	ENERGY (keV)	OLD NET COUNTS	OLD UN- CERTAINTY	OLD CR.LEVEL	NEW NET COUNTS	NEW UN- CERTAINTY	NEW CR.LEVEL	FLAG
11	510.34	452	205	165	336	206	167	
22	1460.94	46	52	41	21	52	42	NET<CL

Paragon Analytics, Div. of DataChem Lab
 GammaScan

Geo 1 / Water

Sample ID: 0613004-6A GEO 1 CAL VER (798)

```
-----
Sampling Start: 07/01/2005 12:00:00 | Counting Start: 09/11/2006 11:24:12
Sampling Stop: 07/01/2005 12:00:00 | Decay Time . . . . . 1.05e+004 Hrs
Buildup Time . . . . . 0.00e+000 Hrs | Live Time . . . . . 5400 Sec
Sample Size . . . . . 1.00e+000 L | Real Time . . . . . 5543 Sec
Collection Efficiency . . . . . 1.0000 | Spectrum File . . . . . 062302D06.SPC
Cr. Level Confidence Interval: 95 % | Det. Limit Confidence Interval: 95 %
-----
```

Detector #: 6 (Detector 6)

Efficiency File: (D06) (Sh01).EFF (Geo 1 Eff Cal)

Eff=10[^][-2.87E+01 +2.52E+01*L +-5.85E+00*L[^]2 +0.00E+00*L[^]3] 09/11/2006

Eff.=10[^][-1.40E+00 +1.00E+00*L +-6.55E-01*L[^]2 +7.72E-02*L[^]3] Above 180.00 keV

Library File:ANALYTICAL.LIB (Analytical)

MEASURED or MDA CONCENTRATIONS

Nuclide	ENERGY E (keV)	Concentration (pCi/L)	MDA	Critical Level	Halflife (hrs)
Am-241	59.54	9.70E+04 +- 1.59E+03	1.70E+03	8.46E+02	3.79E+06
Cd-109	88.02	1.49E+06 +- 1.19E+04	8.90E+03	4.43E+03	1.11E+04
Co-57	122.07	3.20E+04 +- 3.69E+02	3.35E+02	1.67E+02	6.50E+03
Ce-139	165.85	4.85E+04 +- 8.83E+02	9.47E+02	4.70E+02	3.30E+03
Sn-113	391.68	8.82E+04 +- 2.20E+03	2.43E+03	1.20E+03	2.76E+03
Cs-137	661.62	4.10E+04 +- 3.48E+02	1.74E+02	8.60E+01	2.64E+05
Y-88	Average:x	1.45E+05 +- 2.58E+03	2.56E+03
	898.02	1.46E+05 +- 3.64E+03	3.89E+03	1.92E+03	2.56E+03
	1836.01	1.44E+05 +- 3.67E+03	1.21E+03	5.74E+02	2.56E+03
Co-60	Average:x	6.84E+04 +- 3.98E+02	4.62E+04
	1173.21	6.86E+04 +- 5.52E+02	2.01E+02	9.90E+01	4.62E+04
	1332.48	6.81E+04 +- 5.73E+02	1.53E+02	7.49E+01	4.62E+04
Hg-203	279.18	MDA	5.19E+04r	2.57E+04	1.12E+03

MEASURED TOTAL: 2.01E+06 +- 2.03E+04 pCi/L

UNKNOWN, SUM or ESCAPE PEAKS

PK. #	ENERGY (keV)	ADDRESS CHANNEL	NET COUNTS	UN-CERTAINTY	C.L. COUNTS	BKG COUNTS	FWHM (keV)	FLAG
2	86.98	175.03	3233	643	521	33393	2.65	Unknown
5	136.45	273.81	5439	271	187	7422	1.00	Unknown

062302D06.SPC Analyzed by

=====

UNKNOWN, SUM or ESCAPE PEAKS

=====

PK. #	ENERGY (keV)	ADDRESS CHANNEL	NET COUNTS	UN- CERTAINTY	C.L. COUNTS	BKG COUNTS	FWHM (keV)	FLAG
7	255.21	510.93	358	180	145	4784	0.82	Unknown
8	279.26	558.95	328	168	135	4143	0.77	Unknown
10	432.05	864.01	65	158	129	3513	1.00	Deleted
11	510.34	1020.35	336	206	167	4781	1.82	Unknown
12	512.17	1024.00	224	133	106	2608	1.05	Unknown
13	611.66	1222.65	131	130	105	2407	1.22	Unknown
15	813.20	1625.08	362	282	230	6496	3.23	1836DEsc
16	815.26	1629.18	-99	97	81	1624	0.86	Deleted
17	836.24	1671.08	248	176	143	3578	1.88	Unknown
20	1324.43	2645.91	169	88	69	780	2.54	1836SEsc
22	1460.94	2918.51	21	52	42	325	1.76	Deleted

c:\SEEKER\BIN\062302d06.res Analysis Results Saved.

062650D08.SPC Analyzed by *W*

SEEKER G A M M A A N A L Y S I S R E S U L T S P S Version 1.8.4

Paragon Analytics, Div. of DataChem Lab
GammaScan

Geo 1 / Water

Sample ID: 0613004-8 GEO 1 EFF CAL (824)

```

-----
Sampling Start: 07/01/2006 12:00:00 | Counting Start: 09/11/2006 14:14:16
Sampling Stop: 07/01/2006 12:00:00 | Decay Time. . . . . 1.73E+003 Hrs
Buildup Time. . . . . 0.00E+000 Hrs | Live Time . . . . . 1800 Sec
Sample Size . . . . . 1.00E+000 L | Real Time . . . . . 2049 Sec
Collection Efficiency . . . . . 1.0000 | Spc. File . . . . . 062650D08.SPC
-----

```

Detector #: 8 (Detector 8)

Energy(keV) = -0.62 + 0.500*Ch + 0.00E+00*Ch^2 + 0.00E+00*Ch^3 09/11/2006

FWHM(keV) = 0.72 + 0.007*En + 8.11E-04*En^2 + 0.00E+00*En^3 08/17/2006

Where En = Sqrt(Energy in keV)

Search Sensitivity: 1.00 | Sigma Multiplier: 2.00 | Search Start/End: 80/4000

=====

PEAK SEARCH RESULTS

=====

PK. #	ENERGY (keV)	ADDRESS CHANNEL	NET/MDA COUNTS	UN- CERTAINTY	C.L. COUNTS	BKG COUNTS	FWHM (keV)	FLAG
1	49.61	100.41	7127	827	666	49210	2.31	a Wide Pk
2	59.28	119.74	86857	697	305	18791	0.81	a
3	67.15	135.46	7399	791	635	42570	2.50	a Wide Pk
4	70.44	142.04	8866	709	562	36894	2.05	b
5	72.65	146.46	7162	406	304	17028	0.98	c
6	82.26	165.68	1162	306	245	13315	0.72	a
7	87.83	176.81	118764	766	274	15139	0.85	a
8	121.85	244.80	58833	546	206	8546	0.88	a
9	135.93	272.96	1240	572	467	22990	2.38	a Wide Pk
10	136.29	273.67	6261	285	195	7663	0.82	b
11	165.69	332.44	60012	549	204	7690	0.93	a HiResid
12	198.98	398.98	1062	259	206	7259	1.11	a
13	255.03	511.01	1843	214	161	4782	0.99	a
14	279.11	559.15	47051	469	146	3944	1.04	a
15	306.95	614.81	81	158	129	3091	0.91	a NET< CL
16	391.68	784.17	42631	439	122	2923	1.14	a
17	502.80	1006.30	87	130	105	2194	1.17	a NET< CL
18	511.08	1022.85	871	229	182	4431	2.37	a Wide Pk
19	661.79	1324.10	26514	360	126	2914	1.33	a
20	697.35	1395.17	89	162	132	3044	1.41	a NET< CL
21	698.28	1397.03	112	107	86	1691	0.83	b
22	813.99	1628.32	720	111	80	1376	1.22	a
23	898.11	1796.49	48656	459	104	2086	1.53	a
24	1006.86	2013.86	66	85	69	1093	0.94	a NET< CL

000100

=====

PEAK SEARCH RESULTS

=====

PK. #	ENERGY (keV)	ADDRESS CHANNEL	NET/MDA COUNTS	UN- CERTAINTY	C.L. COUNTS	BKG COUNTS	FWHM (keV)	FLAG
25	1173.24	2346.45	30086	360	80	1224	1.76	a
26	1324.94	2649.68	1091	137	99	1397	3.10	a HiResid
27	1332.32	2664.43	27403	340	65	809	1.87	b HiResid
28	1465.23	2930.10	56	68	55	661	1.36	a
29	1835.26	3669.76	28566	341	39	272	2.14	a HiResid
30	1849.43	3698.08	46	54	43	242	3.65	a
31	1856.91	3713.03	43	28	20	88	1.36	b

062650D08.SPC Analyzed by

SEEKER BACKGROUND SUBTRACT RESULTS Version 1.8.2

Paragon Analytics, Div. of DataChem Lab
GammaScan

Background File: DET080908.BKG (060809-8 WEEKLY BKGD)

Bkg.File Detector #: 8

=====

BACKGROUND SUBTRACT RESULTS

=====

PK#	ENERGY (keV)	OLD NET COUNTS	OLD UN- CERTAINTY	OLD CR.LEVEL	NEW NET COUNTS	NEW UN- CERTAINTY	NEW CR.LEVEL	FLAG
3	67.15	7399	791	635	7394	791	635	
12	198.98	1062	259	206	1056	259	206	
18	511.08	871	229	182	820	229	182	
23	898.11	48656	459	104	48654	459	104	

062650D08.SPC Analyzed by

SEEKER CALIBRATION RESULTS Version 2.0.4

Sample ID: 0613004-8 GEO 1 EFF CAL (824)
Stds. Match Tolerance: 2.00 keV

Detector Number: 08 Calibration Date. . . 09/11/2006 14:14:16
Geometry File (D08)(Sh01).EFF ID. Geo 1 Eff Cal
Amount of Std. in Calib. Source: 1.000000 gm

Eff = 1 / [6.36e-01*En^-1.19e+00 + 1.20e+02*En^ 9.04e-01]
(Where En = Energy in MeV) (Exponential)

Pk. #	Energy (keV)	Measured Efficiency	% Difference	Calculated Efficiency	% Difference	Prev.Calc. Efficiency
1	59.50	3.65e-02	-0.92	3.62e-02	0.52	3.63e-02
2	88.04	3.93e-02	2.60	4.03e-02	0.69	4.06e-02
3	122.06	3.99e-02	-2.42	3.89e-02	0.25	3.90e-02
4	165.85	3.45e-02	0.02	3.45e-02	-0.24	3.44e-02
5	279.00	2.48e-02	-0.71	2.46e-02	-0.61	2.44e-02
6	391.68	1.86e-02	1.17	1.88e-02	-0.51	1.87e-02
7	661.64	1.18e-02	1.67	1.20e-02	0.01	1.20e-02
8	898.02	8.90e-03	2.72	9.14e-03	0.42	9.18e-03
9	1173.21	7.22e-03	-0.21	7.20e-03	0.82	7.26e-03
10	1332.48	6.58e-03	-2.44	6.43e-03	1.01	6.49e-03
11	1836.01	4.99e-03	-3.63	4.82e-03	1.52	4.89e-03

Calibration Results Saved.

OK
JJB
9/11/06

Geometry 01 Calibration Verification: Gamma Mixed Nuclide Source																					
Std. #	798	Detector	8																		
NEW SOURCE :	798	Half Life(Y)	432.0000	Gammas/Sec.	1353	Gamma Fraction:	0.3590	Mass of Standard	1	REF DATE :	7/1/2005	count date	9/11/2006								
FROM CALIBRATION CERTIFICATE		KeV	59.9	1353		FROM ANALYTICS.LIB		EXPECTED ACTIVITY		DPS		pCi/g		Recovery		Activity		Pass/Fail		# of half-lives	
Isotope	Am-241	59.9	432.0000	1353	0.3590	L	Am-241	3768.8	101869.5	104000	102%	104000	102%	104000	Pass	0.00					
	Cd-109	88	1.2666	1912	0.0361		Cd-109	52964.0	1431459.2	1400000	98%	1400000	98%	1400000	Pass	0.94					
	Co-57	122	0.7441	1048	0.8551		Co-57	1225.6	33124.0	33900	102%	33900	102%	33900	Pass	1.61					
	Ce-139	166	0.3768	1426	0.8035		Ce-139	1774.7	47965.8	47700	99%	47700	99%	47700	Pass	3.17					
	Hg-203	279	0.1276	3260	0.7730		Hg-203	4217.3	113982.0	NA	>5 h-lives	NA	>5 h-lives	>5 h-lives	Pass	9.38					
	Sn-113	392	0.3151	2038	0.6490		Sn-113	3140.2	84870.7	81700	96%	81700	96%	81700	Pass	3.80					
	Cs-137	662	30.0000	1268	0.8512		Cs-137	1489.7	40261.1	39300	98%	39300	98%	39300	Pass	0.04					
	Y-88	898	0.2919	5008	0.9340		Y-88	5361.9	144915.8	144000	99%	144000	99%	144000	Pass	4.10					
	Co-60	1173	5.2714	2475	1.0000		Co-60	2475.0	66891.9	67500	101%	67500	101%	67500	Pass	0.23					
	Co-60	1332	5.2714	2499	1.0000		Co-60	2499.0	67540.5	68000	101%	68000	101%	68000	Pass	0.23					
	Y-88	1836	0.2919	5224	0.9938		Y-88	5256.6	142070.0	151000	106%	151000	106%	151000	Pass	4.10					

R:\inst\gamma\gammamms17798.xls

000104

 SEEKER G A M M A A N A L Y S I S R E S U L T S P S Version 1.8.4

Paragon Analytics, Div. of DataChem Lab
 GammaScan

Geo 1 / Water

Sample ID: 0613004-8A GEO 1 CAL VER (798)

```
-----
Sampling Start: 07/01/2005 12:00:00 | Counting Start: 09/11/2006 14:56:51
Sampling Stop: 07/01/2005 12:00:00 | Decay Time. . . . . 1.05E+004 Hrs
Buildup Time. . . . . 0.00E+000 Hrs | Live Time . . . . . 1800 Sec
Sample Size . . . . . 1.00E+000 L | Real Time . . . . . 1921 Sec
Collection Efficiency . . . . . 1.0000 | Spc. File . . . . . 062651D08.SPC
-----
```

Detector #: 8 (Detector 8)

Energy(keV) = -0.62 + 0.500*Ch + 0.00E+00*Ch^2 + 0.00E+00*Ch^3 09/11/2006

FWHM(keV) = 0.72 + 0.007*En + 8.11E-04*En^2 + 0.00E+00*En^3 08/17/2006

Where En = Sqrt(Energy in keV)

 Search Sensitivity: 1.00 | Sigma Multiplier: 2.00 | Search Start/End: 80/4000

=====

PEAK SEARCH RESULTS

=====

PK. #	ENERGY (keV)	ADDRESS CHANNEL	NET/MDA COUNTS	UN- CERTAINTY	C.L. COUNTS	BKG COUNTS	FWHM (keV)	FLAG
1	43.03	87.26	399	315	257	11298	1.15	a
2	48.88	98.95	2906	406	322	17672	1.09	a HiResid Wide Pk
3	50.16	101.51	3433	512	410	24876	1.50	b HiResid
4	51.68	104.54	523	186	148	6097	0.41	c HiResid
5	59.26	119.70	89666	661	230	10678	0.83	a
6	65.67	132.50	726	296	239	9759	1.15	a HiResid Wide Pk
7	67.33	135.83	3785	555	445	20913	2.50	b HiResid
8	87.81	176.76	70894	578	184	6817	0.84	a
9	121.82	244.75	24618	345	118	2828	0.88	a
10	136.24	273.57	2838	168	107	2308	0.84	a
11	165.66	332.38	9742	240	113	2344	0.93	a
12	255.06	511.09	204	115	91	1686	0.89	a
13	279.10	559.13	240	120	96	1692	1.00	a
14	391.62	784.06	4771	175	89	1549	1.10	a
15	510.79	1022.26	57	75	60	831	0.83	a NET< CL
16	661.73	1323.98	25997	336	76	1079	1.37	a
17	898.03	1796.32	4783	172	84	1373	1.55	a
18	1115.21	2230.44	80	101	82	1238	2.03	a NET< CL
19	1173.14	2346.24	27650	340	59	676	1.73	a
20	1323.29	2646.38	18	35	28	194	1.02	a NET< CL
21	1332.20	2664.18	24848	318	36	252	1.82	a HiResid
22	1835.09	3669.41	2799	108	16	47	2.20	a

062651D08.SPC Analyzed by

SEEKER BACKGROUND SUBTRACT RESULTS Vers. 2.2.1

Paragon Analytics, Div. of DataChem Lab
GammaScan

Background File: DET080908.BKG (060809-8 WEEKLY BKGD)

Bkg.File Detector #: 8

=====

BACKGROUND SUBTRACT RESULTS

=====

PK#	ENERGY (keV)	OLD NET COUNTS	OLD UN- CERTAINTY	OLD CR.LEVEL	NEW NET COUNTS	NEW UN- CERTAINTY	NEW CR.LEVEL	FLAG
4	51.68	523	186	148	522	186	148	
6	65.67	726	296	239	721	296	239	
15	510.79	57	75	60	7	75	62	NET<CL
17	898.03	4783	172	84	4781	172	84	

 SEEKER F I N A L A C T I V I T Y R E P O R T Version 2.2.1

Paragon Analytics, Div. of DataChem Lab
 GammaScan

Geo 1 / Water

Sample ID: 0613004-8A GEO 1 CAL VER (798)

```
-----
Sampling Start: 07/01/2005 12:00:00 | Counting Start: 09/11/2006 14:56:51
Sampling Stop: 07/01/2005 12:00:00 | Decay Time. . . . . 1.05e+004 Hrs
Buildup Time. . . . . 0.00e+000 Hrs | Live Time . . . . . 1800 Sec
Sample Size . . . . . 1.00e+000 L | Real Time . . . . . 1921 Sec
Collection Efficiency . . . . . 1.0000 | Spectrum File . . . . . 062651D08.SPC
Cr. Level Confidence Interval: 95 % | Det. Limit Confidence Interval: 95 %
-----
```

Detector #: 8 (Detector 8)

Efficiency File: (D08) (Sh01).EFF (Geo 1 Eff Cal)

Eff.=1/[6.36E-01*En^-1.19E+00 + 1.20E+02*En^9.04E-01] 09/11/2006

Library File:ANALYTICAL.LIB (Analytical)

=====

MEASURED or MDA CONCENTRATIONS

=====

Nuclide	ENERGY E (keV)	N T	Concentration (pCi/L)	MDA	Critical Level	Halflife (hrs)
Am-241	59.54	1.04E+05 +- 7.66E+02	5.37E+02	2.67E+02	3.79E+06	
Cd-109	88.02	1.40E+06 +- 1.14E+04	7.34E+03	3.64E+03	1.11E+04	
Co-57	122.07	3.39E+04 +- 4.76E+02	3.30E+02	1.63E+02	6.50E+03	
Ce-139	165.85	4.77E+04 +- 1.18E+03	1.12E+03	5.51E+02	3.30E+03	
Sn-113	391.68	8.17E+04 +- 3.00E+03	3.08E+03	1.52E+03	2.76E+03	
Cs-137	661.62	3.93E+04 +- 5.07E+02	2.35E+02	1.16E+02	2.64E+05	
Y-88	Average:x	1.47E+05 +- 3.86E+03	2.56E+03	
	898.02	1.44E+05 +- 5.18E+03	5.15E+03	2.53E+03	2.56E+03	
	1836.01	1.51E+05 +- 5.79E+03	1.89E+03	8.72E+02	2.56E+03	
Co-60	Average:x	6.77E+04 +- 6.01E+02	4.62E+04	
	1173.21	6.75E+04 +- 8.31E+02	2.96E+02	1.44E+02	4.62E+04	
	1332.48	6.80E+04 +- 8.71E+02	2.05E+02	9.89E+01	4.62E+04	
Hg-203	279.18	MDA	7.24E+04r	3.55E+04 1.12E+03	

MEASURED TOTAL: 1.93E+06 +- 2.18E+04 pCi/L

=====

UNKNOWN, SUM or ESCAPE PEAKS

=====

PK. #	ENERGY (keV)	ADDRESS CHANNEL	NET COUNTS	UN- CERTAINTY	C.L. COUNTS	BKG COUNTS	FWHM (keV)	FLAG
1	43.03	87.26	399	315	257	11298	1.15	Unknown
2	48.88	98.95	2906	406	322	17672	1.09	Unknown
3	50.16	101.51	3433	512	410	24876	1.50	Unknown

062651D08.SPC Analyzed by

=====

UNKNOWN, SUM or ESCAPE PEAKS

=====

PK. #	ENERGY (keV)	ADDRESS CHANNEL	NET COUNTS	UN- CERTAINTY	C.L. COUNTS	BKG COUNTS	FWHM (keV)	FLAG
4	51.68	104.54	522	186	148	6097	0.41	Unknown
6	65.67	132.50	721	296	239	9759	1.15	Unknown
7	67.33	135.83	3785	555	445	20913	2.50	Unknown
10	136.24	273.57	2838	168	107	2308	0.84	Unknown
12	255.06	511.09	204	115	91	1686	0.89	Unknown
13	279.10	559.13	240	120	96	1692	1.00	Unknown
15	510.79	1022.26	7	75	62	831	0.83	Deleted
18	1115.21	2230.44	80	101	82	1238	2.03	Deleted
20	1323.29	2646.38	18	35	28	194	1.02	Deleted

=====

c:\SEEKER\BIN\062651d08.res Analysis Results Saved.

Gamma Spectroscopy

Quality Control Data

Weekly Background Calibrations

000109

Paragon Analytics

Gamma Spectrometer Calibration Log

Date: 10/27/06

Reviewed By/Date: JP 10/27/06

Det. No.	Out. Of Service	Background		Source Check			Repeat Source Check			
		Started	OK	Started	OK	Failed Parameter(s)	OK	Failed Parameter(s)	Corrective Action Taken **	Removed from Service
1.	JP	/	/	/	/					
2.		Q	JP	JP	/	602 FWHM	JP			
3.	JP	/	/	/	/					
4.		Q	*	JP	/	1332 Centroid	JP			
5.	JP	/	/	/	/					
6.		Q	JP	JP	JP					
7.	JP	/	/	/	/					
8.		Q	JP	JP	/	60 FWHM	JP			
9.		Q	JP	JP	JP					
10.	JP	/	/	/	/					
11.										
12.										

** Corrective Action:

* Recount. Peak Shift JP 10/30/06

318711 A

Form 754r11a.doc (6/13/2005)

017718

000110

062512D02.SPC Analyzed by *A*

SEEKER G A M M A A N A L Y S I S R E S U L T S PS Version 1.8.4

Paragon Analytics, Div. of DataChem Lab
GammaScan

Weekly Background Check

Sample ID: 061027-2 WEEKLY BKG

```

-----
Sampling Start: 10/27/2006 18:03:56 | Counting Start: 10/27/2006 18:03:56
Sampling Stop: 10/27/2006 18:03:56 | Decay Time. . . . . 0.00E+000 Hrs
Buildup Time. . . . . 0.00E+000 Hrs | Live Time . . . . . 60000 Sec
Sample Size . . . . . 1.00E+000 L | Real Time . . . . . 60060 Sec
Collection Efficiency . . . . . 1.0000 | Spc. File . . . . . 062512D02.SPC
-----

```

Detector #: 2 (Detector 2)

Energy(keV) = -0.72 + 0.500*Ch + 9.30E-09*Ch^2 + 0.00E+00*Ch^3 10/27/2006

FWHM(keV) = 0.69 + 0.006*En + 1.18E-03*En^2 + 0.00E+00*En^3 05/16/2006

Where En = Sqrt(Energy in keV)

Search Sensitivity: 1.00 | Sigma Multiplier: 2.00 | Search Start/End: 80/4000

=====

PEAK SEARCH RESULTS

=====

PK. #	ENERGY (keV)	ADDRESS CHANNEL	NET/MDA COUNTS	UN- CERTAINTY	C.L. COUNTS	BKG COUNTS	FWHM (keV)	FLAG
1	46.60	94.54	78	81	65	790	1.01	a
2	63.30	127.90	200	75	57	650	0.77	a
3	66.26	133.82	160	83	65	780	0.97	b
4	74.97	151.23	101	75	59	709	0.76	a
5	77.13	155.54	169	57	42	426	0.42	b
6	84.33	169.93	69	79	63	740	0.97	a
7	87.36	175.99	13	48	39	370	0.43	b NET< CL
8	92.71	186.67	436	92	67	776	1.10	a
9	140.17	281.50	164	75	58	621	0.96	a
10	144.03	289.21	49	45	36	311	0.47	b
11	169.79	340.69	51	77	62	656	1.09	a NET< CL
12	185.76	372.58	232	68	50	512	0.80	a
13	198.39	397.82	191	67	50	509	0.76	a
14	238.67	478.30	310	73	53	510	1.00	a
15	241.95	484.86	87	74	59	595	1.14	b
16	268.30	537.51	21	76	62	609	1.24	a NET< CL
17	295.28	591.41	92	57	44	386	0.88	a
18	351.81	704.35	216	63	46	421	1.15	a
19	511.12	1022.65	1648	126	79	733	2.77	a Wide Pk
20	558.51	1117.33	157	49	35	235	1.21	a
21	570.32	1140.93	86	54	42	339	1.43	a
22	583.45	1167.15	169	65	49	407	2.00	a
23	609.40	1219.01	197	61	44	399	1.31	a
24	669.46	1339.00	36	34	27	163	1.00	a

000111

=====

PEAK SEARCH RESULTS

=====

PK. #	ENERGY (keV)	ADDRESS CHANNEL	NET/MDA COUNTS	UN- CERTAINTY	C.L. COUNTS	BKG COUNTS	FWHM (keV)	FLAG
25	692.63	1385.29	33	55	44	375	1.43	a NET< CL
26	803.47	1606.73	159	51	37	252	1.67	a
27	911.75	1823.06	92	40	29	168	1.56	a
28	962.18	1923.82	42	37	29	168	1.60	a
29	1120.54	2240.21	36	32	24	122	1.41	a
30	1461.13	2920.64	221	46	29	126	2.87	a

062512D02.SPC Analyzed by


 SEEKER B A C K G R O U N D Q . C . A N A L Y S I S Version 2.2.2

ID: 061027-2 WEEKLY BKG

Detector # 2 Background Q.C. Analysis for 10/27/2006 18:03:56

#	Parameter	Value	n Sigma Test	Bounds Test	T- Test
10	50-> 150 keV Bkg	23.881	N.A.	Pass	N.A.
11	150-> 250 keV Bkg	19.662	N.A.	Pass	N.A.
12	250-> 500 keV Bkg	29.145	N.A.	Pass	N.A.
13	500->1000 keV Bkg	30.364	N.A.	Pass	N.A.
14	1000->2000 keV Bkg	16.949	N.A.	Pass	N.A.
15	40-> 50 keV Bkg	2.939	N.A.	Pass	N.A.

Q.C. Results Saved.



000114

062598D06.SPC Analyzed by *Q*

SEEKER G A M M A A N A L Y S I S R E S U L T S PS Version 1.8.4

Paragon Analytics, Div. of DataChem Lab
GammaScan

Weekly Background Check

Sample ID: 061027-6 WEEKLY BKG

```

-----
Sampling Start: 10/27/2006 18:04:33 | Counting Start: 10/27/2006 18:04:33
Sampling Stop: 10/27/2006 18:04:33 | Decay Time. . . . . 0.00E+000 Hrs
Buildup Time. . . . . 0.00E+000 Hrs | Live Time . . . . . 60000 Sec
Sample Size . . . . . 1.00E+000 L | Real Time . . . . . 60022 Sec
Collection Efficiency . . . . . 1.0000 | Spc. File . . . . . 062598D06.SPC
-----

```

Detector #: 6 (Detector 6)

Energy(keV) = -0.65 + 0.501*Ch +-2.40E-08*Ch^2 + 0.00E+00*Ch^3 10/27/2006

FWHM(keV) = 1.19 + -0.002*En + 7.29E-04*En^2 + 0.00E+00*En^3 07/24/2006

Where En = Sqrt(Energy in keV)

Search Sensitivity: 1.00 | Sigma Multiplier: 2.00 | Search Start/End: 80/4000

=====

PEAK SEARCH RESULTS

=====

PK. #	ENERGY (keV)	ADDRESS CHANNEL	NET/MDA COUNTS	UN- CERTAINTY	C.L. COUNTS	BKG COUNTS	FWHM (keV)	FLAG
1	53.57	108.25	65	71	57	739	0.80	a
2	66.40	133.87	126	73	58	754	0.80	a
3	74.79	150.61	57	63	50	615	0.69	a
4	92.58	186.13	178	74	57	680	1.00	a
5	139.80	280.39	149	73	57	677	0.98	a
6	185.86	372.34	98	56	43	450	0.72	a
7	198.30	397.18	182	74	57	677	1.03	a
8	238.68	477.79	107	68	53	594	0.97	a
9	295.28	590.78	51	45	35	301	0.69	a
10	352.13	704.29	136	62	47	444	1.17	a
11	511.08	1021.62	1240	112	71	752	2.38	a Wide Pk
12	558.53	1116.34	127	48	35	268	1.07	a
13	570.37	1139.98	48	42	32	242	1.06	a
14	583.17	1165.54	70	40	30	203	1.05	a
15	596.82	1192.80	77	58	45	396	1.40	a
16	598.88	1196.90	65	57	45	396	1.50	b
17	609.28	1217.67	121	61	46	419	1.53	a
18	803.10	1604.64	114	40	28	161	1.24	a
19	911.46	1821.00	55	44	34	192	2.00	a
20	962.06	1922.01	38	29	22	112	0.92	a
21	1460.97	2918.22	326	45	22	88	1.91	a
22	1765.12	3525.56	49	32	23	84	2.73	a

062598D06.SPC Analyzed by


SEEKER B A C K G R O U N D Q . C . A N A L Y S I S V e r s i o n 2 . 2 . 2

ID: 061027-6 WEEKLY BKG

Detector # 6 Background Q.C. Analysis for 10/27/2006 18:04:33

#	Parameter	Value	n Sigma Test	Bounds Test	T- Test
10	50-> 150 keV Bkg	25.076	N.A.	Pass	N.A.
11	150-> 250 keV Bkg	21.437	N.A.	Pass	N.A.
12	250-> 500 keV Bkg	31.223	N.A.	Pass	N.A.
13	500->1000 keV Bkg	30.163	N.A.	Pass	N.A.
14	1000->2000 keV Bkg	16.574	N.A.	Pass	N.A.
15	40-> 50 keV Bkg	3.502	N.A.	Pass	N.A.

Q.C. Results Saved.



000117

Paragon Analytics

Gamma Spectrometer Calibration Log

Date: 11/3/06

Reviewed By/Date: JP 11/3/06

Det. No.	Out Of Service	Background		Source Check			Repeat Source Check			
		Started	OK	Started	OK	Failed Parameter(s)	OK	Failed Parameter(s)	Corrective Action Taken **	Removed from Service
1.	JP	/	/	/	/					
2.		JP	JP	JP	JP					
3.	JP	/	/	/	/					
4.	JP	/	/	/	/					
5.	JP	/	/	/	/					
6.		JP	JP	JP	JP					
7.	JP	/	/	/	/					
8.		JP	JP	JP	/	1332 Centrad	JP			
9.		JP	JP	JP	JP	662 FWHM	JP			
10.	JP	/	/	/	/					
11.										
12.										

** Corrective Action:

318721 A

Form 754r11a.doc (6/13/2005)

057818

000118

 SEEKER G A M M A A N A L Y S I S R E S U L T S PS Version 1.8.4

Paragon Analytics, Div. of DataChem Lab
 GammaScan

Weekly Background Check

Sample ID: 061103-2 WEEKLY BKG

```
-----
Sampling Start: 11/03/2006 17:00:00 | Counting Start: 11/03/2006 17:03:10
Sampling Stop: 11/03/2006 17:00:00 | Decay Time. . . . . 5.28E-002 Hrs
Buildup Time. . . . . 0.00E+000 Hrs | Live Time . . . . . 60000 Sec
Sample Size . . . . . 1.00E+000 L | Real Time . . . . . 60021 Sec
Collection Efficiency . . . . . 1.0000 | Spc. File . . . . . 062549D02.SPC
-----
```

Detector #: 2 (Detector 2)

Energy(keV) = -0.77 + 0.500*Ch + 4.01E-08*Ch^2 + 0.00E+00*Ch^3 11/03/2006
 FWHM(keV) = 0.69 + 0.006*En + 1.18E-03*En^2 + 0.00E+00*En^3 05/16/2006
 Where En = Sqrt(Energy in keV)

 Search Sensitivity: 1.00 | Sigma Multiplier: 2.00 | Search Start/End: 80/4000

=====

PEAK SEARCH RESULTS

=====

PK. #	ENERGY (keV)	ADDRESS CHANNEL	NET/MDA COUNTS	UN- CERTAINTY	C.L. COUNTS	BKG COUNTS	FWHM (keV)	FLAG
1	46.26	94.00	97	83	66	810	0.95	a
2	53.55	108.58	47	69	56	632	0.75	a NET< CL
3	56.27	114.00	13	37	30	253	0.40	b NET< CL
4	63.05	127.56	197	126	101	1331	1.77	a Wide Pk
5	66.31	134.08	295	84	63	726	1.00	b
6	69.73	140.92	107	70	55	605	0.89	c
7	74.14	149.73	358	164	131	1815	2.48	d
8	77.08	155.59	184	72	55	605	0.85	e
9	87.07	175.57	31	49	39	384	0.43	a NET< CL
10	92.55	186.53	427	85	61	688	0.94	a
11	129.13	259.63	37	45	35	310	0.46	a
12	139.06	279.48	56	46	35	310	0.46	a
13	140.02	281.39	115	48	35	310	0.47	b
14	167.60	336.51	69	79	64	694	1.17	a
15	171.69	344.69	27	61	50	496	0.79	b NET< CL
16	185.66	372.62	339	101	77	879	1.44	a
17	192.81	386.91	53	62	50	497	0.82	a
18	198.36	397.99	199	83	64	695	1.09	b
19	238.56	478.34	269	87	66	691	1.26	a
20	241.80	484.82	99	83	66	691	1.31	b
21	295.02	591.17	113	49	36	285	0.73	a
22	338.46	677.99	41	56	45	419	0.97	a NET< CL
23	351.63	704.32	214	64	47	427	1.16	a
24	510.88	1022.58	1667	123	75	698	2.69	a Wide Pk

=====

PEAK SEARCH RESULTS

=====

PK. #	ENERGY (keV)	ADDRESS CHANNEL	NET/MDA COUNTS	UN- CERTAINTY	C.L. COUNTS	BKG COUNTS	FWHM (keV)	FLAG
25	558.24	1117.22	173	54	39	278	1.29	a
26	569.33	1139.36	76	52	41	339	1.27	a
27	582.97	1166.63	141	58	43	349	1.62	a
28	609.28	1219.21	176	65	49	460	1.40	a
29	692.70	1385.91	90	60	47	389	1.82	a
30	695.09	1390.67	112	68	54	460	2.21	b
31	802.84	1605.99	132	48	35	225	1.64	a
32	897.97	1796.08	73	47	36	234	2.07	a
33	911.51	1823.14	83	39	29	167	1.63	a
34	961.63	1923.29	44	32	24	134	1.11	a
35	1014.29	2028.51	40	47	37	236	2.29	a
36	1120.28	2240.28	42	29	21	104	1.11	a
37	1460.68	2920.35	178	41	25	105	2.43	a
38	1764.85	3528.00	37	26	19	65	2.02	a

062550D02.SPC Analyzed by

SEEKER CALIBRATION RESULTS Version 2.0.4

Sample ID: DAILY CHECK

Stds. Match Tolerance: 2.00 keV

Detector Number: 02 Calibration Date. . . 11/04/2006 09:49:55

Energy(keV) = -0.76 + 0.500*Ch + 4.70e-08*Ch^2 + 0.00e+00*Ch^3

Pk. #	Measured Centroid	Calculated Energy	Energy (keV)	% Difference
1	120.43	59.51	59.50	0.01
2	1323.37	661.62	661.64	-0.00
3	2345.36	1173.26	1173.21	0.00
4	2663.28	1332.45	1332.48	-0.00

Calibration Results Saved.

062549D02.SPC Analyzed by

SEEKER B A C K G R O U N D Q . C . A N A L Y S I S V e r s i o n 2 . 2 . 2


ID: 061103-2 WEEKLY BKG

Detector # 2 Background Q.C. Analysis for 11/03/2006 17:03:10

#	Parameter	Value	n Sigma Test	Bounds Test	T- Test
10	50-> 150 keV Bkg	24.032	N.A.	Pass	N.A.
11	150-> 250 keV Bkg	20.055	N.A.	Pass	N.A.
12	250-> 500 keV Bkg	29.655	N.A.	Pass	N.A.
13	500->1000 keV Bkg	30.929	N.A.	Pass	N.A.
14	1000->2000 keV Bkg	17.392	N.A.	Pass	N.A.
15	40-> 50 keV Bkg	3.030	N.A.	Pass	N.A.

Q.C. Results Saved.

RE-CALC QC HEADER for Overwritten Results.



000123

SEEKER GAMMA ANALYSIS RESULTS PS Version 1.8.4

Paragon Analytics, Div. of DataChem Lab
GammaScan

Weekly Background Check

Sample ID: 061103-8 WEEKLY BKG

```
-----
Sampling Start: 11/03/2006 17:00:00 | Counting Start: 11/03/2006 17:03:53
Sampling Stop: 11/03/2006 17:00:00 | Decay Time: . . . . . 6.47E-002 Hrs
Buildup Time: . . . . . 0.00E+000 Hrs | Live Time . . . . . 60000 Sec
Sample Size . . . . . 1.00E+000 L | Real Time . . . . . 60109 Sec
Collection Efficiency . . . . . 1.0000 | Spc. File . . . . . 063142D08.SPC
-----
```

Detector #: 8 (Detector 8)

Energy(keV) = -0.50 + 0.500*Ch + 2.17E-07*Ch^2 + 0.00E+00*Ch^3 11/03/2006
 FWHM(keV) = 0.72 + 0.007*En + 8.11E-04*En^2 + 0.00E+00*En^3 08/17/2006
 Where En = Sqrt(Energy in keV)

Search Sensitivity: 1.00 | Sigma Multiplier: 2.00 | Search Start/End: 80/4000

PEAK SEARCH RESULTS

PK. #	ENERGY (keV)	ADDRESS CHANNEL	NET/MDA COUNTS	UN-CERTAINTY	C.L. COUNTS	BKG COUNTS	FWHM (keV)	FLAG
1	46.41	93.78	437	76	53	611	0.71	a
2	63.18	127.31	560	88	61	755	0.80	a
3	66.15	133.25	244	91	70	906	1.06	b
4	74.78	150.50	216	81	62	783	0.80	a
5	77.05	155.03	219	81	62	783	0.81	b
6	84.45	169.83	85	75	60	717	0.88	a
7	87.26	175.43	80	54	42	430	0.43	b
8	92.58	186.07	773	98	66	813	1.01	a
9	139.71	280.28	184	61	45	453	0.68	a
10	178.44	357.69	50	55	44	424	0.65	a
11	185.72	372.24	360	80	58	624	0.97	a
12	198.26	397.31	216	76	58	621	0.98	a
13	202.53	405.83	51	72	58	621	0.90	b NET< CL
14	205.80	412.36	51	54	43	414	0.68	c
15	238.54	477.80	320	70	50	495	0.88	a
16	241.93	484.57	42	44	35	297	0.52	b
17	295.03	590.68	121	67	52	497	1.06	a
18	351.93	704.37	224	72	54	497	1.11	a
19	511.10	1022.35	1699	128	80	825	2.47	a Wide Pk
20	537.58	1075.25	40	39	31	214	0.84	a
21	558.51	1117.06	294	63	43	322	1.52	a
22	569.89	1139.79	60	61	49	410	1.47	a
23	583.40	1166.76	144	56	42	321	1.26	a
24	609.27	1218.43	172	69	53	480	1.54	a

=====

PEAK SEARCH RESULTS

=====

PK. #	ENERGY (keV)	ADDRESS CHANNEL	NET/MDA COUNTS	UN- CERTAINTY	C.L. COUNTS	BKG COUNTS	FWHM (keV)	FLAG
25	670.16	1340.03	39	41	32	219	0.96	a
26	691.87	1383.38	35	34	26	168	0.73	a Wide Pk
27	694.16	1387.96	145	88	70	630	2.45	b
28	726.90	1453.34	36	42	33	218	1.12	a
29	803.21	1605.70	156	53	38	270	1.65	a
30	898.82	1796.57	47	38	29	178	1.35	a
31	911.55	1821.98	58	42	33	207	1.54	a
32	962.38	1923.43	39	53	42	296	2.02	a NET< CL
33	1460.99	2918.19	175	41	26	128	1.76	a
34	1765.56	3525.39	54	35	26	104	2.88	a

063142D08.SPC Analyzed by


SEEKER B A C K G R O U N D Q . C . A N A L Y S I S V e r s i o n 2 . 2 . 2

ID: 061103-8 WEEKLY BKG

Detector # 8 Background Q.C. Analysis for 11/03/2006 17:03:53

#	Parameter	Value	n Sigma Test	Bounds Test	T- Test
10	50-> 150 keV Bkg	28.284	N.A.	Pass	N.A.
11	150-> 250 keV Bkg	21.703	N.A.	Pass	N.A.
12	250-> 500 keV Bkg	32.931	N.A.	Pass	N.A.
13	500->1000 keV Bkg	35.848	N.A.	Pass	N.A.
14	1000->2000 keV Bkg	20.175	N.A.	Pass	N.A.
15	40-> 50 keV Bkg	3.705	N.A.	Pass	N.A.

Q.C. Results Saved.



000127

Gamma Spectroscopy

Quality Control Data

Daily Instrument Performance
Checks

000128



ATID 0126
 Analytcs, Inc.
 1380 Senhouse Industrial Boulevard
 Atlanta, Georgia 30318
 404 352-6677

Recd 12-29-94

CERTIFICATE OF CALIBRATION
 Standard Radionuclide Source

49500-307

50 mL Liquid in Flame Sealed Vial

This standard radionuclide source was prepared using aliquots measured gravimetrically from master radionuclide solution sources. The Am-241 was calibrated by 4 pi alpha liquid scintillation counting. All other radionuclides were calibrated using a germanium gamma spectrometer system. Calibration and purity were checked using a germanium gamma spectrometer system. At the time of calibration no interfering gamma-ray emitting impurities were detected. The gamma-ray emission rates for the most intense gamma-ray lines are given. Analytcs maintains traceability to the National Institute of Standards and Technology through a Measurements Assurance Program as described in USNRC Regulatory Guide 4.15, Rev. 1, February, 1979.

Calibration date: October 1, 1994 12:00 EST

ISOTOPE	GAMMA-RAY ENERGY	HALF-LIFE	GAMMA-RAYS PER SECOND	TOTAL UNCERTAINTY %
Am-241	59.5	432 y	1759	5.0
Cd-109	88	462.6 d	2622	4.5
Co-57	122	271.79 d	1472	4.2
Ce-139	166	137.64 d	2037	4.6
Hg-203	279	46.595 d	3978	4.5
Sn-113	392	115.09 d	2862	4.7
Cs-137	662	30.0 y	1735	4.7
Y-88	898	106.63 d	7290	4.4
Co-60	1173	5.2714 y	3347	4.3
Co-60	1332	5.2714 y	3355	4.8
Y-88	1836	106.63 d	7630	4.3

50.8520 grams solution 4M HCl.
 P O NUMBER 45864, Item 1

SOURCE PREPARED BY:

M. D. Currie
 M. D. Currie, Radiochemist

QA APPROVED:

J. M. Martyn 12-23-94

This standard will expire one year after the calibration date.

000129



ATI ID 0126

Analytics, Inc.
1380 Seaboard Industrial Boulevard
Atlanta, Georgia 30318
404 372-8677

ANALYSIS OF UNCERTAINTY

BATCH 78 MIXED GAMMA STANDARDS WITH Am-241
CALIBRATION DATE: October 1, 1994 12:00 EST

GAMMA RAY ENERGY (keV)	RANDOM ERROR % (99 % CL)	SYSTEMATIC ERROR %	TOTAL %
59.5	2.0	3.0	5.0
88	1.5	3.0	4.5
122	1.6	2.6	4.2
166	2.1	2.5	4.6
279	1.5	3.0	4.5
392	1.9	2.8	4.7
662	1.9	2.8	4.7
898	1.4	3.0	4.4
1173	1.5	2.8	4.3
1332	2.0	2.8	4.8
1836	1.2	3.1	4.3

The systematic error includes the error in calibration standards, weighing error, and estimated counting corrections. All uncertainties are stated at the 99% confidence level.

No interfering gamma emitting impurities were detected during calibration. Depending on the resolution and energy dispersion (keV/channel) of the measuring system, the following spectral conflicts may occur: (1) between the 88 keV gamma-ray and the X-rays emitted in the decay of Hg-203, (2) between the 1333 keV gamma-ray and the 1325 keV single escape peak from the 1836 keV gamma-ray.

000130



U.S. DEPARTMENT OF COMMERCE
National Institute of Standards & Technology
Gaithersburg, MD 20899

Certificate of Participation

*Analytics, Incorporated
Atlanta, Georgia*

is a participant for the period January 1, 1994, through December 31, 1994, in a radioactivity measurements assurance program conducted by the National Institute of Standards and Technology, in cooperation with the U.S. Council for Energy Awareness. Continued participation is evidenced by dated Reports of Traceability issued for particular radionuclides, which indicate the deviation of the participant's reported value from that measured by the National Institute of Standards and Technology. The significance of these Reports is addressed on the back of this certificate.

For the Director,

A handwritten signature in cursive script, reading "J.M. Robin Hutchinson", is written over a horizontal line.


J.M. Robin Hutchinson, Acting Group Leader
Radioactivity Group
Physics Laboratory
(over)

000131

PROJECT Stds 178 A thru L

(ATI 10) (26)
 Analytcs Std. 49500-307 was diluted 12 ways
 for gamma spectrometer daily calibration standards.
 For each standard, 4 ml of 49500-307 was added
 to 100 ml in a 100 ml volumetric with dilutes.
 The solutions were agitated and transferred to a
 minimal 250 cc pre-chemical right cylinder. Activity
 calculations (as of 10-1-94) are attached below.
 The remaining 96 ml of standard was also counted down
 for chain. *Healthy*

Examination Date:
 For primary calibration: 10-1-95
 For instrument performance
 checks daily calibration
 10-31-94



ATI Working Standards 178-A through L
 ATI Inventory Number 126
 Analytcs Standard ID: 49500-307
 Prepared 01/08/95

Isotope	Energy	GPS	Apr.	Original Vol. (ml)	Aliquot ml	Final ml	Final pCi/ml	Final pCi/Total	Final pCi/l	Final GPS	Final DPS
Am-241	59.5	1759	0.359	50	4	100	105.9	10594	105940	140.7	392.0
Cd-109	88	2622	0.0372	50	4	100	1524.0	152398	1523976	209.8	5638.7
Co-57	122	1472	0.8551	50	4	100	37.2	3722	37220	117.8	137.7
Ce-139	166	2037	0.8035	50	4	100	54.8	5481	54814	163.0	202.8
Hg-203	279	3978	0.773	50	4	100	111.3	11127	111269	318.2	411.7
Sn-113	392	2862	0.649	50	4	100	95.3	9535	95348	229.0	352.8
Cs-137	662	1735	0.8512	50	4	100	44.1	4407	44071	138.8	163.1
Y-88	898	7290	0.934	50	4	100	168.8	16876	168760	583.2	624.4
Co-60	1173	3347	1	50	4	100	72.4	7237	72368	267.8	267.8
Co-60	1332	3355	1	50	4	100	72.5	7254	72541	268.4	268.4
Y-88	1836	7630	0.9938	50	4	100	166.0	16600	166002	610.4	614.2

Continued on Page _____

Read and Understood By

Healthy
 Signed

1/8/95
 Date

D. Lab
 Signed

2/6/91
 000132

Paragon Analytics

Gamma Spectrometer Calibration Log

Date: 11/2/06

Reviewed By/Date: JP 11/2/06

Det. No.	Out Of Service	Background		Source Check			Repeat Source Check			
		Started	OK	Started	OK	Failed Parameter(s)	OK	Failed Parameter(s)	Corrective Action Taken **	Removed from Service
1.	JP			/	/					
2.				JP	JP					
3.	JP			/	/					
4.	JP			/	/					
5.	JP			/	/					
6.				JP	/	60 EFF	JP			
7.	JP			/	/					
8.				JP	JP					
9.				JP	JP					
10.	JP			/	/					
11.										
12.										

** Corrective Action:

318719 A

Form 754r11a.doc (6/13/2005)

817011

000133

062625D06.SPC Analyzed by JP

SEEKER D E T E C T O R Q . C . A N A L Y S I S V e r s i o n 2 . 2 . 2

ID: DAILY CHECK

Detector # 6 Detector Q.C. Analysis for 11/02/2006 07:46:05

Standards File #: 98 (Daily Performance Check 49500-307)

#	Parameter	Value	n Sigma Test	Bounds Test	T- Test
1	60 keV Centroid	120.077	N.A.	Pass	N.A.
2	60 keV FWHM	7.683E-01	N.A.	Pass	N.A.
3	60 keV Efficiency	5.507E-03	N.A.	<FAIL>	N.A.
4	662 keV Centroid	1322.177	N.A.	Pass	N.A.
5	662 keV FWHM	1.426	N.A.	Pass	N.A.
6	662 keV Efficiency	1.696E-02	N.A.	Pass	N.A.
7	1332 keV Centroid	2661.473	N.A.	Pass	N.A.
8	1332 keV FWHM	2.040	N.A.	Pass	N.A.
9	1332 keV Efficiency	8.894E-03	N.A.	Pass	N.A.

062626D06.SPC Analyzed by JP

SEEKER D E T E C T O R Q . C . A N A L Y S I S V e r s i o n 2 . 2 . 2

ID: DAILY CHECK

Detector # 6 Detector Q.C. Analysis for 11/02/2006 08:07:21

Standards File #: 98 (Daily Performance Check 49500-307)

#	Parameter	Value	n Sigma Test	Bounds Test	T-Test
1	60 keV Centroid	120.079	N.A.	Pass	N.A.
2	60 keV FWHM	8.741E-01	N.A.	Pass	N.A.
3	60 keV Efficiency	5.961E-03	N.A.	Pass	N.A.
4	662 keV Centroid	1322.155	N.A.	Pass	N.A.
5	662 keV FWHM	1.388	N.A.	Pass	N.A.
6	662 keV Efficiency	1.625E-02	N.A.	Pass	N.A.
7	1332 keV Centroid	2661.584	N.A.	Pass	N.A.
8	1332 keV FWHM	1.901	N.A.	Pass	N.A.
9	1332 keV Efficiency	9.371E-03	N.A.	Pass	N.A.

Paragon Analytics

Gamma Spectrometer Calibration Log

Date: 11/3/06

Reviewed By/Date: JP 11/3/06

Det. No.	Out Of Service	Background		Source Check			Repeat Source Check			
		Started	OK	Started	OK	Failed Parameter(s)	OK	Failed Parameter(s)	Corrective Action Taken **	Removed from Service
1.	JP	/	/	/	/					
2.				JP	JP					
3.	JP	/	/	/	/					
4.	JP	/	/	/	/					
5.	JP	/	/	/	/					
6.				JP	JP					
7.	JP	/	/	/	/					
8.				JP	/	133Z Centroid	JP			
9.				JP	/	66Z FWHM	JP			
10.	JP	/	/	/	/					
11.										
12.										

** Corrective Action:

318721 A

Form 754r11a.doc (6/13/2005)

OSTAD

000136

062542D02.SPC Analyzed by JP

SEEKER D E T E C T O R Q . C . A N A L Y S I S Version 2.2.2

ID: DAILY CHECK

Detector # 2 Detector Q.C. Analysis for 11/03/2006 06:54:58

Standards File #: 98 (Daily Performance Check 49500-307)

#	Parameter	Value	n Sigma Test	Bounds Test	T- Test
1	60 keV Centroid	120.471	N.A.	Pass	N.A.
2	60 keV FWHM	8.471E-01	N.A.	Pass	N.A.
3	60 keV Efficiency	1.297E-02	N.A.	Pass	N.A.
4	662 keV Centroid	1323.814	N.A.	Pass	N.A.
5	662 keV FWHM	1.745	N.A.	Pass	N.A.
6	662 keV Efficiency	2.131E-02	N.A.	Pass	N.A.
7	1332 keV Centroid	2664.197	N.A.	Pass	N.A.
8	1332 keV FWHM	2.713	N.A.	Pass	N.A.
9	1332 keV Efficiency	1.067E-02	N.A.	Pass	N.A.

062635D06.SPC Analyzed by JP

SEEKER D E T E C T O R Q . C . A N A L Y S I S V e r s i o n 2 . 2 . 2

ID: DAILY CHECK

Detector # 6 Detector Q.C. Analysis for 11/03/2006 06:55:13

Standards File #: 98 (Daily Performance Check 49500-307)

#	Parameter	Value	n Sigma Test	Bounds Test	T-Test
1	60 keV Centroid	120.118	N.A.	Pass	N.A.
2	60 keV FWHM	8.148E-01	N.A.	Pass	N.A.
3	60 keV Efficiency	6.477E-03	N.A.	Pass	N.A.
4	662 keV Centroid	1322.274	N.A.	Pass	N.A.
5	662 keV FWHM	1.422	N.A.	Pass	N.A.
6	662 keV Efficiency	1.678E-02	N.A.	Pass	N.A.
7	1332 keV Centroid	2661.734	N.A.	Pass	N.A.
8	1332 keV FWHM	1.869	N.A.	Pass	N.A.
9	1332 keV Efficiency	8.997E-03	N.A.	Pass	N.A.

Paragon Analytics

Gamma Spectrometer Calibration Log

Date: 11/4/06

Reviewed By/Date: JP 11/4/06

Det. No.	Out Of Service	Background		Source Check			Repeat Source Check			
		Started	OK	Started	OK	Failed Parameter(s)	OK	Failed Parameter(s)	Corrective Action Taken **	Removed from Service
1.	JP			/	/					
2.				JP	JP					
3.	JP			/	/					
4.	JP			/	/					
5.	JP			/	/					
6.				JP	JP					
7.	JP			/	/					
8.				JP	JP					
9.				JP	JP					
10.	JP			/	/					
11.										
12.										

** Corrective Action:

318722 A

157912

062550D02.SPC Analyzed by JP

SEEKER D E T E C T O R Q . C . A N A L Y S I S V e r s i o n 2 . 2 . 2

ID: DAILY CHECK

Detector # 2 Detector Q.C. Analysis for 11/04/2006 09:49:55

Standards File #: 98 (Daily Performance Check 49500-307)

#	Parameter	Value	n Sigma Test	Bounds Test	T- Test
1	60 keV Centroid	120.427	N.A.	Pass	N.A.
2	60 keV FWHM	7.985E-01	N.A.	Pass	N.A.
3	60 keV Efficiency	1.355E-02	N.A.	Pass	N.A.
4	662 keV Centroid	1323.373	N.A.	Pass	N.A.
5	662 keV FWHM	1.739	N.A.	Pass	N.A.
6	662 keV Efficiency	2.159E-02	N.A.	Pass	N.A.
7	1332 keV Centroid	2663.284	N.A.	Pass	N.A.
8	1332 keV FWHM	2.492	N.A.	Pass	N.A.
9	1332 keV Efficiency	1.050E-02	N.A.	Pass	N.A.

Paragon Analytics

Gamma Spectrometer Calibration Log

Date: 11/5/06

Reviewed By/Date: JP 11/5/06

Det. No.	Out Of Service	Background		Source Check			Repeat Source Check			
		Started	OK	Started	OK	Failed Parameter(s)	OK	Failed Parameter(s)	Corrective Action Taken **	Removed from Service
1.	JP			/	/					
2.				JP	JP					
3.	JP			/	/					
4.	JP			/	/					
5.	JP			/	/					
6.				JP	JP					
7.	JP			/	/					
8.				JP	JP					
9.				JP	/	66Z FWHM	JP			
10.	JP			/	/					
11.										
12.										

** Corrective Action:

318723 A

Form 754r11a.doc (6/13/2005)

38718

000141

062553D02.SPC Analyzed by JP

SEEKER D E T E C T O R Q . C . A N A L Y S I S Version 2.2.2

ID: DAILY CHECK
Detector # 2 Detector Q.C. Analysis for 11/05/2006 09:01:31
Standards File #: 98 (Daily Performance Check 49500-307)

#	Parameter	Value	n Sigma Test	Bounds Test	T-Test
1	60 keV Centroid	120.419	N.A.	Pass	N.A.
2	60 keV FWHM	8.361E-01	N.A.	Pass	N.A.
3	60 keV Efficiency	1.334E-02	N.A.	Pass	N.A.
4	662 keV Centroid	1323.295	N.A.	Pass	N.A.
5	662 keV FWHM	1.749	N.A.	Pass	N.A.
6	662 keV Efficiency	2.143E-02	N.A.	Pass	N.A.
7	1332 keV Centroid	2663.271	N.A.	Pass	N.A.
8	1332 keV FWHM	2.709	N.A.	Pass	N.A.
9	1332 keV Efficiency	1.083E-02	N.A.	Pass	N.A.

 SEEKER D E T E C T O R Q . C . A N A L Y S I S V e r s i o n 2 . 2 . 2

ID: DAILY CHECK

Detector # 8 Detector Q.C. Analysis for 11/05/2006 08:55:13

Standards File #: 98 (Daily Performance Check 49500-307)

#	Parameter	Value	n Sigma Test	Bounds Test	T-Test
1	60 keV Centroid	120.018	N.A.	Pass	N.A.
2	60 keV FWHM	8.282E-01	N.A.	Pass	N.A.
3	60 keV Efficiency	7.248E-02	N.A.	Pass	N.A.
4	662 keV Centroid	1323.965	N.A.	Pass	N.A.
5	662 keV FWHM	1.399	N.A.	Pass	N.A.
6	662 keV Efficiency	1.784E-02	N.A.	Pass	N.A.
7	1332 keV Centroid	2663.809	N.A.	Pass	N.A.
8	1332 keV FWHM	1.890	N.A.	Pass	N.A.
9	1332 keV Efficiency	9.478E-03	N.A.	Pass	N.A.

Paragon Analytics

Gamma Spectrometer Calibration Log

Date: 11/2/06 ⁶⁹ _{11/2/06}

Reviewed By/Date: [Signature]

Det. No.	Out Of Service	Background		Source Check			Repeat Source Check			
		Started	OK	Started	OK	Failed Parameter(s)	OK	Failed Parameter(s)	Corrective Action Taken **	Removed from Service
1.	[Signature]									
2.				[Signature]		SI				
3.	[Signature]									
4.	[Signature]									
5.	[Signature]									
6.				[Signature]		SI				
7.	[Signature]									
8.				[Signature]		SI				
9.				[Signature]		SI				
10.	[Signature]									
11.										
12.										

** Corrective Action:

318724 A

Form 754r11a.doc (6/13/2005)

000144

SEEKER D E T E C T O R Q . C . A N A L Y S I S V e r s i o n 2 . 2 . 2

ID: DAILY CHECK

Detector # 2 Detector Q.C. Analysis for 11/06/2006 09:16:57

Standards File #: 98 (Daily Performance Check 49500-307)

#	Parameter	Value	n Sigma Test	Bounds Test	T- Test
1	60 keV Centroid	120.483	N.A.	Pass	N.A.
2	60 keV FWHM	8.519E-01	N.A.	Pass	N.A.
3	60 keV Efficiency	1.365E-02	N.A.	Pass	N.A.
4	662 keV Centroid	1323.462	N.A.	Pass	N.A.
5	662 keV FWHM	1.747	N.A.	Pass	N.A.
6	662 keV Efficiency	2.166E-02	N.A.	Pass	N.A.
7	1332 keV Centroid	2663.626	N.A.	Pass	N.A.
8	1332 keV FWHM	2.736	N.A.	Pass	N.A.
9	1332 keV Efficiency	1.117E-02	N.A.	Pass	N.A.

063149D08.SPC Analyzed by ^Q

SEEKER DETECTOR Q.C. ANALYSIS Version 2.2.2

ID: DAILY CHECK

Detector # 8 Detector Q.C. Analysis for 11/06/2006 09:17:24
Standards File #: 98 (Daily Performance Check 49500-307)

#	Parameter	Value	n Sigma Test	Bounds Test	T-Test
1	60 keV Centroid	119.945	N.A.	Pass	N.A.
2	60 keV FWHM	8.311E-01	N.A.	Pass	N.A.
3	60 keV Efficiency	7.270E-02	N.A.	Pass	N.A.
4	662 keV Centroid	1323.320	N.A.	Pass	N.A.
5	662 keV FWHM	1.408	N.A.	Pass	N.A.
6	662 keV Efficiency	1.863E-02	N.A.	Pass	N.A.
7	1332 keV Centroid	2662.700	N.A.	Pass	N.A.
8	1332 keV FWHM	1.958	N.A.	Pass	N.A.
9	1332 keV Efficiency	9.710E-03	N.A.	Pass	N.A.